

Data Path : Z:\SVOASRV\HPCHEM1\BNA_F\DATA\BF082318\
 Data File : BF108451.D
 Acq On : 24 Aug 2018 5:58
 Operator : JU/SJ
 Sample : J4582-02
 Misc :
 ALS Vial : 27 Sample Multiplier: 1

Instrument :
 BNA_F
 ClientSampleId :
 082018-DBRI-E-D-M

Quant Time: Aug 24 07:57:52 2018
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_F\METHODS\8270-BF080818.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Wed Aug 22 11:01:45 2018
 Response via : Initial Calibration


Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	7.00	152	130662	20.00	ng	0.00
21) Naphthalene-d8	8.28	136	480545	20.00	ng	0.00
38) Acenaphthene-d10	10.04	164	221605	20.00	ng	-0.01
63) Phenanthrene-d10	11.54	188	358038	20.00	ng	-0.01
75) Chrysene-d12	14.19	240	279602	20.00	ng	-0.01
86) Perylene-d12	15.74	264	243443	20.00	ng	-0.02
System Monitoring Compounds						
5) 2-Fluorophenol	5.62	112	299644	41.52	ng	0.00
7) Phenol-d6	6.61	99	254097	26.49	ng	-0.02
23) Nitrobenzene-d5	7.57	82	801619	93.08	ng	0.00
41) 2,4,6-Tribromophenol	10.84	330	175495	79.39	ng	-0.01
44) 2-Fluorobiphenyl	9.36	172	1455689	105.46	ng	0.00
78) Terphenyl-d14	13.12	244	1051076	95.58	ng	0.00
Target Compounds						
19) n-Nitroso-di-n-propylamine	7.57	70	108274	16.695	ng	# 82
25) Isophorone	7.57	82	801612	48.836	ng	# 64
49) Dimethylphthalate	9.75	163	60814	3.940	ng	# 99
50) 2,6-Dinitrotoluene	10.04	165	28547	8.839	ng	# 26
56) 2,4-Dinitrotoluene	10.04	165	28547	6.867	ng	# 23
64) 4,6-Dinitro-2-methylphenol	10.84	198	200	5.491	ng	# 1
66) 4-Bromophenyl-phenylether	10.84	248	11385	2.926	ng	# 1


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\SVOASRV\HPCHEM1\BNA_F\DATA\BF082318\
Data File : BF108451.D
Acq On : 24 Aug 2018 5:58
Operator : JU/SJ
Sample : J4582-02
Misc :
ALS Vial : 27 Sample Multiplier: 1

Instrument :
BNA_F
ClientSampleId :
082018-DBRI-E-D-M

Quant Time: Aug 24 07:57:52 2018
Quant Method : Z:\SVOASRV\HPCHEM1\BNA_F\METHODS\8270-BF080818.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Wed Aug 22 11:01:45 2018
Response via : Initial Calibration






#1




1,4-Dichlorobenzene-d4
Concen: 20.000 ng
RT: 7.00 min Scan# 793
Delta R.T. -0.01 min
Lab File: BF108451.D
Acq: 24 Aug 2018 5:58

Instrument :
BNA_F
ClientSampleId :
082018-DBRI-E-D-M

Tgt Ion:152 Resp: 130662
Ion Ratio Lower Upper
152 100
150 157.0 124.6 186.8
115 57.9 50.1 75.1




Abundance Ion 152.00 (151.70 to 152.70): E
Ion 150.00 (149.70 to 150.70): E
Ion 115.00 (114.70 to 115.70): E





#5


2-Fluorophenol
Concen: 41.519 ng
RT: 5.62 min Scan# 559
Delta R.T. -0.01 min
Lab File: BF108451.D
Acq: 24 Aug 2018 5:58

Tgt Ion:112 Resp: 299644
Ion Ratio Lower Upper
112 100
64 65.7 47.9 71.9
63 35.1 23.7 35.5



Abundance Ion 112.00 (111.70 to 112.70): E
Ion 64.00 (63.70 to 64.70): BF1
Ion 63.00 (62.70 to 63.70): BF1





#7

Phenol-d6

Concen: 26.495 ng

RT: 6.61 min Scan# 727

Delta R.T. -0.02 min

Lab File: BF108451.D

Acq: 24 Aug 2018 5:58

Instrument :

BNA_F

ClientSampleId :

082018-DBRI-E-D-M


Tgt Ion: 99 Resp: 254097

Ion Ratio Lower Upper

99 100

42 20.9 14.5 21.7




71 34.8 25.4 38.0



Abundance Ion 99.00 (98.70 to 99.70): BF1

Ion 42.00 (41.70 to 42.70): BF1

Ion 71.00 (70.70 to 71.70): BF1



#19

n-Nitroso-di-n-propylamine

Concen: 16.695 ng

RT: 7.57 min Scan# 889

Delta R.T. 0.15 min

Lab File: BF108451.D

Acq: 24 Aug 2018 5:58

Tgt Ion: 70 Resp: 108274


Ion Ratio Lower Upper

70 100

42 52.8 47.5 71.3

101 0.0 7.4 11.2#

130 2.1 16.6 25.0#






Abundance Ion 70.00 (69.70 to 70.70): BF1

Ion 42.00 (41.70 to 42.70): BF1

Ion 101.00 (100.70 to 101.70): BF1

Ion 130.00 (129.70 to 130.70): BF1









#21
Naphthalene-d8
Concen: 20.000 ng
RT: 8.28 min Scan# 1011
Delta R.T. -0.01 min
Lab File: BF108451.D
Acq: 24 Aug 2018 5:58




Instrument :
BNA_F
ClientSampleId :
082018-DBRI-E-D-M


Tgt Ion:	136	Resp:	480545
Ion	Ratio	Lower	Upper
136	100		
137	10.9	8.9	13.3
54	9.2	6.6	9.8
68	5.9	5.0	7.4



#23
Nitrobenzene-d5
Concen: 93.085 ng
RT: 7.57 min Scan# 889
Delta R.T. -0.01 min
Lab File: BF108451.D
Acq: 24 Aug 2018 5:58

Tgt Ion:	82	Resp:	801619
Ion	Ratio	Lower	Upper
82	100		
128	43.8	36.1	54.1
54	56.8	41.4	62.2





#25

Isophorone

Concen: 48.836 ng

RT: 7.57 min Scan# 889

Delta R.T. -0.26 min

Lab File: BF108451.D

Acq: 24 Aug 2018 5:58

Instrument :

BNA_F

ClientSampleId :

082018-DBRI-E-D-M


Tgt Ion: 82 Resp: 801612

Ion Ratio Lower Upper

82 100

95 0.0 5.2 7.8#




138 0.0 14.9 22.3#



Abundance Ion 82.00 (81.70 to 82.70): BF1

Ion 95.00 (94.70 to 95.70): BF1

Ion 138.00 (137.70 to 138.70): BF1



#38

Acenaphthene-d10

Concen: 20.000 ng

RT: 10.04 min Scan# 1310

Delta R.T. -0.01 min

Lab File: BF108451.D

Acq: 24 Aug 2018 5:58


Tgt Ion: 164 Resp: 221605

Ion Ratio Lower Upper

164 100

162 103.3 86.1 129.1



160 46.2 38.3 57.5




Abundance Ion 164.00 (163.70 to 164.70): E

Ion 162.00 (161.70 to 162.70): E

Ion 160.00 (159.70 to 160.70): E






#41
 2,4,6-Tribromophenol
 Concen: 79.393 ng
 RT: 10.84 min Scan# 1445
 Delta R.T. -0.01 min
 Lab File: BF108451.D
 Acq: 24 Aug 2018 5:58

Instrument :
 BNA_F
 ClientSampleId :
 082018-DBRI-E-D-M

Tgt Ion	Ratio	Lower	Upper
330	100		
332	94.6	77.0	115.6
141	32.2	29.9	44.9






Abundance

Ion 329.80 (329.50 to 330.50): E


Ion 331.80 (331.50 to 332.50): E

Ion 141.00 (140.70 to 141.70): E



#44
 2-Fluorobiphenyl
 Concen: 105.461 ng
 RT: 9.36 min Scan# 1194
 Delta R.T. -0.01 min
 Lab File: BF108451.D
 Acq: 24 Aug 2018 5:58

Tgt Ion	Ratio	Lower	Upper
172	100		
171	37.3	29.7	44.5
170	25.0	19.6	29.4






Abundance

Ion 172.00 (171.70 to 172.70): E

Ion 171.00 (170.70 to 171.70): E

Ion 170.00 (169.70 to 170.70): E









#49
Dimethylphthalate
Concen: 3.940 ng
RT: 9.75 min Scan# 1260
Delta R.T. -0.02 min
Lab File: BF108451.D
Acq: 24 Aug 2018 5:58




Instrument :
BNA_F
ClientSampleId :
082018-DBRI-E-D-M


Tgt Ion	Ratio	Lower	Upper
163	100		
194	4.1	2.7	4.1#
164	9.9	8.2	12.2



#50
2,6-Dinitrotoluene
Concen: 8.839 ng
RT: 10.04 min Scan# 1310
Delta R.T. 0.21 min
Lab File: BF108451.D
Acq: 24 Aug 2018 5:58

Tgt Ion	Ratio	Lower	Upper
165	100		
63	1.1	44.7	67.1#
89	1.7	44.0	66.0#









#56
 2,4-Dinitrotoluene
 Concen: 6.867 ng
 RT: 10.04 min Scan# 1310
 Delta R.T. -0.19 min
 Lab File: BF108451.D
 Acq: 24 Aug 2018 5:58




Instrument :
 BNA_F
 ClientSampleId :
 082018-DBRI-E-D-M


Tgt Ion:	165	Resp:	28547
Ion	Ratio	Lower	Upper
165	100		
63	1.1	36.4	54.6#
89	1.7	57.8	86.6#
182	0.0	1.6	2.4#



#63
 Phenanthrene-d10
 Concen: 20.000 ng
 RT: 11.54 min Scan# 1564
 Delta R.T. -0.01 min
 Lab File: BF108451.D
 Acq: 24 Aug 2018 5:58

Tgt Ion:	188	Resp:	358038
Ion	Ratio	Lower	Upper
188	100		
94	8.7	8.5	12.7
80	10.0	9.2	13.8









#64
 4,6-Dinitro-2-methylphenol
 Concen: 5.491 ng
 RT: 10.84 min Scan# 1445
 Delta R.T. 0.19 min
 Lab File: BF108451.D
 Acq: 24 Aug 2018 5:58

Instrument :
 BNA_F
 ClientSampleId :
 082018-DBRI-E-D-M

Tgt Ion:198 Resp: 200
 Ion Ratio Lower Upper
 198 100
 51 167.2 37.7 77.7#
 105 161.6 36.3 76.3#




Abundance Ion 198.00 (197.70 to 198.70): E
 Ion 51.00 (50.70 to 51.70): BF1
 Ion 105.00 (104.70 to 105.70): E






#66
 4-Bromophenyl-phenylether
 Concen: 2.926 ng
 RT: 10.84 min Scan# 1445
 Delta R.T. -0.25 min
 Lab File: BF108451.D
 Acq: 24 Aug 2018 5:58

Tgt Ion:248 Resp: 11385
 Ion Ratio Lower Upper
 248 100
 250 196.3 76.9 115.3#
 141 473.2 74.4 111.6#



Abundance Ion 248.00 (247.70 to 248.70): E
 Ion 250.00 (249.70 to 250.70): E
 Ion 141.00 (140.70 to 141.70): E





#75

Chrysene-d12

Concen: 20.000 ng

RT: 14.19 min Scan# 2015

Delta R.T. -0.01 min

Lab File: BF108451.D

Acq: 24 Aug 2018 5:58

Instrument :

BNA_F

ClientSampleId :

082018-DBRI-E-D-M


Tgt Ion:240 Resp: 279602

Ion Ratio Lower Upper

240 100

120 9.8 9.5 14.3




236 26.9 20.7 31.1



Abundance Ion 240.00 (239.70 to 240.70): E

Ion 120.00 (119.70 to 120.70): E

Ion 236.00 (235.70 to 236.70): E



#78

Terphenyl-d14

Concen: 95.580 ng

RT: 13.12 min Scan# 1834

Delta R.T. -0.01 min

Lab File: BF108451.D

Acq: 24 Aug 2018 5:58


Tgt Ion:244 Resp: 1051076

Ion Ratio Lower Upper

244 100

212 7.3 5.4 8.0



122 9.8 11.0 16.4#




Abundance Ion 244.00 (243.70 to 244.70): E

Ion 212.00 (211.70 to 212.70): E

Ion 122.00 (121.70 to 122.70): E





#86
Perylene-d12
Concen: 20.000 ng
RT: 15.74 min Scan# 2279
Delta R.T. -0.02 min
Lab File: BF108451.D
Acq: 24 Aug 2018 5:58

Instrument :
BNA_F
ClientSampleId :
082018-DBRI-E-D-M

Tgt Ion	Ratio	Lower	Upper
264	100		
260	24.1	19.2	28.8
265	21.6	17.5	26.3

