

Data Path : Z:\SVOASRV\HPCHEM1\BNA F\DATA\BF082318\  
 Data File : BF108441.D  
 Acq On : 24 Aug 2018 00:37  
 Operator : JU/SJ  
 Sample : J4278-11 10X  
 Misc :  
 ALS Vial : 19 Sample Multiplier: 1

Instrument :  
 BNA\_F  
 ClientSampleId :  
 EO-01-082118-C

Quant Time: Aug 24 04:22:08 2018  
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA F\METHODS\8270-BF080818.M  
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION  
 QLast Update : Wed Aug 22 11:01:45 2018  
 Response via : Initial Calibration

| Internal Standards          | R.T.  | QIon | Response | Conc  | Units | Dev(Min) |
|-----------------------------|-------|------|----------|-------|-------|----------|
| 1) 1,4-Dichlorobenzene-d4   | 7.00  | 152  | 130617   | 20.00 | ng    | 0.00     |
| 21) Naphthalene-d8          | 8.28  | 136  | 490970   | 20.00 | ng    | 0.00     |
| 38) Acenaphthene-d10        | 10.04 | 164  | 222066   | 20.00 | ng    | -0.01    |
| 63) Phenanthrene-d10        | 11.54 | 188  | 357850   | 20.00 | ng    | 0.00     |
| 75) Chrysene-d12            | 14.19 | 240  | 271371   | 20.00 | ng    | 0.00     |
| 86) Perylene-d12            | 15.75 | 264  | 240455   | 20.00 | ng    | -0.01    |
| System Monitoring Compounds |       |      |          |       |       |          |
| 5) 2-Fluorophenol           | 5.62  | 112  | 66560    | 9.23  | ng    | -0.01    |
| 7) Phenol-d6                | 6.61  | 99   | 91595    | 9.55  | ng    | -0.02    |
| 23) Nitrobenzene-d5         | 7.56  | 82   | 56716    | 6.45  | ng    | -0.01    |
| 41) 2,4,6-Tribromophenol    | 10.84 | 330  | 14233    | 6.43  | ng    | -0.01    |
| 44) 2-Fluorobiphenyl        | 9.35  | 172  | 118085   | 8.54  | ng    | -0.01    |
| 78) Terphenyl-d14           | 13.12 | 244  | 84175    | 7.89  | ng    | -0.01    |

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\SVOASRV\HPCHEM1\BNA F\DATA\BF082318\  
Data File : BF108441.D  
Acq On : 24 Aug 2018 00:37  
Operator : JU/SJ  
Sample : J4278-11 10X  
Misc :  
ALS Vial : 19 Sample Multiplier: 1

Instrument :  
BNA\_F  
ClientSampleId :  
EO-01-082118-C

Quant Time: Aug 24 04:22:08 2018  
Quant Method : Z:\SVOASRV\HPCHEM1\BNA F\METHODS\8270-BF080818.M  
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION  
QLast Update : Wed Aug 22 11:01:45 2018  
Response via : Initial Calibration

