

Data Path : Z:\svoasrv\HPCHEM1\BNA F\Data\BF082318\
 Data File : BF108419.D
 Acq On : 23 Aug 2018 10:09
 Operator : JU/SJ
 Sample : SSTDCCC040
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_F
 LabSampleId :
 SSTDCCC040

Quant Time: Aug 23 12:48:38 2018
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA F\METHODS\8270-BF080818.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Wed Aug 22 11:01:45 2018
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvqRF	CCRF	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	55	0.00
2	1,4-Dioxane	0.568	0.540	4.9	52	0.02
3	Pyridine	1.462	1.384	5.3	51	0.01
4	n-Nitrosodimethylamine	0.823	0.746	9.4	48#	0.01
5 S	2-Fluorophenol	1.105	1.129	-2.2	56	0.00
6	Aniline	1.997	2.039	-2.1	56	0.00
7 S	Phenol-d6	1.468	1.511	-2.9	56	0.00
8	2-Chlorophenol	1.244	1.295	-4.1	56	0.00
9	Benzaldehyde	1.138	1.079	5.2	51	0.00
10 C	Phenol	1.518	1.530	-0.8	56	0.00
11	bis(2-Chloroethyl)ether	1.228	1.239	-0.9	55	0.00
12	1,3-Dichlorobenzene	1.439	1.475	-2.5	56	0.00
13 C	1,4-Dichlorobenzene	1.445	1.472	-1.9	55	0.00
14	1,2-Dichlorobenzene	1.344	1.369	-1.9	56	0.00
15	Benzyl Alcohol	1.236	1.186	4.0	51	0.00
16	2,2'-oxybis(1-Chloropropane	2.529	2.362	6.6	50	0.00
17	2-Methylphenol	1.067	1.136	-6.5	57	0.00
18	Hexachloroethane	0.515	0.534	-3.7	55	0.00
19 P	n-Nitroso-di-n-propylamine	0.993	0.984	0.9	54	0.00
20	3+4-Methylphenols	1.367	1.383	-1.2	55	0.00
21 I	Naphthalene-d8	1.000	1.000	0.0	55	0.00
22	Acetophenone	0.468	0.461	1.5	54	0.00
23 S	Nitrobenzene-d5	0.358	0.369	-3.1	56	0.00
24	Nitrobenzene	0.362	0.363	-0.3	54	0.00
25	Isophorone	0.683	0.676	1.0	55	0.00
26 C	2-Nitrophenol	0.137	0.160	-16.8	63	0.00
27	2,4-Dimethylphenol	0.303	0.303	0.0	55	0.00
28	bis(2-Chloroethoxy)methane	0.399	0.398	0.3	55	0.00
29 C	2,4-Dichlorophenol	0.279	0.293	-5.0	56	0.00
30	1,2,4-Trichlorobenzene	0.308	0.312	-1.3	56	0.00
31	Naphthalene	0.910	0.940	-3.3	57	0.00
32	Benzoic acid	0.161	0.143	11.2	48#	0.01
33	4-Chloroaniline	0.396	0.401	-1.3	55	0.00
34 C	Hexachlorobutadiene	0.195	0.203	-4.1	57	0.00
35	Caprolactam	0.087	0.088	-1.1	53	0.00
36 C	4-Chloro-3-methylphenol	0.301	0.308	-2.3	55	0.00
37	2-Methylnaphthalene	0.644	0.648	-0.6	56	0.00
38 I	Acenaphthene-d10	1.000	1.000	0.0	55	0.00
39	1,2,4,5-Tetrachlorobenzene	0.614	0.637	-3.7	56	0.00
40 P	Hexachlorocyclopentadiene	0.397	0.402	-1.3	52	0.00
41 S	2,4,6-Tribromophenol	0.199	0.226	-13.6	58	0.00
42 C	2,4,6-Trichlorophenol	0.381	0.407	-6.8	56	0.00
43	2,4,5-Trichlorophenol	0.420	0.451	-7.4	56	0.00
44 S	2-Fluorobiphenyl	1.246	1.326	-6.4	60	0.00

Data Path : Z:\svoasrv\HPCHEM1\BNA F\Data\BF082318\
 Data File : BF108419.D
 Acq On : 23 Aug 2018 10:09
 Operator : JU/SJ
 Sample : SSTDCCC040
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_F
 LabSampleId :
 SSTDCCC040

Quant Time: Aug 23 12:48:38 2018
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA F\METHODS\8270-BF080818.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Wed Aug 22 11:01:45 2018
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
45	1,1'-Biphenyl	1.475	1.513	-2.6	56	0.00
46	2-Chloronaphthalene	1.165	1.196	-2.7	56	0.00
47	2-Nitroaniline	0.377	0.401	-6.4	55	0.00
48	Acenaphthylene	1.843	1.925	-4.4	57	0.00
49	Dimethylphthalate	1.393	1.431	-2.7	57	0.00
50	2,6-Dinitrotoluene	0.291	0.316	-8.6	57	0.00
51 C	Acenaphthene	1.129	1.167	-3.4	57	0.00
52	3-Nitroaniline	0.319	0.337	-5.6	56	0.00
53 P	2,4-Dinitrophenol	0.092	0.112	-21.7	65	0.00
54	Dibenzofuran	1.635	1.672	-2.3	56	0.00
55 P	4-Nitrophenol	0.241	0.239	0.8	53	0.00
56	2,4-Dinitrotoluene	0.375	0.420	-12.0	57	0.00
57	Fluorene	1.294	1.346	-4.0	58	0.00
58	2,3,4,6-Tetrachlorophenol	0.351	0.370	-5.4	54	0.00
59	Diethylphthalate	1.349	1.401	-3.9	57	0.00
60	4-Chlorophenyl-phenylether	0.674	0.690	-2.4	56	0.00
61	4-Nitroaniline	0.315	0.346	-9.8	57	0.00
62	Azobenzene	1.393	1.393	0.0	54	0.00
63 I	Phenanthrene-d10	1.000	1.000	0.0	55	0.00
64	4,6-Dinitro-2-methylphenol	0.069	0.078	-13.0	59	0.00
65 c	n-Nitrosodiphenylamine	0.591	0.604	-2.2	56	0.00
66	4-Bromophenyl-phenylether	0.217	0.226	-4.1	57	0.00
67	Hexachlorobenzene	0.229	0.233	-1.7	56	0.00
68	Atrazine	0.198	0.211	-6.6	58	0.00
69 C	Pentachlorophenol	0.109	0.108	0.9	52	0.00
70	Phenanthrene	0.943	0.974	-3.3	58	0.00
71	Anthracene	0.967	1.000	-3.4	58	0.00
72	Carbazole	0.859	0.890	-3.6	57	0.00
73	Di-n-butylphthalate	0.973	1.076	-10.6	60	0.00
74 C	Fluoranthene	1.030	1.076	-4.5	59	0.00
75 I	Chrysene-d12	1.000	1.000	0.0	55	0.00
76	Benzidine	0.747	0.828	-10.8	57	0.00
77	Pyrene	1.266	1.349	-6.6	59	0.00
78 S	Terphenyl-d14	0.787	0.847	-7.6	60	0.00
79	Butylbenzylphthalate	0.503	0.579	-15.1	59	0.00
80	Benzo(a)anthracene	1.148	1.183	-3.0	56	0.00
81	3,3'-Dichlorobenzidine	0.411	0.427	-3.9	53	0.00
82	Chrysene	1.052	1.081	-2.8	56	0.00
83	Bis(2-ethylhexyl)phthalate	0.681	0.743	-9.1	57	0.00
84 c	Di-n-octyl phthalate	1.038	1.209	-16.5	60	0.00
85	Indeno(1,2,3-cd)pyrene	0.912	0.783	14.1	48#	0.00
86 I	Perylene-d12	1.000	1.000	0.0	50	0.00
87	Benzo(b)fluoranthene	1.195	1.273	-6.5	51	0.00

Data Path : Z:\svoasrv\HPCHEM1\BNA F\Data\BF082318\
Data File : BF108419.D
Acq On : 23 Aug 2018 10:09
Operator : JU/SJ
Sample : SSTDCCC040
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
BNA_F
LabSampleId :
SSTDCCC040

Quant Time: Aug 23 12:48:38 2018
Quant Method : Z:\SVOASRV\HPCHEM1\BNA F\METHODS\8270-BF080818.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Wed Aug 22 11:01:45 2018
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
88	Benzo(k)fluoranthene	1.099	1.161	-5.6	54	0.00
89 C	Benzo(a)pyrene	1.070	1.109	-3.6	51	0.00
90	Dibenzo(a,h)anthracene	0.885	0.839	5.2	47#	0.00
91	Benzo(a,h,i)perylene	0.872	0.826	5.3	48#	0.00

(#) = Out of Range

SPCC's out = 0 CCC's out = 0