


Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF082319\
 Data File : BF116479.D
 Acq On : 23 Aug 2019 16:09
 Operator : HP/JU
 Sample : DFTPP
 Misc :
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 BNA_F
 ClientSampleId :
 DFTPP

Integration File: rteint.p

Method : Z:\SVOASRV\HPCHEM1\BNA_F\METHODS\8270-BF082319.M
 Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 Last Update : Fri Aug 23 13:38:31 2019



AutoFind: Scans 1626, 1627, 1628; Background Corrected with Scan 1622

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
51	198	10	80	38.5	27738	PASS
68	69	0.00	2	2.0	614	PASS
69	198	0.00	100	42.4	30509	PASS
70	69	0.00	2	0.0	0	PASS
127	198	10	80	51.5	37040	PASS
197	198	0.00	2	0.0	0	PASS
198	198	100	100	100.0	71976	PASS
199	198	5	9	6.6	4775	PASS
275	198	10	60	25.2	18142	PASS
365	198	1	100	2.7	1917	PASS
441	198	0.01	100	10.3	7445	PASS
442	198	50	100	65.2	46946	PASS
443	442	15	24	18.9	8871	PASS