

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF082423\
 Data File : BF134921.D
 Acq On : 24 Aug 2023 13:32
 Operator : CG\JU
 Sample : 04099-01
 Misc :
 ALS Vial : 8 Sample Multiplier: 1

Instrument :
 BNA_F
 ClientSampleId :
 CONC-WC-1

Manual Integrations
 APPROVED

Reviewed By :Yogesh Patel 08/25/2023
 Supervised By :mohammad ahmed 08/25/2023

Quant Time: Aug 24 15:56:31 2023
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF082223.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Tue Aug 22 19:33:25 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	6.793	152	90895	20.000	ng	0.00
21) Naphthalene-d8	8.075	136	327502	20.000	ng	0.00
39) Acenaphthene-d10	9.845	164	94783	20.000	ng	0.01
64) Phenanthrene-d10	11.428	188	84027m	20.000	ng	0.10
76) Chrysene-d12	14.098	240	129898m	20.000	ng	0.12
86) Perylene-d12	15.515	264	182134	20.000	ng	0.06
System Monitoring Compounds						
5) 2-Fluorophenol	5.416	112	281455	49.378	ng	0.00
7) Phenol-d6	6.428	99	412668	56.658	ng	-0.01
23) Nitrobenzene-d5	7.351	82	243923	41.041	ng	-0.01
42) 2,4,6-Tribromophenol	10.698	330	42870	38.403	ng	0.07
45) 2-Fluorobiphenyl	9.151	172	469687	76.605	ng	0.00
79) Terphenyl-d14	13.063	244	237048	25.099	ng	0.14
Target Compounds						
71) Phenanthrene	11.451	178	183907m	41.767	ng	Qvalue
81) Benzo(a)anthracene	14.116	228	98485m	11.112	ng	

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF082423\
Data File : BF134921.D
Acq On : 24 Aug 2023 13:32
Operator : CG\JU
Sample : 04099-01
Misc :
ALS Vial : 8 Sample Multiplier: 1

Instrument :
BNA_F
ClientSampleId :
CONC-WC-1

Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 08/25/2023
Supervised By :mohammad ahmed 08/25/2023

Quant Time: Aug 24 15:56:31 2023
Quant Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF082223.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Tue Aug 22 19:33:25 2023
Response via : Initial Calibration

