

Data Path : Z:\SVOASRV\HPCHEM1\BNA F\DATA\BF082519\
 Data File : BF116527.D
 Acq On : 25 Aug 2019 13:05
 Operator : HP/JU
 Sample : K4475-13
 Misc :
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 BNA_F
 ClientSampleId :
 T1-2-B

Manual Integrations
 APPROVED

mohammad
 8/29/2019 7:53:00 AM

Quant Time: Aug 26 06:47:43 2019
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA F\METHODS\8270-BF082319.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Fri Aug 23 13:38:31 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	6.88	152	160825	20.00	ng	0.00
21) Naphthalene-d8	8.16	136	638838	20.00	ng	0.00
39) Acenaphthene-d10	9.92	164	330908	20.00	ng	0.00
64) Phenanthrene-d10	11.40	188	508116	20.00	ng	0.00
76) Chrysene-d12	14.03	240	293141	20.00	ng	0.00
87) Perylene-d12	15.50	264	265687	20.00	ng	-0.01
System Monitoring Compounds						
5) 2-Fluorophenol	5.51	112	1021530	92.81	ng	0.01
7) Phenol-d6	6.51	99	1398015	103.59	ng	0.00
23) Nitrobenzene-d5	7.44	82	897024	76.82	ng	0.00
42) 2,4,6-Tribromophenol	10.70	330	314356	110.63	ng	0.00
45) 2-Fluorobiphenyl	9.24	172	1553464	75.60	ng	0.00
79) Terphenyl-d14	12.99	244	1320758	84.16	ng	0.00
Target Compounds						
50) Dimethylphthalate	9.62	163	123860	5.365	ng	99
75) Fluoranthene	12.60	202	67928	2.374	ng	93
78) Pyrene	12.83	202	72423	3.089	ng	97
84) Bis(2-ethylhexyl)phthalate	14.02	149	73538	5.447	ng	98
88) Benzo(b)fluoranthene	15.06	252	41516m	2.529	ng	

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\SVOASRV\HPCHEM1\BNA F\DATA\BF082519\
Data File : BF116527.D
Acq On : 25 Aug 2019 13:05
Operator : HP/JU
Sample : K4475-13
Misc :
ALS Vial : 3 Sample Multiplier: 1

Instrument :
BNA_F
ClientSampleId :
T1-2-B

Manual Integrations
APPROVED

mohammad
8/29/2019 7:53:00 AM

Quant Time: Aug 26 06:47:43 2019
Quant Method : Z:\SVOASRV\HPCHEM1\BNA F\METHODS\8270-BF082319.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Fri Aug 23 13:38:31 2019
Response via : Initial Calibration

