

Data Path : Z:\SVOASRV\HPCHEM1\BNA F\DATA\BF082718\
 Data File : BF108510.D
 Acq On : 27 Aug 2018 15:08
 Operator : JU/SJ
 Sample : J4646-10
 Misc :
 ALS Vial : 14 Sample Multiplier: 1

Instrument :
 BNA_F
 ClientSampleId :
 EP1-Q4-C6

Manual Integrations
 APPROVED

Sohil
 8/27/2018 5:09:54 PM

Quant Time: Aug 27 16:18:17 2018
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA F\METHODS\8270-BF080818.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Wed Aug 22 11:01:45 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	7.00	152	135246	20.00	ng	0.00
21) Naphthalene-d8	8.28	136	519263	20.00	ng	0.00
38) Acenaphthene-d10	10.04	164	253900	20.00	ng	-0.01
63) Phenanthrene-d10	11.54	188	443251	20.00	ng	-0.01
75) Chrysene-d12	14.20	240	305170	20.00	ng	0.00
86) Perylene-d12	15.75	264	291342	20.00	ng	-0.01
System Monitoring Compounds						
5) 2-Fluorophenol	5.63	112	693887	92.89	ng	0.00
7) Phenol-d6	6.62	99	943483	95.04	ng	0.00
23) Nitrobenzene-d5	7.56	82	627214	67.40	ng	-0.01
41) 2,4,6-Tribromophenol	10.84	330	251870	99.45	ng	-0.01
44) 2-Fluorobiphenyl	9.35	172	1187262	75.07	ng	-0.01
78) Terphenyl-d14	13.12	244	918859	76.56	ng	0.00
Target Compounds						
49) Dimethylphthalate	9.75	163	147458	8.337	ng	Qvalue # 99
70) Phenanthrene	11.56	178	47719	2.282	ng	98
74) Fluoranthene	12.76	202	78580	3.444	ng	93
77) Pyrene	12.99	202	71602	3.706	ng	100
87) Benzo(b)fluoranthene	15.28	252	43967m	2.526	ng	

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\SVOASRV\HPCHEM1\BNA F\DATA\BF082718\
Data File : BF108510.D
Acq On : 27 Aug 2018 15:08
Operator : JU/SJ
Sample : J4646-10
Misc :
ALS Vial : 14 Sample Multiplier: 1

Instrument :
BNA_F
ClientSampleId :
EP1-Q4-C6

Manual Integrations
APPROVED

Sohil
8/27/2018 5:09:54 PM

Quant Time: Aug 27 16:18:17 2018
Quant Method : Z:\SVOASRV\HPCHEM1\BNA F\METHODS\8270-BF080818.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Wed Aug 22 11:01:45 2018
Response via : Initial Calibration

