

Data Path : Z:\SVOASRV\HPCHEM1\BNA F\DATA\BF082719\
 Data File : BF116583.D
 Acq On : 27 Aug 2019 14:19
 Operator : HP/JU
 Sample : K4490-03RE 2X
 Misc :
 ALS Vial : 7 Sample Multiplier: 1

Instrument :
 BNA_F
 ClientSampleId :
 STUCCORE

Quant Time: Aug 27 18:00:46 2019
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA F\METHODS\8270-BF082319.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Fri Aug 23 13:38:31 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	6.87	152	43401	20.00	ng	0.00
21) Naphthalene-d8	8.16	136	172155	20.00	ng	0.00
39) Acenaphthene-d10	9.90	164	88391	20.00	ng	-0.01
64) Phenanthrene-d10	11.39	188	139331	20.00	ng	-0.01
76) Chrysene-d12	14.02	240	81479	20.00	ng	-0.01
87) Perylene-d12	15.49	264	91560	20.00	ng	-0.02
System Monitoring Compounds						
5) 2-Fluorophenol	5.48	112	62738	21.12	ng	-0.02
7) Phenol-d6	6.49	99	106379	29.21	ng	-0.01
23) Nitrobenzene-d5	7.43	82	69566	22.11	ng	-0.01
42) 2,4,6-Tribromophenol	10.69	330	18498	24.37	ng	-0.01
45) 2-Fluorobiphenyl	9.23	172	136298	24.83	ng	-0.01
79) Terphenyl-d14	12.97	244	110112	25.24	ng	-0.01
Target Compounds						
50) Dimethylphthalate	9.62	163	14173	2.298	ng	Qvalue # 86

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\SVOASRV\HPCHEM1\BNA F\DATA\BF082719\
Data File : BF116583.D
Acq On : 27 Aug 2019 14:19
Operator : HP/JU
Sample : K4490-03RE 2X
Misc :
ALS Vial : 7 Sample Multiplier: 1

Instrument :
BNA_F
ClientSampleId :
STUCCORE

Quant Time: Aug 27 18:00:46 2019
Quant Method : Z:\SVOASRV\HPCHEM1\BNA F\METHODS\8270-BF082319.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Fri Aug 23 13:38:31 2019
Response via : Initial Calibration

