

Data Path : Z:\SVOASRV\HPCHEM1\BNA F\DATA\BF082818\  
 Data File : BF108573.D  
 Acq On : 28 Aug 2018 23:17  
 Operator : JU/SJ  
 Sample : J4648-09  
 Misc :  
 ALS Vial : 23 Sample Multiplier: 1

Instrument :  
 BNA\_F  
 ClientSampleId :  
 TMW-1

Quant Time: Aug 29 04:34:26 2018  
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA F\METHODS\8270-BF080818.M  
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION  
 QLast Update : Wed Aug 22 11:01:45 2018  
 Response via : Initial Calibration

| Internal Standards          | R.T.  | QIon | Response | Conc   | Units | Dev(Min) |
|-----------------------------|-------|------|----------|--------|-------|----------|
| 1) 1,4-Dichlorobenzene-d4   | 7.00  | 152  | 105997   | 20.00  | ng    | 0.00     |
| 21) Naphthalene-d8          | 8.28  | 136  | 391613   | 20.00  | ng    | 0.00     |
| 38) Acenaphthene-d10        | 10.04 | 164  | 167680   | 20.00  | ng    | -0.01    |
| 63) Phenanthrene-d10        | 11.54 | 188  | 289716   | 20.00  | ng    | 0.00     |
| 75) Chrysene-d12            | 14.20 | 240  | 304693   | 20.00  | ng    | 0.00     |
| 86) Perylene-d12            | 15.75 | 264  | 243938   | 20.00  | ng    | -0.01    |
| System Monitoring Compounds |       |      |          |        |       |          |
| 5) 2-Fluorophenol           | 5.65  | 112  | 57346    | 9.79   | ng    | 0.02     |
| 7) Phenol-d6                | 6.64  | 99   | 67606    | 8.69   | ng    | 0.01     |
| 23) Nitrobenzene-d5         | 7.56  | 82   | 127588   | 18.18  | ng    | -0.01    |
| 41) 2,4,6-Tribromophenol    | 10.84 | 330  | 17348    | 10.37  | ng    | -0.01    |
| 44) 2-Fluorobiphenyl        | 9.35  | 172  | 230660   | 22.08  | ng    | -0.01    |
| 78) Terphenyl-d14           | 13.12 | 244  | 243866   | 20.35  | ng    | -0.01    |
| Target Compounds            |       |      |          |        |       | Qvalue   |
| 49) Dimethylphthalate       | 9.75  | 163  | 177993   | 15.239 | ng    | 99       |

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\SVOASRV\HPCHEM1\BNA F\DATA\BF082818\  
Data File : BF108573.D  
Acq On : 28 Aug 2018 23:17  
Operator : JU/SJ  
Sample : J4648-09  
Misc :  
ALS Vial : 23 Sample Multiplier: 1

Instrument :  
BNA\_F  
ClientSampleId :  
TMW-1

Quant Time: Aug 29 04:34:26 2018  
Quant Method : Z:\SVOASRV\HPCHEM1\BNA F\METHODS\8270-BF080818.M  
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION  
QLast Update : Wed Aug 22 11:01:45 2018  
Response via : Initial Calibration

