

Data Path : Z:\svoasrv\HPCHEM1\BNA F\Data\BF082818\
 Data File : BF108574.D
 Acq On : 28 Aug 2018 23:45
 Operator : JU/SJ
 Sample : SSTDCCC040
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_F
 LabSampleId :
 SSTDCCC040

Quant Time: Aug 29 00:50:39 2018
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA F\METHODS\8270-BF080818.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Wed Aug 22 11:01:45 2018
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvqRF	CCRF	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	41#	0.00
2	1,4-Dioxane	0.568	0.512	9.9	37#	-0.02
3	Pyridine	1.462	1.231	15.8	34#	-0.01
4	n-Nitrosodimethylamine	0.823	0.681	17.3	33#	-0.02
5 S	2-Fluorophenol	1.105	1.175	-6.3	43#	0.00
6	Aniline	1.997	1.960	1.9	41#	0.00
7 S	Phenol-d6	1.468	1.532	-4.4	43#	0.00
8	2-Chlorophenol	1.244	1.313	-5.5	43#	0.00
9	Benzaldehyde	1.138	1.082	4.9	38#	0.00
10 C	Phenol	1.518	1.650	-8.7	45#	0.00
11	bis(2-Chloroethyl)ether	1.228	1.199	2.4	40#	0.00
12	1,3-Dichlorobenzene	1.439	1.494	-3.8	42#	0.00
13 C	1,4-Dichlorobenzene	1.445	1.562	-8.1	44#	0.00
14	1,2-Dichlorobenzene	1.344	1.433	-6.6	44#	0.00
15	Benzyl Alcohol	1.236	1.212	1.9	39#	0.00
16	2,2'-oxybis(1-Chloropropane	2.529	2.291	9.4	37#	0.00
17	2-Methylphenol	1.067	1.093	-2.4	41#	0.00
18	Hexachloroethane	0.515	0.508	1.4	39#	-0.01
19 P	n-Nitroso-di-n-propylamine	0.993	0.981	1.2	40#	0.00
20	3+4-Methylphenols	1.367	1.421	-4.0	42#	0.00
21 I	Naphthalene-d8	1.000	1.000	0.0	41#	0.00
22	Acetophenone	0.468	0.483	-3.2	42#	0.00
23 S	Nitrobenzene-d5	0.358	0.376	-5.0	42#	0.00
24	Nitrobenzene	0.362	0.367	-1.4	40#	0.00
25	Isophorone	0.683	0.656	4.0	39#	0.00
26 C	2-Nitrophenol	0.137	0.161	-17.5	47#	0.00
27	2,4-Dimethylphenol	0.303	0.290	4.3	39#	0.00
28	bis(2-Chloroethoxy)methane	0.399	0.405	-1.5	41#	0.00
29 C	2,4-Dichlorophenol	0.279	0.290	-3.9	41#	0.00
30	1,2,4-Trichlorobenzene	0.308	0.315	-2.3	42#	0.00
31	Naphthalene	0.910	0.947	-4.1	43#	0.00
32	Benzoic acid	0.161	0.123	23.6	30#	0.01
33	4-Chloroaniline	0.396	0.403	-1.8	41#	0.00
34 C	Hexachlorobutadiene	0.195	0.202	-3.6	42#	0.00
35	Caprolactam	0.087	0.080	8.0	36#	0.00
36 C	4-Chloro-3-methylphenol	0.301	0.294	2.3	39#	0.00
37	2-Methylnaphthalene	0.644	0.648	-0.6	41#	0.00
38 I	Acenaphthene-d10	1.000	1.000	0.0	35#	0.00
39	1,2,4,5-Tetrachlorobenzene	0.614	0.716	-16.6	41#	0.00
40 P	Hexachlorocyclopentadiene	0.397	0.085	78.6#	7#	0.00
41 S	2,4,6-Tribromophenol	0.199	0.195	2.0	32#	0.00
42 C	2,4,6-Trichlorophenol	0.381	0.439	-15.2	39#	0.00
43	2,4,5-Trichlorophenol	0.420	0.456	-8.6	36#	0.00
44 S	2-Fluorobiphenyl	1.246	1.520	-22.0	44#	0.00

Data Path : Z:\svoasrv\HPCHEM1\BNA F\Data\BF082818\
 Data File : BF108574.D
 Acq On : 28 Aug 2018 23:45
 Operator : JU/SJ
 Sample : SSTDCCC040
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_F
 LabSampleId :
 SSTDCCC040

Quant Time: Aug 29 00:50:39 2018
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA F\METHODS\8270-BF080818.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Wed Aug 22 11:01:45 2018
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvqRF	CCRF	%Dev	Area%	Dev(min)
45	1,1'-Biphenyl	1.475	1.706	-15.7	41#	0.00
46	2-Chloronaphthalene	1.165	1.287	-10.5	39#	0.00
47	2-Nitroaniline	0.377	0.420	-11.4	37#	0.00
48	Acenaphthylene	1.843	1.999	-8.5	38#	0.00
49	Dimethylphthalate	1.393	1.552	-11.4	39#	0.00
50	2,6-Dinitrotoluene	0.291	0.321	-10.3	37#	0.00
51 C	Acenaphthene	1.129	1.140	-1.0	36#	0.00
52	3-Nitroaniline	0.319	0.326	-2.2	35#	0.00
53 P	2,4-Dinitrophenol	0.092	0.014#	84.8#	5#	0.00
54	Dibenzofuran	1.635	1.697	-3.8	37#	0.00
55 P	4-Nitrophenol	0.241	0.184	23.7	26#	0.02
56	2,4-Dinitrotoluene	0.375	0.385	-2.7	34#	0.00
57	Fluorene	1.294	1.323	-2.2	36#	0.00
58	2,3,4,6-Tetrachlorophenol	0.351	0.337	4.0	32#	0.00
59	Diethylphthalate	1.349	1.438	-6.6	37#	0.00
60	4-Chlorophenyl-phenylether	0.674	0.694	-3.0	36#	0.00
61	4-Nitroaniline	0.315	0.299	5.1	32#	0.00
62	Azobenzene	1.393	1.337	4.0	34#	0.00
63 I	Phenanthrene-d10	1.000	1.000	0.0	28#	0.00
64	4,6-Dinitro-2-methylphenol	0.069	0.022	68.1#	9#	0.00
65 c	n-Nitrosodiphenylamine	0.591	0.717	-21.3#	34#	0.00
66	4-Bromophenyl-phenylether	0.217	0.255	-17.5	33#	0.00
67	Hexachlorobenzene	0.229	0.252	-10.0	31#	0.00
68	Atrazine	0.198	0.221	-11.6	31#	0.00
69 C	Pentachlorophenol	0.109	0.099	9.2	25#	0.00
70	Phenanthrene	0.943	1.018	-8.0	31#	0.00
71	Anthracene	0.967	1.042	-7.8	31#	0.00
72	Carbazole	0.859	0.873	-1.6	29#	0.00
73	Di-n-butylphthalate	0.973	1.124	-15.5	33#	0.00
74 C	Fluoranthene	1.030	1.065	-3.4	30#	0.00
75 I	Chrysene-d12	1.000	1.000	0.0	32#	0.00
76	Benzidine	0.747	0.576	22.9	23#	0.00
77	Pyrene	1.266	1.190	6.0	30#	0.00
78 S	Terphenyl-d14	0.787	0.770	2.2	32#	0.00
79	Butylbenzylphthalate	0.503	0.509	-1.2	30#	0.00
80	Benzo(a)anthracene	1.148	1.192	-3.8	33#	0.00
81	3,3'-Dichlorobenzidine	0.411	0.450	-9.5	33#	0.00
82	Chrysene	1.052	1.102	-4.8	33#	0.00
83	Bis(2-ethylhexyl)phthalate	0.681	0.684	-0.4	30#	0.00
84 c	Di-n-octyl phthalate	1.038	1.274	-22.7#	36#	-0.01
85	Indeno(1,2,3-cd)pyrene	0.912	0.791	13.3	28#	0.00
86 I	Perylene-d12	1.000	1.000	0.0	31#	0.00
87	Benzo(b)fluoranthene	1.195	1.268	-6.1	31#	0.00

Data Path : Z:\svoasrv\HPCHEM1\BNA F\Data\BF082818\
Data File : BF108574.D
Acq On : 28 Aug 2018 23:45
Operator : JU/SJ
Sample : SSTDCCC040
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
BNA_F
LabSampleId :
SSTDCCC040

Quant Time: Aug 29 00:50:39 2018
Quant Method : Z:\SVOASRV\HPCHEM1\BNA F\METHODS\8270-BF080818.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Wed Aug 22 11:01:45 2018
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
88	Benzo(k)fluoranthene	1.099	1.220	-11.0	35#	0.00
89 C	Benzo(a)pyrene	1.070	1.119	-4.6	31#	0.00
90	Dibenzo(a,h)anthracene	0.885	0.821	7.2	28#	0.00
91	Benzo(a,h,i)perylene	0.872	0.782	10.3	28#	0.00

(#) = Out of Range

SPCC's out = 1 CCC's out = 2