

Data Path : Z:\SVOASRV\HPCHEM1\BNA F\DATA\BF083018\  
 Data File : BF108678.D  
 Acq On : 31 Aug 2018 9:01  
 Operator : JU/SJ  
 Sample : J4700-16  
 Misc :  
 ALS Vial : 37 Sample Multiplier: 1

Instrument :  
 BNA\_F  
 ClientSampleId :  
 082818-JETT-F-U-S

Manual Integrations  
 APPROVED

Sohil  
 8/31/2018 12:19:14 PM

Quant Time: Aug 31 10:49:31 2018  
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA F\METHODS\8270-BF083018.M  
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION  
 QLast Update : Thu Aug 30 15:06:55 2018  
 Response via : Initial Calibration

| Internal Standards          | R.T.  | QIon | Response | Conc   | Units | Dev(Min)       |
|-----------------------------|-------|------|----------|--------|-------|----------------|
| 1) 1,4-Dichlorobenzene-d4   | 7.01  | 152  | 82105    | 20.00  | ng    | 0.00           |
| 21) Naphthalene-d8          | 8.28  | 136  | 339877   | 20.00  | ng    | 0.00           |
| 38) Acenaphthene-d10        | 10.04 | 164  | 154363   | 20.00  | ng    | 0.00           |
| 63) Phenanthrene-d10        | 11.54 | 188  | 287549   | 20.00  | ng    | 0.00           |
| 75) Chrysene-d12            | 14.19 | 240  | 297323   | 20.00  | ng    | -0.01          |
| 86) Perylene-d12            | 15.74 | 264  | 247314   | 20.00  | ng    | -0.01          |
| System Monitoring Compounds |       |      |          |        |       |                |
| 5) 2-Fluorophenol           | 5.67  | 112  | 208750m  | 44.17  | ng    | 0.02           |
| 7) Phenol-d6                | 6.65  | 99   | 185263m  | 27.37  | ng    | 0.00           |
| 23) Nitrobenzene-d5         | 7.57  | 82   | 686325   | 102.50 | ng    | 0.00           |
| 41) 2,4,6-Tribromophenol    | 10.84 | 330  | 158304   | 77.68  | ng    | -0.01          |
| 44) 2-Fluorobiphenyl        | 9.35  | 172  | 1310244  | 114.13 | ng    | 0.00           |
| 78) Terphenyl-d14           | 13.12 | 244  | 1184879  | 77.43  | ng    | -0.01          |
| Target Compounds            |       |      |          |        |       |                |
| 49) Dimethylphthalate       | 9.75  | 163  | 86351    | 7.104  | ng    | Qvalue<br># 98 |

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\SVOASRV\HPCHEM1\BNA F\DATA\BF083018\  
Data File : BF108678.D  
Acq On : 31 Aug 2018 9:01  
Operator : JU/SJ  
Sample : J4700-16  
Misc :  
ALS Vial : 37 Sample Multiplier: 1

Instrument :  
BNA\_F  
ClientSampleId :  
082818-JETT-F-U-S

Manual Integrations  
APPROVED

Sohil  
8/31/2018 12:19:14 PM

Quant Time: Aug 31 10:49:31 2018  
Quant Method : Z:\SVOASRV\HPCHEM1\BNA F\METHODS\8270-BF083018.M  
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION  
QLast Update : Thu Aug 30 15:06:55 2018  
Response via : Initial Calibration

