

Data Path : Z:\SVOASRV\HPCHEM1\BNA F\DATA\BF083018\  
 Data File : BF108641.D  
 Acq On : 30 Aug 2018 15:31  
 Operator : JU/SJ  
 Sample : SSTDCCC040  
 Misc :  
 ALS Vial : 2 Sample Multiplier: 1

Instrument :  
 BNA\_F  
 LabSampleId :  
 SSTDCCC040

Quant Time: Aug 30 16:33:07 2018  
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA F\METHODS\8270-BF083018.M  
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION  
 QLast Update : Thu Aug 30 15:06:55 2018  
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min  
 Max. RRF Dev : 25% Max. Rel. Area : 150%

|      | Compound                     | Amount | Calc.  | %Dev | Area% | Dev(min) |
|------|------------------------------|--------|--------|------|-------|----------|
| 1 I  | 1,4-Dichlorobenzene-d4       | 20.000 | 20.000 | 0.0  | 102   | 0.00     |
| 2    | 1,4-Dioxane                  | 40.000 | 36.336 | 9.2  | 95    | 0.00     |
| 3    | Pyridine                     | 40.000 | 36.568 | 8.6  | 94    | 0.00     |
| 4    | n-Nitrosodimethylamine       | 40.000 | 33.711 | 15.7 | 84    | 0.00     |
| 5 S  | 2-Fluorophenol               | 80.000 | 71.419 | 10.7 | 91    | 0.00     |
| 6    | Aniline                      | 40.000 | 38.978 | 2.6  | 96    | 0.00     |
| 7 S  | Phenol-d6                    | 80.000 | 75.999 | 5.0  | 97    | 0.00     |
| 8    | 2-Chlorophenol               | 40.000 | 38.995 | 2.5  | 97    | 0.00     |
| 9    | Benzaldehyde                 | 40.000 | 38.990 | 2.5  | 99    | 0.00     |
| 10 C | Phenol                       | 40.000 | 38.936 | 2.7  | 98    | 0.00     |
| 11   | bis(2-Chloroethyl)ether      | 40.000 | 37.675 | 5.8  | 95    | 0.00     |
| 12   | 1,3-Dichlorobenzene          | 40.000 | 38.213 | 4.5  | 96    | 0.00     |
| 13 C | 1,4-Dichlorobenzene          | 40.000 | 38.660 | 3.4  | 97    | 0.00     |
| 14   | 1,2-Dichlorobenzene          | 40.000 | 38.579 | 3.6  | 100   | 0.00     |
| 15   | Benzyl Alcohol               | 40.000 | 38.100 | 4.7  | 94    | 0.00     |
| 16   | 2,2'-oxybis(1-Chloropropane) | 40.000 | 38.029 | 4.9  | 96    | 0.00     |
| 17   | 2-Methylphenol               | 40.000 | 40.443 | -1.1 | 97    | 0.00     |
| 18   | Hexachloroethane             | 40.000 | 38.143 | 4.6  | 97    | 0.00     |
| 19 P | n-Nitroso-di-n-propylamine   | 40.000 | 39.048 | 2.4  | 96    | 0.00     |
| 20   | 3+4-Methylphenols            | 40.000 | 42.227 | -5.6 | 99    | 0.00     |
| 21 I | Naphthalene-d8               | 20.000 | 20.000 | 0.0  | 102   | 0.00     |
| 22   | Acetophenone                 | 40.000 | 37.841 | 5.4  | 97    | 0.00     |
| 23 S | Nitrobenzene-d5              | 80.000 | 76.481 | 4.4  | 98    | 0.00     |
| 24   | Nitrobenzene                 | 40.000 | 39.054 | 2.4  | 98    | 0.00     |
| 25   | Isophorone                   | 40.000 | 38.140 | 4.6  | 98    | 0.00     |
| 26 C | 2-Nitrophenol                | 40.000 | 39.776 | 0.6  | 99    | 0.00     |
| 27   | 2,4-Dimethylphenol           | 40.000 | 36.931 | 7.7  | 96    | 0.00     |
| 28   | bis(2-Chloroethoxy)methane   | 40.000 | 38.354 | 4.1  | 98    | 0.00     |
| 29 C | 2,4-Dichlorophenol           | 40.000 | 38.552 | 3.6  | 98    | 0.00     |
| 30   | 1,2,4-Trichlorobenzene       | 40.000 | 37.742 | 5.6  | 97    | 0.00     |
| 31   | Naphthalene                  | 40.000 | 41.065 | -2.7 | 105   | 0.00     |
| 32   | Benzoic acid                 | 40.000 | 36.175 | 9.6  | 92    | 0.00     |
| 33   | 4-Chloroaniline              | 40.000 | 38.977 | 2.6  | 100   | 0.00     |
| 34 C | Hexachlorobutadiene          | 40.000 | 37.990 | 5.0  | 97    | 0.00     |
| 35   | Caprolactam                  | 40.000 | 39.197 | 2.0  | 98    | 0.00     |
| 36 C | 4-Chloro-3-methylphenol      | 40.000 | 38.432 | 3.9  | 97    | 0.00     |
| 37   | 2-Methylnaphthalene          | 40.000 | 38.234 | 4.4  | 98    | 0.00     |
| 38 I | Acenaphthene-d10             | 20.000 | 20.000 | 0.0  | 104   | 0.00     |
| 39   | 1,2,4,5-Tetrachlorobenzene   | 40.000 | 37.916 | 5.2  | 98    | 0.00     |
| 40 P | Hexachlorocyclopentadiene    | 40.000 | 36.750 | 8.1  | 97    | 0.00     |
| 41 S | 2,4,6-Tribromophenol         | 80.000 | 75.419 | 5.7  | 98    | 0.00     |
| 42 C | 2,4,6-Trichlorophenol        | 40.000 | 38.750 | 3.1  | 97    | 0.00     |
| 43   | 2,4,5-Trichlorophenol        | 40.000 | 38.572 | 3.6  | 101   | 0.00     |
| 44 S | 2-Fluorobiphenyl             | 80.000 | 74.718 | 6.6  | 99    | 0.00     |

Data Path : Z:\SVOASRV\HPCHEM1\BNA F\DATA\BF083018\  
 Data File : BF108641.D  
 Acq On : 30 Aug 2018 15:31  
 Operator : JU/SJ  
 Sample : SSTDCCC040  
 Misc :  
 ALS Vial : 2 Sample Multiplier: 1

Instrument :  
 BNA\_F  
 LabSampleId :  
 SSTDCCC040

Quant Time: Aug 30 16:33:07 2018  
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA F\METHODS\8270-BF083018.M  
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION  
 QLast Update : Thu Aug 30 15:06:55 2018  
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min  
 Max. RRF Dev : 25% Max. Rel. Area : 150%

|      | Compound                   | Amount | Calc.  | %Dev  | Area% | Dev(min) |
|------|----------------------------|--------|--------|-------|-------|----------|
| 45   | 1,1'-Biphenyl              | 40.000 | 37.589 | 6.0   | 99    | 0.00     |
| 46   | 2-Chloronaphthalene        | 40.000 | 37.798 | 5.5   | 99    | 0.00     |
| 47   | 2-Nitroaniline             | 40.000 | 38.082 | 4.8   | 98    | 0.00     |
| 48   | Acenaphthylene             | 40.000 | 37.579 | 6.1   | 98    | 0.00     |
| 49   | Dimethylphthalate          | 40.000 | 37.673 | 5.8   | 99    | 0.00     |
| 50   | 2,6-Dinitrotoluene         | 40.000 | 38.075 | 4.8   | 98    | 0.00     |
| 51 C | Acenaphthene               | 40.000 | 37.109 | 7.2   | 100   | 0.00     |
| 52   | 3-Nitroaniline             | 40.000 | 38.252 | 4.4   | 99    | 0.00     |
| 53 P | 2,4-Dinitrophenol          | 40.000 | 38.449 | 3.9   | 97    | 0.00     |
| 54   | Dibenzofuran               | 40.000 | 37.447 | 6.4   | 98    | 0.00     |
| 55 P | 4-Nitrophenol              | 40.000 | 33.807 | 15.5  | 81    | 0.00     |
| 56   | 2,4-Dinitrotoluene         | 40.000 | 38.165 | 4.6   | 99    | 0.00     |
| 57   | Fluorene                   | 40.000 | 37.285 | 6.8   | 98    | 0.00     |
| 58   | 2,3,4,6-Tetrachlorophenol  | 40.000 | 36.862 | 7.8   | 95    | 0.00     |
| 59   | Diethylphthalate           | 40.000 | 37.546 | 6.1   | 99    | 0.00     |
| 60   | 4-Chlorophenyl-phenylether | 40.000 | 37.450 | 6.4   | 99    | 0.00     |
| 61   | 4-Nitroaniline             | 40.000 | 38.570 | 3.6   | 101   | 0.00     |
| 62   | Azobenzene                 | 40.000 | 37.177 | 7.1   | 99    | 0.00     |
| 63 I | Phenanthrene-d10           | 20.000 | 20.000 | 0.0   | 104   | 0.00     |
| 64   | 4,6-Dinitro-2-methylphenol | 40.000 | 39.169 | 2.1   | 100   | 0.00     |
| 65 c | n-Nitrosodiphenylamine     | 40.000 | 37.380 | 6.5   | 99    | 0.00     |
| 66   | 4-Bromophenyl-phenylether  | 40.000 | 37.841 | 5.4   | 99    | 0.00     |
| 67   | Hexachlorobenzene          | 40.000 | 38.088 | 4.8   | 101   | 0.00     |
| 68   | Atrazine                   | 40.000 | 36.862 | 7.8   | 99    | 0.00     |
| 69 C | Pentachlorophenol          | 40.000 | 40.374 | -0.9  | 98    | 0.00     |
| 70   | Phenanthrene               | 40.000 | 37.731 | 5.7   | 99    | 0.00     |
| 71   | Anthracene                 | 40.000 | 37.805 | 5.5   | 100   | 0.00     |
| 72   | Carbazole                  | 40.000 | 38.127 | 4.7   | 101   | 0.00     |
| 73   | Di-n-butylphthalate        | 40.000 | 37.400 | 6.5   | 99    | 0.00     |
| 74 C | Fluoranthene               | 40.000 | 37.896 | 5.3   | 101   | 0.00     |
| 75 I | Chrysene-d12               | 20.000 | 20.000 | 0.0   | 105   | 0.00     |
| 76   | Benzidine                  | 40.000 | 45.717 | -14.3 | 117   | 0.00     |
| 77   | Pyrene                     | 40.000 | 37.888 | 5.3   | 100   | 0.00     |
| 78 S | Terphenyl-d14              | 80.000 | 75.507 | 5.6   | 101   | 0.00     |
| 79   | Butylbenzylphthalate       | 40.000 | 38.636 | 3.4   | 102   | 0.00     |
| 80   | Benzo(a)anthracene         | 40.000 | 37.036 | 7.4   | 100   | 0.00     |
| 81   | 3,3'-Dichlorobenzidine     | 40.000 | 37.913 | 5.2   | 99    | 0.00     |
| 82   | Chrysene                   | 40.000 | 37.890 | 5.3   | 101   | 0.00     |
| 83   | Bis(2-ethylhexyl)phthalate | 40.000 | 38.210 | 4.5   | 101   | 0.00     |
| 84 c | Di-n-octyl phthalate       | 40.000 | 38.450 | 3.9   | 103   | 0.00     |
| 85   | Indeno(1,2,3-cd)pyrene     | 40.000 | 36.416 | 9.0   | 100   | 0.00     |
| 86 I | Perylene-d12               | 20.000 | 20.000 | 0.0   | 108   | 0.00     |
| 87   | Benzo(b)fluoranthene       | 40.000 | 36.551 | 8.6   | 94    | 0.00     |

Data Path : Z:\SVOASRV\HPCHEM1\BNA F\DATA\BF083018\  
Data File : BF108641.D  
Acq On : 30 Aug 2018 15:31  
Operator : JU/SJ  
Sample : SSTDCCC040  
Misc :  
ALS Vial : 2 Sample Multiplier: 1

Instrument :  
BNA\_F  
LabSampleId :  
SSTDCCC040

Quant Time: Aug 30 16:33:07 2018  
Quant Method : Z:\SVOASRV\HPCHEM1\BNA F\METHODS\8270-BF083018.M  
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION  
QLast Update : Thu Aug 30 15:06:55 2018  
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min  
Max. RRF Dev : 25% Max. Rel. Area : 150%

|      | Compound               | Amount | Calc.  | %Dev | Area% | Dev(min) |
|------|------------------------|--------|--------|------|-------|----------|
| 88   | Benzo(k)fluoranthene   | 40.000 | 39.839 | 0.4  | 111   | 0.00     |
| 89 C | Benzo(a)pyrene         | 40.000 | 37.680 | 5.8  | 102   | 0.00     |
| 90   | Dibenzo(a,h)anthracene | 40.000 | 38.014 | 5.0  | 100   | 0.00     |
| 91   | Benzo(a,h,i)perylene   | 40.000 | 38.326 | 4.2  | 99    | 0.00     |

(#) = Out of Range

SPCC's out = 0 CCC's out = 0