

Data Path : Z:\SVOASRV\HPCHEM1\BNA F\DATA\BF083019\
 Data File : BF116663.D
 Acq On : 30 Aug 2019 21:20
 Operator : HP/JU
 Sample : SSTDCCC040
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_F
 LabSampleId :
 SSTDCCC040

Quant Time: Aug 31 03:39:24 2019
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA F\METHODS\8270-BF083019.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Fri Aug 30 20:02:30 2019
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvqRF	CCRF	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	97	0.00
2	1,4-Dioxane	0.570	0.576	-1.1	97	0.00
3	Pyridine	1.650	1.681	-1.9	97	0.00
4	n-Nitrosodimethylamine	0.567	0.572	-0.9	97	0.00
5 S	2-Fluorophenol	1.235	1.253	-1.5	98	0.00
6	Aniline	2.245	2.318	-3.3	97	0.00
7 S	Phenol-d6	1.621	1.652	-1.9	97	0.00
8	2-Chlorophenol	1.348	1.372	-1.8	97	0.00
9	Benzaldehyde	1.068	1.052	1.5	100	0.00
10 C	Phenol	1.795	1.854	-3.3	98	0.00
11	bis(2-Chloroethyl)ether	1.398	1.419	-1.5	97	0.00
12	1,3-Dichlorobenzene	1.523	1.532	-0.6	97	0.00
13 C	1,4-Dichlorobenzene	1.516	1.550	-2.2	97	0.00
14	1,2-Dichlorobenzene	1.405	1.433	-2.0	97	0.00
15	Benzyl Alcohol	1.316	1.362	-3.5	98	0.00
16	2,2'-oxybis(1-Chloropropane	1.640	1.675	-2.1	96	0.00
17	2-Methylphenol	1.180	1.205	-2.1	99	0.00
18	Hexachloroethane	0.589	0.603	-2.4	97	0.00
19 P	n-Nitroso-di-n-propylamine	1.067	1.089	-2.1	98	0.00
20	3+4-Methylphenols	1.372	1.412	-2.9	97	0.00
21 I	Naphthalene-d8	1.000	1.000	0.0	97	0.00
22	Acetophenone	0.482	0.493	-2.3	99	0.00
23 S	Nitrobenzene-d5	0.350	0.370	-5.7	103	0.00
24	Nitrobenzene	0.356	0.378	-6.2	101	0.00
25	Isophorone	0.710	0.713	-0.4	98	0.00
26 C	2-Nitrophenol	0.132	0.145	-9.8	107	0.00
27	2,4-Dimethylphenol	0.268	0.276	-3.0	99	0.00
28	bis(2-Chloroethoxy)methane	0.434	0.442	-1.8	98	0.00
29 C	2,4-Dichlorophenol	0.257	0.268	-4.3	100	0.00
30	1,2,4-Trichlorobenzene	0.279	0.282	-1.1	97	0.00
31	Naphthalene	0.951	0.970	-2.0	97	0.00
32	Benzoic acid	0.151	0.155	-2.6	116	0.00
33	4-Chloroaniline	0.433	0.441	-1.8	99	0.00
34 C	Hexachlorobutadiene	0.152	0.156	-2.6	100	0.00
35	Caprolactam	0.096	0.096	0.0	98	0.00
36 C	4-Chloro-3-methylphenol	0.296	0.308	-4.1	100	0.00
37	2-Methylnaphthalene	0.633	0.650	-2.7	98	0.00
38	1-Methylnaphthalene	0.597	0.614	-2.8	99	0.00
39 I	Acenaphthene-d10	1.000	1.000	0.0	99	0.00
40	1,2,4,5-Tetrachlorobenzene	0.479	0.485	-1.3	99	0.00
41 P	Hexachlorocyclopentadiene	0.277	0.279	-0.7	98	0.00
42 S	2,4,6-Tribromophenol	0.134	0.142	-6.0	106	0.00
43 C	2,4,6-Trichlorophenol	0.329	0.334	-1.5	99	0.00
44	2,4,5-Trichlorophenol	0.341	0.354	-3.8	103	0.00

Data Path : Z:\SVOASRV\HPCHEM1\BNA F\DATA\BF083019\
 Data File : BF116663.D
 Acq On : 30 Aug 2019 21:20
 Operator : HP/JU
 Sample : SSTDCCC040
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_F
 LabSampleId :
 SSTDCCC040

Quant Time: Aug 31 03:39:24 2019
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA F\METHODS\8270-BF083019.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Fri Aug 30 20:02:30 2019
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvqRF	CCRF	%Dev	Area%	Dev(min)
45 S	2-Fluorobiphenyl	1.186	1.197	-0.9	100	0.00
46	1,1'-Biphenyl	1.518	1.536	-1.2	99	0.00
47	2-Chloronaphthalene	1.202	1.225	-1.9	98	0.00
48	2-Nitroaniline	0.348	0.383	-10.1	109	0.00
49	Acenaphthylene	1.817	1.851	-1.9	99	0.00
50	Dimethylphthalate	1.381	1.386	-0.4	99	0.00
51	2,6-Dinitrotoluene	0.263	0.291	-10.6	107	0.00
52 C	Acenaphthene	1.116	1.145	-2.6	101	0.00
53	3-Nitroaniline	0.330	0.356	-7.9	104	0.00
54 P	2,4-Dinitrophenol	0.077	0.088	-14.3	117	0.00
55	Dibenzofuran	1.574	1.608	-2.2	99	0.00
56 P	4-Nitrophenol	0.226	0.248	-9.7	107	0.00
57	2,4-Dinitrotoluene	0.322	0.368	-14.3	111	0.00
58	Fluorene	1.199	1.206	-0.6	100	0.00
59	2,3,4,6-Tetrachlorophenol	0.246	0.258	-4.9	104	0.00
60	Diethylphthalate	1.383	1.407	-1.7	99	0.00
61	4-Chlorophenyl-phenylether	0.525	0.537	-2.3	101	0.00
62	4-Nitroaniline	0.320	0.348	-8.7	108	0.00
63	Azobenzene	1.442	1.478	-2.5	101	0.00
64 I	Phenanthrene-d10	1.000	1.000	0.0	101	0.00
65	4,6-Dinitro-2-methylphenol	0.068	0.078	-14.7	119	0.00
66 c	n-Nitrosodiphenylamine	0.692	0.694	-0.3	100	0.00
67	4-Bromophenyl-phenylether	0.203	0.203	0.0	99	0.00
68	Hexachlorobenzene	0.213	0.215	-0.9	101	0.00
69	Atrazine	0.194	0.197	-1.5	103	0.00
70 C	Pentachlorophenol	0.103	0.112	-8.7	109	0.00
71	Phenanthrene	1.113	1.109	0.4	99	0.00
72	Anthracene	1.121	1.132	-1.0	101	0.00
73	Carbazole	1.068	1.062	0.6	100	0.00
74	Di-n-butylphthalate	1.369	1.377	-0.6	101	0.00
75 C	Fluoranthene	1.094	1.101	-0.6	102	0.00
76 I	Chrysene-d12	1.000	1.000	0.0	108	0.00
77	Benzidine	0.784	0.814	-3.8	114	0.00
78	Pyrene	1.778	1.792	-0.8	102	0.00
79 S	Terphenyl-d14	1.081	1.095	-1.3	104	0.00
80	Butylbenzylphthalate	0.865	0.892	-3.1	109	0.00
81	Benzo(a)anthracene	1.266	1.282	-1.3	108	0.00
82	3,3'-Dichlorobenzidine	0.428	0.440	-2.8	106	0.00
83	Chrysene	1.304	1.325	-1.6	108	0.00
84	Bis(2-ethylhexyl)phthalate	1.067	1.063	0.4	107	0.00
85 c	Di-n-octyl phthalate	1.747	1.780	-1.9	112	0.00
86	Indeno(1,2,3-cd)pyrene	1.047	0.959	8.4	98	0.00
87 I	Perylene-d12	1.000	1.000	0.0	105	0.00

Data Path : Z:\SVOASRV\HPCHEM1\BNA F\DATA\BF083019\
Data File : BF116663.D
Acq On : 30 Aug 2019 21:20
Operator : HP/JU
Sample : SSTDCCC040
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
BNA_F
LabSampleId :
SSTDCCC040

Quant Time: Aug 31 03:39:24 2019
Quant Method : Z:\SVOASRV\HPCHEM1\BNA F\METHODS\8270-BF083019.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Fri Aug 30 20:02:30 2019
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
88	Benzo(b)fluoranthene	1.209	1.190	1.6	107	0.00
89	Benzo(k)fluoranthene	1.102	1.176	-6.7	105	0.00
90 C	Benzo(a)pyrene	1.052	1.075	-2.2	107	0.00
91	Dibenzo(a,h)anthracene	0.921	0.904	1.8	100	0.00
92	Benzo(q,h,i)perylene	0.939	0.914	2.7	99	0.00

(#) = Out of Range

SPCC's out = 0 CCC's out = 0