

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF083023\
 Data File : BF135014.D
 Acq On : 30 Aug 2023 20:23
 Operator : CG\JU
 Sample : SSTDCCC040
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_F
 LabSampleId :
 SSTDCCC040

Quant Time: Aug 31 05:51:08 2023
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF083023.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Thu Aug 31 03:48:08 2023
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	20.000	20.000	0.0	90	0.00
2	1,4-Dioxane	40.000	37.595	6.0	91	0.00
3	Pyridine	40.000	38.192	4.5	91	0.00
4	n-Nitrosodimethylamine	40.000	38.417	4.0	93	0.00
5 S	2-Fluorophenol	80.000	74.157	7.3	93	0.00
6	Aniline	40.000	37.624	5.9	93	0.00
7 S	Phenol-d6	80.000	74.855	6.4	93	0.00
8	2-Chlorophenol	40.000	37.926	5.2	94	0.00
9	Benzaldehyde	40.000	35.038	12.4	91	0.00
10 C	Phenol	40.000	37.669	5.8	93	0.00
11	bis(2-Chloroethyl)ether	40.000	36.761	8.1	90	0.00
12	1,3-Dichlorobenzene	40.000	36.895	7.8	91	0.00
13 C	1,4-Dichlorobenzene	40.000	37.453	6.4	92	0.00
14	1,2-Dichlorobenzene	40.000	36.905	7.7	92	0.00
15	Benzyl Alcohol	40.000	38.210	4.5	93	0.00
16	2,2'-oxybis(1-Chloropropane	40.000	36.399	9.0	89	0.00
17	2-Methylphenol	40.000	37.519	6.2	92	0.00
18	Hexachloroethane	40.000	36.790	8.0	90	0.00
19 P	n-Nitroso-di-n-propylamine	40.000	35.633	10.9	92	0.00
20	3+4-Methylphenols	40.000	37.130	7.2	93	0.00
21 I	Naphthalene-d8	20.000	20.000	0.0	91	0.00
22	Acetophenone	40.000	36.665	8.3	93	0.00
23 S	Nitrobenzene-d5	80.000	74.629	6.7	94	0.00
24	Nitrobenzene	40.000	37.348	6.6	92	0.00
25	Isophorone	40.000	36.358	9.1	92	0.00
26 C	2-Nitrophenol	40.000	38.241	4.4	93	0.00
27	2,4-Dimethylphenol	40.000	37.271	6.8	93	0.00
28	bis(2-Chloroethoxy)methane	40.000	35.925	10.2	90	0.00
29 C	2,4-Dichlorophenol	40.000	37.555	6.1	94	0.00
30	1,2,4-Trichlorobenzene	40.000	36.761	8.1	92	0.00
31	Naphthalene	40.000	36.452	8.9	92	0.00
32	Benzoic acid	40.000	40.573	-1.4	98	0.00
33	4-Chloroaniline	40.000	36.680	8.3	92	0.00
34 C	Hexachlorobutadiene	40.000	36.093	9.8	89	0.00
35	Caprolactam	40.000	38.692	3.3	95	0.00
36 C	4-Chloro-3-methylphenol	40.000	37.660	5.9	94	0.00
37	2-Methylnaphthalene	40.000	36.686	8.3	92	0.00
38	1-Methylnaphthalene	40.000	36.576	8.6	93	0.00
39 I	Acenaphthene-d10	20.000	20.000	0.0	91	0.00
40	1,2,4,5-Tetrachlorobenzene	40.000	35.815	10.5	92	0.00
41 P	Hexachlorocyclopentadiene	40.000	36.852	7.9	89	0.00
42 S	2,4,6-Tribromophenol	80.000	73.629	8.0	95	0.00
43 C	2,4,6-Trichlorophenol	40.000	37.110	7.2	94	0.00
44	2,4,5-Trichlorophenol	40.000	36.790	8.0	93	0.00
45 S	2-Fluorobiphenyl	80.000	71.011	11.2	92	0.00
46	1,1'-Biphenyl	40.000	35.839	10.4	92	0.00
47	2-Chloronaphthalene	40.000	36.346	9.1	92	0.00

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF083023\
 Data File : BF135014.D
 Acq On : 30 Aug 2023 20:23
 Operator : CG\JU
 Sample : SSTDCCC040
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_F
 LabSampleId :
 SSTDCCC040

Quant Time: Aug 31 05:51:08 2023
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF083023.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Thu Aug 31 03:48:08 2023
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
48	2-Nitroaniline	40.000	38.234	4.4	95	0.00
49	Acenaphthylene	40.000	36.793	8.0	94	0.00
50	Dimethylphthalate	40.000	35.984	10.0	93	0.00
51	2,6-Dinitrotoluene	40.000	37.515	6.2	94	0.00
52 C	Acenaphthene	40.000	36.056	9.9	93	0.00
53	3-Nitroaniline	40.000	37.028	7.4	93	0.00
54 P	2,4-Dinitrophenol	40.000	43.286	-8.2	99	0.00
55	Dibenzofuran	40.000	36.079	9.8	93	0.00
56 P	4-Nitrophenol	40.000	38.680	3.3	95	0.00
57	2,4-Dinitrotoluene	40.000	38.405	4.0	95	0.00
58	Fluorene	40.000	36.017	10.0	95	0.00
59	2,3,4,6-Tetrachlorophenol	40.000	36.647	8.4	92	0.00
60	Diethylphthalate	40.000	35.840	10.4	92	0.00
61	4-Chlorophenyl-phenylether	40.000	35.371	11.6	92	0.00
62	4-Nitroaniline	40.000	37.342	6.6	93	0.00
63	Azobenzene	40.000	36.132	9.7	92	0.00
64 I	Phenanthrene-d10	20.000	20.000	0.0	93	0.00
65	4,6-Dinitro-2-methylphenol	40.000	39.948	0.1	94	0.00
66 c	n-Nitrosodiphenylamine	40.000	35.893	10.3	93	0.00
67	4-Bromophenyl-phenylether	40.000	35.689	10.8	92	0.00
68	Hexachlorobenzene	40.000	36.683	8.3	95	0.00
69	Atrazine	40.000	39.601	1.0	93	0.00
70 C	Pentachlorophenol	40.000	37.908	5.2	94	0.00
71	Phenanthrene	40.000	35.792	10.5	94	0.00
72	Anthracene	40.000	35.896	10.3	93	0.00
73	Carbazole	40.000	36.899	7.8	95	0.00
74	Di-n-butylphthalate	40.000	35.597	11.0	92	0.00
75 C	Fluoranthene	40.000	36.334	9.2	94	0.00
76 I	Chrysene-d12	20.000	20.000	0.0	96	0.00
77	Benzydine	40.000	44.378	-10.9	108	0.00
78	Pyrene	40.000	35.600	11.0	94	0.00
79 S	Terphenyl-d14	80.000	70.766	11.5	94	0.00
80	Butylbenzylphthalate	40.000	38.551	3.6	98	0.00
81	Benzo(a)anthracene	40.000	35.398	11.5	95	0.00
82	3,3'-Dichlorobenzidine	40.000	37.556	6.1	102	0.00
83	Chrysene	40.000	35.923	10.2	97	0.00
84	Bis(2-ethylhexyl)phthalate	40.000	40.015	-0.0	103	0.00
85 c	Di-n-octyl phthalate	40.000	38.706	3.2	102	0.00
86 I	Perylene-d12	20.000	20.000	0.0	99	0.00
87	Indeno(1,2,3-cd)pyrene	40.000	36.789	8.0	95	0.00
88	Benzo(b)fluoranthene	40.000	37.273	6.8	106	0.00
89	Benzo(k)fluoranthene	40.000	36.040	9.9	96	0.00
90 C	Benzo(a)pyrene	40.000	36.596	8.5	100	0.00
91	Dibenzo(a,h)anthracene	40.000	37.163	7.1	94	0.00
92	Benzo(g,h,i)perylene	40.000	36.985	7.5	95	0.00

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF083023\
Data File : BF135014.D
Acq On : 30 Aug 2023 20:23
Operator : CG\JU
Sample : SSTDCCC040
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
BNA_F
LabSampleId :
SSTDCCC040

Quant Time: Aug 31 05:51:08 2023
Quant Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF083023.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Thu Aug 31 03:48:08 2023
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

Compound	Amount	Calc.	%Dev	Area%	Dev(min)
----------	--------	-------	------	-------	----------

(#) = Out of Range

SPCC's out = 0 CCC's out = 0