

Data Path : Z:\SVOASRV\HPCHEM1\BNA F\DATA\BF083118\
 Data File : BF108727.D
 Acq On : 1 Sep 2018 9:11
 Operator : JU/SJ
 Sample : J4729-13
 Misc :
 ALS Vial : 39 Sample Multiplier: 1

Instrument :
 BNA_F
 ClientSampleId :
 082918-SIDI-F-U-B

Manual Integrations
 APPROVED

Sohil
 9/4/2018 3:54:48 PM

Quant Time: Sep 04 02:58:18 2018
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA F\METHODS\8270-BF083018.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Thu Aug 30 15:06:55 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	7.01	152	60220	20.00	ng	0.00
21) Naphthalene-d8	8.28	136	226043	20.00	ng	0.00
38) Acenaphthene-d10	10.04	164	95713	20.00	ng	-0.01
63) Phenanthrene-d10	11.54	188	215123	20.00	ng	0.00
75) Chrysene-d12	14.20	240	216660	20.00	ng	0.00
86) Perylene-d12	15.75	264	146868	20.00	ng	0.00
System Monitoring Compounds						
5) 2-Fluorophenol	5.67	112	106369m	30.68	ng	0.02
7) Phenol-d6	6.66	99	117492m	23.66	ng	0.01
23) Nitrobenzene-d5	7.57	82	487712	109.52	ng	0.00
41) 2,4,6-Tribromophenol	10.84	330	109194	86.42	ng	-0.01
44) 2-Fluorobiphenyl	9.35	172	898489	126.22	ng	0.00
78) Terphenyl-d14	13.12	244	993021	89.05	ng	-0.01
Target Compounds						
49) Dimethylphthalate	9.74	163	32429	4.303	ng	Qvalue # 99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\SVOASRV\HPCHEM1\BNA F\DATA\BF083118\
Data File : BF108727.D
Acq On : 1 Sep 2018 9:11
Operator : JU/SJ
Sample : J4729-13
Misc :
ALS Vial : 39 Sample Multiplier: 1

Instrument :
BNA_F
ClientSampleId :
082918-SIDI-F-U-B

Manual Integrations
APPROVED

Sohil
9/4/2018 3:54:48 PM

Quant Time: Sep 04 02:58:18 2018
Quant Method : Z:\SVOASRV\HPCHEM1\BNA F\METHODS\8270-BF083018.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Thu Aug 30 15:06:55 2018
Response via : Initial Calibration

