

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF083122\
 Data File : BF130051.D
 Acq On : 31 Aug 2022 14:32
 Operator : CG\JU
 Sample : SSTDIC010
 Misc :
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 BNA_F
 ClientSampleId :

Quant Time: Aug 31 16:00:01 2022
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF083122.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Wed Aug 31 15:58:13 2022
 Response via : Initial Calibration

Manual Integrations APPROVED

Reviewed By : Christian Giraldo 09/01/2022
 Supervised By : Jagrut Upadhyay 09/01/2022

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)	
Internal Standards							09/01/2022
1) 1,4-Dichlorobenzene-d4	6.704	152	226090	20.000	ng	0.00	Supervised By : Jagrut Upadhyay
21) Naphthalene-d8	7.986	136	885544	20.000	ng	# 0.00	
39) Acenaphthene-d10	9.739	164	462928	20.000	ng	0.00	
64) Phenanthrene-d10	11.216	188	768446	20.000	ng	# 0.00	
76) Chrysene-d12	13.845	240	547483	20.000	ng	#-0.01	
86) Perylene-d12	15.251	264	464494	20.000	ng	0.00	09/01/2022
System Monitoring Compounds							
5) 2-Fluorophenol	5.292	112	305240	22.353	ng	-0.02	
7) Phenol-d6	6.322	99	380997	22.280	ng	-0.02	
23) Nitrobenzene-d5	7.263	82	276858	16.736	ng	-0.02	
42) 2,4,6-Tribromophenol	10.522	330	84386	18.162	ng	-0.01	
45) 2-Fluorobiphenyl	9.063	172	688297	22.582	ng	0.00	
79) Terphenyl-d14	12.798	244	656498	19.953	ng	-0.01	
Target Compounds							Qvalue
2) 1,4-Dioxane	2.369	88	68780	12.209	ng		94
3) Pyridine	3.075	79	181144	12.287	ng		93
4) n-Nitrosodimethylamine	3.004	42	69417	10.217	ng	#	88
6) Aniline	6.363	93	228769	11.738	ng	#	51
8) 2-Chlorophenol	6.481	128	167989	11.051	ng		99
9) Benzaldehyde	6.251	77	122011	12.086	ng		97
10) Phenol	6.339	94	218749m	11.662	ng		
11) bis(2-Chloroethyl)ether	6.439	93	162618	11.417	ng		98
12) 1,3-Dichlorobenzene	6.639	146	186027	11.335	ng		97
13) 1,4-Dichlorobenzene	6.722	146	188211	11.224	ng		97
14) 1,2-Dichlorobenzene	6.875	146	175655	11.279	ng		97
15) Benzyl Alcohol	6.845	79	131470	10.184	ng		92
16) 2,2'-oxybis(1-Chloropr...	6.986	45	221472	11.604	ng	#	50
17) 2-Methylphenol	6.951	107	137858	11.294	ng	#	85
18) Hexachloroethane	7.216	117	64464	10.237	ng		97
19) n-Nitroso-di-n-propyla...	7.116	70	111326	10.300	ng		95
20) 3+4-Methylphenols	7.104	107	178225	10.878	ng	#	85
22) Acetophenone	7.116	105	234798	10.880	ng		94
24) Nitrobenzene	7.286	77	153455	9.200	ng		96
25) Isophorone	7.528	82	296011	10.397	ng		97
26) 2-Nitrophenol	7.604	139	46967	5.724	ng		91
27) 2,4-Dimethylphenol	7.639	122	125267	10.646	ng		96
28) bis(2-Chloroethoxy)met...	7.739	93	185493	11.158	ng		99
29) 2,4-Dichlorophenol	7.839	162	134690	10.468	ng		97
30) 1,2,4-Trichlorobenzene	7.928	180	148376	10.931	ng		98
31) Naphthalene	8.010	128	523017	11.296	ng		98
32) Benzoic acid	7.716	122	54233	10.851	ng		98
33) 4-Chloroaniline	8.057	127	202412	11.117	ng		97
34) Hexachlorobutadiene	8.128	225	81409	9.850	ng		99
35) Caprolactam	8.404	113	38729	9.679	ng		90
36) 4-Chloro-3-methylphenol	8.528	107	136095	9.805	ng		98
37) 2-Methylnaphthalene	8.698	142	334770	11.016	ng		100
38) 1-Methylnaphthalene	8.798	142	325635	11.026	ng		98
40) 1,2,4,5-Tetrachloroben...	8.863	216	135994	10.562	ng		98
41) Hexachlorocyclopentadiene	8.851	237	61553	22.644	ng		99
43) 2,4,6-Trichlorophenol	8.969	196	90192	10.781	ng		96

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF083122\
 Data File : BF130051.D
 Acq On : 31 Aug 2022 14:32
 Operator : CG\JU
 Sample : SSTDICC010
 Misc :
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 BNA_F
 ClientSampleId :


Quant Time: Aug 31 16:00:01 2022
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF083122.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Wed Aug 31 15:58:13 2022
 Response via : Initial Calibration

Manual Integrations APPROVED

Reviewed By : Christian Giraldo 09/01/2022
 Supervised By : Jagrut Upadhyay 09/01/2022

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)	
44) 2,4,5-Trichlorophenol	9.004	196	99744	11.082	ng	97	09/01/2022
46) 1,1'-Biphenyl	9.163	154	396711	11.280	ng	100	Supervised By :Jagrut Upadhyay
47) 2-Chloronaphthalene	9.180	162	310322	11.152	ng	99	
48) 2-Nitroaniline	9.280	65	60043	6.828	ng	91	
49) Acenaphthylene	9.598	152	481408	11.409	ng	98	
50) Dimethylphthalate	9.463	163	345739	10.418	ng	99	
51) 2,6-Dinitrotoluene	9.522	165	58820	8.171	ng	98	09/01/2022
52) Acenaphthene	9.769	154	294589	11.507	ng	99	
53) 3-Nitroaniline	9.686	138	69701	8.755	ng	# 95	
54) 2,4-Dinitrophenol	9.786	184	12321	4.510	ng	89	
55) Dibenzofuran	9.939	168	435436	11.322	ng	97	
56) 4-Nitrophenol	9.833	139	51449	14.527	ng	99	
57) 2,4-Dinitrotoluene	9.922	165	65987	7.061	ng	# 82	
58) Fluorene	10.280	166	340448	10.601	ng	100	
59) 2,3,4,6-Tetrachlorophenol	10.057	232	78858	11.939	ng	# 81	
60) Diethylphthalate	10.163	149	334659	10.114	ng	99	
61) 4-Chlorophenyl-phenyle...	10.280	204	147409	10.229	ng	97	
62) 4-Nitroaniline	10.292	138	66738	8.310	ng	98	
63) Azobenzene	10.439	77	339879	10.746	ng	92	
65) 4,6-Dinitro-2-methylph...	10.322	198	20165	4.470	ng	98	
66) n-Nitrosodiphenylamine	10.392	169	291940	11.256	ng	100	
67) 4-Bromophenyl-phenylether	10.769	248	92646	10.293	ng	91	
68) Hexachlorobenzene	10.822	284	103026	10.567	ng	97	
69) Atrazine	10.922	200	78372	9.845	ng	97	
70) Pentachlorophenol	11.016	266	49998	20.115	ng	94	
71) Phenanthrene	11.239	178	491596	11.454	ng	100	
72) Anthracene	11.292	178	482726	11.311	ng	98	
73) Carbazole	11.445	167	444040	11.461	ng	99	
74) Di-n-butylphthalate	11.786	149	500424	10.036	ng	99	
75) Fluoranthene	12.421	202	488701	11.128	ng	96	
77) Benzidine	12.551	184	164491	11.351	ng	100	
78) Pyrene	12.651	202	504897	10.612	ng	99	
80) Butylbenzylphthalate	13.280	149	168523	8.371	ng	97	
81) Benzo(a)anthracene	13.833	228	412743	10.959	ng	98	
82) 3,3'-Dichlorobenzidine	13.804	252	123853	10.618	ng	98	
83) Chrysene	13.874	228	400189	11.352	ng	100	
84) Bis(2-ethylhexyl)phtha...	13.845	149	234031	9.799	ng	100	
85) Di-n-octyl phthalate	14.451	149	333351	10.111	ng	98	
87) Indeno(1,2,3-cd)pyrene	16.586	276	276194	8.661	ng	# 93	
88) Benzo(b)fluoranthene	14.851	252	337656	10.616	ng	100	
89) Benzo(k)fluoranthene	14.874	252	331519m	10.869	ng		
90) Benzo(a)pyrene	15.186	252	257501	10.281	ng	# 97	
91) Dibenzo(a,h)anthracene	16.609	278	229227	8.712	ng	# 96	
92) Benzo(g,h,i)perylene	16.992	276	223648	8.520	ng	# 92	

(#) = qualifier out of range (m) = manual integration (+) = signals summed




#1
1,4-Dichlorobenzene-d4
Concen: 20.000 ng
RT: 6.704 min Scan# 71
Delta R.T. -0.006 min
Lab File: BF130051.D
Acq: 31 Aug 2022 14:32

Instrument :

BNA_F



ClientSampleId :



Tgt Ion:152 Resp: 226090
Ion Ratio Lower Upper
152 100
150 151.9 126.9 190.3
115 60.8 53.4 80.0


Manual Integrations
APPROVED

Reviewed By :Christian Giraldo 09/01/2022
Supervised By :Jagrut Upadhyay 09/01/2022






09/01/2022
Supervised By :Jagrut
Upadhyay


09/01/2022



#2
1,4-Dioxane
Concen: 12.209 ng
RT: 2.369 min Scan# 14
Delta R.T. -0.012 min
Lab File: BF130051.D
Acq: 31 Aug 2022 14:32


Tgt Ion: 88 Resp: 68780
Ion Ratio Lower Upper
88 100
58 71.6 60.9 91.3
43 25.2 23.6 35.4





#3
Pyridine
Concen: 12.287 ng
RT: 3.075 min Scan# 11
Delta R.T. -0.012 min
Lab File: BF130051.D
Acq: 31 Aug 2022 14:32



Instrument :
BNA_F
ClientSampleId :



Tgt Ion: 79 Resp: 181144
Ion Ratio Lower Upper
79 100
52 67.4 59.4 89.2
51 34.1 29.5 44.3


Manual Integrations
APPROVED

Reviewed By :Christian Giraldo 09/01/2022
Supervised By :Jagrut Upadhyay 09/01/2022




09/01/2022
Supervised By :Jagrut
Upadhyay



09/01/2022




#4
n-Nitrosodimethylamine
Concen: 10.217 ng
RT: 3.004 min Scan# 122
Delta R.T. -0.041 min
Lab File: BF130051.D
Acq: 31 Aug 2022 14:32



Tgt Ion: 42 Resp: 69417
Ion Ratio Lower Upper
42 100
74 143.7 103.1 154.7
44 9.3 5.4 8.2#






#5
2-Fluorophenol
Concen: 22.353 ng
RT: 5.292 min Scan# 514
Delta R.T. -0.018 min
Lab File: BF130051.D
Acq: 31 Aug 2022 14:32

Instrument :

BNA_F



ClientSampleId :



Tgt Ion:112 Resp: 305240
Ion Ratio Lower Upper
112 100
64 58.6 47.6 71.4
63 28.9 24.6 37.0


Manual Integrations
APPROVED

Reviewed By :Christian Giraldo 09/01/2022
Supervised By :Jagrut Upadhyay 09/01/2022






09/01/2022
Supervised By :Jagrut
Upadhyay

09/01/2022



#6
Aniline
Concen: 11.738 ng
RT: 6.363 min Scan# 693
Delta R.T. -0.012 min
Lab File: BF130051.D
Acq: 31 Aug 2022 14:32


Tgt Ion: 93 Resp: 228769
Ion Ratio Lower Upper
93 100
66 39.1 68.7 103.1#
65 19.8 41.6 62.4#





#7
Phenol-d6
Concen: 22.280 ng
RT: 6.322 min Scan# 61
Delta R.T. -0.024 min
Lab File: BF130051.D
Acq: 31 Aug 2022 14:32



Instrument :
BNA_F
ClientSampleId :



Tgt Ion: 99 Resp: 38099
Ion Ratio Lower Upper
99 100
42 16.5 15.0 22.4
71 30.0 28.8 43.2


Manual Integrations
APPROVED

Reviewed By :Christian Giraldo 09/01/2022
Supervised By :Jagrut Upadhyay 09/01/2022




09/01/2022
Supervised By :Jagrut
Upadhyay



09/01/2022




#8
2-Chlorophenol
Concen: 11.051 ng
RT: 6.481 min Scan# 713
Delta R.T. -0.012 min
Lab File: BF130051.D
Acq: 31 Aug 2022 14:32



Tgt Ion:128 Resp: 167989
Ion Ratio Lower Upper
128 100
130 31.2 12.5 52.5
64 45.2 24.9 64.9






#9
Benzaldehyde
Concen: 12.086 ng
RT: 6.251 min Scan# 61
Delta R.T. -0.006 min
Lab File: BF130051.D
Acq: 31 Aug 2022 14:32

Instrument :

BNA_F



ClientSampleId :



Tgt Ion: 77 Resp: 12201
Ion Ratio Lower Upper
77 100
105 103.4 80.3 120.3
106 97.6 74.3 114.3


Manual Integrations
APPROVED

Reviewed By :Christian Giraldo 09/01/2022
Supervised By :Jagrut Upadhyay 09/01/2022




09/01/2022
Supervised By :Jagrut
Upadhyay



09/01/2022




#10
Phenol
Concen: 11.662 ng m
RT: 6.339 min Scan# 689
Delta R.T. -0.018 min
Lab File: BF130051.D
Acq: 31 Aug 2022 14:32




Tgt Ion: 94 Resp: 218749
Ion Ratio Lower Upper
94 100
65 25.2 24.9 64.9
66 31.5 54.2 94.2#





#11
bis(2-Chloroethyl)ether
Concen: 11.417 ng
RT: 6.439 min Scan# 708
Delta R.T. -0.012 min
Lab File: BF130051.D
Acq: 31 Aug 2022 14:32



Instrument :
BNA_F
ClientSampleId :



Tgt Ion: 93 Resp: 162618
Ion Ratio Lower Upper
93 100
63 77.5 55.9 95.9
95 33.0 12.3 52.3


Manual Integrations
APPROVED

Reviewed By :Christian Giraldo 09/01/2022
Supervised By :Jagrut Upadhyay 09/01/2022






09/01/2022
Supervised By :Jagrut
Upadhyay


09/01/2022



#12
1,3-Dichlorobenzene
Concen: 11.335 ng
RT: 6.639 min Scan# 740
Delta R.T. -0.006 min
Lab File: BF130051.D
Acq: 31 Aug 2022 14:32


Tgt Ion:146 Resp: 186027
Ion Ratio Lower Upper
146 100
148 61.5 51.2 76.8
75 31.7 26.5 39.7





#13
1,4-Dichlorobenzene
Concen: 11.224 ng
RT: 6.722 min Scan# 755
Delta R.T. -0.006 min
Lab File: BF130051.D
Acq: 31 Aug 2022 14:32



Instrument :
BNA_F
ClientSampleId :



Tgt Ion:146 Resp: 18821
Ion Ratio Lower Upper
146 100
148 64.0 50.6 75.8
111 41.1 36.2 54.4


Manual Integrations
APPROVED

Reviewed By :Christian Giraldo 09/01/2022
Supervised By :Jagrut Upadhyay 09/01/2022




09/01/2022
Supervised By :Jagrut
Upadhyay



09/01/2022




#14
1,2-Dichlorobenzene
Concen: 11.279 ng
RT: 6.875 min Scan# 780
Delta R.T. -0.006 min
Lab File: BF130051.D
Acq: 31 Aug 2022 14:32




Tgt Ion:146 Resp: 175655
Ion Ratio Lower Upper
146 100
148 65.6 51.8 77.6
111 42.8 37.2 55.8





#15
Benzyl Alcohol
Concen: 10.184 ng
RT: 6.845 min Scan# 777
Delta R.T. -0.012 min
Lab File: BF130051.D
Acq: 31 Aug 2022 14:32



Instrument :
BNA_F
ClientSampleId :



Tgt Ion: 79 Resp: 131470
Ion Ratio Lower Upper
79 100
108 85.5 59.2 88.8
77 63.5 51.9 77.9


Manual Integrations
APPROVED

Reviewed By :Christian Giraldo 09/01/2022
Supervised By :Jagrut Upadhyay 09/01/2022




09/01/2022
Supervised By :Jagrut
Upadhyay



09/01/2022




#16
2,2'-oxybis(1-Chloropropane)
Concen: 11.604 ng
RT: 6.986 min Scan# 799
Delta R.T. -0.006 min
Lab File: BF130051.D
Acq: 31 Aug 2022 14:32




Tgt Ion: 45 Resp: 221472
Ion Ratio Lower Upper
45 100
77 14.4 27.4 67.4#
79 11.1 24.4 64.4#





#17
2-Methylphenol
Concen: 11.294 ng
RT: 6.951 min Scan# 795
Delta R.T. -0.012 min
Lab File: BF130051.D
Acq: 31 Aug 2022 14:32



Instrument :
BNA_F
ClientSampleId :



Tgt Ion:107 Resp: 13785
Ion Ratio Lower Upper
107 100
108 114.4 90.2 135.2
77 42.1 51.8 77.8
79 40.4 47.4 71.2


Manual Integrations
APPROVED

Reviewed By :Christian Giraldo 09/01/2022
Supervised By :Jagrut Upadhyay 09/01/2022






09/01/2022
Supervised By :Jagrut
Upadhyay


09/01/2022



#18
Hexachloroethane
Concen: 10.237 ng
RT: 7.216 min Scan# 838
Delta R.T. -0.006 min
Lab File: BF130051.D
Acq: 31 Aug 2022 14:32


Tgt Ion:117 Resp: 64464
Ion Ratio Lower Upper
117 100
119 96.6 79.1 118.7
201 86.4 66.6 100.0





#19
n-Nitroso-di-n-propylamine
Concen: 10.300 ng
RT: 7.116 min Scan# 819
Delta R.T. -0.024 min
Lab File: BF130051.D
Acq: 31 Aug 2022 14:32



Instrument :
BNA_F
ClientSampleId :



Tgt Ion: 70 Resp: 111320
Ion Ratio Lower Upper
70 100
42 53.7 47.6 71.4
101 10.0 8.1 12.1
130 22.0 17.2 25.8


Manual Integrations
APPROVED

Reviewed By :Christian Giraldo 09/01/2022
Supervised By :Jagrut Upadhyay 09/01/2022




09/01/2022
Supervised By :Jagrut
Upadhyay



09/01/2022




#20
3+4-Methylphenols
Concen: 10.878 ng
RT: 7.104 min Scan# 819
Delta R.T. -0.018 min
Lab File: BF130051.D
Acq: 31 Aug 2022 14:32




Tgt Ion:107 Resp: 178225
Ion Ratio Lower Upper
107 100
108 86.3 71.3 111.3
77 63.9 20.4 60.4#
79 24.5 11.1 51.1





#21
Naphthalene-d8
Concen: 20.000 ng
RT: 7.986 min Scan# 91
Delta R.T. -0.006 min
Lab File: BF130051.D
Acq: 31 Aug 2022 14:32



Instrument :
BNA_F
ClientSampleId :



Tgt Ion:136 Resp: 885544
Ion Ratio Lower Upper
136 100
137 11.0 8.4 12.6
54 7.4 6.2 9.2
68 6.7 4.2 6.4#


Manual Integrations
APPROVED

Reviewed By :Christian Giraldo 09/01/2022
Supervised By :Jagrut Upadhyay 09/01/2022




09/01/2022
Supervised By :Jagrut
Upadhyay



09/01/2022




#22
Acetophenone
Concen: 10.880 ng
RT: 7.116 min Scan# 821
Delta R.T. -0.012 min
Lab File: BF130051.D
Acq: 31 Aug 2022 14:32



Tgt Ion:105 Resp: 234798
Ion Ratio Lower Upper
105 100
71 8.4 6.6 9.8
51 24.1 23.8 35.8
120 23.6 17.9 26.9





#23

Nitrobenzene-d5

Concen: 16.736 ng

RT: 7.263 min Scan# 84

Delta R.T. -0.018 min


Lab File: BF130051.D

Acq: 31 Aug 2022 14:32

Instrument :

BNA_F



ClientSampleId :



Tgt Ion:	82	Resp:	27685
Ion	Ratio	Lower	Upper
82	100		
128	43.2	36.1	54.1
54	50.6	40.0	60.0


Manual Integrations
APPROVED

Reviewed By :Christian Giraldo 09/01/2022
Supervised By :Jagrut Upadhyay 09/01/2022



09/01/2022
Supervised By :Jagrut
Upadhyay

09/01/2022



#24

Nitrobenzene

Concen: 9.200 ng




RT: 7.286 min Scan# 850


Delta R.T. -0.012 min

Lab File: BF130051.D

Acq: 31 Aug 2022 14:32


Tgt Ion:	77	Resp:	153455
Ion	Ratio	Lower	Upper
77	100		
123	52.4	39.2	58.8
65	12.9	11.1	16.7





#25
Isophorone
Concen: 10.397 ng
RT: 7.528 min Scan# 891
Delta R.T. -0.012 min
Lab File: BF130051.D
Acq: 31 Aug 2022 14:32



Instrument :
BNA_F
ClientSampleId :



Tgt Ion: 82 Resp: 29601
Ion Ratio Lower Upper
82 100
95 6.6 5.9 8.9
138 19.7 14.6 21.8


Manual Integrations
APPROVED

Reviewed By :Christian Giraldo 09/01/2022
Supervised By :Jagrut Upadhyay 09/01/2022




09/01/2022
Supervised By :Jagrut
Upadhyay



09/01/2022




#26
2-Nitrophenol
Concen: 5.724 ng
RT: 7.604 min Scan# 904
Delta R.T. -0.006 min
Lab File: BF130051.D
Acq: 31 Aug 2022 14:32



Tgt Ion:139 Resp: 46967
Ion Ratio Lower Upper
139 100
109 26.1 24.4 36.6
65 44.8 40.9 61.3





#27

2,4-Dimethylphenol

Concen: 10.646 ng

RT: 7.639 min Scan# 911

Delta R.T. -0.006 min


Lab File: BF130051.D

Acq: 31 Aug 2022 14:32

Instrument :

BNA_F

ClientSampleId :





Tgt Ion:	122	Resp:	12526
Ion Ratio	Lower	Upper	
122	100		
107	108.6	92.2	138.4
121	57.6	46.1	69.1

Manual Integrations


APPROVED

Reviewed By :Christian Giraldo 09/01/2022
Supervised By :Jagrut Upadhyay 09/01/2022



09/01/2022
Supervised By :Jagrut
Upadhyay

09/01/2022



#28

bis(2-Chloroethoxy)methane

Concen: 11.158 ng




RT: 7.739 min Scan# 927


Delta R.T. -0.012 min

Lab File: BF130051.D

Acq: 31 Aug 2022 14:32

Tgt Ion:	93	Resp:	185493
Ion Ratio	Lower	Upper	
93	100		
95	31.5	25.6	38.4
123	11.6	9.0	13.4





#29

2,4-Dichlorophenol

Concen: 10.468 ng

RT: 7.839 min Scan# 945

Delta R.T. -0.006 min


Lab File: BF130051.D

Acq: 31 Aug 2022 14:32

Instrument :

BNA_F

ClientSampleId :




Tgt Ion:	162	Resp:	134690
Ion Ratio	Lower	Upper	
162	100		
164	62.8	43.7	83.7
98	33.3	16.5	56.5

Manual Integrations

APPROVED


Reviewed By :Christian Giraldo 09/01/2022

Supervised By :Jagrut Upadhyay 09/01/2022



09/01/2022
Supervised By :Jagrut
Upadhyay

09/01/2022



#30

1,2,4-Trichlorobenzene

Concen: 10.931 ng




RT: 7.928 min Scan# 959


Delta R.T. -0.006 min

Lab File: BF130051.D

Acq: 31 Aug 2022 14:32


Tgt Ion:	180	Resp:	148376
Ion Ratio	Lower	Upper	
180	100		
182	94.4	77.6	116.4
145	31.7	25.6	38.4





#31
Naphthalene
Concen: 11.296 ng
RT: 8.010 min Scan# 91
Delta R.T. -0.006 min
Lab File: BF130051.D
Acq: 31 Aug 2022 14:32



Instrument :
BNA_F
ClientSampleId :



Tgt Ion:128 Resp: 52301
Ion Ratio Lower Upper
128 100
129 10.7 8.9 13.3
127 12.6 10.7 16.1


Manual Integrations
APPROVED

Reviewed By :Christian Giraldo 09/01/2022
Supervised By :Jagrut Upadhyay 09/01/2022






09/01/2022
Supervised By :Jagrut
Upadhyay


09/01/2022



#32
Benzoic acid
Concen: 10.851 ng
RT: 7.716 min Scan# 923
Delta R.T. -0.071 min
Lab File: BF130051.D
Acq: 31 Aug 2022 14:32


Tgt Ion:122 Resp: 54233
Ion Ratio Lower Upper
122 100
105 117.6 99.9 139.9
77 84.9 65.9 105.9





#33
4-Chloroaniline
Concen: 11.117 ng
RT: 8.057 min Scan# 91
Delta R.T. -0.006 min
Lab File: BF130051.D
Acq: 31 Aug 2022 14:32



Instrument :
BNA_F
ClientSampleId :



Tgt Ion:127 Resp: 20241
Ion Ratio Lower Upper
127 100
129 32.3 26.7 40.1
65 28.2 24.2 36.2
92 17.5 15.2 22.8


Manual Integrations
APPROVED

Reviewed By :Christian Giraldo 09/01/2022
Supervised By :Jagrut Upadhyay 09/01/2022




09/01/2022
Supervised By :Jagrut
Upadhyay



09/01/2022



#34
Hexachlorobutadiene
Concen: 9.850 ng
RT: 8.128 min Scan# 993
Delta R.T. -0.000 min
Lab File: BF130051.D
Acq: 31 Aug 2022 14:32




Tgt Ion:225 Resp: 81409
Ion Ratio Lower Upper
225 100
223 62.2 50.2 75.2
227 63.9 50.6 75.8





#35
Caprolactam
Concen: 9.679 ng
RT: 8.404 min Scan# 1048
Delta R.T. -0.047 min
Lab File: BF130051.D
Acq: 31 Aug 2022 14:32



Instrument :
BNA_F
ClientSampleId :



Tgt Ion:113 Resp: 38729
Ion Ratio Lower Upper
113 100
55 159.9 148.8 188.8
56 111.5 109.1 149.1


Manual Integrations
APPROVED

Reviewed By :Christian Giraldo 09/01/2022
Supervised By :Jagrut Upadhyay 09/01/2022




09/01/2022
Supervised By :Jagrut
Upadhyay



09/01/2022




#36
4-Chloro-3-methylphenol
Concen: 9.805 ng
RT: 8.528 min Scan# 1061
Delta R.T. -0.018 min
Lab File: BF130051.D
Acq: 31 Aug 2022 14:32




Tgt Ion:107 Resp: 136095
Ion Ratio Lower Upper
107 100
144 24.5 19.9 29.9
142 77.7 60.2 90.4





#37
2-Methylnaphthalene
Concen: 11.016 ng
RT: 8.698 min Scan# 1107
Delta R.T. -0.006 min
Lab File: BF130051.D
Acq: 31 Aug 2022 14:32



Instrument :
BNA_F
ClientSampleId :



Tgt Ion:142 Resp: 334770
Ion Ratio Lower Upper
142 100
141 88.2 70.9 106.3


Manual Integrations
APPROVED

Reviewed By :Christian Giraldo 09/01/2022
Supervised By :Jagrut Upadhyay 09/01/2022




09/01/2022
Supervised By :Jagrut
Upadhyay



09/01/2022




#38
1-Methylnaphthalene
Concen: 11.026 ng
RT: 8.798 min Scan# 1107
Delta R.T. -0.006 min
Lab File: BF130051.D
Acq: 31 Aug 2022 14:32




Tgt Ion:142 Resp: 325635
Ion Ratio Lower Upper
142 100
141 89.9 73.5 110.3





#39
Acenaphthene-d10
Concen: 20.000 ng
RT: 9.739 min Scan# 1118
Delta R.T. -0.006 min
Lab File: BF130051.D
Acq: 31 Aug 2022 14:32



Instrument :
BNA_F
ClientSampleId :



Tgt Ion:164 Resp: 462928
Ion Ratio Lower Upper
164 100
162 100.7 82.0 123.0
160 44.8 35.4 53.0


Manual Integrations
APPROVED

Reviewed By :Christian Giraldo 09/01/2022
Supervised By :Jagrut Upadhyay 09/01/2022




09/01/2022
Supervised By :Jagrut
Upadhyay



09/01/2022




#40
1,2,4,5-Tetrachlorobenzene
Concen: 10.562 ng
RT: 8.863 min Scan# 1118
Delta R.T. -0.006 min
Lab File: BF130051.D
Acq: 31 Aug 2022 14:32




Tgt Ion:216 Resp: 135994
Ion Ratio Lower Upper
216 100
214 80.4 63.1 94.7
179 22.5 18.1 27.1
108 21.6 15.9 23.9





#41
Hexachlorocyclopentadiene
Concen: 22.644 ng
RT: 8.851 min Scan# 1116
Delta R.T. -0.006 min
Lab File: BF130051.D
Acq: 31 Aug 2022 14:32



Instrument :
BNA_F
ClientSampleId :



Tgt Ion:237 Resp: 6155
Ion Ratio Lower Upper
237 100
235 60.6 41.5 81.5
272 14.2 0.0 33.3


Manual Integrations
APPROVED

Reviewed By :Christian Giraldo 09/01/2022
Supervised By :Jagrut Upadhyay 09/01/2022




09/01/2022
Supervised By :Jagrut
Upadhyay



09/01/2022




#42
2,4,6-Tribromophenol
Concen: 18.162 ng
RT: 10.522 min Scan# 1400
Delta R.T. -0.012 min
Lab File: BF130051.D
Acq: 31 Aug 2022 14:32



Tgt Ion:330 Resp: 84386
Ion Ratio Lower Upper
330 100
332 103.7 82.6 124.0
141 38.0 27.6 41.4






#43
2,4,6-Trichlorophenol
Concen: 10.781 ng
RT: 8.969 min Scan# 1136
Delta R.T. -0.012 min
Lab File: BF130051.D
Acq: 31 Aug 2022 14:32

Instrument :

BNA_F



ClientSampleId :



Tgt Ion:196 Resp: 90192
Ion Ratio Lower Upper
196 100
198 94.2 79.0 118.4
200 30.8 25.1 37.7


Manual Integrations
APPROVED

Reviewed By :Christian Giraldo 09/01/2022
Supervised By :Jagrut Upadhyay 09/01/2022




09/01/2022
Supervised By :Jagrut
Upadhyay



09/01/2022




#44
2,4,5-Trichlorophenol
Concen: 11.082 ng
RT: 9.004 min Scan# 1142
Delta R.T. -0.012 min
Lab File: BF130051.D
Acq: 31 Aug 2022 14:32




Tgt Ion:196 Resp: 99744
Ion Ratio Lower Upper
196 100
198 96.1 76.4 114.6
97 52.2 46.5 69.7
132 29.9 25.2 37.8





#45
2-Fluorobiphenyl
Concen: 22.582 ng
RT: 9.063 min Scan# 1153
Delta R.T. -0.006 min
Lab File: BF130051.D
Acq: 31 Aug 2022 14:32



Instrument :
BNA_F
ClientSampleId :



Tgt Ion:172 Resp: 68829
Ion Ratio Lower Upper
172 100
171 34.2 28.4 42.6
170 23.2 19.1 28.7


Manual Integrations
APPROVED

Reviewed By :Christian Giraldo 09/01/2022
Supervised By :Jagrut Upadhyay 09/01/2022




09/01/2022
Supervised By :Jagrut
Upadhyay



09/01/2022




#46
1,1'-Biphenyl
Concen: 11.280 ng
RT: 9.163 min Scan# 1169
Delta R.T. -0.006 min
Lab File: BF130051.D
Acq: 31 Aug 2022 14:32




Tgt Ion:154 Resp: 396711
Ion Ratio Lower Upper
154 100
153 40.2 20.2 60.2
76 12.8 0.0 32.1





#47
2-Chloronaphthalene
Concen: 11.152 ng
RT: 9.180 min Scan# 1172
Delta R.T. -0.012 min
Lab File: BF130051.D
Acq: 31 Aug 2022 14:32



Instrument :
BNA_F
ClientSampleId :



Tgt Ion:162 Resp: 31032
Ion Ratio Lower Upper
162 100
127 40.7 33.1 49.7
164 32.5 26.7 40.1


Manual Integrations
APPROVED

Reviewed By :Christian Giraldo 09/01/2022
Supervised By :Jagrut Upadhyay 09/01/2022






09/01/2022
Supervised By :Jagrut
Upadhyay


09/01/2022



#48
2-Nitroaniline
Concen: 6.828 ng
RT: 9.280 min Scan# 1189
Delta R.T. -0.006 min
Lab File: BF130051.D
Acq: 31 Aug 2022 14:32

Tgt Ion: 65 Resp: 60043
Ion Ratio Lower Upper
65 100
92 70.2 55.0 82.4
138 115.7 81.3 121.9






#49
Acenaphthylene
Concen: 11.409 ng
RT: 9.598 min Scan# 1243
Delta R.T. -0.006 min
Lab File: BF130051.D
Acq: 31 Aug 2022 14:32

Instrument :

BNA_F



ClientSampleId :



Tgt Ion:152 Resp: 481408
Ion Ratio Lower Upper
152 100
151 19.5 16.1 24.1
153 12.5 10.6 16.0


Manual Integrations
APPROVED

Reviewed By :Christian Giraldo 09/01/2022
Supervised By :Jagrut Upadhyay 09/01/2022




09/01/2022
Supervised By :Jagrut
Upadhyay



09/01/2022




#50
Dimethylphthalate
Concen: 10.418 ng
RT: 9.463 min Scan# 1220
Delta R.T. -0.018 min
Lab File: BF130051.D
Acq: 31 Aug 2022 14:32



Tgt Ion:163 Resp: 345739
Ion Ratio Lower Upper
163 100
194 4.0 3.3 4.9
164 9.9 8.2 12.2





#51
2,6-Dinitrotoluene
Concen: 8.171 ng
RT: 9.522 min Scan# 11
Delta R.T. -0.012 min
Lab File: BF130051.D
Acq: 31 Aug 2022 14:32

Instrument :



BNA_F

ClientSampleId :

Tgt Ion:165 Resp: 58820
Ion Ratio Lower Upper
165 100
63 56.7 47.8 71.8
89 58.8 47.2 70.8



Manual Integrations
APPROVED

Reviewed By :Christian Giraldo 09/01/2022
Supervised By :Jagrut Upadhyay 09/01/2022






09/01/2022
Supervised By :Jagrut
Upadhyay


09/01/2022



#52
Acenaphthene
Concen: 11.507 ng
RT: 9.769 min Scan# 1272
Delta R.T. -0.006 min
Lab File: BF130051.D
Acq: 31 Aug 2022 14:32

Tgt Ion:154 Resp: 294589
Ion Ratio Lower Upper
154 100
153 111.3 89.8 134.6
152 53.3 42.1 63.1






#53
3-Nitroaniline
Concen: 8.755 ng
RT: 9.686 min Scan# 1258
Delta R.T. -0.018 min
Lab File: BF130051.D
Acq: 31 Aug 2022 14:32

Instrument :

BNA_F



ClientSampleId :



Tgt Ion:138 Resp: 6970
Ion Ratio Lower Upper
138 100
108 13.9 8.4 12.6
92 117.3 89.8 134.8


Manual Integrations
APPROVED

Reviewed By :Christian Giraldo 09/01/2022
Supervised By :Jagrut Upadhyay 09/01/2022




09/01/2022
Supervised By :Jagrut
Upadhyay



09/01/2022




#54
2,4-Dinitrophenol
Concen: 4.510 ng
RT: 9.786 min Scan# 1275
Delta R.T. -0.018 min
Lab File: BF130051.D
Acq: 31 Aug 2022 14:32




Tgt Ion:184 Resp: 12321
Ion Ratio Lower Upper
184 100
63 83.1 59.6 89.4
154 60.0 55.3 82.9





#55
Dibenzenofuran
Concen: 11.322 ng
RT: 9.939 min Scan# 11
Delta R.T. -0.012 min
Lab File: BF130051.D
Acq: 31 Aug 2022 14:32



Instrument :
BNA_F
ClientSampleId :



Tgt Ion:168 Resp: 435430
Ion Ratio Lower Upper
168 100
139 36.9 31.3 46.9
169 13.6 10.7 16.1


Manual Integrations
APPROVED

Reviewed By :Christian Giraldo 09/01/2022
Supervised By :Jagrut Upadhyay 09/01/2022






09/01/2022
Supervised By :Jagrut
Upadhyay


09/01/2022



#56
4-Nitrophenol
Concen: 14.527 ng
RT: 9.833 min Scan# 1283
Delta R.T. -0.024 min
Lab File: BF130051.D
Acq: 31 Aug 2022 14:32

Tgt Ion:139 Resp: 51449
Ion Ratio Lower Upper
139 100
109 72.6 52.9 92.9
65 100.5 79.5 119.5






#57
2,4-Dinitrotoluene
Concen: 7.061 ng
RT: 9.922 min Scan# 1300
Delta R.T. -0.012 min
Lab File: BF130051.D
Acq: 31 Aug 2022 14:32

Instrument :

BNA_F



ClientSampleId :



Tgt Ion:165 Resp: 6598
Ion Ratio Lower Upper
165 100
63 38.0 48.4 72.6
89 68.9 62.3 93.5
182 2.7 2.1 3.1


Manual Integrations
APPROVED

Reviewed By :Christian Giraldo 09/01/2022
Supervised By :Jagrut Upadhyay 09/01/2022




09/01/2022
Supervised By :Jagrut
Upadhyay



09/01/2022




#58
Fluorene
Concen: 10.601 ng
RT: 10.280 min Scan# 1359
Delta R.T. -0.012 min
Lab File: BF130051.D
Acq: 31 Aug 2022 14:32




Tgt Ion:166 Resp: 340448
Ion Ratio Lower Upper
166 100
165 97.3 77.9 116.9
167 13.5 11.0 16.4





#59
2,3,4,6-Tetrachlorophenol
Concen: 11.939 ng
RT: 10.057 min Scan# 1322
Delta R.T. -0.006 min
Lab File: BF130051.D
Acq: 31 Aug 2022 14:32



Instrument :
BNA_F
ClientSampleId :



Tgt Ion:232 Resp: 7885
Ion Ratio Lower Upper
232 100
131 48.5 29.9 44.9
130 2.4 1.5 2.3
166 32.3 17.8 26.8


Manual Integrations
APPROVED

Reviewed By :Christian Giraldo 09/01/2022
Supervised By :Jagrut Upadhyay 09/01/2022




09/01/2022
Supervised By :Jagrut
Upadhyay



09/01/2022




#60
Diethylphthalate
Concen: 10.114 ng
RT: 10.163 min Scan# 1339
Delta R.T. -0.012 min
Lab File: BF130051.D
Acq: 31 Aug 2022 14:32



Tgt Ion:149 Resp: 334659
Ion Ratio Lower Upper
149 100
177 21.5 17.7 26.5
150 12.3 10.1 15.1






#61
4-Chlorophenyl-phenylether
Concen: 10.229 ng
RT: 10.280 min Scan# 1359
Delta R.T. -0.006 min
Lab File: BF130051.D
Acq: 31 Aug 2022 14:32

Instrument :

BNA_F



ClientSampleId :



Tgt Ion:204 Resp: 147409
Ion Ratio Lower Upper
204 100
206 34.0 26.5 39.7
141 63.4 52.6 79.0


Manual Integrations
APPROVED

Reviewed By :Christian Giraldo 09/01/2022
Supervised By :Jagrut Upadhyay 09/01/2022




09/01/2022
Supervised By :Jagrut
Upadhyay



09/01/2022




#62
4-Nitroaniline
Concen: 8.310 ng
RT: 10.292 min Scan# 1361
Delta R.T. -0.024 min
Lab File: BF130051.D
Acq: 31 Aug 2022 14:32




Tgt Ion:138 Resp: 66738
Ion Ratio Lower Upper
138 100
92 53.5 30.3 70.3
108 76.0 55.7 95.7





#63
Azobenzene
Concen: 10.746 ng
RT: 10.439 min Scan# 11
Delta R.T. -0.006 min
Lab File: BF130051.D
Acq: 31 Aug 2022 14:32



Instrument :
BNA_F
ClientSampleId :



Tgt Ion: 77 Resp: 339879
Ion Ratio Lower Upper
77 100
182 27.2 2.2 42.2
105 22.9 0.8 40.8
51 25.7 10.3 50.3


Manual Integrations
APPROVED

Reviewed By :Christian Giraldo 09/01/2022
Supervised By :Jagrut Upadhyay 09/01/2022






09/01/2022
Supervised By :Jagrut
Upadhyay


09/01/2022



#64
Phenanthrene-d10
Concen: 20.000 ng
RT: 11.216 min Scan# 1518
Delta R.T. -0.006 min
Lab File: BF130051.D
Acq: 31 Aug 2022 14:32


Tgt Ion:188 Resp: 768446
Ion Ratio Lower Upper
188 100
94 12.0 7.4 11.2#
80 12.3 7.9 11.9#





#65
4,6-Dinitro-2-methylphenol
Concen: 4.470 ng
RT: 10.322 min Scan# 1378
Delta R.T. -0.018 min
Lab File: BF130051.D
Acq: 31 Aug 2022 14:32



Instrument :
BNA_F
ClientSampleId :



Tgt Ion:198 Resp: 2016
Ion Ratio Lower Upper
198 100
51 46.7 29.4 69.4
105 49.9 29.2 69.2


Manual Integrations
APPROVED

Reviewed By :Christian Giraldo 09/01/2022
Supervised By :Jagrut Upadhyay 09/01/2022






09/01/2022
Supervised By :Jagrut
Upadhyay


09/01/2022



#66
n-Nitrosodiphenylamine
Concen: 11.256 ng
RT: 10.392 min Scan# 1378
Delta R.T. -0.012 min
Lab File: BF130051.D
Acq: 31 Aug 2022 14:32


Tgt Ion:169 Resp: 291940
Ion Ratio Lower Upper
169 100
168 66.9 53.4 80.2
167 36.1 29.1 43.7





#67
4-Bromophenyl-phenylether
Concen: 10.293 ng
RT: 10.769 min Scan# 1442
Delta R.T. -0.006 min
Lab File: BF130051.D
Acq: 31 Aug 2022 14:32



Instrument :
BNA_F
ClientSampleId :



Tgt Ion:248 Resp: 92640
Ion Ratio Lower Upper
248 100
250 98.6 77.8 116.6
141 66.5 66.2 99.2


Manual Integrations
APPROVED

Reviewed By :Christian Giraldo 09/01/2022
Supervised By :Jagrut Upadhyay 09/01/2022




09/01/2022
Supervised By :Jagrut
Upadhyay



09/01/2022




#68
Hexachlorobenzene
Concen: 10.567 ng
RT: 10.822 min Scan# 1451
Delta R.T. -0.012 min
Lab File: BF130051.D
Acq: 31 Aug 2022 14:32



Tgt Ion:284 Resp: 103026
Ion Ratio Lower Upper
284 100
142 41.4 31.4 47.0
249 30.6 25.7 38.5





#69

Atrazine

Concen: 9.845 ng

RT: 10.922 min Scan# 1468

Delta R.T. -0.012 min


Lab File: BF130051.D

Acq: 31 Aug 2022 14:32

Instrument :

BNA_F



ClientSampleId :



Tgt Ion:	200	Resp:	78372
Ion Ratio	Lower	Upper	
200	100		
173	28.5	6.4	46.4
215	48.7	27.1	67.1


Manual Integrations
APPROVED

Reviewed By :Christian Giraldo 09/01/2022
Supervised By :Jagrut Upadhyay 09/01/2022



09/01/2022
Supervised By :Jagrut
Upadhyay

09/01/2022



#70

Pentachlorophenol

Concen: 20.115 ng




RT: 11.016 min Scan# 1484


Delta R.T. -0.006 min

Lab File: BF130051.D

Acq: 31 Aug 2022 14:32


Tgt Ion:	266	Resp:	49998
Ion Ratio	Lower	Upper	
266	100		
268	64.3	49.0	73.4
264	65.4	47.9	71.9





#71
Phenanthrene
Concen: 11.454 ng
RT: 11.239 min Scan# 1531
Delta R.T. -0.012 min
Lab File: BF130051.D
Acq: 31 Aug 2022 14:32



Instrument :
BNA_F
ClientSampleId :



Tgt Ion:178 Resp: 491590
Ion Ratio Lower Upper
178 100
176 19.2 15.5 23.3
179 15.5 12.5 18.7


Manual Integrations
APPROVED

Reviewed By :Christian Giraldo 09/01/2022
Supervised By :Jagrut Upadhyay 09/01/2022






09/01/2022
Supervised By :Jagrut
Upadhyay


09/01/2022



#72
Anthracene
Concen: 11.311 ng
RT: 11.292 min Scan# 1531
Delta R.T. -0.006 min
Lab File: BF130051.D
Acq: 31 Aug 2022 14:32


Tgt Ion:178 Resp: 482726
Ion Ratio Lower Upper
178 100
176 18.2 15.1 22.7
179 15.1 12.6 18.8





#73
 Carbazole
 Concen: 11.461 ng
 RT: 11.445 min Scan# 11
 Delta R.T. -0.012 min
 Lab File: BF130051.D
 Acq: 31 Aug 2022 14:32



Instrument :
 BNA_F
 ClientSampleId :



Tgt Ion:167 Resp: 444040
 Ion Ratio Lower Upper
 167 100
 166 21.4 17.1 25.7
 139 13.0 11.0 16.6


Manual Integrations
 APPROVED

Reviewed By :Christian Giraldo 09/01/2022
 Supervised By :Jagrut Upadhyay 09/01/2022




09/01/2022
 Supervised By :Jagrut
 Upadhyay



09/01/2022




#74
 Di-n-butylphthalate
 Concen: 10.036 ng
 RT: 11.786 min Scan# 1615
 Delta R.T. -0.006 min
 Lab File: BF130051.D
 Acq: 31 Aug 2022 14:32




Tgt Ion:149 Resp: 500424
 Ion Ratio Lower Upper
 149 100
 150 9.2 7.8 11.6
 104 4.9 4.2 6.4





#75
Fluoranthene
Concen: 11.128 ng
RT: 12.421 min Scan# 1725
Delta R.T. -0.012 min
Lab File: BF130051.D
Acq: 31 Aug 2022 14:32



Instrument :
BNA_F
ClientSampleId :



Tgt Ion:202 Resp: 48870
Ion Ratio Lower Upper
202 100
101 14.4 0.0 30.7
203 17.6 0.0 37.5


Manual Integrations
APPROVED

Reviewed By :Christian Giraldo 09/01/2022
Supervised By :Jagrut Upadhyay 09/01/2022




09/01/2022
Supervised By :Jagrut
Upadhyay



09/01/2022




#76
Chrysene-d12
Concen: 20.000 ng
RT: 13.845 min Scan# 1965
Delta R.T. -0.012 min
Lab File: BF130051.D
Acq: 31 Aug 2022 14:32




Tgt Ion:240 Resp: 547483
Ion Ratio Lower Upper
240 100
120 15.1 9.2 13.8#
236 26.1 20.7 31.1





#77
Benzidine
Concen: 11.351 ng
RT: 12.551 min Scan# 1745
Delta R.T. -0.006 min
Lab File: BF130051.D
Acq: 31 Aug 2022 14:32



Instrument :
BNA_F
ClientSampleId :



Tgt Ion:184 Resp: 16449
Ion Ratio Lower Upper
184 100
185 14.5 11.4 17.2
183 11.7 9.4 14.2


Manual Integrations
APPROVED

Reviewed By :Christian Giraldo 09/01/2022
Supervised By :Jagrut Upadhyay 09/01/2022






09/01/2022
Supervised By :Jagrut
Upadhyay


09/01/2022



#78
Pyrene
Concen: 10.612 ng
RT: 12.651 min Scan# 1762
Delta R.T. -0.006 min
Lab File: BF130051.D
Acq: 31 Aug 2022 14:32

Tgt Ion:202 Resp: 504897
Ion Ratio Lower Upper
202 100
200 20.7 17.0 25.6
203 17.2 14.1 21.1





#79

Terphenyl-d14

Concen: 19.953 ng

RT: 12.798 min Scan# 1789

Delta R.T. -0.012 min


Lab File: BF130051.D

Acq: 31 Aug 2022 14:32

Instrument :

BNA_F

ClientSampleId :



Tgt Ion:244 Resp: 656498

Ion Ratio Lower Upper

244 100

212 7.1 5.9 8.9


122 15.1 9.9 14.9

Manual Integrations

APPROVED

Reviewed By :Christian Giraldo 09/01/2022


Supervised By :Jagrut Upadhyay 09/01/2022



09/01/2022

Supervised By :Jagrut
Upadhyay

09/01/2022



#80

Butylbenzylphthalate

Concen: 8.371 ng

RT: 13.280 min Scan# 1869

Delta R.T. -0.006 min

Lab File: BF130051.D

Acq: 31 Aug 2022 14:32




Tgt Ion:149 Resp: 168523


Ion Ratio Lower Upper

149 100

91 71.9 55.4 83.0


206 17.8 15.8 23.6





#81
Benzo(a)anthracene
Concen: 10.959 ng
RT: 13.833 min Scan# 1963
Delta R.T. -0.012 min
Lab File: BF130051.D
Acq: 31 Aug 2022 14:32



Instrument :
BNA_F
ClientSampleId :



Tgt Ion:228 Resp: 41274
Ion Ratio Lower Upper
228 100
226 27.8 21.4 32.0
229 19.7 16.1 24.1


Manual Integrations
APPROVED

Reviewed By :Christian Giraldo 09/01/2022
Supervised By :Jagrut Upadhyay 09/01/2022




09/01/2022
Supervised By :Jagrut
Upadhyay



09/01/2022




#82
3,3'-Dichlorobenzidine
Concen: 10.618 ng
RT: 13.804 min Scan# 1958
Delta R.T. -0.012 min
Lab File: BF130051.D
Acq: 31 Aug 2022 14:32




Tgt Ion:252 Resp: 123853
Ion Ratio Lower Upper
252 100
254 63.1 51.4 77.2
126 13.5 10.2 15.4





#83
Chrysene
Concen: 11.352 ng
RT: 13.874 min Scan# 1971
Delta R.T. -0.006 min
Lab File: BF130051.D
Acq: 31 Aug 2022 14:32



Instrument :
BNA_F
ClientSampleId :



Tgt Ion:228 Resp: 400189
Ion Ratio Lower Upper
228 100
226 30.5 24.4 36.6
229 19.4 15.9 23.9


Manual Integrations
APPROVED

Reviewed By :Christian Giraldo 09/01/2022
Supervised By :Jagrut Upadhyay 09/01/2022




09/01/2022
Supervised By :Jagrut
Upadhyay



09/01/2022




#84
Bis(2-ethylhexyl)phthalate
Concen: 9.799 ng
RT: 13.845 min Scan# 1965
Delta R.T. -0.000 min
Lab File: BF130051.D
Acq: 31 Aug 2022 14:32




Tgt Ion:149 Resp: 234031
Ion Ratio Lower Upper
149 100
167 27.6 22.0 33.0
279 4.4 3.6 5.4





#85
Di-n-octyl phthalate
Concen: 10.111 ng
RT: 14.451 min Scan# 2069
Delta R.T. -0.006 min
Lab File: BF130051.D
Acq: 31 Aug 2022 14:32



Instrument :
BNA_F
ClientSampleId :



Tgt Ion:149 Resp: 33335
Ion Ratio Lower Upper
149 100
167 1.4 1.3 1.9
43 9.0 7.8 11.6


Manual Integrations
APPROVED

Reviewed By :Christian Giraldo 09/01/2022
Supervised By :Jagrut Upadhyay 09/01/2022






09/01/2022
Supervised By :Jagrut
Upadhyay


09/01/2022



#86
Perylene-d12
Concen: 20.000 ng
RT: 15.251 min Scan# 2204
Delta R.T. -0.000 min
Lab File: BF130051.D
Acq: 31 Aug 2022 14:32

Tgt Ion:264 Resp: 464494
Ion Ratio Lower Upper
264 100
260 23.8 19.1 28.7
265 21.7 16.7 25.1





#87

Indeno(1,2,3-cd)pyrene

Concen: 8.661 ng

RT: 16.586 min Scan# 2435

Delta R.T. -0.024 min

Lab File: BF130051.D

Acq: 31 Aug 2022 14:32

Instrument :

BNA_F

ClientSampleId :

Tgt Ion:276 Resp: 276194

Ion Ratio Lower Upper

276 100

138 31.0 19.8 29.8



277 24.9 20.4 30.6

Manual Integrations

APPROVED

Reviewed By :Christian Giraldo 09/01/2022



Supervised By :Jagrut Upadhyay 09/01/2022



09/01/2022

Supervised By :Jagrut Upadhyay

09/01/2022



#88

Benzo(b)fluoranthene

Concen: 10.616 ng

RT: 14.851 min Scan# 2136

Delta R.T. -0.006 min

Lab File: BF130051.D

Acq: 31 Aug 2022 14:32




Tgt Ion:252 Resp: 337656


Ion Ratio Lower Upper

252 100

253 21.9 17.5 26.3


125 10.9 8.3 12.5





#89
Benzo(k)fluoranthene
Concen: 10.869 ng m
RT: 14.874 min Scan# 2140
Delta R.T. -0.012 min
Lab File: BF130051.D
Acq: 31 Aug 2022 14:32



Instrument :
BNA_F
ClientSampleId :



Tgt Ion:252 Resp: 331519
Ion Ratio Lower Upper
252 100
253 21.5 17.1 25.7
125 12.7 8.3 12.5


Manual Integrations
APPROVED

Reviewed By :Christian Giraldo 09/01/2022
Supervised By :Jagrut Upadhyay 09/01/2022






09/01/2022
Supervised By :Jagrut
Upadhyay


09/01/2022



#90
Benzo(a)pyrene
Concen: 10.281 ng
RT: 15.186 min Scan# 2193
Delta R.T. -0.012 min
Lab File: BF130051.D
Acq: 31 Aug 2022 14:32


Tgt Ion:252 Resp: 257501
Ion Ratio Lower Upper
252 100
253 21.3 17.2 25.8
125 13.5 8.8 13.2





#91
Dibenzo(a,h)anthracene
Concen: 8.712 ng
RT: 16.609 min Scan# 2435
Delta R.T. -0.018 min
Lab File: BF130051.D
Acq: 31 Aug 2022 14:32



Instrument :
BNA_F
ClientSampleId :



Tgt Ion: 278 Resp: 229227
Ion Ratio Lower Upper
278 100
139 19.6 12.2 18.4
279 23.4 18.7 28.1


Manual Integrations
APPROVED

Reviewed By : Christian Giraldo 09/01/2022
Supervised By : Jagrut Upadhyay 09/01/2022




09/01/2022
Supervised By : Jagrut
Upadhyay

09/01/2022



#92
Benzo(g,h,i)perylene
Concen: 8.520 ng
RT: 16.992 min Scan# 2500
Delta R.T. -0.029 min
Lab File: BF130051.D
Acq: 31 Aug 2022 14:32



Tgt Ion: 276 Resp: 223648
Ion Ratio Lower Upper
276 100
277 23.7 18.2 27.4
138 26.4 15.8 23.8

