

Data Path : Z:\svoasrv\HPCHEM1\BNA\_F\Data\BF083123\  
 Data File : BF135046.D  
 Acq On : 31 Aug 2023 17:40  
 Operator : CG\JU  
 Sample : PB155213BL  
 Misc :  
 ALS Vial : 3 Sample Multiplier: 1

Instrument :  
 BNA\_F  
 ClientSampleId :  
 PB155213BL

Quant Time: Aug 31 19:13:47 2023  
 Quant Method : Z:\svoasrv\HPCHEM1\BNA\_F\Methods\8270-BF083023.M  
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION  
 QLast Update : Thu Aug 31 03:48:08 2023  
 Response via : Initial Calibration

| Compound                    | R.T.   | QIon | Response | Conc    | Units | Dev(Min) |
|-----------------------------|--------|------|----------|---------|-------|----------|
| -----                       |        |      |          |         |       |          |
| Internal Standards          |        |      |          |         |       |          |
| 1) 1,4-Dichlorobenzene-d4   | 6.769  | 152  | 131386   | 20.000  | ng    | 0.00     |
| 21) Naphthalene-d8          | 8.051  | 136  | 540370   | 20.000  | ng    | 0.00     |
| 39) Acenaphthene-d10        | 9.810  | 164  | 270209   | 20.000  | ng    | 0.00     |
| 64) Phenanthrene-d10        | 11.304 | 188  | 518035   | 20.000  | ng    | 0.00     |
| 76) Chrysene-d12            | 13.957 | 240  | 363140   | 20.000  | ng    | 0.00     |
| 86) Perylene-d12            | 15.427 | 264  | 303741   | 20.000  | ng    | 0.00     |
| System Monitoring Compounds |        |      |          |         |       |          |
| 5) 2-Fluorophenol           | 5.416  | 112  | 1094717  | 128.699 | ng    | 0.02     |
| 7) Phenol-d6                | 6.416  | 99   | 1348471  | 127.187 | ng    | 0.00     |
| 23) Nitrobenzene-d5         | 7.334  | 82   | 862010   | 89.621  | ng    | -0.01    |
| 42) 2,4,6-Tribromophenol    | 10.604 | 330  | 389993   | 136.059 | ng    | 0.00     |
| 45) 2-Fluorobiphenyl        | 9.133  | 172  | 1436358  | 86.546  | ng    | 0.00     |
| 79) Terphenyl-d14           | 12.904 | 244  | 1719351  | 77.265  | ng    | 0.00     |

| Target Compounds | Qvalue |
|------------------|--------|
| -----            |        |

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA\_F\Data\BF083123\  
Data File : BF135046.D  
Acq On : 31 Aug 2023 17:40  
Operator : CG\JU  
Sample : PB155213BL  
Misc :  
ALS Vial : 3 Sample Multiplier: 1

Instrument :  
BNA\_F  
ClientSampleId :  
PB155213BL

Quant Time: Aug 31 19:13:47 2023  
Quant Method : Z:\svoasrv\HPCHEM1\BNA\_F\Methods\8270-BF083023.M  
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION  
QLast Update : Thu Aug 31 03:48:08 2023  
Response via : Initial Calibration

