

Data Path : Z:\SVOASRV\HPCHEM1\BNA F\DATA\BF090119\
 Data File : BF116756.D
 Acq On : 1 Sep 2019 21:24
 Operator : HP/JU
 Sample : K4575-14
 Misc :
 ALS Vial : 19 Sample Multiplier: 1

Instrument :
 BNA_F
 ClientSampleId :

Quant Time: Sep 02 05:27:59 2019
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA F\METHODS\8270-BF083019.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Fri Aug 30 20:02:30 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	6.85	152	70262	20.00	ng	0.00
21) Naphthalene-d8	8.13	136	295887	20.00	ng	0.00
39) Acenaphthene-d10	9.88	164	168606	20.00	ng	0.00
64) Phenanthrene-d10	11.36	188	301333	20.00	ng	0.00
76) Chrysene-d12	13.99	240	191813	20.00	ng	0.00
87) Perylene-d12	15.44	264	192169	20.00	ng	0.00

System Monitoring Compounds

5) 2-Fluorophenol	5.46	112	474026	109.30	ng	0.00
7) Phenol-d6	6.47	99	659995	115.89	ng	0.00
23) Nitrobenzene-d5	7.40	82	393398	75.90	ng	0.00
42) 2,4,6-Tribromophenol	10.66	330	140183	124.37	ng	-0.01
45) 2-Fluorobiphenyl	9.20	172	731357	73.17	ng	0.00
79) Terphenyl-d14	12.94	244	753791	72.70	ng	0.00

Target Compounds


						Qvalue
10) Phenol	6.49	94	61018	9.676	ng	91
19) n-Nitroso-di-n-propylamine	7.40	70	51775	13.808	ng	# 70
25) Isophorone	7.40	82	393395	37.457	ng	# 66
50) Dimethylphthalate	9.59	163	92426	7.940	ng	98
52) Acenaphthene	9.91	154	24373	2.590	ng	99
56) 4-Nitrophenol	10.08	139	6654	3.487	ng	# 3
57) 2,4-Dinitrotoluene	9.88	165	21782	8.021	ng	# 20
58) Fluorene	10.42	166	27866	2.756	ng	96
65) 4,6-Dinitro-2-methylphenol	10.66	198	164	7.576	ng	# 1
67) 4-Bromophenyl-phenylether	10.66	248	9177	2.998	ng	# 1
71) Phenanthrene	11.39	178	439828	26.221	ng	100
72) Anthracene	11.43	178	93641	5.546	ng	98
73) Carbazole	11.59	167	61770	3.840	ng	98
75) Fluoranthene	12.57	202	635690	38.562	ng	99
78) Pyrene	12.80	202	526763	30.894	ng	99
81) Benzo(a)anthracene	13.98	228	222999	18.369	ng	98
83) Chrysene	14.02	228	220570	17.636	ng	97
86) Indeno(1,2,3-cd)pyrene	16.89	276	90781	9.037	ng	97
88) Benzo(b)fluoranthene	15.02	252	354029	30.472	ng	# 95
89) Benzo(k)fluoranthene	15.02	252	354029	33.420	ng	# 95
90) Benzo(a)pyrene	15.39	252	167977	16.618	ng	98
91) Dibenzo(a,h)anthracene	16.90	278	22335	2.524	ng	# 80
92) Benzo(g,h,i)perylene	17.32	276	83616	9.267	ng	# 91


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\SVOASRV\HPCHEM1\BNA F\DATA\BF090119\
Data File : BF116756.D
Acq On : 1 Sep 2019 21:24
Operator : HP/JU
Sample : K4575-14
Misc :
ALS Vial : 19 Sample Multiplier: 1

Instrument :
BNA_F
ClientSampled :

Quant Time: Sep 02 05:27:59 2019
Quant Method : Z:\SVOASRV\HPCHEM1\BNA F\METHODS\8270-BF083019.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Fri Aug 30 20:02:30 2019
Response via : Initial Calibration






#1

1,4-Dichlorobenzene-d4
Concen: 20.000 ng
RT: 6.85 min Scan# 809
Delta R.T. -0.01 min
Lab File: BF116756.D
Acq: 1 Sep 2019 21:24

Instrument :
BNA_F
ClientSampleId :

Tot Ion	Ratio	Lower	Upper
152	100		
150	156.7	124.4	186.6
115	65.2	49.9	74.9






Abundance

Ion 152.00 (151.70 to 152.70): E

Ion 150.00 (149.70 to 150.70): E


Ion 115.00 (114.70 to 115.70): E



#5

2-Fluorophenol
Concen: 109.298 ng
RT: 5.46 min Scan# 573
Delta R.T. -0.01 min
Lab File: BF116756.D
Acq: 1 Sep 2019 21:24

Tgt Ion	Ratio	Lower	Upper
112	100		
64	56.5	47.1	70.7
63	29.3	24.9	37.3





Abundance

Ion 112.00 (111.70 to 112.70): E

Ion 64.00 (63.70 to 64.70): BF1

Ion 63.00 (62.70 to 63.70): BF1





#7

Phenol-d6

Concen: 115.891 ng

RT: 6.47 min Scan# 746

Delta R.T. -0.01 min

Lab File: BF116756.D

Acq: 1 Sep 2019 21:24

Instrument :

BNA_F

ClientSampleId :


Tot Ion: 99 Resp: 659995

Ion Ratio Lower Upper

99 100

42 14.8 13.7 20.5




71 33.7 30.8 46.2



Abundance Ion 99.00 (98.70 to 99.70): BF1

Ion 42.00 (41.70 to 42.70): BF1

Ion 71.00 (70.70 to 71.70): BF1



#10

Phenol

Concen: 9.676 ng

RT: 6.49 min Scan# 748

Delta R.T. -0.01 min

Lab File: BF116756.D

Acq: 1 Sep 2019 21:24


Tgt Ion: 94 Resp: 61018

Ion Ratio Lower Upper

94 100

65 25.3 10.8 50.8



66 35.6 21.1 61.1




Abundance Ion 94.00 (93.70 to 94.70): BF1

Ion 65.00 (64.70 to 65.70): BF1

Ion 66.00 (65.70 to 66.70): BF1









#19
n-Nitroso-di-n-propylamine
Concen: 13.808 ng
RT: 7.40 min Scan# 904
Delta R.T. 0.14 min
Lab File: BF116756.D
Acq: 1 Sep 2019 21:24




Instrument :
BNA_F
ClientSampleId :


Tot Ion	Ratio	Lower	Upper
70	100		
42	35.3	43.5	65.3#
101	0.0	7.8	11.6#
130	2.5	16.6	24.8#



#21
Naphthalene-d8
Concen: 20.000 ng
RT: 8.13 min Scan# 1027
Delta R.T. -0.01 min
Lab File: BF116756.D
Acq: 1 Sep 2019 21:24

Tgt Ion	Ratio	Lower	Upper
136	100		
137	10.9	8.9	13.3
54	6.0	4.8	7.2
68	5.7	4.0	6.0









#23
Nitrobenzene-d5
Concen: 75.901 ng
RT: 7.40 min Scan# 904
Delta R.T. -0.01 min
Lab File: BF116756.D
Acq: 1 Sep 2019 21:24




Instrument :
BNA_F
ClientSampled :


Tot Ion	Ratio	Lower	Upper
82	100		
128	46.0	32.8	49.2
54	39.7	36.9	55.3



#25
Isophorone
Concen: 37.457 ng
RT: 7.40 min Scan# 904
Delta R.T. -0.26 min
Lab File: BF116756.D
Acq: 1 Sep 2019 21:24

Tgt Ion	Ratio	Lower	Upper
82	100		
95	0.1	5.8	8.6#
138	0.0	13.5	20.3#









#39
Acenaphthene-d10
Concen: 20.000 ng
RT: 9.88 min Scan# 1325
Delta R.T. -0.01 min
Lab File: BF116756.D
Acq: 1 Sep 2019 21:24




Instrument :
BNA_F
ClientSampleId :


Tot Ion:164 Resp: 168606
Ion Ratio Lower Upper
164 100
162 102.1 81.4 122.0
160 44.2 35.4 53.2



#42
2,4,6-Tribromophenol
Concen: 124.368 ng
RT: 10.66 min Scan# 1458
Delta R.T. -0.01 min
Lab File: BF116756.D
Acq: 1 Sep 2019 21:24

Tgt Ion:330 Resp: 140183
Ion Ratio Lower Upper
330 100
332 97.1 76.9 115.3
141 43.3 31.4 47.2









#45
 2-Fluorobiphenyl
 Concen: 73.172 ng
 RT: 9.20 min Scan# 1210
 Delta R.T. -0.00 min
 Lab File: BF116756.D
 Acq: 1 Sep 2019 21:24

Instrument :
 BNA_F
 ClientSampled :

Tot Ion:172 Resp: 731357




Ion	Ratio	Lower	Upper
172	100		
171	34.5	27.8	41.6
170	23.0	18.9	28.3




#50
 Dimethylphthalate
 Concen: 7.940 ng
 RT: 9.59 min Scan# 1276
 Delta R.T. -0.02 min
 Lab File: BF116756.D
 Acq: 1 Sep 2019 21:24

Tgt Ion:163 Resp: 92426

Ion	Ratio	Lower	Upper
163	100		
194	3.3	2.8	4.2
164	9.7	8.6	12.8





#52

Acenaphthene

Concen: 2.590 ng

RT: 9.91 min Scan# 1330

Delta R.T. -0.01 min

Lab File: BF116756.D

Acq: 1 Sep 2019 21:24

Instrument :

BNA_F

ClientSampleId :





Tot Ion:154 Resp: 24373

Ion Ratio Lower Upper

154 100

153 111.1 89.5 134.3

152 53.5 41.9 62.9



#56

4-Nitrophenol

Concen: 3.487 ng

RT: 10.08 min Scan# 1359

Delta R.T. 0.09 min

Lab File: BF116756.D

Acq: 1 Sep 2019 21:24




Tgt Ion:139 Resp: 6654


Ion Ratio Lower Upper

139 100

109 0.0 61.7 101.7#

65 3.7 89.1 129.1#









#57
2,4-Dinitrotoluene
Concen: 8.021 ng
RT: 9.88 min Scan# 1325
Delta R.T. -0.19 min
Lab File: BF116756.D
Acq: 1 Sep 2019 21:24

Instrument :
BNA_F
ClientSampleId :

Tot Ion	Ratio	Lower	Upper
165	100		
63	0.0	36.4	54.6#
89	1.4	62.6	94.0#
182	0.0	2.1	3.1#




Abundance Ion 165.00 (164.70 to 165.70): E
Ion 63.00 (62.70 to 63.70): BF1
Ion 89.00 (88.70 to 89.70): BF1
Ion 182.00 (181.70 to 182.70): E




#58
Fluorene
Concen: 2.756 ng
RT: 10.42 min Scan# 1417
Delta R.T. -0.01 min
Lab File: BF116756.D
Acq: 1 Sep 2019 21:24

Tgt Ion	Ratio	Lower	Upper
166	100		
165	102.9	78.8	118.2
167	13.8	10.0	15.0

Abundance Ion 166.00 (165.70 to 166.70): E
Ion 165.00 (164.70 to 165.70): E
Ion 167.00 (166.70 to 167.70): E





#64

Phenanthrene-d10

Concen: 20.000 ng

RT: 11.36 min Scan# 1577

Delta R.T. -0.01 min

Lab File: BF116756.D

Acq: 1 Sep 2019 21:24

Instrument :

BNA_F

ClientSampleId :





Tot Ion:188 Resp: 301333

Ion Ratio Lower Upper

188 100

94 10.4 7.3 10.9

80 11.0 8.4 12.6



#65

4,6-Dinitro-2-methylphenol

Concen: 7.576 ng

RT: 10.66 min Scan# 1458

Delta R.T. 0.19 min

Lab File: BF116756.D

Acq: 1 Sep 2019 21:24




Tgt Ion:198 Resp: 164


Ion Ratio Lower Upper

198 100

51 269.7 18.4 58.4#

105 322.2 22.9 62.9#





#67

4-Bromophenyl-phenylether

Concen: 2.998 ng

RT: 10.66 min Scan# 1458

Delta R.T. -0.25 min

Lab File: BF116756.D

Acq: 1 Sep 2019 21:24

Instrument :

BNA_F

ClientSampled :


Tot Ion:248 Resp: 9177

Ion Ratio Lower Upper

248 100

250 196.3 77.8 116.8#




141 650.5 75.9 113.9#



Abundance Ion 248.00 (247.70 to 248.70): E

Ion 250.00 (249.70 to 250.70): E

Ion 141.00 (140.70 to 141.70): E



#71

Phenanthrene

Concen: 26.221 ng

RT: 11.39 min Scan# 1581

Delta R.T. -0.01 min

Lab File: BF116756.D

Acq: 1 Sep 2019 21:24


Tgt Ion:178 Resp: 439828

Ion Ratio Lower Upper

178 100

176 19.0 15.4 23.0



179 15.2 12.2 18.2




Abundance Ion 178.00 (177.70 to 178.70): E

Ion 176.00 (175.70 to 176.70): E

Ion 179.00 (178.70 to 179.70): E





#72

Anthracene

Concen: 5.546 ng

RT: 11.43 min Scan# 1589

Delta R.T. -0.01 min

Lab File: BF116756.D

Acq: 1 Sep 2019 21:24

Instrument :

BNA_F

ClientSampleId :


Tot Ion:178 Resp: 93641

Ion Ratio Lower Upper

178 100

176 18.3 15.4 23.0




179 14.8 12.5 18.7



Abundance Ion 178.00 (177.70 to 178.70): E

Ion 176.00 (175.70 to 176.70): E

Ion 179.00 (178.70 to 179.70): E



#73

Carbazole

Concen: 3.840 ng

RT: 11.59 min Scan# 1615

Delta R.T. -0.01 min

Lab File: BF116756.D

Acq: 1 Sep 2019 21:24


Tgt Ion:167 Resp: 61770

Ion Ratio Lower Upper

167 100

166 20.7 17.8 26.6



139 13.7 11.1 16.7




Abundance Ion 167.00 (166.70 to 167.70): E

Ion 166.00 (165.70 to 166.70): E

Ion 139.00 (138.70 to 139.70): E





#75

Fluoranthene

Concen: 38.562 ng

RT: 12.57 min Scan# 1783

Delta R.T. 0.00 min

Lab File: BF116756.D

Acq: 1 Sep 2019 21:24

Instrument :

BNA_F

ClientSampleId :


Tot Ion:202 Resp: 635690

Ion	Ratio	Lower	Upper
202	100		
101	11.5	0.0	31.3
203	17.3	0.0	37.6

202 100

101 11.5 0.0 31.3




203 17.3 0.0 37.6



Abundance Ion 202.00 (201.70 to 202.70): E

Ion 101.00 (100.70 to 101.70): E

Ion 203.00 (202.70 to 203.70): E



#76

Chrysene-d12

Concen: 20.000 ng

RT: 13.99 min Scan# 2024

Delta R.T. -0.01 min

Lab File: BF116756.D

Acq: 1 Sep 2019 21:24


Tgt Ion:240 Resp: 191813

Ion	Ratio	Lower	Upper
240	100		
120	12.8	9.1	13.7
236	25.9	21.1	31.7

240 100

120 12.8 9.1 13.7



236 25.9 21.1 31.7




Abundance Ion 240.00 (239.70 to 240.70): E

Ion 120.00 (119.70 to 120.70): E

Ion 236.00 (235.70 to 236.70): E









#78
Pvrene
Concen: 30.894 ng
RT: 12.80 min Scan# 1821
Delta R.T. -0.01 min
Lab File: BF116756.D
Acq: 1 Sep 2019 21:24




Instrument :
BNA_F
ClientSampleId :


Tot Ion	Ratio	Lower	Upper
202	100		
200	20.8	16.5	24.7
203	17.8	13.7	20.5



#79
Terphenyl-d14
Concen: 72.699 ng
RT: 12.94 min Scan# 1846
Delta R.T. -0.01 min
Lab File: BF116756.D
Acq: 1 Sep 2019 21:24

Tgt Ion	Ratio	Lower	Upper
244	100		
212	6.8	6.1	9.1
122	12.0	10.2	15.4









#81
Benzo(a)anthracene
Concen: 18.369 ng
RT: 13.98 min Scan# 2022
Delta R.T. -0.01 min
Lab File: BF116756.D
Acq: 1 Sep 2019 21:24




Instrument :
BNA_F
ClientSampled :


Tot Ion	Ratio	Lower	Upper
228	100		
226	26.6	22.0	33.0
229	19.7	16.4	24.6



#83
Chrysene
Concen: 17.636 ng
RT: 14.02 min Scan# 2028
Delta R.T. -0.01 min
Lab File: BF116756.D
Acq: 1 Sep 2019 21:24

Tgt Ion	Ratio	Lower	Upper
228	100		
226	29.7	24.0	36.0
229	21.8	15.4	23.0









#86
Indeno(1,2,3-cd)pyrene
Concen: 9.037 ng
RT: 16.89 min Scan# 2516
Delta R.T. -0.01 min
Lab File: BF116756.D
Acq: 1 Sep 2019 21:24




Instrument :
BNA_F
ClientSampleId :


Tot Ion	Ratio	Lower	Upper
276	100		
138	24.9	19.4	29.2
277	22.9	20.5	30.7



#87
Perylene-d12
Concen: 20.000 ng
RT: 15.44 min Scan# 2271
Delta R.T. 0.00 min
Lab File: BF116756.D
Acq: 1 Sep 2019 21:24

Tgt Ion	Ratio	Lower	Upper
264	100		
260	22.0	18.3	27.5
265	22.7	16.5	24.7









#88
Benzo(b)fluoranthene
Concen: 30.472 ng
RT: 15.02 min Scan# 2199
Delta R.T. -0.01 min
Lab File: BF116756.D
Acq: 1 Sep 2019 21:24




Instrument :
BNA_F
ClientSampleId :


Tot Ion	Ratio	Lower	Upper
252	100		
253	22.7	17.7	26.5
125	15.2	8.2	12.2



#89
Benzo(k)fluoranthene
Concen: 33.420 ng
RT: 15.02 min Scan# 2199
Delta R.T. -0.04 min
Lab File: BF116756.D
Acq: 1 Sep 2019 21:24

Tgt Ion	Ratio	Lower	Upper
252	100		
253	22.7	17.7	26.5
125	15.2	8.6	13.0









#90
Benzo(a)pyrene
Concen: 16.618 ng
RT: 15.39 min Scan# 2261
Delta R.T. 0.00 min
Lab File: BF116756.D
Acq: 1 Sep 2019 21:24




Instrument :
BNA_F
ClientSampleId :


Tot Ion:252 Resp: 167977
Ion Ratio Lower Upper
252 100
253 23.3 18.1 27.1
125 13.4 9.8 14.8



#91
Dibenzo(a,h)anthracene
Concen: 2.524 ng
RT: 16.90 min Scan# 2518
Delta R.T. -0.01 min
Lab File: BF116756.D
Acq: 1 Sep 2019 21:24

Tgt Ion:278 Resp: 22335
Ion Ratio Lower Upper
278 100
139 28.4 12.4 18.6#
279 28.8 18.2 27.4#





#92

Benzo(a,h,i)perylene

Concen: 9.267 ng

RT: 17.32 min Scan# 2590

Delta R.T. -0.01 min

Lab File: BF116756.D

Acq: 1 Sep 2019 21:24

Instrument :

BNA_F

ClientSampleId :


Tot Ion:276 Resp: 83616

Ion Ratio Lower Upper

276 100

277 25.6 18.5 27.7

138 26.4 16.2 24.2#



Abundance

Ion 276.00 (275.70 to 276.70): E

Ion 277.00 (276.70 to 277.70): E

Ion 138.00 (137.70 to 138.70): E

