

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF090123\
 Data File : BF135083.D
 Acq On : 01 Sep 2023 18:46
 Operator : CG\JU
 Sample : 04189-02DL 10X
 Misc :
 ALS Vial : 18 Sample Multiplier: 1

Instrument :
 BNA_F
 ClientSampleId :
 SB-06(16-18)DL

Quant Time: Sep 02 00:28:17 2023
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF083023.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Thu Aug 31 03:48:08 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	6.769	152	75350	20.000	ng	0.00
21) Naphthalene-d8	8.051	136	259574	20.000	ng	# 0.00
39) Acenaphthene-d10	9.810	164	127967	20.000	ng	0.00
64) Phenanthrene-d10	11.298	188	218058	20.000	ng	-0.01
76) Chrysene-d12	13.951	240	188389	20.000	ng	-0.01
86) Perylene-d12	15.421	264	203057	20.000	ng	0.00
System Monitoring Compounds						
5) 2-Fluorophenol	5.387	112	49198	10.085	ng	-0.01
7) Phenol-d6	6.398	99	59844	9.842	ng	-0.02
23) Nitrobenzene-d5	7.328	82	57298	12.401	ng	-0.02
42) 2,4,6-Tribromophenol	10.598	330	9326	6.870	ng	-0.01
45) 2-Fluorobiphenyl	9.128	172	62454	7.946	ng	-0.01
79) Terphenyl-d14	12.892	244	65238	5.651	ng	-0.01
Target Compounds						
						Qvalue
37) 2-Methylnaphthalene	8.769	142	419680	49.605	ng	99
38) 1-Methylnaphthalene	8.869	142	326133	41.198	ng	99
58) Fluorene	10.357	166	19740	2.534	ng	# 91

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF090123\
Data File : BF135083.D
Acq On : 01 Sep 2023 18:46
Operator : CG\JU
Sample : 04189-02DL 10X
Misc :
ALS Vial : 18 Sample Multiplier: 1

Instrument :
BNA_F
ClientSampleId :
SB-06(16-18)DL

Quant Time: Sep 02 00:28:17 2023
Quant Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF083023.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Thu Aug 31 03:48:08 2023
Response via : Initial Calibration

