



Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF090221\
 Data File : BF125317.D
 Acq On : 2 Sep 2021 13:08
 Operator : JU/CG
 Sample : DFTPP
 Misc :
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 BNA_F
 ClientSampleId :
 DFTPP

Integration File: rteint.p

Method : Z:\SVOASRV\HPCHEM1\BNA_F\METHODS\8270-BF081821.M
 Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 Last Update : Fri Sep 03 11:56:34 2021






AutoFind: Scans 1631, 1632, 1633; Background Corrected with Scan 1626

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
51	198	10	80	41.8	96997	PASS
68	69	0.00	2	1.6	1288	PASS
69	198	0.00	100	35.3	82032	PASS
70	69	0.00	2	0.4	332	PASS
127	198	10	80	43.1	100221	PASS
197	198	0.00	2	0.4	948	PASS
198	198	100	100	100.0	232299	PASS
199	198	5	9	6.6	15353	PASS
275	198	10	60	27.9	64800	PASS
365	198	1	100	3.5	8023	PASS
441	198	0.01	100	17.6	40862	PASS
442	442	50	100	100.0	261411	PASS
443	442	15	24	18.8	49272	PASS

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF090221\
 Data File : BF125317.D
 Acq On : 2 Sep 2021 13:08
 Operator : JU/CG
 Sample : DFTPP
 Misc :
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 BNA_F
 ClientSampleId :
 DFTPP

Quant Time: Sep 02 13:46:24 2021
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_F\METHODS\8270-BF081821.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Tue Aug 24 12:39:21 2021
 Response via : Initial Calibration



TIC: BF125317.D\data.ms

(70) Pentachlorophenol (C)

11.245min (+ 0.006) 29912.43 ng




response 200225

Ion	Exp%	Act%
265.70	100.00	100.00
268.00	61.10	70.92
264.00	69.90	63.14
0.00	0.00	0.00

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF090221\
 Data File : BF125317.D
 Acq On : 2 Sep 2021 13:08
 Operator : JU/CG
 Sample : DFTPP
 Misc :
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 BNA_F
 ClientSampleId :
 DFTPP

Quant Time: Sep 02 13:46:24 2021
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_F\METHODS\8270-BF081821.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Tue Aug 24 12:39:21 2021
 Response via : Initial Calibration



TIC: BF125317.D\data.ms

(77) Benzidine

12.780min (+ 0.006) 269392.90 ng m

response 1016117

Ion	Exp%	Act%
184.00	100.00	100.00
185.00	13.50	14.36
183.00	10.10	11.60
0.00	0.00	0.00

DDT Breakdown

Date	Instrument Name	DFTPP Data File
9/2/2021	BNA_F	<u>BF125317.D</u>
Compound Name	Response	Retention Time
DDT	541144	13.6
DDD	9055	13.3
DDE	1506	12.96
SUM(DDD+DDE)	SUM(DDT+DDD+DDE)	% Breakdown Of DDT
10561	551705	1.91