

Data Path : Z:\SVOASRV\HPCHEM1\BNA F\DATA\BF090319\
 Data File : BF116784.D
 Acq On : 3 Sep 2019 20:33
 Operator : HP/JU
 Sample : K4554-19 2X
 Misc :
 ALS Vial : 22 Sample Multiplier: 1

Instrument :
 BNA_F
 ClientSampleId :

Quant Time: Sep 04 06:04:00 2019
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA F\METHODS\8270-BF083019.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Fri Aug 30 20:02:30 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	6.85	152	115472	20.00	ng	0.00
21) Naphthalene-d8	8.13	136	474535	20.00	ng	0.00
39) Acenaphthene-d10	9.88	164	250848	20.00	ng	0.00
64) Phenanthrene-d10	11.36	188	436863	20.00	ng	0.00
76) Chrysene-d12	13.99	240	312052	20.00	ng	0.00
87) Perylene-d12	15.45	264	332610	20.00	ng	0.00

System Monitoring Compounds

5) 2-Fluorophenol	5.46	112	420643	59.02	ng	0.00
7) Phenol-d6	6.47	99	580816	62.06	ng	0.00
23) Nitrobenzene-d5	7.40	82	340343	40.94	ng	0.00
42) 2,4,6-Tribromophenol	10.67	330	126120	75.21	ng	0.00
45) 2-Fluorobiphenyl	9.20	172	632635	42.54	ng	0.00
79) Terphenyl-d14	12.94	244	668072	39.61	ng	0.00

Target Compounds


						Qvalue
10) Phenol	6.48	94	55734	5.378	ng	99
19) n-Nitroso-di-n-propylamine	7.40	70	44708	7.255	ng	# 74
25) Isophorone	7.40	82	340339	20.206	ng	# 66
50) Dimethylphthalate	9.59	163	56969	3.289	ng	98
51) 2,6-Dinitrotoluene	9.88	165	34115	10.325	ng	# 21
56) 4-Nitrophenol	10.08	139	6726	2.369	ng	# 1
57) 2,4-Dinitrotoluene	9.88	165	34115	8.444	ng	# 21
71) Phenanthrene	11.39	178	388899	15.992	ng	99
72) Anthracene	11.44	178	133953	5.472	ng	99
75) Fluoranthene	12.57	202	470130	19.671	ng	99
78) Pyrene	12.80	202	412550	14.872	ng	98
81) Benzo(a)anthracene	13.99	228	191139	9.678	ng	96
83) Chrysene	14.02	228	166095	8.163	ng	98
86) Indeno(1,2,3-cd)pyrene	16.89	276	78457	4.801	ng	95
88) Benzo(b)fluoranthene	15.03	252	258066	12.833	ng	# 95
89) Benzo(k)fluoranthene	15.03	252	258066	14.075	ng	# 95
90) Benzo(a)pyrene	15.39	252	142188	8.127	ng	# 96
92) Benzo(g,h,i)perylene	17.33	276	63464	4.064	ng	# 91

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\SVOASRV\HPCHEM1\BNA F\DATA\BF090319\
Data File : BF116784.D
Acq On : 3 Sep 2019 20:33
Operator : HP/JU
Sample : K4554-19 2X
Misc :
ALS Vial : 22 Sample Multiplier: 1

Instrument :
BNA_F
ClientSampleId :

Quant Time: Sep 04 06:04:00 2019
Quant Method : Z:\SVOASRV\HPCHEM1\BNA F\METHODS\8270-BF083019.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Fri Aug 30 20:02:30 2019
Response via : Initial Calibration






#1

1,4-Dichlorobenzene-d4
Concen: 20.000 ng
RT: 6.85 min Scan# 809
Delta R.T. -0.01 min
Lab File: BF116784.D
Acq: 3 Sep 2019 20:33

Instrument :
BNA_F
ClientSampleId :

Tot Ion	Ratio	Lower	Upper
152	100		
150	152.9	124.4	186.6
115	61.9	49.9	74.9





Abundance

Ion 152.00 (151.70 to 152.70): E

Ion 150.00 (149.70 to 150.70): E


Ion 115.00 (114.70 to 115.70): E



#5

2-Fluorophenol
Concen: 59.016 ng
RT: 5.46 min Scan# 573
Delta R.T. -0.01 min
Lab File: BF116784.D
Acq: 3 Sep 2019 20:33

Tgt Ion	Ratio	Lower	Upper
112	100		
64	54.3	47.1	70.7
63	27.6	24.9	37.3






Abundance

Ion 112.00 (111.70 to 112.70): E

Ion 64.00 (63.70 to 64.70): BF1

Ion 63.00 (62.70 to 63.70): BF1





#7

Phenol-d6

Concen: 62.057 ng

RT: 6.47 min Scan# 746

Delta R.T. -0.01 min

Lab File: BF116784.D

Acq: 3 Sep 2019 20:33





Instrument :

BNA_F

ClientSampleId :

Tot Ion: 99 Resp: 580816

Ion	Ratio	Lower	Upper
99	100		
42	13.0	13.7	20.5#
71	30.9	30.8	46.2



#10

Phenol

Concen: 5.378 ng

RT: 6.48 min Scan# 747




Delta R.T. -0.01 min


Lab File: BF116784.D

Acq: 3 Sep 2019 20:33

Tgt Ion: 94 Resp: 55734

Ion	Ratio	Lower	Upper
94	100		
65	30.4	10.8	50.8
66	41.8	21.1	61.1








#19
n-Nitroso-di-n-propylamine
Concen: 7.255 ng
RT: 7.40 min Scan# 903
Delta R.T. 0.14 min
Lab File: BF116784.D
Acq: 3 Sep 2019 20:33




Instrument :
BNA_F
ClientSampleId :


Tot Ion	Ratio	Lower	Upper
70	100		
42	40.0	43.5	65.3#
101	0.0	7.8	11.6#
130	1.9	16.6	24.8#



#21
Naphthalene-d8
Concen: 20.000 ng
RT: 8.13 min Scan# 1027
Delta R.T. -0.01 min
Lab File: BF116784.D
Acq: 3 Sep 2019 20:33

Tgt Ion	Ratio	Lower	Upper
136	100		
137	10.9	8.9	13.3
54	6.2	4.8	7.2
68	6.0	4.0	6.0#









#23
Nitrobenzene-d5
Concen: 40.944 ng
RT: 7.40 min Scan# 904
Delta R.T. -0.01 min
Lab File: BF116784.D
Acq: 3 Sep 2019 20:33

Instrument :
BNA_F
ClientSampled :

Tot Ion	Ratio	Lower	Upper
82	100		
128	46.5	32.8	49.2
54	38.8	36.9	55.3




Abundance Ion 82.00 (81.70 to 82.70): BF1
Ion 128.00 (127.70 to 128.70): BF1
Ion 54.00 (53.70 to 54.70): BF1





#25
Isophorone
Concen: 20.206 ng
RT: 7.40 min Scan# 904
Delta R.T. -0.26 min
Lab File: BF116784.D
Acq: 3 Sep 2019 20:33

Tgt Ion	Ratio	Lower	Upper
82	100		
95	0.0	5.8	8.6#
138	0.0	13.5	20.3#



Abundance Ion 82.00 (81.70 to 82.70): BF1
Ion 95.00 (94.70 to 95.70): BF1
Ion 138.00 (137.70 to 138.70): BF1








#39
 Acenaphthene-d10
 Concen: 20.000 ng
 RT: 9.88 min Scan# 1325
 Delta R.T. -0.01 min
 Lab File: BF116784.D
 Acq: 3 Sep 2019 20:33




Instrument :
 BNA_F
 ClientSampleId :


Tot Ion:164 Resp: 250848
 Ion Ratio Lower Upper
 164 100
 162 103.2 81.4 122.0
 160 44.5 35.4 53.2



#42
 2,4,6-Tribromophenol
 Concen: 75.207 ng
 RT: 10.67 min Scan# 1459
 Delta R.T. -0.01 min
 Lab File: BF116784.D
 Acq: 3 Sep 2019 20:33

Tgt Ion:330 Resp: 126120
 Ion Ratio Lower Upper
 330 100
 332 97.0 76.9 115.3
 141 42.8 31.4 47.2









#45
2-Fluorobiphenyl
Concen: 42.543 ng
RT: 9.20 min Scan# 1210
Delta R.T. -0.01 min
Lab File: BF116784.D
Acq: 3 Sep 2019 20:33

Instrument :
BNA_F
ClientSampleId :

Tot Ion	Ratio	Lower	Upper
172	100		
171	35.5	27.8	41.6
170	23.4	18.9	28.3




Abundance Ion 172.00 (171.70 to 172.70): E
Ion 171.00 (170.70 to 171.70): E
Ion 170.00 (169.70 to 170.70): E






#50
Dimethylphthalate
Concen: 3.289 ng
RT: 9.59 min Scan# 1276
Delta R.T. -0.02 min
Lab File: BF116784.D
Acq: 3 Sep 2019 20:33

Tgt Ion	Ratio	Lower	Upper
163	100		
194	3.6	2.8	4.2
164	9.9	8.6	12.8



Abundance Ion 163.00 (162.70 to 163.70): E
Ion 194.00 (193.70 to 194.70): E
Ion 164.00 (163.70 to 164.70): E









#51
 2,6-Dinitrotoluene
 Concen: 10.325 ng
 RT: 9.88 min Scan# 1325
 Delta R.T. 0.22 min
 Lab File: BF116784.D
 Acq: 3 Sep 2019 20:33

Instrument :
 BNA_F
 ClientSampled :

Tot Ion	Ratio	Lower	Upper
165	100		
63	1.2	47.3	70.9#
89	1.0	52.5	78.7#




Abundance Ion 165.00 (164.70 to 165.70): E
 Ion 63.00 (62.70 to 63.70): BF1
 Ion 89.00 (88.70 to 89.70): BF1






#56
 4-Nitrophenol
 Concen: 2.369 ng
 RT: 10.08 min Scan# 1359
 Delta R.T. 0.09 min
 Lab File: BF116784.D
 Acq: 3 Sep 2019 20:33

Tgt Ion	Ratio	Lower	Upper
139	100		
109	0.0	61.7	101.7#
65	0.0	89.1	129.1#



Abundance Ion 139.00 (138.70 to 139.70): E
 Ion 109.00 (108.70 to 109.70): E
 Ion 65.00 (64.70 to 65.70): BF1






#57
 2,4-Dinitrotoluene
 Concen: 8.444 ng
 RT: 9.88 min Scan# 1325
 Delta R.T. -0.19 min
 Lab File: BF116784.D
 Acq: 3 Sep 2019 20:33

Instrument :
 BNA_F
 ClientSampleId :

Tot Ion	Ratio	Lower	Upper
165	100		
63	1.2	36.4	54.6#
89	1.0	62.6	94.0#
182	0.0	2.1	3.1#






Abundance Ion 165.00 (164.70 to 165.70): E

Ion 63.00 (62.70 to 63.70): BF1


Ion 89.00 (88.70 to 89.70): BF1

Ion 182.00 (181.70 to 182.70): E



#64
 Phenanthrene-d10
 Concen: 20.000 ng
 RT: 11.36 min Scan# 1577
 Delta R.T. -0.01 min
 Lab File: BF116784.D
 Acq: 3 Sep 2019 20:33



Tgt Ion	Ratio	Lower	Upper
188	100		
94	11.1	7.3	10.9#
80	12.6	8.4	12.6




Abundance Ion 188.00 (187.70 to 188.70): E

Ion 94.00 (93.70 to 94.70): BF1

Ion 80.00 (79.70 to 80.70): BF1









#71
Phenanthrene
Concen: 15.992 ng
RT: 11.39 min Scan# 1581
Delta R.T. -0.01 min
Lab File: BF116784.D
Acq: 3 Sep 2019 20:33




Instrument :
BNA_F
ClientSampleId :


Tot Ion:178 Resp: 388899
Ion Ratio Lower Upper
178 100
176 18.9 15.4 23.0
179 15.5 12.2 18.2



#72
Anthracene
Concen: 5.472 ng
RT: 11.44 min Scan# 1590
Delta R.T. -0.00 min
Lab File: BF116784.D
Acq: 3 Sep 2019 20:33

Tgt Ion:178 Resp: 133953
Ion Ratio Lower Upper
178 100
176 19.1 15.4 23.0
179 15.1 12.5 18.7





#75

Fluoranthene

Concen: 19.671 ng

RT: 12.57 min Scan# 1783

Delta R.T. 0.00 min

Lab File: BF116784.D

Acq: 3 Sep 2019 20:33

Instrument :

BNA_F

ClientSampleId :


Tot Ion:202 Resp: 470130

Ion Ratio Lower Upper

202 100

101 12.6 0.0 31.3




203 17.7 0.0 37.6



Abundance Ion 202.00 (201.70 to 202.70): E

Ion 101.00 (100.70 to 101.70): E

Ion 203.00 (202.70 to 203.70): E



#76

Chrysene-d12

Concen: 20.000 ng

RT: 13.99 min Scan# 2024

Delta R.T. -0.01 min

Lab File: BF116784.D

Acq: 3 Sep 2019 20:33


Tgt Ion:240 Resp: 312052

Ion Ratio Lower Upper

240 100

120 14.1 9.1 13.7#



236 25.5 21.1 31.7




Abundance Ion 240.00 (239.70 to 240.70): E

Ion 120.00 (119.70 to 120.70): E

Ion 236.00 (235.70 to 236.70): E






#78
Pvrene
Concen: 14.872 ng
RT: 12.80 min Scan# 1821
Delta R.T. -0.01 min
Lab File: BF116784.D
Acq: 3 Sep 2019 20:33

Instrument :
BNA_F
ClientSampled :

Tot Ion	Ratio	Lower	Upper
202	100		
200	21.0	16.5	24.7
203	18.3	13.7	20.5






Abundance

Ion 202.00 (201.70 to 202.70): E


Ion 200.00 (199.70 to 200.70): E

Ion 203.00 (202.70 to 203.70): E



#79
Terphenyl-d14
Concen: 39.605 ng
RT: 12.94 min Scan# 1846
Delta R.T. -0.01 min
Lab File: BF116784.D
Acq: 3 Sep 2019 20:33

Tgt Ion	Ratio	Lower	Upper
244	100		
212	6.6	6.1	9.1
122	12.6	10.2	15.4






Abundance

Ion 244.00 (243.70 to 244.70): E

Ion 212.00 (211.70 to 212.70): E

Ion 122.00 (121.70 to 122.70): E









#81
Benzo(a)anthracene
Concen: 9.678 ng
RT: 13.99 min Scan# 2023
Delta R.T. -0.01 min
Lab File: BF116784.D
Acq: 3 Sep 2019 20:33




Instrument :
BNA_F
ClientSampleId :


Tot Ion	Ratio	Lower	Upper
228	100		
226	29.8	22.0	33.0
229	19.0	16.4	24.6



#83
Chrysene
Concen: 8.163 ng
RT: 14.02 min Scan# 2029
Delta R.T. -0.01 min
Lab File: BF116784.D
Acq: 3 Sep 2019 20:33

Tgt Ion	Ratio	Lower	Upper
228	100		
226	30.3	24.0	36.0
229	21.3	15.4	23.0









#86
Indeno(1,2,3-cd)pyrene
Concen: 4.801 ng
RT: 16.89 min Scan# 2517
Delta R.T. 0.00 min
Lab File: BF116784.D
Acq: 3 Sep 2019 20:33

Instrument :
BNA_F
ClientSampled :

Tot Ion	Ratio	Lower	Upper
276	100		
138	26.1	19.4	29.2
277	22.8	20.5	30.7




Abundance Ion 276.00 (275.70 to 276.70): E
Ion 138.00 (137.70 to 138.70): E
Ion 277.00 (276.70 to 277.70): E






#87
Perylene-d12
Concen: 20.000 ng
RT: 15.45 min Scan# 2272
Delta R.T. 0.01 min
Lab File: BF116784.D
Acq: 3 Sep 2019 20:33

Tgt Ion	Ratio	Lower	Upper
264	100		
260	23.4	18.3	27.5
265	21.7	16.5	24.7



Abundance Ion 264.00 (263.70 to 264.70): E
Ion 260.00 (259.70 to 260.70): E
Ion 265.00 (264.70 to 265.70): E









#88
Benzo(b)fluoranthene
Concen: 12.833 ng
RT: 15.03 min Scan# 2200
Delta R.T. 0.00 min
Lab File: BF116784.D
Acq: 3 Sep 2019 20:33

Instrument :
BNA_F
ClientSampleId :

Tot Ion	Ratio	Lower	Upper
252	100		
253	22.9	17.7	26.5
125	15.0	8.2	12.2




Abundance Ion 252.00 (251.70 to 252.70): E
Ion 253.00 (252.70 to 253.70): E
Ion 125.00 (124.70 to 125.70): E






#89
Benzo(k)fluoranthene
Concen: 14.075 ng
RT: 15.03 min Scan# 2200
Delta R.T. -0.03 min
Lab File: BF116784.D
Acq: 3 Sep 2019 20:33

Tgt Ion	Ratio	Lower	Upper
252	100		
253	22.9	17.7	26.5
125	15.0	8.6	13.0



Abundance Ion 252.00 (251.70 to 252.70): E
Ion 253.00 (252.70 to 253.70): E
Ion 125.00 (124.70 to 125.70): E









#90
Benzo(a)pyrene
Concen: 8.127 ng
RT: 15.39 min Scan# 2261
Delta R.T. 0.00 min
Lab File: BF116784.D
Acq: 3 Sep 2019 20:33

Instrument :
BNA_F
ClientSampleId :

Tot Ion	Ratio	Lower	Upper
252	100		
253	23.4	18.1	27.1
125	15.1	9.8	14.8



#92
Benzo(g,h,i)perylene
Concen: 4.064 ng
RT: 17.33 min Scan# 2591
Delta R.T. 0.00 min
Lab File: BF116784.D
Acq: 3 Sep 2019 20:33

Tgt Ion	Ratio	Lower	Upper
276	100		
277	23.1	18.5	27.7
138	29.5	16.2	24.2

