

Data Path : Z:\SVOASRV\HPCHEM1\BNA F\DATA\BF090419\
 Data File : BF116806.D
 Acq On : 4 Sep 2019 11:03
 Operator : HP/JU
 Sample : K4623-08
 Misc :
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 BNA_F
 ClientSampleId :
 PAR-11-D1-D4-(0-4)

Quant Time: Sep 05 02:29:11 2019
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA F\METHODS\8270-BF083019.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Fri Aug 30 20:02:30 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	6.85	152	65743	20.00	ng	0.00
21) Naphthalene-d8	8.13	136	263202	20.00	ng	0.00
39) Acenaphthene-d10	9.88	164	138028	20.00	ng	0.00
64) Phenanthrene-d10	11.36	188	225095	20.00	ng	0.00
76) Chrysene-d12	13.99	240	187990	20.00	ng	0.00
87) Perylene-d12	15.44	264	183698	20.00	ng	0.00
System Monitoring Compounds						
5) 2-Fluorophenol	5.46	112	434878	107.16	ng	0.00
7) Phenol-d6	6.47	99	596139	111.87	ng	0.00
23) Nitrobenzene-d5	7.40	82	366033	79.39	ng	0.00
42) 2,4,6-Tribromophenol	10.66	330	127760	138.46	ng	-0.01
45) 2-Fluorobiphenyl	9.20	172	518343	63.35	ng	-0.01
79) Terphenyl-d14	12.95	244	508430	50.03	ng	0.00
Target Compounds						
50) Dimethylphthalate	9.59	163	90562	9.503	ng	99
75) Fluoranthene	12.57	202	29494	2.395	ng	85

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\SVOASRV\HPCHEM1\BNA F\DATA\BF090419\
Data File : BF116806.D
Acq On : 4 Sep 2019 11:03
Operator : HP/JU
Sample : K4623-08
Misc :
ALS Vial : 3 Sample Multiplier: 1

Instrument :
BNA_F
ClientSampleId :
PAR-11-D1-D4-(0-4)

Quant Time: Sep 05 02:29:11 2019
Quant Method : Z:\SVOASRV\HPCHEM1\BNA F\METHODS\8270-BF083019.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Fri Aug 30 20:02:30 2019
Response via : Initial Calibration

