

Data Path : Z:\SVOASRV\HPCHEM1\BNA F\DATA\BF090419\
 Data File : BF116824.D
 Acq On : 4 Sep 2019 19:09
 Operator : HP/JU
 Sample : K4655-03
 Misc :
 ALS Vial : 21 Sample Multiplier: 1

Instrument :
 BNA_F
 ClientSampleId :
 TP-15-S

Quant Time: Sep 05 04:24:17 2019
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA F\METHODS\8270-BF083019.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Fri Aug 30 20:02:30 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	6.85	152	59587	20.00	ng	0.00
21) Naphthalene-d8	8.13	136	240000	20.00	ng	0.00
39) Acenaphthene-d10	9.88	164	124496	20.00	ng	0.00
64) Phenanthrene-d10	11.36	188	203858	20.00	ng	0.00
76) Chrysene-d12	14.00	240	143983	20.00	ng	0.00
87) Perylene-d12	15.46	264	139849	20.00	ng	0.01

System Monitoring Compounds	R.T.	QIon	Response	Conc	Units	Dev(Min)
5) 2-Fluorophenol	5.46	112	328242	89.24	ng	0.00
7) Phenol-d6	6.48	99	444234	91.98	ng	0.00
23) Nitrobenzene-d5	7.40	82	271347	64.54	ng	0.00
42) 2,4,6-Tribromophenol	10.67	330	86730	104.21	ng	0.00
45) 2-Fluorobiphenyl	9.20	172	473244	64.12	ng	0.00
79) Terphenyl-d14	12.94	244	442775	56.89	ng	0.00

Target Compounds	R.T.	QIon	Response	Conc	Units	Qvalue
10) Phenol	6.49	94	13708	2.563	ng	93
50) Dimethylphthalate	9.59	163	50905	5.923	ng	# 99
52) Acenaphthene	9.92	154	22070	3.176	ng	96
58) Fluorene	10.43	166	34867	4.671	ng	95
71) Phenanthrene	11.39	178	88132	7.766	ng	98
72) Anthracene	11.44	178	37446	3.278	ng	96
75) Fluoranthene	12.57	202	107603	9.648	ng	100
78) Pyrene	12.80	202	97540	7.621	ng	99
81) Benzo(a)anthracene	13.99	228	31504	3.457	ng	93
83) Chrysene	14.02	228	27498	2.929	ng	# 93
88) Benzo(b)fluoranthene	15.03	252	23646m	2.797	ng	
90) Benzo(a)pyrene	15.39	252	19184	2.608	ng	# 75

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\SVOASRV\HPCHEM1\BNA F\DATA\BF090419\
Data File : BF116824.D
Acq On : 4 Sep 2019 19:09
Operator : HP/JU
Sample : K4655-03
Misc :
ALS Vial : 21 Sample Multiplier: 1

Instrument :
BNA_F
ClientSampleId :
TP-15-S

Quant Time: Sep 05 04:24:17 2019
Quant Method : Z:\SVOASRV\HPCHEM1\BNA F\METHODS\8270-BF083019.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Fri Aug 30 20:02:30 2019
Response via : Initial Calibration

