

Data Path : Z:\SVOASRV\HPCHEM1\BNA F\DATA\BF090518\
 Data File : BF108789.D
 Acq On : 5 Sep 2018 22:07
 Operator : JU/SJ
 Sample : J4741-19 10X
 Misc :
 ALS Vial : 9 Sample Multiplier: 1

Instrument :
 BNA_F
 ClientSampleId :

Quant Time: Sep 06 05:27:17 2018
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA F\METHODS\8270-BF090518.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Wed Sep 05 18:03:31 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	6.73	152	293825	20.00	ng	0.00
21) Naphthalene-d8	8.01	136	1094409	20.00	ng	0.00
38) Acenaphthene-d10	9.77	164	456914	20.00	ng	0.00
63) Phenanthrene-d10	11.24	188	734252	20.00	ng	0.00
75) Chrysene-d12	13.86	240	623326	20.00	ng	0.00
86) Perylene-d12	15.27	264	491150	20.00	ng	0.00

System Monitoring Compounds

5) 2-Fluorophenol	5.32	112	110993	6.26	ng	-0.01
7) Phenol-d6	6.35	99	142829	6.26	ng	-0.02
23) Nitrobenzene-d5	7.28	82	74940	4.29	ng	-0.02
41) 2,4,6-Tribromophenol	10.54	330	18768	4.30	ng	-0.01
44) 2-Fluorobiphenyl	9.09	172	158948	5.93	ng	0.00
78) Terphenyl-d14	12.82	244	127826	4.78	ng	0.00

Target Compounds

						Qvalue
2) 1,4-Dioxane	2.44	88	122	0.014	ng	# 1
3) Pyridine	2.87	79	117	0.005	ng	# 1
4) n-Nitrosodimethylamine	3.08	42	118	0.013	ng	# 1
6) Aniline	6.50	93	122	0.004	ng	# 1
9) Benzaldehyde	6.27	77	119	0.007	ng	# 51
10) Phenol	6.37	94	543	0.021	ng	# 79
11) bis(2-Chloroethyl)ether	6.50	93	122	0.006	ng	# 1
15) Benzyl Alcohol	6.88	79	1614	0.092	ng	# 48
16) 2,2'-oxybis(1-Chloropropan	6.97	45	123	0.004	ng	# 70
19) n-Nitroso-di-n-propylamine	7.28	70	11132	0.702	ng	# 75
22) Acetophenone	7.14	105	548	0.022	ng	# 54
24) Nitrobenzene	7.27	77	114	0.006	ng	# 38
25) Isophorone	7.28	82	74748	2.143	ng	# 64
28) bis(2-Chloroethoxy)methane	8.02	93	150	0.006	ng	# 1
31) Naphthalene	8.04	128	1312	0.027	ng	# 68
32) Benzoic acid	8.01	122	120	4.030	ng	# 1
36) 4-Chloro-3-methylphenol	8.62	107	223	0.014	ng	# 20
37) 2-Methylnaphthalene	8.72	142	939	0.029	ng	# 91
45) 1,1'-Biphenyl	9.19	154	701	0.020	ng	# 81
46) 2-Chloronaphthalene	9.48	162	230	0.008	ng	# 1
47) 2-Nitroaniline	9.48	65	818	0.105	ng	# 1
48) Acenaphthylene	9.62	152	4797	0.113	ng	# 92
49) Dimethylphthalate	9.48	163	41031	1.284	ng	# 98
50) 2,6-Dinitrotoluene	9.76	165	58038	9.531	ng	# 25
51) Acenaphthene	9.80	154	3060	0.117	ng	# 84
54) Dibenzofuran	9.97	168	2417	0.063	ng	# 98
55) 4-Nitrophenol	9.97	139	896	0.165	ng	# 9
56) 2,4-Dinitrotoluene	9.76	165	58038	7.453	ng	# 23
57) Fluorene	10.31	166	4847	0.197	ng	# 95
59) Diethylphthalate	10.18	149	6560	0.199	ng	# 93
62) Azobenzene	10.47	77	450	0.013	ng	# 1
65) n-Nitrosodiphenylamine	10.42	169	261	0.011	ng	# 24
66) 4-Bromophenyl-phenylether	10.54	248	1336	0.161	ng	# 1
70) Phenanthrene	11.26	178	54098	1.543	ng	# 99

Data Path : Z:\SVOASRV\HPCHEM1\BNA F\DATA\BF090518\
 Data File : BF108789.D
 Acq On : 5 Sep 2018 22:07
 Operator : JU/SJ
 Sample : J4741-19 10X
 Misc :
 ALS Vial : 9 Sample Multiplier: 1

Instrument :
 BNA_F
 ClientSampleId :

Quant Time: Sep 06 05:27:17 2018
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA F\METHODS\8270-BF090518.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Wed Sep 05 18:03:31 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)


71) Anthracene	11.31	178	14971	0.446	nq	97
72) Carbazole	11.46	167	5340	0.151	nq	99
73) Di-n-butylphthalate	11.81	149	2322	0.059	nq	# 66
74) Fluoranthene	12.44	202	164503	4.738	nq	98
76) Benzidine	12.55	184	4348	0.147	nq	# 69
77) Pyrene	12.67	202	159409	3.324	nq	98
79) Butylbenzylphthalate	13.30	149	6395	0.296	nq	89
80) Benzo(a)anthracene	13.85	228	87465	2.189	nq	94
81) 3,3'-Dichlorobenzidine	13.79	252	810	0.052	nq	# 29
82) Chrysene	13.89	228	92051	2.380	nq	97
83) Bis(2-ethylhexyl)phthalate	13.87	149	217742	8.022	nq	100
84) Di-n-octyl phthalate	14.48	149	751	0.017	nq	# 47
85) Indeno(1,2,3-cd)pyrene	16.61	276	42937	1.254	nq	95
87) Benzo(b)fluoranthene	14.87	252	144116	5.396	nq	# 89
88) Benzo(k)fluoranthene	14.87	252	144116	5.860	nq	# 92
89) Benzo(a)pyrene	15.21	252	66766	2.771	nq	# 90
90) Dibenzo(a,h)anthracene	16.63	278	13346	0.629	nq	# 57
91) Benzo(g,h,i)perylene	17.02	276	44169	2.078	nq	# 94


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\SVOASRV\HPCHEM1\BNA F\DATA\BF090518\
Data File : BF108789.D
Acq On : 5 Sep 2018 22:07
Operator : JU/SJ
Sample : J4741-19 10X
Misc :
ALS Vial : 9 Sample Multiplier: 1

Instrument :
BNA_F
ClientSampled :

Quant Time: Sep 06 05:27:17 2018
Quant Method : Z:\SVOASRV\HPCHEM1\BNA F\METHODS\8270-BF090518.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Wed Sep 05 18:03:31 2018
Response via : Initial Calibration






#1




1,4-Dichlorobenzene-d4
Concen: 20.000 ng
RT: 6.73 min Scan# 747
Delta R.T. -0.01 min
Lab File: BF108789.D
Acq: 5 Sep 2018 22:07

Instrument :
BNA_F
ClientSampleId :

Tot Ion:152 Resp: 293825
Ion Ratio Lower Upper
152 100
150 159.5 124.6 186.8
115 64.1 50.1 75.1




Abundance Ion 152.00 (151.70 to 152.70): E
Ion 150.00 (149.70 to 150.70): E
Ion 115.00 (114.70 to 115.70): E





#2


1,4-Dioxane
Concen: 0.014 ng
RT: 2.44 min Scan# 18
Delta R.T. -0.00 min
Lab File: BF108789.D
Acq: 5 Sep 2018 22:07

Tgt Ion: 88 Resp: 122
Ion Ratio Lower Upper
88 100
58 0.0 62.6 93.8#
43 289.3 24.6 37.0#



Abundance Ion 88.00 (87.70 to 88.70): BF1
Ion 58.00 (57.70 to 58.70): BF1
Ion 43.00 (42.70 to 43.70): BF1





#3

Pvridine

Concen: 0.005 ng

RT: 2.87 min Scan# 91

Delta R.T. -0.26 min

Lab File: BF108789.D

Acq: 5 Sep 2018 22:07

Instrument :

BNA_F

ClientSampleId :


Tot Ion: 79 Resp: 117

Ion Ratio Lower Upper




79 100

52 0.0 59.8 89.6#

51 165.4 29.8 44.8#



Abundance Ion 79.00 (78.70 to 79.70): BF1
Ion 52.00 (51.70 to 52.70): BF1
Ion 51.00 (50.70 to 51.70): BF1



#4

n-Nitrosodimethylamine

Concen: 0.013 ng

RT: 3.08 min Scan# 126

Delta R.T. -0.00 min

Lab File: BF108789.D

Acq: 5 Sep 2018 22:07


Tgt Ion: 42 Resp: 118

Ion Ratio Lower Upper



42 100


74 0.0 104.9 157.3#

44 670.4 6.9 10.3#



Abundance Ion 42.00 (41.70 to 42.70): BF1
Ion 74.00 (73.70 to 74.70): BF1
Ion 44.00 (43.70 to 44.70): BF1





#5

2-Fluorophenol

Concen: 6.258 ng

RT: 5.32 min Scan# 508

Delta R.T. -0.01 min

Lab File: BF108789.D

Acq: 5 Sep 2018 22:07

Instrument :
BNA_F
ClientSampled :


Tot Ion:112 Resp: 110993

Ion Ratio Lower Upper

112 100

64 58.5 47.9 71.9




63 28.3 23.7 35.5



Abundance Ion 112.00 (111.70 to 112.70): E

Ion 64.00 (63.70 to 64.70): BF1

Ion 63.00 (62.70 to 63.70): BF1



#6

Aniline

Concen: 0.004 ng

RT: 6.50 min Scan# 707

Delta R.T. 0.10 min

Lab File: BF108789.D

Acq: 5 Sep 2018 22:07


Tgt Ion: 93 Resp: 122

Ion Ratio Lower Upper

93 100

66 10918.7 32.2 48.2#



65 879.8 16.4 24.6#




Abundance Ion 93.00 (92.70 to 93.70): BF1

Ion 66.00 (65.70 to 66.70): BF1

Ion 65.00 (64.70 to 65.70): BF1





#7

Phenol-d6

Concen: 6.259 ng

RT: 6.35 min Scan# 682

Delta R.T. -0.02 min

Lab File: BF108789.D

Acq: 5 Sep 2018 22:07

Instrument :

BNA_F

ClientSampled :


Tot Ion: 99 Resp: 142829

Ion Ratio Lower Upper




99 100

42 17.8 14.5 21.7

71 31.6 25.4 38.0



Abundance Ion 99.00 (98.70 to 99.70): BF1
Ion 42.00 (41.70 to 42.70): BF1
Ion 71.00 (70.70 to 71.70): BF1



#9

Benzaldehyde

Concen: 0.007 ng

RT: 6.27 min Scan# 669

Delta R.T. -0.01 min

Lab File: BF108789.D

Acq: 5 Sep 2018 22:07


Tgt Ion: 77 Resp: 119

Ion Ratio Lower Upper



77 100


105 97.3 81.1 121.1

106 0.0 74.5 114.5#



Abundance Ion 77.00 (76.70 to 77.70): BF1
Ion 105.00 (104.70 to 105.70): BF1
Ion 106.00 (105.70 to 106.70): BF1





#10

Phenol

Concen: 0.021 ng

RT: 6.37 min Scan# 685

Delta R.T. -0.01 min

Lab File: BF108789.D

Acq: 5 Sep 2018 22:07

Instrument :

BNA_F

ClientSampleId :


Tot Ion: 94 Resp: 543

Ion Ratio Lower Upper

94 100

65 44.4 7.9 47.9




66 43.9 16.2 56.2



Abundance Ion 94.00 (93.70 to 94.70): BF1

Ion 65.00 (64.70 to 65.70): BF1

Ion 66.00 (65.70 to 66.70): BF1



#11

bis(2-Chloroethyl)ether

Concen: 0.006 ng

RT: 6.50 min Scan# 707

Delta R.T. 0.02 min

Lab File: BF108789.D

Acq: 5 Sep 2018 22:07


Tgt Ion: 93 Resp: 122

Ion Ratio Lower Upper

93 100

63 373.2 58.6 98.6#



95 150.7 11.8 51.8#




Abundance Ion 93.00 (92.70 to 93.70): BF1

Ion 63.00 (62.70 to 63.70): BF1

Ion 95.00 (94.70 to 95.70): BF1





#15

Benzyl Alcohol

Concen: 0.092 ng

RT: 6.88 min Scan# 773

Delta R.T. 0.01 min

Lab File: BF108789.D

Acq: 5 Sep 2018 22:07

Instrument :

BNA_F

ClientSampleId :


Tot Ion: 79 Resp: 1614

Ion Ratio Lower Upper




79 100

108 23.0 65.1 97.7#

77 89.7 50.6 75.8#



Abundance Ion 79.00 (78.70 to 79.70): BF1
Ion 108.00 (107.70 to 108.70): BF1
Ion 77.00 (76.70 to 77.70): BF1



#16

2,2'-oxybis(1-Chloropropane)

Concen: 0.004 ng

RT: 6.97 min Scan# 787

Delta R.T. -0.05 min

Lab File: BF108789.D

Acq: 5 Sep 2018 22:07


Tgt Ion: 45 Resp: 123

Ion Ratio Lower Upper



45 100


77 0.0 0.0 32.9

79 0.0 0.0 29.6



Abundance Ion 45.00 (44.70 to 45.70): BF1
Ion 77.00 (76.70 to 77.70): BF1
Ion 79.00 (78.70 to 79.70): BF1









#19
n-Nitroso-di-n-propylamine
Concen: 0.702 ng
RT: 7.28 min Scan# 841
Delta R.T. 0.12 min
Lab File: BF108789.D
Acq: 5 Sep 2018 22:07




Instrument :
BNA_F
ClientSampleId :


Tot Ion:	70	Resp:	11132
Ion	Ratio	Lower	Upper
70	100		
42	44.9	47.5	71.3#
101	0.0	7.4	11.2#
130	2.1	16.6	25.0#



#21
Naphthalene-d8
Concen: 20.000 ng
RT: 8.01 min Scan# 965
Delta R.T. -0.01 min
Lab File: BF108789.D
Acq: 5 Sep 2018 22:07

Tgt Ion:	136	Resp:	1094409
Ion	Ratio	Lower	Upper
136	100		
137	11.6	8.9	13.3
54	8.3	6.6	9.8
68	7.1	5.0	7.4





#22

Acetophenone

Concen: 0.022 ng

RT: 7.14 min Scan# 816

Delta R.T. -0.01 min

Lab File: BF108789.D

Acq: 5 Sep 2018 22:07

Instrument :

BNA_F

ClientSampleId :

Tot Ion:105 Resp: 548



Ion Ratio Lower Upper

105 100

71 0.0 4.4 6.6#

51 0.0 22.3 33.5#

120 0.0 16.9 25.3#





Abundance Ion 105.00 (104.70 to 105.70): E

Ion 71.00 (70.70 to 71.70): BF1

Ion 51.00 (50.70 to 51.70): BF1

Ion 120.00 (119.70 to 120.70): E



#23

Nitrobenzene-d5

Concen: 4.285 ng

RT: 7.28 min Scan# 841

Delta R.T. -0.02 min

Lab File: BF108789.D

Acq: 5 Sep 2018 22:07



Tgt Ion: 82 Resp: 74940

Ion Ratio Lower Upper

82 100

128 41.2 36.1 54.1


54 47.3 41.4 62.2




Abundance Ion 82.00 (81.70 to 82.70): BF1

Ion 128.00 (127.70 to 128.70): E

Ion 54.00 (53.70 to 54.70): BF1





#24

Nitrobenzene

Concen: 0.006 ng

RT: 7.27 min Scan# 839

Delta R.T. -0.05 min

Lab File: BF108789.D

Acq: 5 Sep 2018 22:07

Instrument :

BNA_F

ClientSampleId :


Tot Ion: 77 Resp: 114

Ion Ratio Lower Upper

77 100

123 0.0 37.9 56.9#


65 0.0 11.0 16.4#





Abundance Ion 77.00 (76.70 to 77.70): BF1

Ion 123.00 (122.70 to 123.70): BF1

Ion 65.00 (64.70 to 65.70): BF1



Time-->



#25

Isophorone

Concen: 2.143 ng

RT: 7.28 min Scan# 841

Delta R.T. -0.28 min

Lab File: BF108789.D

Acq: 5 Sep 2018 22:07


Tgt Ion: 82 Resp: 74748

Ion Ratio Lower Upper

82 100

95 0.0 5.2 7.8#


138 0.0 14.9 22.3#




Abundance Ion 82.00 (81.70 to 82.70): BF1


Ion 95.00 (94.70 to 95.70): BF1

Ion 138.00 (137.70 to 138.70): BF1



Time-->





#28

bis(2-Chloroethoxy)methane

Concen: 0.006 ng

RT: 8.02 min Scan# 966

Delta R.T. 0.25 min

Lab File: BF108789.D

Acq: 5 Sep 2018 22:07

Instrument :

BNA_F

ClientSampled :


Tot Ion: 93 Resp: 150

Ion Ratio Lower Upper

93 100

95 164.4 26.3 39.5#




123 0.0 9.8 14.6#



Abundance Ion 93.00 (92.70 to 93.70): BF1

Ion 95.00 (94.70 to 95.70): BF1

Ion 123.00 (122.70 to 123.70): BF1



#31

Naphthalene

Concen: 0.027 ng

RT: 8.04 min Scan# 969

Delta R.T. -0.01 min

Lab File: BF108789.D

Acq: 5 Sep 2018 22:07


Tgt Ion:128 Resp: 1312

Ion Ratio Lower Upper

128 100

129 0.0 8.8 13.2#



127 0.0 10.9 16.3#




Abundance Ion 128.00 (127.70 to 128.70): E

Ion 129.00 (128.70 to 129.70): E

Ion 127.00 (126.70 to 127.70): E









#32
Benzoic acid
Concen: 4.030 ng
RT: 8.01 min Scan# 965
Delta R.T. 0.21 min
Lab File: BF108789.D
Acq: 5 Sep 2018 22:07

Instrument :
BNA_F
ClientSampleId :

Tot Ion:122 Resp: 120
Ion Ratio Lower Upper
122 100
105 295.3 101.1 141.1#
77 1031.9 70.7 110.7#




Abundance Ion 122.00 (121.70 to 122.70): E
Ion 105.00 (104.70 to 105.70): E
Ion 77.00 (76.70 to 77.70): BF1






#36
4-Chloro-3-methylphenol
Concen: 0.014 ng
RT: 8.62 min Scan# 1069
Delta R.T. 0.06 min
Lab File: BF108789.D
Acq: 5 Sep 2018 22:07

Tgt Ion:107 Resp: 223
Ion Ratio Lower Upper
107 100
144 0.0 20.5 30.7#
142 0.0 62.0 93.0#



Abundance Ion 107.00 (106.70 to 107.70): E
Ion 144.00 (143.70 to 144.70): E
Ion 142.00 (141.70 to 142.70): E









#37
2-Methylnaphthalene
Concen: 0.029 ng
RT: 8.72 min Scan# 1086
Delta R.T. -0.01 min
Lab File: BF108789.D
Acq: 5 Sep 2018 22:07

Instrument :
BNA_F
ClientSampleId :

Tot Ion:142 Resp: 939
Ion Ratio Lower Upper
142 100
141 94.3 70.8 106.2
115 41.5 26.1 39.1#




Abundance Ion 142.00 (141.70 to 142.70): E
Ion 141.00 (140.70 to 141.70): E
Ion 115.00 (114.70 to 115.70): E






#38
Acenaphthene-d10
Concen: 20.000 ng
RT: 9.77 min Scan# 1263
Delta R.T. -0.00 min
Lab File: BF108789.D
Acq: 5 Sep 2018 22:07

Tgt Ion:164 Resp: 456914
Ion Ratio Lower Upper
164 100
162 102.3 86.1 129.1
160 48.3 38.3 57.5



Abundance Ion 164.00 (163.70 to 164.70): E
Ion 162.00 (161.70 to 162.70): E
Ion 160.00 (159.70 to 160.70): E









#41
 2,4,6-Tribromophenol
 Concen: 4.303 ng
 RT: 10.54 min Scan# 1395
 Delta R.T. -0.01 min
 Lab File: BF108789.D
 Acq: 5 Sep 2018 22:07

Instrument :
 BNA_F
 ClientSampleId :

Tot Ion:330 Resp: 18768
 Ion Ratio Lower Upper
 330 100
 332 96.7 77.0 115.6
 141 35.8 29.9 44.9




Abundance Ion 329.80 (329.50 to 330.50): E
 Ion 331.80 (331.50 to 332.50): E
 Ion 141.00 (140.70 to 141.70): E






#44
 2-Fluorobiphenyl
 Concen: 5.935 ng
 RT: 9.09 min Scan# 1148
 Delta R.T. -0.01 min
 Lab File: BF108789.D
 Acq: 5 Sep 2018 22:07

Tgt Ion:172 Resp: 158948
 Ion Ratio Lower Upper
 172 100
 171 35.8 29.7 44.5
 170 22.9 19.6 29.4



Abundance Ion 172.00 (171.70 to 172.70): E
 Ion 171.00 (170.70 to 171.70): E
 Ion 170.00 (169.70 to 170.70): E









#45
 1,1'-Biphenyl
 Concen: 0.020 ng
 RT: 9.19 min Scan# 1165
 Delta R.T. -0.01 min
 Lab File: BF108789.D
 Acq: 5 Sep 2018 22:07

Instrument :
 BNA_F
 ClientSampleId :

Tot Ion:154 Resp: 701
 Ion Ratio Lower Upper
 154 100
 153 32.5 21.3 61.3
 76 0.0 0.0 34.0




Abundance Ion 154.00 (153.70 to 154.70): E
 Ion 153.00 (152.70 to 153.70): E
 Ion 76.00 (75.70 to 76.70): BF1






#46
 2-Chloronaphthalene
 Concen: 0.008 ng
 RT: 9.48 min Scan# 1214
 Delta R.T. 0.26 min
 Lab File: BF108789.D
 Acq: 5 Sep 2018 22:07

Tgt Ion:162 Resp: 230
 Ion Ratio Lower Upper
 162 100
 127 0.0 33.9 50.9#
 164 704.8 26.5 39.7#



Abundance Ion 162.00 (161.70 to 162.70): E
 Ion 127.00 (126.70 to 127.70): E
 Ion 164.00 (163.70 to 164.70): E









#47
2-Nitroaniline
Concen: 0.105 ng
RT: 9.48 min Scan# 1214
Delta R.T. 0.17 min
Lab File: BF108789.D
Acq: 5 Sep 2018 22:07

Instrument :
BNA_F
ClientSampleId :

Tot Ion: 65 Resp: 818
Ion Ratio Lower Upper
65 100
92 653.3 55.8 83.6#
138 0.0 91.2 136.8#




Abundance Ion 65.00 (64.70 to 65.70): BF1
Ion 92.00 (91.70 to 92.70): BF1
Ion 138.00 (137.70 to 138.70): BF1






#48
Acenaphthylene
Concen: 0.113 ng
RT: 9.62 min Scan# 1239
Delta R.T. -0.01 min
Lab File: BF108789.D
Acq: 5 Sep 2018 22:07

Tgt Ion: 152 Resp: 4797
Ion Ratio Lower Upper
152 100
151 18.0 16.3 24.5
153 18.6 10.7 16.1#



Abundance Ion 152.00 (151.70 to 152.70): E
Ion 151.00 (150.70 to 151.70): E
Ion 153.00 (152.70 to 153.70): E









#49
Dimethylphthalate
Concen: 1.284 ng
RT: 9.48 min Scan# 1214
Delta R.T. -0.02 min
Lab File: BF108789.D
Acq: 5 Sep 2018 22:07

Instrument :
BNA_F
ClientSampleId :

Tot Ion:163 Resp: 41031
Ion Ratio Lower Upper
163 100
194 3.7 2.7 4.1
164 9.4 8.2 12.2




Abundance Ion 163.00 (162.70 to 163.70): E
Ion 194.00 (193.70 to 194.70): E
Ion 164.00 (163.70 to 164.70): E






#50
2,6-Dinitrotoluene
Concen: 9.531 ng
RT: 9.76 min Scan# 1262
Delta R.T. 0.21 min
Lab File: BF108789.D
Acq: 5 Sep 2018 22:07

Tgt Ion:165 Resp: 58038
Ion Ratio Lower Upper
165 100
63 1.3 44.7 67.1#
89 1.0 44.0 66.0#



Abundance Ion 165.00 (164.70 to 165.70): E
Ion 63.00 (62.70 to 63.70): BF1
Ion 89.00 (88.70 to 89.70): BF1





#51

Acenaphthene

Concen: 0.117 ng

RT: 9.80 min Scan# 1268

Delta R.T. -0.01 min

Lab File: BF108789.D

Acq: 5 Sep 2018 22:07

Instrument :

BNA_F

ClientSampleId :


Tot Ion:154 Resp: 3060

Ion Ratio Lower Upper




154 100

153 138.4 91.2 136.8#

152 55.4 43.8 65.8



Abundance Ion 154.00 (153.70 to 154.70): E
Ion 153.00 (152.70 to 153.70): E
Ion 152.00 (151.70 to 152.70): E



#54

Dibenzofuran

Concen: 0.063 ng

RT: 9.97 min Scan# 1297

Delta R.T. -0.01 min

Lab File: BF108789.D

Acq: 5 Sep 2018 22:07


Tgt Ion:168 Resp: 2417

Ion Ratio Lower Upper



168 100


139 35.8 30.1 45.1

169 13.7 10.9 16.3



Abundance Ion 168.00 (167.70 to 168.70): E
Ion 139.00 (138.70 to 139.70): E
Ion 169.00 (168.70 to 169.70): E





#55

4-Nitrophenol

Concen: 0.165 ng

RT: 9.97 min Scan# 1297

Delta R.T. 0.10 min

Lab File: BF108789.D

Acq: 5 Sep 2018 22:07

Instrument :
BNA_F
ClientSampleId :


Tot Ion:139 Resp: 896

Ion Ratio Lower Upper




139 100

109 0.0 48.4 88.4#

65 0.0 72.6 112.6#



Abundance Ion 139.00 (138.70 to 139.70): E
Ion 109.00 (108.70 to 109.70): E
Ion 65.00 (64.70 to 65.70): BF1



#56

2,4-Dinitrotoluene

Concen: 7.453 ng

RT: 9.76 min Scan# 1262

Delta R.T. -0.19 min

Lab File: BF108789.D

Acq: 5 Sep 2018 22:07

Tgt Ion:165 Resp: 58038


Ion Ratio Lower Upper

165 100



63 1.3 36.4 54.6#


89 1.0 57.8 86.6#

182 0.0 1.6 2.4#



Abundance Ion 165.00 (164.70 to 165.70): E
Ion 63.00 (62.70 to 63.70): BF1
Ion 89.00 (88.70 to 89.70): BF1
Ion 182.00 (181.70 to 182.70): E









#57
 Fluorene
 Concen: 0.197 ng
 RT: 10.31 min Scan# 1355
 Delta R.T. -0.01 min
 Lab File: BF108789.D
 Acq: 5 Sep 2018 22:07




Instrument :
 BNA_F
 ClientSampleId :


Tot Ion:166 Resp: 4847
 Ion Ratio Lower Upper
 166 100
 165 103.9 79.8 119.8
 167 17.0 10.6 16.0#



#59
 Diethylphthalate
 Concen: 0.199 ng
 RT: 10.18 min Scan# 1334
 Delta R.T. -0.01 min
 Lab File: BF108789.D
 Acq: 5 Sep 2018 22:07

Tgt Ion:149 Resp: 6560
 Ion Ratio Lower Upper
 149 100
 177 22.8 16.1 24.1
 150 16.2 10.1 15.1#





#62

Azobenzene

Concen: 0.013 ng

RT: 10.47 min Scan# 1382

Delta R.T. -0.00 min

Lab File: BF108789.D

Acq: 5 Sep 2018 22:07

Instrument :

BNA_F

ClientSampleId :

Tot Ion: 77 Resp: 450


Ion Ratio Lower Upper

77 100

182 255.2 1.7 41.7#

105 66.5 3.0 43.0#

51 0.0 9.9 49.9#






Abundance Ion 77.00 (76.70 to 77.70): BF1

Ion 182.00 (181.70 to 182.70): E

Ion 105.00 (104.70 to 105.70): E

Ion 51.00 (50.70 to 51.70): BF1



#63

Phenanthrene-d10

Concen: 20.000 ng

RT: 11.24 min Scan# 1513

Delta R.T. -0.01 min

Lab File: BF108789.D

Acq: 5 Sep 2018 22:07


Tgt Ion: 188 Resp: 734252

Ion Ratio Lower Upper

188 100

94 11.4 8.5 12.7



80 12.6 9.2 13.8




Abundance Ion 188.00 (187.70 to 188.70): E

Ion 94.00 (93.70 to 94.70): BF1

Ion 80.00 (79.70 to 80.70): BF1









#65
 n-Nitrosodiphenylamine
 Concen: 0.011 ng
 RT: 10.42 min Scan# 1374
 Delta R.T. -0.01 min
 Lab File: BF108789.D
 Acq: 5 Sep 2018 22:07




Instrument :
 BNA_F
 ClientSampled :


Tot Ion:169 Resp: 261
 Ion Ratio Lower Upper
 169 100
 168 0.0 54.4 81.6#
 167 0.0 29.8 44.6#



#66
 4-Bromophenyl-phenylether
 Concen: 0.161 ng
 RT: 10.54 min Scan# 1395
 Delta R.T. -0.25 min
 Lab File: BF108789.D
 Acq: 5 Sep 2018 22:07

Tgt Ion:248 Resp: 1336
 Ion Ratio Lower Upper
 248 100
 250 140.8 76.9 115.3#
 141 311.3 74.4 111.6#









#70
Phenanthrene
Concen: 1.543 ng
RT: 11.26 min Scan# 1517
Delta R.T. -0.01 min
Lab File: BF108789.D
Acq: 5 Sep 2018 22:07




Instrument :
BNA_F
ClientSampleId :


Tot Ion:178 Resp: 54098
Ion Ratio Lower Upper
178 100
176 20.2 15.8 23.8
179 16.5 13.0 19.6



#71
Anthracene
Concen: 0.446 ng
RT: 11.31 min Scan# 1526
Delta R.T. -0.01 min
Lab File: BF108789.D
Acq: 5 Sep 2018 22:07

Tgt Ion:178 Resp: 14971
Ion Ratio Lower Upper
178 100
176 17.3 15.4 23.2
179 15.5 12.6 19.0





#72

Carbazole

Concen: 0.151 ng

RT: 11.46 min Scan# 1551

Delta R.T. -0.01 min

Lab File: BF108789.D

Acq: 5 Sep 2018 22:07

Instrument :

BNA_F

ClientSampleId :


Tot Ion:167 Resp: 5340

Ion Ratio Lower Upper

167 100

166 22.2 17.6 26.4




139 13.9 10.9 16.3



Abundance Ion 167.00 (166.70 to 167.70): E

Ion 166.00 (165.70 to 166.70): E

Ion 139.00 (138.70 to 139.70): E



#73

Di-n-butylphthalate

Concen: 0.059 ng

RT: 11.81 min Scan# 1610

Delta R.T. -0.01 min

Lab File: BF108789.D

Acq: 5 Sep 2018 22:07


Tgt Ion:149 Resp: 2322

Ion Ratio Lower Upper

149 100

150 25.5 7.8 11.6#



104 0.0 3.8 5.8#




Abundance Ion 149.00 (148.70 to 149.70): E

Ion 150.00 (149.70 to 150.70): E

Ion 104.00 (103.70 to 104.70): E





#74

Fluoranthene

Concen: 4.738 ng

RT: 12.44 min Scan# 1718

Delta R.T. -0.01 min

Lab File: BF108789.D

Acq: 5 Sep 2018 22:07

Instrument :

BNA_F

ClientSampled :


Tot Ion:202 Resp: 164503

Ion Ratio Lower Upper

202 100

101 13.0 0.0 34.1




203 17.6 0.0 38.1



Abundance Ion 202.00 (201.70 to 202.70): E

Ion 101.00 (100.70 to 101.70): E

Ion 203.00 (202.70 to 203.70): E



#75

Chrysene-d12

Concen: 20.000 ng

RT: 13.86 min Scan# 1959

Delta R.T. -0.01 min

Lab File: BF108789.D

Acq: 5 Sep 2018 22:07


Tgt Ion:240 Resp: 623326

Ion Ratio Lower Upper

240 100

120 13.9 9.5 14.3



236 26.0 20.7 31.1




Abundance Ion 240.00 (239.70 to 240.70): E

Ion 120.00 (119.70 to 120.70): E

Ion 236.00 (235.70 to 236.70): E





#76

Benzidine

Concen: 0.147 ng

RT: 12.55 min Scan# 1736

Delta R.T. -0.02 min

Lab File: BF108789.D

Acq: 5 Sep 2018 22:07

Instrument :

BNA_F

ClientSampled :


Tot Ion:184 Resp: 4348

Ion Ratio Lower Upper

184 100

185 27.4 12.6 18.8#




183 25.1 9.3 13.9#



Abundance Ion 184.00 (183.70 to 184.70): E

Ion 185.00 (184.70 to 185.70): E

Ion 183.00 (182.70 to 183.70): E



#77

Pyrene

Concen: 3.324 ng

RT: 12.67 min Scan# 1757

Delta R.T. -0.01 min

Lab File: BF108789.D

Acq: 5 Sep 2018 22:07


Tgt Ion:202 Resp: 159409

Ion Ratio Lower Upper

202 100

200 20.3 17.4 26.2



203 17.8 14.3 21.5




Abundance Ion 202.00 (201.70 to 202.70): E

Ion 200.00 (199.70 to 200.70): E

Ion 203.00 (202.70 to 203.70): E





#78

Terphenyl-d14

Concen: 4.782 ng

RT: 12.82 min Scan# 1782

Delta R.T. -0.01 min

Lab File: BF108789.D

Acq: 5 Sep 2018 22:07

Instrument :

BNA_F

ClientSampleId :


Tot Ion:244 Resp: 127826

Ion Ratio Lower Upper

244 100

212 6.4 5.4 8.0




122 10.9 11.0 16.4#



Abundance Ion 244.00 (243.70 to 244.70): E

Ion 212.00 (211.70 to 212.70): E

Ion 122.00 (121.70 to 122.70): E



#79

Butylbenzylphthalate

Concen: 0.296 ng

RT: 13.30 min Scan# 1864

Delta R.T. -0.00 min

Lab File: BF108789.D

Acq: 5 Sep 2018 22:07


Tgt Ion:149 Resp: 6395

Ion Ratio Lower Upper

149 100

91 80.7 56.8 85.2



206 21.0 14.1 21.1




Abundance Ion 149.00 (148.70 to 149.70): E

Ion 91.00 (90.70 to 91.70): BF1

Ion 206.00 (205.70 to 206.70): E





#80

Benzo(a)anthracene

Concen: 2.189 ng

RT: 13.85 min Scan# 1958

Delta R.T. -0.00 min

Lab File: BF108789.D

Acq: 5 Sep 2018 22:07

Instrument :

BNA_F

ClientSampleId :


Tot Ion:228 Resp: 87465

Ion Ratio Lower Upper

228 100

226 31.4 22.6 33.8




229 22.1 15.8 23.6



Abundance Ion 228.00 (227.70 to 228.70): E

Ion 226.00 (225.70 to 226.70): E

Ion 229.00 (228.70 to 229.70): E



#81

3,3'-Dichlorobenzidine

Concen: 0.052 ng

RT: 13.79 min Scan# 1948

Delta R.T. -0.03 min

Lab File: BF108789.D

Acq: 5 Sep 2018 22:07


Tgt Ion:252 Resp: 810

Ion Ratio Lower Upper

252 100

254 31.0 50.4 75.6#



126 98.2 11.3 16.9#




Abundance Ion 252.00 (251.70 to 252.70): E

Ion 254.00 (253.70 to 254.70): E

Ion 126.00 (125.70 to 126.70): E









#82
 Chrysene
 Concen: 2.380 ng
 RT: 13.89 min Scan# 1964
 Delta R.T. -0.01 min
 Lab File: BF108789.D
 Acq: 5 Sep 2018 22:07




Instrument :
 BNA_F
 ClientSampleId :


Tot Ion:228 Resp: 92051
 Ion Ratio Lower Upper
 228 100
 226 30.4 25.0 37.6
 229 23.0 16.2 24.4



#83
 Bis(2-ethylhexyl)phthalate
 Concen: 8.022 ng
 RT: 13.87 min Scan# 1960
 Delta R.T. -0.00 min
 Lab File: BF108789.D
 Acq: 5 Sep 2018 22:07

Tgt Ion:149 Resp: 217742
 Ion Ratio Lower Upper
 149 100
 167 26.8 21.3 31.9
 279 3.5 3.0 4.4





#84

Di-n-octyl phthalate

Concen: 0.017 ng

RT: 14.48 min Scan# 2064

Delta R.T. -0.00 min

Lab File: BF108789.D

Acq: 5 Sep 2018 22:07

Instrument :

BNA_F

ClientSampled :


Tot Ion:149 Resp: 751

Ion Ratio Lower Upper

149 100

167 88.0 1.2 1.8#




43 0.0 8.4 12.6#



Abundance Ion 149.00 (148.70 to 149.70): E

Ion 167.00 (166.70 to 167.70): E

Ion 43.00 (42.70 to 43.70): BF1



#85

Indeno(1,2,3-cd)pyrene

Concen: 1.254 ng

RT: 16.61 min Scan# 2427

Delta R.T. -0.01 min

Lab File: BF108789.D

Acq: 5 Sep 2018 22:07


Tgt Ion:276 Resp: 42937

Ion Ratio Lower Upper

276 100

138 27.5 25.0 37.6



277 23.4 20.1 30.1




Abundance Ion 276.00 (275.70 to 276.70): E

Ion 138.00 (137.70 to 138.70): E

Ion 277.00 (276.70 to 277.70): E









#86
Pervlene-d12
Concen: 20.000 ng
RT: 15.27 min Scan# 2199
Delta R.T. 0.01 min
Lab File: BF108789.D
Acq: 5 Sep 2018 22:07




Instrument :
BNA_F
ClientSampleId :


Tot Ion	Ratio	Lower	Upper
264	100		
260	23.0	19.2	28.8
265	21.4	17.5	26.3



#87
Benzo(b)fluoranthene
Concen: 5.396 ng
RT: 14.87 min Scan# 2131
Delta R.T. -0.00 min
Lab File: BF108789.D
Acq: 5 Sep 2018 22:07

Tgt Ion	Ratio	Lower	Upper
252	100		
253	24.0	17.0	25.6
125	18.9	9.0	13.6









#88
Benzo(k)fluoranthene
Concen: 5.860 ng
RT: 14.87 min Scan# 2131
Delta R.T. -0.03 min
Lab File: BF108789.D
Acq: 5 Sep 2018 22:07




Instrument :
BNA_F
ClientSampled :


Tot Ion	Ratio	Lower	Upper
252	100		
253	24.0	17.9	26.9
125	18.9	10.2	15.2



#89
Benzo(a)pyrene
Concen: 2.771 ng
RT: 15.21 min Scan# 2188
Delta R.T. -0.01 min
Lab File: BF108789.D
Acq: 5 Sep 2018 22:07

Tgt Ion	Ratio	Lower	Upper
252	100		
253	23.9	17.1	25.7
125	18.8	9.8	14.6









#90
Dibenzo(a,h)anthracene
Concen: 0.629 ng
RT: 16.63 min Scan# 2430
Delta R.T. -0.01 min
Lab File: BF108789.D
Acq: 5 Sep 2018 22:07

Instrument :
BNA_F
ClientSampled :

Tot Ion:278 Resp: 13346
Ion Ratio Lower Upper
278 100
139 47.7 15.9 23.9#
279 37.7 19.0 28.4#



#91
Benzo(g,h,i)perylene
Concen: 2.078 ng
RT: 17.02 min Scan# 2496
Delta R.T. -0.01 min
Lab File: BF108789.D
Acq: 5 Sep 2018 22:07

Tgt Ion:276 Resp: 44169
Ion Ratio Lower Upper
276 100
277 28.8 17.9 26.9#
138 27.0 21.8 32.8

