

Data Path : Z:\SVOASRV\HPCHEM1\BNA F\DATA\BF090518\
 Data File : BF108787.D
 Acq On : 5 Sep 2018 21:12
 Operator : JU/SJ
 Sample : J4741-16 10X
 Misc :
 ALS Vial : 7 Sample Multiplier: 1

Instrument :
 BNA_F
 ClientSampleId :
 AOC10-01-D

Manual Integrations
 APPROVED

Sohil
 9/6/2018 4:44:58 PM

Quant Time: Sep 06 06:41:24 2018
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA F\METHODS\8270-BF090518.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Wed Sep 05 18:03:31 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	6.73	152	298883	20.00	ng	0.00
21) Naphthalene-d8	8.01	136	1116393	20.00	ng	0.00
38) Acenaphthene-d10	9.76	164	480056	20.00	ng	0.00
63) Phenanthrene-d10	11.24	188	765029	20.00	ng	0.00
75) Chrysene-d12	13.86	240	638080	20.00	ng	0.00
86) Perylene-d12	15.27	264	517360	20.00	ng	0.00
System Monitoring Compounds						
5) 2-Fluorophenol	5.32	112	148271	8.22	ng	-0.01
7) Phenol-d6	6.35	99	194193	8.37	ng	-0.02
23) Nitrobenzene-d5	7.28	82	102216	5.73	ng	-0.02
41) 2,4,6-Tribromophenol	10.54	330	26485	5.78	ng	-0.01
44) 2-Fluorobiphenyl	9.09	172	206134	7.33	ng	0.00
78) Terphenyl-d14	12.82	244	153152	5.60	ng	0.00
Target Compounds						
70) Phenanthrene	11.26	178	79785	2.184	ng	99
74) Fluoranthene	12.44	202	146797	4.058	ng	98
77) Pyrene	12.67	202	127049	2.588	ng	99
79) Butylbenzylphthalate	13.29	149	58004	2.623	ng	96
87) Benzo(b)fluoranthene	14.87	252	69395m	2.467	ng	

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\SVOASRV\HPCHEM1\BNA F\DATA\BF090518\
Data File : BF108787.D
Acq On : 5 Sep 2018 21:12
Operator : JU/SJ
Sample : J4741-16 10X
Misc :
ALS Vial : 7 Sample Multiplier: 1

Instrument :
BNA_F
ClientSampleId :
AOC10-01-D

Manual Integrations
APPROVED

Sohil
9/6/2018 4:44:58 PM

Quant Time: Sep 06 06:41:24 2018
Quant Method : Z:\SVOASRV\HPCHEM1\BNA F\METHODS\8270-BF090518.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Wed Sep 05 18:03:31 2018
Response via : Initial Calibration

