




Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF090523\
Data File : BF135087.D
Acq On : 05 Sep 2023 09:40
Operator : CG\JU
Sample : DFTPP
Misc :
ALS Vial : 1 Sample Multiplier: 1

Quant Time: Sep 05 10:03:26 2023
Quant Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF083023.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Thu Aug 31 03:48:08 2023
Response via : Initial Calibration



TIC: BF135087.D\data.ms

(70) Pentachlorophenol (C)




11.110min (-0.000) 34226.01 ng

response 154424

Ion	Exp%	Act%
265.70	100.00	100.00
268.00	66.60	67.89
264.00	61.60	63.40
0.00	0.00	0.00

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF090523\
 Data File : BF135087.D
 Acq On : 05 Sep 2023 09:40
 Operator : CG\JU
 Sample : DFTPP
 Misc :
 ALS Vial : 1 Sample Multiplier: 1

Quant Time: Sep 05 10:03:26 2023
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF083023.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Thu Aug 31 03:48:08 2023
 Response via : Initial Calibration



TIC: BF135087.D\data.ms

(77) Benzidine

12.651min (-12.651) 0.00 ng

response 0

Ion	Exp%	Act%
184.00	100.00	0.00
185.00	15.00	0.00#
183.00	11.30	0.00#
0.00	0.00	0.00