

Data Path : Z:\SVOASRV\HPCHEM1\BNA F\DATA\BF090618\
 Data File : BF108828.D
 Acq On : 6 Sep 2018 17:31
 Operator : JU/SJ
 Sample : J4803-01
 Misc :
 ALS Vial : 6 Sample Multiplier: 1

Instrument :
 BNA_F
 ClientSampleId :
 EP1-Q2-E5

Quant Time: Sep 06 18:37:30 2018
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA F\METHODS\8270-BF090518.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Wed Sep 05 18:03:31 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	6.74	152	263574	20.00	ng	0.00
21) Naphthalene-d8	8.02	136	997271	20.00	ng	0.00
38) Acenaphthene-d10	9.77	164	456257	20.00	ng	0.00
63) Phenanthrene-d10	11.24	188	761762	20.00	ng	0.00
75) Chrysene-d12	13.87	240	531573	20.00	ng	0.00
86) Perylene-d12	15.28	264	702697	20.00	ng	0.01
System Monitoring Compounds						
5) 2-Fluorophenol	5.34	112	987809	62.09	ng	0.00
7) Phenol-d6	6.37	99	1267150	61.90	ng	0.00
23) Nitrobenzene-d5	7.29	82	787538	49.42	ng	-0.01
41) 2,4,6-Tribromophenol	10.55	330	272914	62.66	ng	0.00
44) 2-Fluorobiphenyl	9.10	172	1388787	51.93	ng	0.00
78) Terphenyl-d14	12.82	244	1070096	46.94	ng	0.00
Target Compounds						
49) Dimethylphthalate	9.48	163	497978	15.605	ng	99
74) Fluoranthene	12.45	202	103335	2.869	ng	98
77) Pyrene	12.67	202	97185	2.376	ng	98
83) Bis(2-ethylhexyl)phthalate	13.87	149	893570	38.602	ng	100

(#) = qualifier out of range (m) = manual integration (+) = signals summed

