

Data Path : Z:\SVOASRV\HPCHEM1\BNA F\DATA\BF090919\
 Data File : BF116915.D
 Acq On : 9 Sep 2019 23:40
 Operator : HP/JU
 Sample : K4758-03
 Misc :
 ALS Vial : 22 Sample Multiplier: 1

Instrument :
 BNA_F
 ClientSampleId :
 B-4-8

Quant Time: Sep 10 03:59:52 2019
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA F\METHODS\8270-BF083019.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Fri Aug 30 20:02:30 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	6.83	152	74827	20.00	ng	-0.02
21) Naphthalene-d8	8.12	136	302118	20.00	ng	-0.02
39) Acenaphthene-d10	9.87	164	157903	20.00	ng	-0.02
64) Phenanthrene-d10	11.35	188	271415	20.00	ng	-0.02
76) Chrysene-d12	13.98	240	223908	20.00	ng	-0.02
87) Perylene-d12	15.43	264	214742	20.00	ng	-0.01
System Monitoring Compounds						
5) 2-Fluorophenol	5.45	112	483894	104.77	ng	-0.01
7) Phenol-d6	6.47	99	671260	110.68	ng	-0.01
23) Nitrobenzene-d5	7.39	82	413644	78.16	ng	-0.02
42) 2,4,6-Tribromophenol	10.66	330	147434	139.67	ng	-0.02
45) 2-Fluorobiphenyl	9.19	172	722806	77.22	ng	-0.02
79) Terphenyl-d14	12.93	244	779222	64.38	ng	-0.02
Target Compounds						
50) Dimethylphthalate	9.58	163	93608	8.587	ng	Qvalue 99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\SVOASRV\HPCHEM1\BNA F\DATA\BF090919\
Data File : BF116915.D
Acq On : 9 Sep 2019 23:40
Operator : HP/JU
Sample : K4758-03
Misc :
ALS Vial : 22 Sample Multiplier: 1

Instrument :
BNA_F
ClientSampleId :
B-4-8

Quant Time: Sep 10 03:59:52 2019
Quant Method : Z:\SVOASRV\HPCHEM1\BNA F\METHODS\8270-BF083019.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Fri Aug 30 20:02:30 2019
Response via : Initial Calibration

