




Data Path : Z:\svoasrv\HPCHEM1\BNA F\Data\BF091218\
Data File : BF108948.D
Acq On : 12 Sep 2018 13:06
Operator : JU/SJ
Sample : DFTPP
Misc :
ALS Vial : 1 Sample Multiplier: 1

Quant Time: Sep 12 18:08:38 2018
Quant Method : Z:\SVOASRV\HPCHEM1\BNA F\METHODS\8270-BF090518.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Wed Sep 05 18:03:31 2018
Response via : Initial Calibration



TIC: BF108948.D

(76) Benzidine




12.571min (+0.006) 48783.24ng

response 642773

Ion	Exp%	Act%
184.00	100	100
185.00	15.70	14.88
183.00	11.60	11.60
0.00	0.00	0.00

Data Path : Z:\svoasrv\HPCHEM1\BNA F\Data\BF091218\
Data File : BF108948.D
Acq On : 12 Sep 2018 13:06
Operator : JU/SJ
Sample : DFTPP
Misc :
ALS Vial : 1 Sample Multiplier: 1

Quant Time: Sep 12 18:08:38 2018
Quant Method : Z:\SVOASRV\HPCHEM1\BNA F\METHODS\8270-BF090518.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Wed Sep 05 18:03:31 2018
Response via : Initial Calibration



(69) Pentachlorophenol (C)

11.054min (+0.006) 43523.19ng

response 116780

Ion	Exp%	Act%
265.70	100	100
268.00	63.90	62.74
264.00	61.90	62.34
0.00	0.00	0.00