

Data Path : Z:\svoasrv\HPCHEM1\BNA\_F\Data\BF091223\  
 Data File : BF135170.D  
 Acq On : 12 Sep 2023 08:53  
 Operator : CG\JU  
 Sample : SSTDCCC040  
 Misc :  
 ALS Vial : 2 Sample Multiplier: 1

Instrument :  
 BNA\_F  
 LabSampleId :  
 SSTDCCC040

Quant Time: Sep 13 01:54:08 2023  
 Quant Method : Z:\svoasrv\HPCHEM1\BNA\_F\Methods\8270-BF091123.M  
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION  
 QLast Update : Tue Sep 12 00:19:07 2023  
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min  
 Max. RRF Dev : 25% Max. Rel. Area : 150%

|      | Compound                    | AvgRF | CCRF  | %Dev | Area% | Dev(min) |
|------|-----------------------------|-------|-------|------|-------|----------|
| 1 I  | 1,4-Dichlorobenzene-d4      | 1.000 | 1.000 | 0.0  | 105   | 0.00     |
| 2    | 1,4-Dioxane                 | 0.539 | 0.536 | 0.6  | 105   | 0.00     |
| 3    | Pyridine                    | 1.451 | 1.475 | -1.7 | 104   | 0.00     |
| 4    | n-Nitrosodimethylamine      | 0.770 | 0.774 | -0.5 | 104   | 0.00     |
| 5 S  | 2-Fluorophenol              | 1.230 | 1.194 | 2.9  | 102   | 0.00     |
| 6    | Aniline                     | 2.050 | 2.023 | 1.3  | 104   | 0.00     |
| 7 S  | Phenol-d6                   | 1.564 | 1.520 | 2.8  | 102   | 0.00     |
| 8    | 2-Chlorophenol              | 1.313 | 1.287 | 2.0  | 103   | 0.00     |
| 9    | Benzaldehyde                | 0.891 | 0.852 | 4.4  | 103   | 0.00     |
| 10 C | Phenol                      | 1.715 | 1.661 | 3.1  | 101   | 0.00     |
| 11   | bis(2-Chloroethyl)ether     | 1.286 | 1.272 | 1.1  | 103   | 0.00     |
| 12   | 1,3-Dichlorobenzene         | 1.334 | 1.323 | 0.8  | 105   | 0.00     |
| 13 C | 1,4-Dichlorobenzene         | 1.357 | 1.318 | 2.9  | 103   | 0.00     |
| 14   | 1,2-Dichlorobenzene         | 1.260 | 1.218 | 3.3  | 103   | 0.00     |
| 15   | Benzyl Alcohol              | 1.122 | 1.102 | 1.8  | 102   | 0.00     |
| 16   | 2,2'-oxybis(1-Chloropropane | 2.424 | 2.347 | 3.2  | 101   | 0.00     |
| 17   | 2-Methylphenol              | 1.158 | 1.139 | 1.6  | 102   | 0.00     |
| 18   | Hexachloroethane            | 0.502 | 0.489 | 2.6  | 102   | 0.00     |
| 19 P | n-Nitroso-di-n-propylamine  | 0.968 | 0.912 | 5.8  | 101   | 0.00     |
| 20   | 3+4-Methylphenols           | 1.393 | 1.331 | 4.5  | 102   | 0.00     |
| 21 I | Naphthalene-d8              | 1.000 | 1.000 | 0.0  | 103   | 0.00     |
| 22   | Acetophenone                | 0.457 | 0.433 | 5.3  | 101   | 0.00     |
| 23 S | Nitrobenzene-d5             | 0.368 | 0.366 | 0.5  | 103   | 0.00     |
| 24   | Nitrobenzene                | 0.364 | 0.362 | 0.5  | 102   | 0.00     |
| 25   | Isophorone                  | 0.676 | 0.656 | 3.0  | 101   | 0.00     |
| 26 C | 2-Nitrophenol               | 0.175 | 0.178 | -1.7 | 105   | 0.00     |
| 27   | 2,4-Dimethylphenol          | 0.294 | 0.289 | 1.7  | 104   | 0.00     |
| 28   | bis(2-Chloroethoxy)methane  | 0.405 | 0.398 | 1.7  | 104   | 0.00     |
| 29 C | 2,4-Dichlorophenol          | 0.272 | 0.267 | 1.8  | 102   | 0.00     |
| 30   | 1,2,4-Trichlorobenzene      | 0.284 | 0.276 | 2.8  | 101   | 0.00     |
| 31   | Naphthalene                 | 0.972 | 0.952 | 2.1  | 103   | 0.00     |
| 32   | Benzoic acid                | 0.221 | 0.224 | -1.4 | 100   | 0.00     |
| 33   | 4-Chloroaniline             | 0.426 | 0.426 | 0.0  | 104   | 0.00     |
| 34 C | Hexachlorobutadiene         | 0.171 | 0.169 | 1.2  | 104   | 0.00     |
| 35   | Caprolactam                 | 0.091 | 0.092 | -1.1 | 106   | 0.00     |
| 36 C | 4-Chloro-3-methylphenol     | 0.302 | 0.300 | 0.7  | 103   | 0.00     |
| 37   | 2-Methylnaphthalene         | 0.653 | 0.635 | 2.8  | 103   | 0.00     |
| 38   | 1-Methylnaphthalene         | 0.612 | 0.587 | 4.1  | 102   | 0.00     |
| 39 I | Acenaphthene-d10            | 1.000 | 1.000 | 0.0  | 104   | 0.00     |
| 40   | 1,2,4,5-Tetrachlorobenzene  | 0.579 | 0.575 | 0.7  | 104   | 0.00     |
| 41 P | Hexachlorocyclopentadiene   | 0.210 | 0.217 | -3.3 | 102   | 0.00     |
| 42 S | 2,4,6-Tribromophenol        | 0.199 | 0.199 | 0.0  | 106   | 0.00     |
| 43 C | 2,4,6-Trichlorophenol       | 0.379 | 0.380 | -0.3 | 104   | 0.00     |
| 44   | 2,4,5-Trichlorophenol       | 0.437 | 0.431 | 1.4  | 102   | 0.00     |
| 45 S | 2-Fluorobiphenyl            | 1.326 | 1.271 | 4.1  | 104   | 0.00     |
| 46   | 1,1'-Biphenyl               | 1.537 | 1.511 | 1.7  | 103   | 0.00     |
| 47   | 2-Chloronaphthalene         | 1.115 | 1.089 | 2.3  | 103   | 0.00     |

Data Path : Z:\svoasrv\HPCHEM1\BNA\_F\Data\BF091223\  
 Data File : BF135170.D  
 Acq On : 12 Sep 2023 08:53  
 Operator : CG\JU  
 Sample : SSTDCCC040  
 Misc :  
 ALS Vial : 2 Sample Multiplier: 1

Instrument :  
 BNA\_F  
 LabSampleId :  
 SSTDCCC040

Quant Time: Sep 13 01:54:08 2023  
 Quant Method : Z:\svoasrv\HPCHEM1\BNA\_F\Methods\8270-BF091123.M  
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION  
 QLast Update : Tue Sep 12 00:19:07 2023  
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min  
 Max. RRF Dev : 25% Max. Rel. Area : 150%

|      | Compound                   | AvgRF | CCRF  | %Dev  | Area% | Dev(min) |
|------|----------------------------|-------|-------|-------|-------|----------|
| 48   | 2-Nitroaniline             | 0.433 | 0.434 | -0.2  | 104   | 0.00     |
| 49   | Acenaphthylene             | 1.853 | 1.840 | 0.7   | 104   | 0.00     |
| 50   | Dimethylphthalate          | 1.352 | 1.329 | 1.7   | 105   | 0.00     |
| 51   | 2,6-Dinitrotoluene         | 0.296 | 0.303 | -2.4  | 107   | 0.00     |
| 52 C | Acenaphthene               | 1.095 | 1.071 | 2.2   | 104   | 0.00     |
| 53   | 3-Nitroaniline             | 0.348 | 0.358 | -2.9  | 107   | 0.00     |
| 54 P | 2,4-Dinitrophenol          | 0.143 | 0.146 | -2.1  | 104   | 0.00     |
| 55   | Dibenzofuran               | 1.671 | 1.638 | 2.0   | 104   | 0.00     |
| 56 P | 4-Nitrophenol              | 0.221 | 0.235 | -6.3  | 106   | 0.00     |
| 57   | 2,4-Dinitrotoluene         | 0.371 | 0.378 | -1.9  | 104   | 0.00     |
| 58   | Fluorene                   | 1.284 | 1.253 | 2.4   | 104   | 0.00     |
| 59   | 2,3,4,6-Tetrachlorophenol  | 0.336 | 0.346 | -3.0  | 108   | 0.00     |
| 60   | Diethylphthalate           | 1.357 | 1.337 | 1.5   | 104   | 0.00     |
| 61   | 4-Chlorophenyl-phenylether | 0.617 | 0.595 | 3.6   | 104   | 0.00     |
| 62   | 4-Nitroaniline             | 0.334 | 0.339 | -1.5  | 104   | 0.00     |
| 63   | Azobenzene                 | 1.328 | 1.311 | 1.3   | 104   | 0.00     |
| 64 I | Phenanthrene-d10           | 1.000 | 1.000 | 0.0   | 103   | 0.00     |
| 65   | 4,6-Dinitro-2-methylphenol | 0.106 | 0.108 | -1.9  | 106   | 0.00     |
| 66 c | n-Nitrosodiphenylamine     | 0.605 | 0.593 | 2.0   | 105   | 0.00     |
| 67   | 4-Bromophenyl-phenylether  | 0.199 | 0.197 | 1.0   | 106   | 0.00     |
| 68   | Hexachlorobenzene          | 0.203 | 0.202 | 0.5   | 106   | 0.00     |
| 69   | Atrazine                   | 0.163 | 0.165 | -1.2  | 109   | 0.00     |
| 70 C | Pentachlorophenol          | 0.121 | 0.131 | -8.3  | 106   | 0.00     |
| 71   | Phenanthrene               | 0.946 | 0.926 | 2.1   | 104   | 0.00     |
| 72   | Anthracene                 | 0.972 | 0.957 | 1.5   | 105   | 0.00     |
| 73   | Carbazole                  | 0.885 | 0.881 | 0.5   | 105   | 0.00     |
| 74   | Di-n-butylphthalate        | 1.043 | 1.063 | -1.9  | 107   | 0.00     |
| 75 C | Fluoranthene               | 0.986 | 0.961 | 2.5   | 103   | 0.00     |
| 76 I | Chrysene-d12               | 1.000 | 1.000 | 0.0   | 110   | 0.00     |
| 77   | Benzydine                  | 0.410 | 0.463 | -12.9 | 138   | 0.00     |
| 78   | Pyrene                     | 1.904 | 1.835 | 3.6   | 106   | 0.00     |
| 79 S | Terphenyl-d14              | 1.290 | 1.194 | 7.4   | 104   | 0.00     |
| 80   | Butylbenzylphthalate       | 0.670 | 0.685 | -2.2  | 109   | 0.00     |
| 81   | Benzo(a)anthracene         | 1.291 | 1.244 | 3.6   | 107   | 0.00     |
| 82   | 3,3'-Dichlorobenzidine     | 0.403 | 0.401 | 0.5   | 109   | 0.00     |
| 83   | Chrysene                   | 1.284 | 1.232 | 4.0   | 107   | 0.00     |
| 84   | Bis(2-ethylhexyl)phthalate | 0.688 | 0.687 | 0.1   | 108   | 0.00     |
| 85 c | Di-n-octyl phthalate       | 1.093 | 1.055 | 3.5   | 103   | 0.00     |
| 86 I | Perylene-d12               | 1.000 | 1.000 | 0.0   | 100   | 0.00     |
| 87   | Indeno(1,2,3-cd)pyrene     | 1.311 | 1.335 | -1.8  | 97    | 0.00     |
| 88   | Benzo(b)fluoranthene       | 1.083 | 1.085 | -0.2  | 100   | 0.00     |
| 89   | Benzo(k)fluoranthene       | 1.069 | 1.059 | 0.9   | 100   | 0.00     |
| 90 C | Benzo(a)pyrene             | 1.032 | 1.043 | -1.1  | 99    | 0.00     |
| 91   | Dibenzo(a,h)anthracene     | 1.073 | 1.087 | -1.3  | 97    | 0.00     |
| 92   | Benzo(g,h,i)perylene       | 1.121 | 1.152 | -2.8  | 97    | 0.00     |

Data Path : Z:\svoasrv\HPCHEM1\BNA\_F\Data\BF091223\  
Data File : BF135170.D  
Acq On : 12 Sep 2023 08:53  
Operator : CG\JU  
Sample : SSTDCCC040  
Misc :  
ALS Vial : 2 Sample Multiplier: 1

Instrument :  
BNA\_F  
LabSampleId :  
SSTDCCC040

Quant Time: Sep 13 01:54:08 2023  
Quant Method : Z:\svoasrv\HPCHEM1\BNA\_F\Methods\8270-BF091123.M  
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION  
QLast Update : Tue Sep 12 00:19:07 2023  
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min  
Max. RRF Dev : 25% Max. Rel. Area : 150%

| Compound | AvgRF | CCRF | %Dev | Area% | Dev(min) |
|----------|-------|------|------|-------|----------|
|----------|-------|------|------|-------|----------|

(#) = Out of Range

SPCC's out = 0 CCC's out = 0