

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF091223\
 Data File : BF135185.D
 Acq On : 12 Sep 2023 17:07
 Operator : CG\JU
 Sample : SSTDCCC040
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_F
 LabSampleId :
 SSTDCCC040

Quant Time: Sep 13 02:39:25 2023
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF091123.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Tue Sep 12 00:19:07 2023
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	20.000	20.000	0.0	103	0.00
2	1,4-Dioxane	40.000	39.857	0.4	103	0.00
3	Pyridine	40.000	40.640	-1.6	102	0.00
4	n-Nitrosodimethylamine	40.000	40.463	-1.2	102	0.00
5 S	2-Fluorophenol	80.000	78.862	1.4	102	0.00
6	Aniline	40.000	39.236	1.9	101	0.00
7 S	Phenol-d6	80.000	78.912	1.4	101	0.00
8	2-Chlorophenol	40.000	39.783	0.5	102	0.00
9	Benzaldehyde	40.000	38.146	4.6	100	0.00
10 C	Phenol	40.000	39.555	1.1	101	0.00
11	bis(2-Chloroethyl)ether	40.000	39.474	1.3	101	0.00
12	1,3-Dichlorobenzene	40.000	39.415	1.5	102	0.00
13 C	1,4-Dichlorobenzene	40.000	39.447	1.4	102	0.00
14	1,2-Dichlorobenzene	40.000	39.230	1.9	103	0.00
15	Benzyl Alcohol	40.000	40.203	-0.5	103	0.00
16	2,2'-oxybis(1-Chloropropane	40.000	39.860	0.4	102	0.00
17	2-Methylphenol	40.000	39.973	0.1	102	0.00
18	Hexachloroethane	40.000	40.209	-0.5	103	0.00
19 P	n-Nitroso-di-n-propylamine	40.000	38.712	3.2	101	0.00
20	3+4-Methylphenols	40.000	38.910	2.7	102	0.00
21 I	Naphthalene-d8	20.000	20.000	0.0	107	0.00
22	Acetophenone	40.000	37.291	6.8	103	0.00
23 S	Nitrobenzene-d5	80.000	76.449	4.4	103	0.00
24	Nitrobenzene	40.000	39.028	2.4	104	0.00
25	Isophorone	40.000	37.975	5.1	103	0.00
26 C	2-Nitrophenol	40.000	40.041	-0.1	107	0.00
27	2,4-Dimethylphenol	40.000	38.317	4.2	105	0.00
28	bis(2-Chloroethoxy)methane	40.000	39.285	1.8	108	0.00
29 C	2,4-Dichlorophenol	40.000	39.288	1.8	106	0.00
30	1,2,4-Trichlorobenzene	40.000	38.785	3.0	105	0.00
31	Naphthalene	40.000	39.348	1.6	107	0.00
32	Benzoic acid	40.000	42.311	-5.8	108	0.00
33	4-Chloroaniline	40.000	39.434	1.4	107	0.00
34 C	Hexachlorobutadiene	40.000	39.255	1.9	107	0.00
35	Caprolactam	40.000	41.402	-3.5	112	0.00
36 C	4-Chloro-3-methylphenol	40.000	38.681	3.3	104	0.00
37	2-Methylnaphthalene	40.000	37.642	5.9	104	0.00
38	1-Methylnaphthalene	40.000	38.443	3.9	106	0.00
39 I	Acenaphthene-d10	20.000	20.000	0.0	106	0.00
40	1,2,4,5-Tetrachlorobenzene	40.000	40.358	-0.9	107	0.00
41 P	Hexachlorocyclopentadiene	40.000	38.617	3.5	106	0.00
42 S	2,4,6-Tribromophenol	80.000	80.326	-0.4	108	0.00
43 C	2,4,6-Trichlorophenol	40.000	40.604	-1.5	108	0.00
44	2,4,5-Trichlorophenol	40.000	40.779	-1.9	107	0.00
45 S	2-Fluorobiphenyl	80.000	76.823	4.0	106	0.00
46	1,1'-Biphenyl	40.000	39.903	0.2	106	0.00
47	2-Chloronaphthalene	40.000	39.536	1.2	107	0.00

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF0911223\
 Data File : BF135185.D
 Acq On : 12 Sep 2023 17:07
 Operator : CG\JU
 Sample : SSTDCCC040
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_F
 LabSampleId :
 SSTDCCC040

Quant Time: Sep 13 02:39:25 2023
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF091123.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Tue Sep 12 00:19:07 2023
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
48	2-Nitroaniline	40.000	41.656	-4.1	110	0.00
49	Acenaphthylene	40.000	39.799	0.5	106	0.00
50	Dimethylphthalate	40.000	39.945	0.1	109	0.00
51	2,6-Dinitrotoluene	40.000	40.967	-2.4	109	0.00
52 C	Acenaphthene	40.000	39.700	0.7	108	0.00
53	3-Nitroaniline	40.000	41.814	-4.5	111	0.00
54 P	2,4-Dinitrophenol	40.000	41.405	-3.5	116	0.00
55	Dibenzofuran	40.000	39.718	0.7	108	0.00
56 P	4-Nitrophenol	40.000	45.054	-12.6	114	0.00
57	2,4-Dinitrotoluene	40.000	41.296	-3.2	108	0.00
58	Fluorene	40.000	39.067	2.3	107	0.00
59	2,3,4,6-Tetrachlorophenol	40.000	40.629	-1.6	109	0.00
60	Diethylphthalate	40.000	40.307	-0.8	108	0.00
61	4-Chlorophenyl-phenylether	40.000	39.045	2.4	107	0.00
62	4-Nitroaniline	40.000	42.450	-6.1	111	0.00
63	Azobenzene	40.000	40.869	-2.2	110	0.00
64 I	Phenanthrene-d10	20.000	20.000	0.0	104	0.00
65	4,6-Dinitro-2-methylphenol	40.000	44.026	-10.1	116	0.00
66 c	n-Nitrosodiphenylamine	40.000	40.610	-1.5	110	0.00
67	4-Bromophenyl-phenylether	40.000	41.478	-3.7	112	0.00
68	Hexachlorobenzene	40.000	40.805	-2.0	109	0.00
69	Atrazine	40.000	41.041	-2.6	111	0.00
70 C	Pentachlorophenol	40.000	43.372	-8.4	108	0.00
71	Phenanthrene	40.000	39.328	1.7	105	0.00
72	Anthracene	40.000	39.232	1.9	106	0.00
73	Carbazole	40.000	40.006	-0.0	107	0.00
74	Di-n-butylphthalate	40.000	41.820	-4.6	111	0.00
75 C	Fluoranthene	40.000	40.844	-2.1	109	0.00
76 I	Chrysene-d12	20.000	20.000	0.0	118	0.00
77	Benzydine	40.000	40.394	-1.0	132	0.00
78	Pyrene	40.000	37.562	6.1	111	0.00
79 S	Terphenyl-d14	80.000	74.744	6.6	113	0.00
80	Butylbenzylphthalate	40.000	42.705	-6.8	122	0.00
81	Benzo(a)anthracene	40.000	38.924	2.7	116	0.00
82	3,3'-Dichlorobenzidine	40.000	40.352	-0.9	119	0.00
83	Chrysene	40.000	38.671	3.3	116	0.00
84	Bis(2-ethylhexyl)phthalate	40.000	43.553	-8.9	126	0.00
85 c	Di-n-octyl phthalate	40.000	41.545	-3.9	119	0.00
86 I	Perylene-d12	20.000	20.000	0.0	100	0.00
87	Indeno(1,2,3-cd)pyrene	40.000	39.880	0.3	95	0.00
88	Benzo(b)fluoranthene	40.000	42.953	-7.4	108	0.00
89	Benzo(k)fluoranthene	40.000	39.034	2.4	100	0.00
90 C	Benzo(a)pyrene	40.000	40.913	-2.3	101	0.00
91	Dibenzo(a,h)anthracene	40.000	39.907	0.2	97	0.00
92	Benzo(g,h,i)perylene	40.000	40.079	-0.2	96	0.00

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF091223\
Data File : BF135185.D
Acq On : 12 Sep 2023 17:07
Operator : CG\JU
Sample : SSTDCCC040
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
BNA_F
LabSampleId :
SSTDCCC040

Quant Time: Sep 13 02:39:25 2023
Quant Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF091123.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Tue Sep 12 00:19:07 2023
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

Compound	Amount	Calc.	%Dev	Area%	Dev(min)
----------	--------	-------	------	-------	----------

(#) = Out of Range

SPCC's out = 0 CCC's out = 0