

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF091223\
 Data File : BF135200.D
 Acq On : 13 Sep 2023 01:49
 Operator : CG\JU
 Sample : SSTDCCC040
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_F
 LabSampleId :
 SSTDCCC040

Quant Time: Sep 13 02:49:09 2023
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF091123.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Tue Sep 12 00:19:07 2023
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	20.000	20.000	0.0	103	0.00
2	1,4-Dioxane	40.000	40.304	-0.8	105	0.00
3	Pyridine	40.000	40.542	-1.4	102	0.00
4	n-Nitrosodimethylamine	40.000	40.964	-2.4	103	0.00
5 S	2-Fluorophenol	80.000	78.739	1.6	102	0.00
6	Aniline	40.000	39.976	0.1	103	0.00
7 S	Phenol-d6	80.000	79.692	0.4	102	0.00
8	2-Chlorophenol	40.000	39.801	0.5	102	0.00
9	Benzaldehyde	40.000	39.292	1.8	104	0.00
10 C	Phenol	40.000	39.967	0.1	102	0.00
11	bis(2-Chloroethyl)ether	40.000	39.949	0.1	102	0.00
12	1,3-Dichlorobenzene	40.000	40.105	-0.3	104	0.00
13 C	1,4-Dichlorobenzene	40.000	39.520	1.2	103	0.00
14	1,2-Dichlorobenzene	40.000	39.186	2.0	103	0.00
15	Benzyl Alcohol	40.000	39.841	0.4	102	0.00
16	2,2'-oxybis(1-Chloropropane	40.000	39.006	2.5	100	0.00
17	2-Methylphenol	40.000	40.302	-0.8	103	0.00
18	Hexachloroethane	40.000	39.671	0.8	102	0.00
19 P	n-Nitroso-di-n-propylamine	40.000	39.098	2.3	103	0.00
20	3+4-Methylphenols	40.000	39.125	2.2	103	0.00
21 I	Naphthalene-d8	20.000	20.000	0.0	103	0.00
22	Acetophenone	40.000	38.407	4.0	102	0.00
23 S	Nitrobenzene-d5	80.000	79.706	0.4	103	0.00
24	Nitrobenzene	40.000	39.849	0.4	103	0.00
25	Isophorone	40.000	39.304	1.7	103	0.00
26 C	2-Nitrophenol	40.000	40.483	-1.2	104	0.00
27	2,4-Dimethylphenol	40.000	39.172	2.1	103	0.00
28	bis(2-Chloroethoxy)methane	40.000	39.686	0.8	105	0.00
29 C	2,4-Dichlorophenol	40.000	39.880	0.3	104	0.00
30	1,2,4-Trichlorobenzene	40.000	39.031	2.4	102	0.00
31	Naphthalene	40.000	39.277	1.8	103	0.00
32	Benzoic acid	40.000	39.110	2.2	97	0.00
33	4-Chloroaniline	40.000	39.428	1.4	103	0.00
34 C	Hexachlorobutadiene	40.000	39.460	1.3	104	0.00
35	Caprolactam	40.000	41.716	-4.3	109	0.00
36 C	4-Chloro-3-methylphenol	40.000	40.161	-0.4	104	0.00
37	2-Methylnaphthalene	40.000	38.933	2.7	104	0.00
38	1-Methylnaphthalene	40.000	39.050	2.4	104	0.00
39 I	Acenaphthene-d10	20.000	20.000	0.0	106	0.00
40	1,2,4,5-Tetrachlorobenzene	40.000	39.425	1.4	105	0.00
41 P	Hexachlorocyclopentadiene	40.000	34.803	13.0	93	0.00
42 S	2,4,6-Tribromophenol	80.000	80.789	-1.0	109	0.00
43 C	2,4,6-Trichlorophenol	40.000	39.085	2.3	104	0.00
44	2,4,5-Trichlorophenol	40.000	39.829	0.4	105	0.00
45 S	2-Fluorobiphenyl	80.000	76.096	4.9	105	0.00
46	1,1'-Biphenyl	40.000	38.805	3.0	104	0.00
47	2-Chloronaphthalene	40.000	38.796	3.0	105	0.00

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF091223\
 Data File : BF135200.D
 Acq On : 13 Sep 2023 01:49
 Operator : CG\JU
 Sample : SSTDCCC040
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_F
 LabSampleId :
 SSTDCCC040

Quant Time: Sep 13 02:49:09 2023
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF091123.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Tue Sep 12 00:19:07 2023
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
48	2-Nitroaniline	40.000	40.015	-0.0	106	0.00
49	Acenaphthylene	40.000	38.904	2.7	104	0.00
50	Dimethylphthalate	40.000	39.421	1.4	108	0.00
51	2,6-Dinitrotoluene	40.000	40.502	-1.3	108	0.00
52 C	Acenaphthene	40.000	39.922	0.2	108	0.00
53	3-Nitroaniline	40.000	39.748	0.6	106	0.00
54 P	2,4-Dinitrophenol	40.000	33.341	16.6	89	0.00
55	Dibenzofuran	40.000	38.792	3.0	105	0.00
56 P	4-Nitrophenol	40.000	40.907	-2.3	104	0.00
57	2,4-Dinitrotoluene	40.000	40.349	-0.9	105	0.00
58	Fluorene	40.000	39.154	2.1	107	0.00
59	2,3,4,6-Tetrachlorophenol	40.000	39.753	0.6	106	0.00
60	Diethylphthalate	40.000	39.921	0.2	107	0.00
61	4-Chlorophenyl-phenylether	40.000	39.116	2.2	108	0.00
62	4-Nitroaniline	40.000	40.515	-1.3	106	0.00
63	Azobenzene	40.000	39.600	1.0	107	0.00
64 I	Phenanthrene-d10	20.000	20.000	0.0	105	0.00
65	4,6-Dinitro-2-methylphenol	40.000	38.676	3.3	102	0.00
66 c	n-Nitrosodiphenylamine	40.000	39.379	1.6	107	0.00
67	4-Bromophenyl-phenylether	40.000	40.013	-0.0	109	0.00
68	Hexachlorobenzene	40.000	39.692	0.8	107	0.00
69	Atrazine	40.000	40.860	-2.1	111	0.00
70 C	Pentachlorophenol	40.000	39.792	0.5	100	0.00
71	Phenanthrene	40.000	39.069	2.3	105	0.00
72	Anthracene	40.000	39.193	2.0	107	0.00
73	Carbazole	40.000	40.027	-0.1	107	0.00
74	Di-n-butylphthalate	40.000	41.628	-4.1	111	0.00
75 C	Fluoranthene	40.000	39.762	0.6	107	0.00
76 I	Chrysene-d12	20.000	20.000	0.0	104	0.00
77	Benzydine	40.000	45.959	-14.9	133	0.00
78	Pyrene	40.000	41.194	-3.0	108	0.00
79 S	Terphenyl-d14	80.000	83.794	-4.7	111	0.00
80	Butylbenzylphthalate	40.000	45.800	-14.5	116	0.00
81	Benzo(a)anthracene	40.000	39.279	1.8	104	0.00
82	3,3'-Dichlorobenzidine	40.000	42.806	-7.0	111	0.00
83	Chrysene	40.000	39.225	1.9	104	0.00
84	Bis(2-ethylhexyl)phthalate	40.000	49.858	-24.6	128	0.00
85 c	Di-n-octyl phthalate	40.000	45.592	-14.0	116	0.00
86 I	Perylene-d12	20.000	20.000	0.0	106	0.00
87	Indeno(1,2,3-cd)pyrene	40.000	41.521	-3.8	105	0.00
88	Benzo(b)fluoranthene	40.000	39.969	0.1	107	0.00
89	Benzo(k)fluoranthene	40.000	38.834	2.9	105	0.00
90 C	Benzo(a)pyrene	40.000	39.969	0.1	105	0.00
91	Dibenzo(a,h)anthracene	40.000	41.456	-3.6	106	0.00
92	Benzo(g,h,i)perylene	40.000	41.036	-2.6	103	0.00

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF091223\
Data File : BF135200.D
Acq On : 13 Sep 2023 01:49
Operator : CG\JU
Sample : SSTDCCC040
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
BNA_F
LabSampleId :
SSTDCCC040

Quant Time: Sep 13 02:49:09 2023
Quant Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF091123.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Tue Sep 12 00:19:07 2023
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

Compound	Amount	Calc.	%Dev	Area%	Dev(min)
----------	--------	-------	------	-------	----------

(#) = Out of Range

SPCC's out = 0 CCC's out = 0