

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF091323\
 Data File : BF135223.D
 Acq On : 13 Sep 2023 15:18
 Operator : CG\JU
 Sample : SSTDCCC040
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_F
 LabSampleId :
 SSTDCCC040

Quant Time: Sep 13 15:52:56 2023
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF091123.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Tue Sep 12 00:19:07 2023
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	104	0.00
2	1,4-Dioxane	0.539	0.530	1.7	103	0.01
3	Pyridine	1.451	1.476	-1.7	103	0.01
4	n-Nitrosodimethylamine	0.770	0.775	-0.6	103	0.02
5 S	2-Fluorophenol	1.230	1.215	1.2	103	0.00
6	Aniline	2.050	2.024	1.3	103	0.00
7 S	Phenol-d6	1.564	1.543	1.3	102	0.00
8	2-Chlorophenol	1.313	1.317	-0.3	104	0.00
9	Benzaldehyde	0.891	0.848	4.8	101	0.00
10 C	Phenol	1.715	1.686	1.7	102	0.00
11	bis(2-Chloroethyl)ether	1.286	1.272	1.1	102	0.00
12	1,3-Dichlorobenzene	1.334	1.330	0.3	104	0.00
13 C	1,4-Dichlorobenzene	1.357	1.348	0.7	104	0.00
14	1,2-Dichlorobenzene	1.260	1.239	1.7	104	0.00
15	Benzyl Alcohol	1.122	1.126	-0.4	103	0.00
16	2,2'-oxybis(1-Chloropropane	2.424	2.347	3.2	100	0.00
17	2-Methylphenol	1.158	1.141	1.5	102	0.00
18	Hexachloroethane	0.502	0.501	0.2	104	0.00
19 P	n-Nitroso-di-n-propylamine	0.968	0.928	4.1	101	0.00
20	3+4-Methylphenols	1.393	1.360	2.4	104	0.00
21 I	Naphthalene-d8	1.000	1.000	0.0	103	0.00
22	Acetophenone	0.457	0.443	3.1	103	0.00
23 S	Nitrobenzene-d5	0.368	0.364	1.1	103	0.00
24	Nitrobenzene	0.364	0.361	0.8	102	0.00
25	Isophorone	0.676	0.661	2.2	102	0.00
26 C	2-Nitrophenol	0.175	0.178	-1.7	105	0.00
27	2,4-Dimethylphenol	0.294	0.290	1.4	104	0.00
28	bis(2-Chloroethoxy)methane	0.405	0.401	1.0	105	0.00
29 C	2,4-Dichlorophenol	0.272	0.269	1.1	103	0.00
30	1,2,4-Trichlorobenzene	0.284	0.283	0.4	104	0.00
31	Naphthalene	0.972	0.961	1.1	104	0.00
32	Benzoic acid	0.221	0.224	-1.4	101	0.00
33	4-Chloroaniline	0.426	0.417	2.1	103	0.00
34 C	Hexachlorobutadiene	0.171	0.170	0.6	104	0.00
35	Caprolactam	0.091	0.091	0.0	105	0.00
36 C	4-Chloro-3-methylphenol	0.302	0.302	0.0	104	0.00
37	2-Methylnaphthalene	0.653	0.644	1.4	105	0.00
38	1-Methylnaphthalene	0.612	0.596	2.6	104	0.00
39 I	Acenaphthene-d10	1.000	1.000	0.0	103	0.00
40	1,2,4,5-Tetrachlorobenzene	0.579	0.594	-2.6	106	0.00
41 P	Hexachlorocyclopentadiene	0.210	0.173	17.6	81	0.00
42 S	2,4,6-Tribromophenol	0.199	0.202	-1.5	107	0.00
43 C	2,4,6-Trichlorophenol	0.379	0.391	-3.2	107	0.00
44	2,4,5-Trichlorophenol	0.437	0.440	-0.7	103	0.00
45 S	2-Fluorobiphenyl	1.326	1.291	2.6	105	0.00
46	1,1'-Biphenyl	1.537	1.527	0.7	103	0.00
47	2-Chloronaphthalene	1.115	1.112	0.3	105	0.00

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF091323\
 Data File : BF135223.D
 Acq On : 13 Sep 2023 15:18
 Operator : CG\JU
 Sample : SSTDCCC040
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_F
 LabSampleId :
 SSTDCCC040

Quant Time: Sep 13 15:52:56 2023
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF091123.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Tue Sep 12 00:19:07 2023
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
48	2-Nitroaniline	0.433	0.440	-1.6	104	0.00
49	Acenaphthylene	1.853	1.840	0.7	103	0.00
50	Dimethylphthalate	1.352	1.349	0.2	106	0.00
51	2,6-Dinitrotoluene	0.296	0.299	-1.0	105	0.00
52 C	Acenaphthene	1.095	1.089	0.5	105	0.00
53	3-Nitroaniline	0.348	0.348	0.0	104	0.00
54 P	2,4-Dinitrophenol	0.143	0.146	-2.1	103	0.00
55	Dibenzofuran	1.671	1.652	1.1	104	0.00
56 P	4-Nitrophenol	0.221	0.225	-1.8	101	0.00
57	2,4-Dinitrotoluene	0.371	0.374	-0.8	103	0.00
58	Fluorene	1.284	1.265	1.5	105	0.00
59	2,3,4,6-Tetrachlorophenol	0.336	0.350	-4.2	108	0.00
60	Diethylphthalate	1.357	1.336	1.5	103	0.00
61	4-Chlorophenyl-phenylether	0.617	0.611	1.0	106	0.00
62	4-Nitroaniline	0.334	0.335	-0.3	102	0.00
63	Azobenzene	1.328	1.314	1.1	104	0.00
64 I	Phenanthrene-d10	1.000	1.000	0.0	102	0.00
65	4,6-Dinitro-2-methylphenol	0.106	0.108	-1.9	105	0.00
66 c	n-Nitrosodiphenylamine	0.605	0.600	0.8	105	0.00
67	4-Bromophenyl-phenylether	0.199	0.201	-1.0	107	0.00
68	Hexachlorobenzene	0.203	0.205	-1.0	107	0.00
69	Atrazine	0.163	0.159	2.5	104	0.00
70 C	Pentachlorophenol	0.121	0.128	-5.8	103	0.00
71	Phenanthrene	0.946	0.929	1.8	103	0.00
72	Anthracene	0.972	0.957	1.5	104	0.00
73	Carbazole	0.885	0.860	2.8	102	0.00
74	Di-n-butylphthalate	1.043	1.041	0.2	104	0.00
75 C	Fluoranthene	0.986	0.965	2.1	103	0.00
76 I	Chrysene-d12	1.000	1.000	0.0	98	0.00
77	Benzidine	0.410	0.467	-13.9	124	0.00
78	Pyrene	1.904	2.015	-5.8	104	0.00
79 S	Terphenyl-d14	1.290	1.335	-3.5	103	0.00
80	Butylbenzylphthalate	0.670	0.735	-9.7	104	0.00
81	Benzo(a)anthracene	1.291	1.287	0.3	99	0.00
82	3,3'-Dichlorobenzidine	0.403	0.429	-6.5	104	0.00
83	Chrysene	1.284	1.270	1.1	98	0.00
84	Bis(2-ethylhexyl)phthalate	0.688	0.822	-19.5	115	0.00
85 c	Di-n-octyl phthalate	1.093	1.368	-25.2#	119	0.00
86 I	Perylene-d12	1.000	1.000	0.0	113	0.00
87	Indeno(1,2,3-cd)pyrene	1.311	1.238	5.6	101	0.02
88	Benzo(b)fluoranthene	1.083	1.038	4.2	109	0.00
89	Benzo(k)fluoranthene	1.069	1.021	4.5	109	0.00
90 C	Benzo(a)pyrene	1.032	1.026	0.6	110	0.01
91	Dibenzo(a,h)anthracene	1.073	1.010	5.9	102	0.02
92	Benzo(g,h,i)perylene	1.121	1.051	6.2	100	0.02

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF091323\
Data File : BF135223.D
Acq On : 13 Sep 2023 15:18
Operator : CG\JU
Sample : SSTDCCC040
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
BNA_F
LabSampleId :
SSTDCCC040

Quant Time: Sep 13 15:52:56 2023
Quant Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF091123.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Tue Sep 12 00:19:07 2023
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
----------	-------	------	------	-------	----------

(#) = Out of Range

SPCC's out = 0 CCC's out = 1