

Data Path : Z:\SVOASRV\HPCHEM1\BNA F\DATA\BF091619\
 Data File : BF117056.D
 Acq On : 16 Sep 2019 12:11
 Operator : HP/JU
 Sample : K4663-06MSD
 Misc :
 ALS Vial : 9 Sample Multiplier: 1

Instrument :
 BNA_F
 ClientSampleId :

Quant Time: Sep 16 14:57:45 2019
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA F\METHODS\8270-BF091319.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Fri Sep 13 16:23:17 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	6.82	152	60774	20.00	ng	0.00
21) Naphthalene-d8	8.10	136	231297	20.00	ng	0.00
39) Acenaphthene-d10	9.85	164	121668	20.00	ng	0.00
64) Phenanthrene-d10	11.33	188	204516	20.00	ng	0.00
76) Chrysene-d12	13.97	240	161206	20.00	ng	0.00
87) Perylene-d12	15.42	264	152863	20.00	ng	0.00

System Monitoring Compounds

5) 2-Fluorophenol	5.46	112	471841	121.21	ng	0.02
7) Phenol-d6	6.46	99	587679	116.82	ng	0.00
23) Nitrobenzene-d5	7.39	82	433799	98.54	ng	0.00
42) 2,4,6-Tribromophenol	10.65	330	152313	149.79	ng	0.00
45) 2-Fluorobiphenyl	9.17	172	747465	96.06	ng	0.00
79) Terphenyl-d14	12.92	244	795540	92.24	ng	0.00

Target Compounds

						Qvalue
2) 1,4-Dioxane	2.68	88	63262	38.838	ng	# 86
3) Pyridine	3.41	79	123202	27.634	ng	98
4) n-Nitrosodimethylamine	3.32	42	70775	46.258	ng	98
6) Aniline	6.49	93	168431	26.004	ng	90
8) 2-Chlorophenol	6.62	128	195172	46.476	ng	97
9) Benzaldehyde	6.37	77	89482	34.239	ng	98
10) Phenol	6.47	94	218054	39.318	ng	97
11) bis(2-Chloroethyl)ether	6.56	93	183955	45.131	ng	97
12) 1,3-Dichlorobenzene	6.77	146	200671	42.585	ng	98
13) 1,4-Dichlorobenzene	6.84	146	205882	42.812	ng	98
14) 1,2-Dichlorobenzene	7.00	146	193374	43.189	ng	97
15) Benzyl Alcohol	6.97	79	181648	47.086	ng	99
16) 2,2'-oxybis(1-Chloropropan	7.10	45	176309	43.781	ng	95
17) 2-Methylphenol	7.08	107	159194	45.232	ng	94
18) Hexachloroethane	7.33	117	77769	42.037	ng	96
19) n-Nitroso-di-n-propylamine	7.23	70	144239	47.086	ng	94
20) 3+4-Methylphenols	7.23	107	195266	44.542	ng	98
22) Acetophenone	7.23	105	295936	50.359	ng	99
24) Nitrobenzene	7.40	77	222970	51.290	ng	98
25) Isophorone	7.65	82	398294	51.101	ng	100
26) 2-Nitrophenol	7.72	139	100804	53.984	ng	99
27) 2,4-Dimethylphenol	7.76	122	172349	58.203	ng	99
28) bis(2-Chloroethoxy)methane	7.85	93	239800	49.936	ng	99
29) 2,4-Dichlorophenol	7.96	162	162330	52.318	ng	100
30) 1,2,4-Trichlorobenzene	8.05	180	160419	47.856	ng	98
31) Naphthalene	8.13	128	568901	51.142	ng	98
32) Benzoic acid	7.87	122	39001	21.641	ng	98
33) 4-Chloroaniline	8.17	127	93537	18.958	ng	98
34) Hexachlorobutadiene	8.25	225	90355	47.510	ng	97
35) Caprolactam	8.54	113	21257	20.241	ng	90
36) 4-Chloro-3-methylphenol	8.66	107	193635	53.522	ng	96
37) 2-Methylnaphthalene	8.82	142	388541	51.869	ng	99
38) 1-Methylnaphthalene	8.91	142	364415	51.943	ng	99
40) 1,2,4,5-Tetrachlorobenzene	8.98	216	152710	48.606	ng	98


Data Path : Z:\SVOASRV\HPCHEM1\BNA F\DATA\BF091619\
 Data File : BF117056.D
 Acq On : 16 Sep 2019 12:11
 Operator : HP/JU
 Sample : K4663-06MSD
 Misc :
 ALS Vial : 9 Sample Multiplier: 1

Instrument :
 BNA_F
 ClientSampleId :

Quant Time: Sep 16 14:57:45 2019
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA F\METHODS\8270-BF091319.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Fri Sep 13 16:23:17 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
41) Hexachlorocyclopentadiene	8.97	237	142143	100.161	nq	98
43) 2,4,6-Trichlorophenol	9.09	196	106636	50.014	nq	99
44) 2,4,5-Trichlorophenol	9.14	196	117489	51.576	nq	95
46) 1,1'-Biphenyl	9.27	154	466739	48.906	nq	99
47) 2-Chloronaphthalene	9.30	162	366358	48.640	nq	98
48) 2-Nitroaniline	9.40	65	121792	50.413	nq	99
49) Acenaphthylene	9.72	152	579874	51.392	nq	100
50) Dimethylphthalate	9.57	163	434627	50.452	nq	99
51) 2,6-Dinitrotoluene	9.64	165	95524	54.444	nq	96
52) Acenaphthene	9.89	154	332216	49.669	nq	98
53) 3-Nitroaniline	9.80	138	77157	34.854	nq	98
54) 2,4-Dinitrophenol	9.92	184	38799	54.512	nq	93
55) Dibenzofuran	10.06	168	504555	51.128	nq	96
56) 4-Nitrophenol	9.97	139	135064	94.139	nq	97
57) 2,4-Dinitrotoluene	10.05	165	129399	56.817	nq	95
58) Fluorene	10.40	166	411392	51.751	nq	99
59) 2,3,4,6-Tetrachlorophenol	10.18	232	93962	53.902	nq	98
60) Diethylphthalate	10.27	149	425400	50.193	nq	97
61) 4-Chlorophenyl-phenylether	10.39	204	175734	51.094	nq	96
62) 4-Nitroaniline	10.42	138	117448	51.017	nq	98
63) Azobenzene	10.55	77	451538	52.164	nq	97
65) 4,6-Dinitro-2-methylphenol	10.45	198	39047	42.779	nq	85
66) n-Nitrosodiphenylamine	10.51	169	361869	51.233	nq	99
67) 4-Bromophenyl-phenylether	10.88	248	104411	49.991	nq	93
68) Hexachlorobenzene	10.95	284	110836	48.518	nq	96
69) Atrazine	11.03	200	101786	65.625	nq	98
70) Pentachlorophenol	11.15	266	112903	105.436	nq	99
71) Phenanthrene	11.36	178	591492	50.919	nq	100
72) Anthracene	11.41	178	615871	51.931	nq	98
73) Carbazole	11.56	167	587907	52.226	nq	99
74) Di-n-butylphthalate	11.89	149	694357	51.843	nq	98
75) Fluoranthene	12.54	202	629529	52.743	nq	98
77) Benzidine	12.66	184	231719	44.623	nq	99
78) Pyrene	12.77	202	628800	46.487	nq	100
80) Butylbenzylphthalate	13.38	149	305184	47.749	nq	97
81) Benzo(a)anthracene	13.96	228	526151	48.852	nq	99
82) 3,3'-Dichlorobenzidine	13.92	252	104604	30.139	nq	96
83) Chrysene	14.00	228	528762	50.336	nq	98
84) Bis(2-ethylhexyl)phthalate	13.94	149	416408	50.530	nq	98
85) Di-n-octyl phthalate	14.56	149	713591	55.016	nq	99
86) Indeno(1,2,3-cd)pyrene	16.86	276	415837	54.260	nq	99
88) Benzo(b)fluoranthene	15.00	252	470612	50.825	nq	98
89) Benzo(k)fluoranthene	15.03	252	434999	49.594	nq	98
90) Benzo(a)pyrene	15.36	252	424129	52.198	nq	98
91) Dibenzo(a,h)anthracene	16.87	278	347632	53.706	nq	99
92) Benzo(g,h,i)perylene	17.29	276	345643	53.586	nq	97


(#) = qualifier out of range (m) = manual integration (+) = signals summed






#1
 1,4-Dichlorobenzene-d4
 Concen: 20.000 ng
 RT: 6.82 min Scan# 805
 Delta R.T. 0.00 min
 Lab File: BF117056.D
 Acq: 16 Sep 2019 12:11

Instrument :
 BNA_F
 ClientSampleId :

Tot Ion:152 Resp: 60774
 Ion Ratio Lower Upper
 152 100
 150 155.4 127.5 191.3
 115 66.0 55.0 82.4




Abundance Ion 152.00 (151.70 to 152.70): E
 Ion 150.00 (149.70 to 150.70): E
 Ion 115.00 (114.70 to 115.70): E





#2
 1,4-Dioxane
 Concen: 38.838 ng
 RT: 2.68 min Scan# 100
 Delta R.T. 0.11 min
 Lab File: BF117056.D
 Acq: 16 Sep 2019 12:11

Tgt Ion: 88 Resp: 63262
 Ion Ratio Lower Upper
 88 100
 58 60.8 49.3 73.9
 43 0.0 18.8 28.2#



Abundance Ion 88.00 (87.70 to 88.70): BF1
 Ion 58.00 (57.70 to 58.70): BF1
 Ion 43.00 (42.70 to 43.70): BF1





#3

Pvridine

Concen: 27.634 ng

RT: 3.41 min Scan# 225

Delta R.T. 0.11 min

Lab File: BF117056.D

Acq: 16 Sep 2019 12:11

Instrument :

BNA_F

ClientSampleId :


Tot Ion: 79 Resp: 123202

Ion Ratio Lower Upper




79 100

52 53.1 43.5 65.3

51 26.4 21.8 32.6



Abundance Ion 79.00 (78.70 to 79.70): BF1
Ion 52.00 (51.70 to 52.70): BF1
Ion 51.00 (50.70 to 51.70): BF1



#4

n-Nitrosodimethylamine

Concen: 46.258 ng

RT: 3.32 min Scan# 210

Delta R.T. 0.07 min

Lab File: BF117056.D

Acq: 16 Sep 2019 12:11


Tgt Ion: 42 Resp: 70775

Ion Ratio Lower Upper



42 100


74 155.6 127.1 190.7

44 5.2 4.3 6.5



Abundance Ion 42.00 (41.70 to 42.70): BF1
Ion 74.00 (73.70 to 74.70): BF1
Ion 44.00 (43.70 to 44.70): BF1





#5

2-Fluorophenol

Concen: 121.213 ng

RT: 5.46 min Scan# 573

Delta R.T. 0.02 min

Lab File: BF117056.D

Acq: 16 Sep 2019 12:11

Instrument :
BNA_F
ClientSampleId :


Tot Ion:112 Resp: 471841

Ion Ratio Lower Upper




112 100

64 57.4 45.7 68.5

63 30.6 24.9 37.3



Abundance Ion 112.00 (111.70 to 112.70): BF1
Ion 64.00 (63.70 to 64.70): BF1
Ion 63.00 (62.70 to 63.70): BF1



#6

Aniline

Concen: 26.004 ng

RT: 6.49 min Scan# 748

Delta R.T. 0.00 min

Lab File: BF117056.D

Acq: 16 Sep 2019 12:11


Tgt Ion: 93 Resp: 168431

Ion Ratio Lower Upper



93 100


66 47.7 32.6 49.0

65 24.4 16.7 25.1



Abundance Ion 93.00 (92.70 to 93.70): BF1
Ion 66.00 (65.70 to 66.70): BF1
Ion 65.00 (64.70 to 65.70): BF1





#7

Phenol-d6

Concen: 116.818 ng

RT: 6.46 min Scan# 744

Delta R.T. 0.01 min

Lab File: BF117056.D

Acq: 16 Sep 2019 12:11

Instrument :

BNA_F

ClientSampleId :


Tot Ion: 99 Resp: 587679

Ion Ratio Lower Upper

99 100

42 16.2 12.2 18.2




71 38.2 29.8 44.8



Abundance Ion 99.00 (98.70 to 99.70): BF1

Ion 42.00 (41.70 to 42.70): BF1

Ion 71.00 (70.70 to 71.70): BF1



#8

2-Chlorophenol

Concen: 46.476 ng

RT: 6.62 min Scan# 770

Delta R.T. 0.01 min

Lab File: BF117056.D

Acq: 16 Sep 2019 12:11


Tgt Ion: 128 Resp: 195172

Ion Ratio Lower Upper

128 100

130 32.4 11.8 51.8



64 41.6 25.0 65.0




Abundance Ion 128.00 (127.70 to 128.70): E

Ion 130.00 (129.70 to 130.70): E

Ion 64.00 (63.70 to 64.70): BF1





#9

Benzaldehyde

Concen: 34.239 ng

RT: 6.37 min Scan# 729

Delta R.T. 0.00 min

Lab File: BF117056.D

Acq: 16 Sep 2019 12:11

Instrument :

BNA_F

ClientSampleId :


Tot Ion: 77 Resp: 89482

Ion Ratio Lower Upper

77 100

105 99.8 81.8 121.8




106 95.6 77.3 117.3



Abundance Ion 77.00 (76.70 to 77.70): BF1

Ion 105.00 (104.70 to 105.70): BF1

Ion 106.00 (105.70 to 106.70): BF1



#10

Phenol

Concen: 39.318 ng

RT: 6.47 min Scan# 746

Delta R.T. 0.00 min

Lab File: BF117056.D

Acq: 16 Sep 2019 12:11


Tgt Ion: 94 Resp: 218054

Ion Ratio Lower Upper

94 100

65 31.4 12.5 52.5



66 46.4 28.9 68.9



Abundance Ion 94.00 (93.70 to 94.70): BF1

Ion 65.00 (64.70 to 65.70): BF1

Ion 66.00 (65.70 to 66.70): BF1









#11
bis(2-Chloroethyl)ether
Concen: 45.131 ng
RT: 6.56 min Scan# 761
Delta R.T. 0.01 min
Lab File: BF117056.D
Acq: 16 Sep 2019 12:11




Instrument :
BNA_F
ClientSampled :


Tot Ion	Ratio	Lower	Upper
93	100		
63	66.1	49.3	89.3
95	32.0	11.5	51.5



#12
1,3-Dichlorobenzene
Concen: 42.585 ng
RT: 6.77 min Scan# 796
Delta R.T. 0.01 min
Lab File: BF117056.D
Acq: 16 Sep 2019 12:11

Tgt Ion	Ratio	Lower	Upper
146	100		
148	64.2	50.8	76.2
75	32.9	28.7	43.1









#13
 1,4-Dichlorobenzene
 Concen: 42.812 ng
 RT: 6.84 min Scan# 808
 Delta R.T. 0.00 min
 Lab File: BF117056.D
 Acq: 16 Sep 2019 12:11




Instrument :
 BNA_F
 ClientSampleId :




Tot Ion:146 Resp: 205882
 Ion Ratio Lower Upper
 146 100
 148 64.1 50.6 76.0
 111 45.6 35.0 52.6



#14
 1,2-Dichlorobenzene
 Concen: 43.189 ng
 RT: 7.00 min Scan# 835
 Delta R.T. 0.00 min
 Lab File: BF117056.D
 Acq: 16 Sep 2019 12:11

Tgt Ion:146 Resp: 193374
 Ion Ratio Lower Upper
 146 100
 148 64.0 52.6 78.8
 111 42.6 36.2 54.2









#15
Benzyl Alcohol
Concen: 47.086 ng
RT: 6.97 min Scan# 830
Delta R.T. 0.01 min
Lab File: BF117056.D
Acq: 16 Sep 2019 12:11


Instrument :
BNA_F
ClientSampleId :


Tot Ion	Ratio	Lower	Upper
79	100		
108	75.7	59.4	89.0
77	62.1	50.2	75.2



#16
2,2'-oxybis(1-Chloropropane)
Concen: 43.781 ng
RT: 7.10 min Scan# 852
Delta R.T. 0.00 min
Lab File: BF117056.D
Acq: 16 Sep 2019 12:11

Tgt Ion	Ratio	Lower	Upper
45	100		
77	26.1	2.7	42.7
79	19.9	0.0	38.9






#17
2-Methylphenol
Concen: 45.232 ng
RT: 7.08 min Scan# 849
Delta R.T. 0.00 min
Lab File: BF117056.D
Acq: 16 Sep 2019 12:11

Instrument :
BNA_F
ClientSampleId :

Tot Ion	Ratio	Lower	Upper
107	100		
108	116.9	87.7	131.5
77	50.7	38.6	58.0
79	50.5	36.8	55.2






Abundance

Ion 107.00 (106.70 to 107.70): E

Ion 108.00 (107.70 to 108.70): E


Ion 77.00 (76.70 to 77.70): BF1

Ion 79.00 (78.70 to 79.70): BF1



#18
Hexachloroethane
Concen: 42.037 ng
RT: 7.33 min Scan# 892
Delta R.T. 0.00 min
Lab File: BF117056.D
Acq: 16 Sep 2019 12:11

Tgt Ion	Ratio	Lower	Upper
117	100		
119	97.0	75.8	113.8
201	80.6	60.9	91.3






Abundance

Ion 117.00 (116.70 to 117.70): E

Ion 119.00 (118.70 to 119.70): E

Ion 201.00 (200.70 to 201.70): E









#19
n-Nitroso-di-n-propylamine
Concen: 47.086 ng
RT: 7.23 min Scan# 875
Delta R.T. -0.01 min
Lab File: BF117056.D
Acq: 16 Sep 2019 12:11




Instrument :
BNA_F
ClientSampleId :


Tot Ion:	70	Resp:	144239
Ion	Ratio	Lower	Upper
70	100		
42	47.5	34.0	51.0
101	9.0	7.8	11.6
130	21.0	17.8	26.8



#20
3+4-Methylphenols
Concen: 44.542 ng
RT: 7.23 min Scan# 875
Delta R.T. 0.00 min
Lab File: BF117056.D
Acq: 16 Sep 2019 12:11

Tgt Ion:	107	Resp:	195266
Ion	Ratio	Lower	Upper
107	100		
108	90.8	72.6	112.6
77	148.6	132.1	172.1
79	30.3	12.1	52.1





#21

Naphthalene-d8

Concen: 20.000 ng

RT: 8.10 min Scan# 1023

Delta R.T. 0.00 min

Lab File: BF117056.D

Acq: 16 Sep 2019 12:11

Instrument :

BNA_F

ClientSampleId :

Tot Ion:136 Resp: 231297



Ion Ratio Lower Upper

136 100

137 10.4 9.0 13.4

54 5.6 4.5 6.7

68 4.2 4.0 6.0





Abundance Ion 136.00 (135.70 to 136.70): E

Ion 137.00 (136.70 to 137.70): E

Ion 54.00 (53.70 to 54.70): BF1

Ion 68.00 (67.70 to 68.70): BF1



#22

Acetophenone

Concen: 50.359 ng

RT: 7.23 min Scan# 875

Delta R.T. 0.00 min

Lab File: BF117056.D

Acq: 16 Sep 2019 12:11

Tgt Ion:105 Resp: 295936



Ion Ratio Lower Upper

105 100

71 7.5 5.7 8.5

51 28.2 22.1 33.1

120 21.2 17.1 25.7





Abundance Ion 105.00 (104.70 to 105.70): E

Ion 71.00 (70.70 to 71.70): BF1

Ion 51.00 (50.70 to 51.70): BF1

Ion 120.00 (119.70 to 120.70): E









#23
 Nitrobenzene-d5
 Concen: 98.544 ng
 RT: 7.39 min Scan# 901
 Delta R.T. 0.00 min
 Lab File: BF117056.D
 Acq: 16 Sep 2019 12:11




Instrument :
 BNA_F
 ClientSampled :


Tot Ion	82	Resp	433799
Ion Ratio	100	Lower	Upper
128	41.4	32.3	48.5
54	41.5	33.3	49.9



#24
 Nitrobenzene
 Concen: 51.290 ng
 RT: 7.40 min Scan# 904
 Delta R.T. 0.00 min
 Lab File: BF117056.D
 Acq: 16 Sep 2019 12:11

Tgt Ion	77	Resp	222970
Ion Ratio	100	Lower	Upper
123	43.5	33.5	50.3
65	13.0	10.5	15.7





#25

Isophorone

Concen: 51.101 ng

RT: 7.65 min Scan# 945

Delta R.T. 0.00 min

Lab File: BF117056.D

Acq: 16 Sep 2019 12:11

Instrument :

BNA_F

ClientSampleId :





Tot Ion: 82 Resp: 398294

Ion Ratio Lower Upper

82 100

95 7.4 5.8 8.6

138 16.9 13.5 20.3



#26

2-Nitrophenol

Concen: 53.984 ng

RT: 7.72 min Scan# 958

Delta R.T. 0.00 min

Lab File: BF117056.D

Acq: 16 Sep 2019 12:11




Tgt Ion: 139 Resp: 100804


Ion Ratio Lower Upper

139 100

109 31.2 24.6 36.8

65 50.1 40.9 61.3









#27
 2,4-Dimethylphenol
 Concen: 58.203 ng
 RT: 7.76 min Scan# 964
 Delta R.T. 0.00 min
 Lab File: BF117056.D
 Acq: 16 Sep 2019 12:11

Instrument :
 BNA_F
 ClientSampleId :

Tot Ion:122 Resp: 172349
 Ion Ratio Lower Upper
 122 100
 107 122.6 98.9 148.3
 121 55.5 45.3 67.9




Abundance Ion 122.00 (121.70 to 122.70): E
 Ion 107.00 (106.70 to 107.70): E
 Ion 121.00 (120.70 to 121.70): E






#28
 bis(2-Chloroethoxy)methane
 Concen: 49.936 ng
 RT: 7.85 min Scan# 980
 Delta R.T. 0.00 min
 Lab File: BF117056.D
 Acq: 16 Sep 2019 12:11

Tgt Ion: 93 Resp: 239800
 Ion Ratio Lower Upper
 93 100
 95 32.4 26.2 39.4
 123 11.5 9.0 13.6



Abundance Ion 93.00 (92.70 to 93.70): BF1
 Ion 95.00 (94.70 to 95.70): BF1
 Ion 123.00 (122.70 to 123.70): E






#29
 2,4-Dichlorophenol
 Concen: 52.318 ng
 RT: 7.96 min Scan# 999
 Delta R.T. 0.00 min
 Lab File: BF117056.D
 Acq: 16 Sep 2019 12:11

Instrument :
 BNA_F
 ClientSampled :

Tot Ion	Ratio	Lower	Upper
162	100		
164	63.7	43.8	83.8
98	42.5	22.1	62.1






Abundance

Ion 162.00 (161.70 to 162.70): E


Ion 164.00 (163.70 to 164.70): E

Ion 98.00 (97.70 to 98.70): BF1



#30
 1,2,4-Trichlorobenzene
 Concen: 47.856 ng
 RT: 8.05 min Scan# 1013
 Delta R.T. 0.00 min
 Lab File: BF117056.D
 Acq: 16 Sep 2019 12:11

Tgt Ion	Ratio	Lower	Upper
180	100		
182	96.3	75.4	113.0
145	33.6	26.7	40.1






Abundance

Ion 180.00 (179.70 to 180.70): E

Ion 182.00 (181.70 to 182.70): E

Ion 145.00 (144.70 to 145.70): E





#31

Naphthalene

Concen: 51.142 ng

RT: 8.13 min Scan# 1027

Delta R.T. 0.00 min

Lab File: BF117056.D

Acq: 16 Sep 2019 12:11

Instrument :

BNA_F

ClientSampleId :


Tot Ion:128 Resp: 568901

Ion Ratio Lower Upper

128 100

129 10.4 9.0 13.4




127 12.9 10.8 16.2



Abundance Ion 128.00 (127.70 to 128.70): E

Ion 129.00 (128.70 to 129.70): E

Ion 127.00 (126.70 to 127.70): E



#32

Benzoic acid

Concen: 21.641 ng

RT: 7.87 min Scan# 983

Delta R.T. -0.02 min

Lab File: BF117056.D

Acq: 16 Sep 2019 12:11


Tgt Ion:122 Resp: 39001

Ion Ratio Lower Upper

122 100

105 131.3 114.8 154.8



77 99.9 78.7 118.7




Abundance Ion 122.00 (121.70 to 122.70): E

Ion 105.00 (104.70 to 105.70): E

Ion 77.00 (76.70 to 77.70): BF1









#33
4-Chloroaniline
Concen: 18.958 ng
RT: 8.17 min Scan# 1035
Delta R.T. 0.00 min
Lab File: BF117056.D
Acq: 16 Sep 2019 12:11




Instrument :
BNA_F
ClientSampled :


Tot Ion:127	Resp:	93537
Ion Ratio	Lower	Upper
127	100	
129	32.3	26.2 39.2
65	28.1	23.5 35.3
92	18.8	15.9 23.9



#34
Hexachlorobutadiene
Concen: 47.510 ng
RT: 8.25 min Scan# 1047
Delta R.T. 0.00 min
Lab File: BF117056.D
Acq: 16 Sep 2019 12:11

Tgt Ion:225	Resp:	90355
Ion Ratio	Lower	Upper
225	100	
223	63.3	49.1 73.7
227	62.3	52.2 78.2









#35
Caprolactam
Concen: 20.241 ng
RT: 8.54 min Scan# 1097
Delta R.T. 0.00 min
Lab File: BF117056.D
Acq: 16 Sep 2019 12:11

Instrument :
BNA_F
ClientSampled :

Tot Ion:113 Resp: 21257
Ion Ratio Lower Upper
113 100
55 122.4 120.6 160.6
56 104.2 88.8 128.8




Abundance Ion 113.00 (112.70 to 113.70): E
Ion 55.00 (54.70 to 55.70): BF1
Ion 56.00 (55.70 to 56.70): BF1






#36
4-Chloro-3-methylphenol
Concen: 53.522 ng
RT: 8.66 min Scan# 1117
Delta R.T. 0.00 min
Lab File: BF117056.D
Acq: 16 Sep 2019 12:11

Tgt Ion:107 Resp: 193635
Ion Ratio Lower Upper
107 100
144 21.9 19.2 28.8
142 71.2 59.6 89.4



Abundance Ion 107.00 (106.70 to 107.70): E
Ion 144.00 (143.70 to 144.70): E
Ion 142.00 (141.70 to 142.70): E









#37
 2-Methylnaphthalene
 Concen: 51.869 ng
 RT: 8.82 min Scan# 1144
 Delta R.T. 0.00 min
 Lab File: BF117056.D
 Acq: 16 Sep 2019 12:11




Instrument :
 BNA_F
 ClientSampled :


Tot Ion	Ratio	Lower	Upper
142	100		
141	88.9	70.7	106.1
115	35.8	29.4	44.0



#38
 1-Methylnaphthalene
 Concen: 51.943 ng
 RT: 8.91 min Scan# 1160
 Delta R.T. -0.01 min
 Lab File: BF117056.D
 Acq: 16 Sep 2019 12:11

Tgt Ion	Ratio	Lower	Upper
142	100		
141	91.0	71.9	107.9
116	4.2	3.2	4.8






#39
Acenaphthene-d10
Concen: 20.000 ng
RT: 9.85 min Scan# 1320
Delta R.T. -0.01 min
Lab File: BF117056.D
Acq: 16 Sep 2019 12:11




Instrument :
BNA_F
ClientSampleId :

Tot Ion:164 Resp: 121668
Ion Ratio Lower Upper
164 100
162 102.4 83.4 125.2
160 45.0 36.4 54.6




Abundance

Ion 164.00 (163.70 to 164.70): E
Ion 162.00 (161.70 to 162.70): E
Ion 160.00 (159.70 to 160.70): E





#40
1,2,4,5-Tetrachlorobenzene
Concen: 48.606 ng
RT: 8.98 min Scan# 1172
Delta R.T. -0.01 min
Lab File: BF117056.D
Acq: 16 Sep 2019 12:11

Tgt Ion:216 Resp: 152710
Ion Ratio Lower Upper
216 100
214 80.3 63.6 95.4
179 23.8 18.3 27.5
108 25.3 17.8 26.6



Abundance

Ion 216.00 (215.70 to 216.70): E
Ion 214.00 (213.70 to 214.70): E
Ion 179.00 (178.70 to 179.70): E
Ion 108.00 (107.70 to 108.70): E





#41

Hexachlorocyclopentadiene

Concen: 100.161 ng

RT: 8.97 min Scan# 1170

Delta R.T. 0.00 min

Lab File: BF117056.D

Acq: 16 Sep 2019 12:11

Instrument :

BNA_F

ClientSampleId :


Tot Ion:237 Resp: 142143

Ion Ratio Lower Upper

237 100

235 61.5 43.2 83.2




272 12.1 0.0 31.9



Abundance Ion 236.80 (236.50 to 237.50): E

Ion 234.90 (234.60 to 235.60): E

Ion 271.75 (271.45 to 272.45): E



#42

2,4,6-Tribromophenol

Concen: 149.787 ng

RT: 10.65 min Scan# 1455

Delta R.T. 0.00 min

Lab File: BF117056.D

Acq: 16 Sep 2019 12:11


Tgt Ion:330 Resp: 152313

Ion Ratio Lower Upper

330 100

332 97.8 78.0 117.0



141 45.8 35.8 53.6




Abundance Ion 329.80 (329.50 to 330.50): E

Ion 331.80 (331.50 to 332.50): E

Ion 141.00 (140.70 to 141.70): E





#43

2,4,6-Trichlorophenol

Concen: 50.014 ng

RT: 9.09 min Scan# 1191

Delta R.T. 0.00 min

Lab File: BF117056.D

Acq: 16 Sep 2019 12:11

Instrument :

BNA_F

ClientSampleId :


Tot Ion:196 Resp: 106636

Ion Ratio Lower Upper

196 100

198 95.2 75.7 113.5

200 30.8 23.3 34.9






Abundance

Ion 196.00 (195.70 to 196.70): E

Ion 198.00 (197.70 to 198.70): E

Ion 200.00 (199.70 to 200.70): E



#44

2,4,5-Trichlorophenol

Concen: 51.576 ng

RT: 9.14 min Scan# 1199

Delta R.T. 0.01 min

Lab File: BF117056.D

Acq: 16 Sep 2019 12:11

Tgt Ion:196 Resp: 117489

Ion Ratio Lower Upper

196 100

198 96.6 74.6 112.0

97 56.1 49.5 74.3

132 32.1 27.4 41.2




Abundance

Ion 196.00 (195.70 to 196.70): E

Ion 198.00 (197.70 to 198.70): E

Ion 97.00 (96.70 to 97.70): BF1

Ion 132.00 (131.70 to 132.70): E






#45
2-Fluorobiphenyl
Concen: 96.058 ng
RT: 9.17 min Scan# 1205
Delta R.T. 0.00 min
Lab File: BF117056.D
Acq: 16 Sep 2019 12:11

Instrument :
BNA_F
ClientSampleId :

Tot Ion	Ratio	Lower	Upper
172	100		
171	35.7	28.2	42.2
170	23.7	18.9	28.3






Abundance

Ion 172.00 (171.70 to 172.70): E


Ion 171.00 (170.70 to 171.70): E

Ion 170.00 (169.70 to 170.70): E



#46
1,1'-Biphenyl
Concen: 48.906 ng
RT: 9.27 min Scan# 1222
Delta R.T. 0.00 min
Lab File: BF117056.D
Acq: 16 Sep 2019 12:11

Tgt Ion	Ratio	Lower	Upper
154	100		
153	40.1	20.0	60.0
76	12.1	0.0	32.7






Abundance

Ion 154.00 (153.70 to 154.70): E

Ion 153.00 (152.70 to 153.70): E

Ion 76.00 (75.70 to 76.70): BF1









#47
 2-Chloronaphthalene
 Concen: 48.640 ng
 RT: 9.30 min Scan# 1227
 Delta R.T. 0.00 min
 Lab File: BF117056.D
 Acq: 16 Sep 2019 12:11

Instrument :
 BNA_F
 ClientSampled :

Tot Ion	Ratio	Lower	Upper
162	100		
127	41.1	34.1	51.1
164	33.4	25.9	38.9




Abundance Ion 162.00 (161.70 to 162.70): E
 Ion 127.00 (126.70 to 127.70): E
 Ion 164.00 (163.70 to 164.70): E





#48
 2-Nitroaniline
 Concen: 50.413 ng
 RT: 9.40 min Scan# 1243
 Delta R.T. 0.00 min
 Lab File: BF117056.D
 Acq: 16 Sep 2019 12:11

Tgt Ion	Ratio	Lower	Upper
65	100		
92	73.4	57.5	86.3
138	113.8	90.6	136.0



Abundance Ion 65.00 (64.70 to 65.70): BF1
 Ion 92.00 (91.70 to 92.70): BF1
 Ion 138.00 (137.70 to 138.70): E





#49

Acenaphthylene

Concen: 51.392 ng

RT: 9.72 min Scan# 1297

Delta R.T. 0.00 min

Lab File: BF117056.D

Acq: 16 Sep 2019 12:11

Instrument :

BNA_F

ClientSampled :


Tot Ion:152 Resp: 579874

Ion Ratio Lower Upper

152 100

151 20.4 16.2 24.2




153 12.8 10.2 15.4



Abundance Ion 152.00 (151.70 to 152.70): E

Ion 151.00 (150.70 to 151.70): E

Ion 153.00 (152.70 to 153.70): E



#50

Dimethylphthalate

Concen: 50.452 ng

RT: 9.57 min Scan# 1273

Delta R.T. 0.00 min

Lab File: BF117056.D

Acq: 16 Sep 2019 12:11


Tgt Ion:163 Resp: 434627

Ion Ratio Lower Upper

163 100

194 3.7 2.9 4.3



164 9.9 8.2 12.4




Abundance Ion 163.00 (162.70 to 163.70): E

Ion 194.00 (193.70 to 194.70): E

Ion 164.00 (163.70 to 164.70): E









#51
 2,6-Dinitrotoluene
 Concen: 54.444 ng
 RT: 9.64 min Scan# 1284
 Delta R.T. 0.00 min
 Lab File: BF117056.D
 Acq: 16 Sep 2019 12:11

Instrument :
 BNA_F
 ClientSampleId :

Tot Ion:165 Resp: 95524
 Ion Ratio Lower Upper
 165 100
 63 46.9 41.8 62.8
 89 57.2 46.5 69.7




Abundance Ion 165.00 (164.70 to 165.70): E
 Ion 63.00 (62.70 to 63.70): BF1
 Ion 89.00 (88.70 to 89.70): BF1






#52
 Acenaphthene
 Concen: 49.669 ng
 RT: 9.89 min Scan# 1326
 Delta R.T. 0.00 min
 Lab File: BF117056.D
 Acq: 16 Sep 2019 12:11

Tgt Ion:154 Resp: 332216
 Ion Ratio Lower Upper
 154 100
 153 113.6 89.0 133.6
 152 53.0 41.9 62.9



Abundance Ion 154.00 (153.70 to 154.70): E
 Ion 153.00 (152.70 to 153.70): E
 Ion 152.00 (151.70 to 152.70): E









#53
 3-Nitroaniline
 Concen: 34.854 ng
 RT: 9.80 min Scan# 1312
 Delta R.T. -0.01 min
 Lab File: BF117056.D
 Acq: 16 Sep 2019 12:11

Instrument :
 BNA_F
 ClientSampleId :

Tot Ion:138 Resp: 77157
 Ion Ratio Lower Upper
 138 100
 108 11.7 9.5 14.3
 92 120.0 94.1 141.1




Abundance Ion 138.00 (137.70 to 138.70): E
 Ion 108.00 (107.70 to 108.70): E
 Ion 92.00 (91.70 to 92.70): BF1






#54
 2,4-Dinitrophenol
 Concen: 54.512 ng
 RT: 9.92 min Scan# 1331
 Delta R.T. 0.00 min
 Lab File: BF117056.D
 Acq: 16 Sep 2019 12:11

Tgt Ion:184 Resp: 38799
 Ion Ratio Lower Upper
 184 100
 63 85.0 63.8 95.6
 154 83.0 60.4 90.6



Abundance Ion 184.00 (183.70 to 184.70): E
 Ion 63.00 (62.70 to 63.70): BF1
 Ion 154.00 (153.70 to 154.70): E









#55
Dibenzenofuran
Concen: 51.128 ng
RT: 10.06 min Scan# 1355
Delta R.T. -0.01 min
Lab File: BF117056.D
Acq: 16 Sep 2019 12:11

Instrument :
BNA_F
ClientSampleId :

Tot Ion:168 Resp: 504555
Ion Ratio Lower Upper
168 100
139 43.8 32.8 49.2
169 13.5 10.5 15.7




Abundance Ion 168.00 (167.70 to 168.70): E
Ion 139.00 (138.70 to 139.70): E
Ion 169.00 (168.70 to 169.70): E





#56
4-Nitrophenol
Concen: 94.139 ng
RT: 9.97 min Scan# 1341
Delta R.T. 0.01 min
Lab File: BF117056.D
Acq: 16 Sep 2019 12:11

Tgt Ion:139 Resp: 135064
Ion Ratio Lower Upper
139 100
109 82.2 56.4 96.4
65 101.7 82.5 122.5



Abundance Ion 139.00 (138.70 to 139.70): E
Ion 109.00 (108.70 to 109.70): E
Ion 65.00 (64.70 to 65.70): BF1






#57
2,4-Dinitrotoluene
Concen: 56.817 ng
RT: 10.05 min Scan# 1353
Delta R.T. 0.00 min
Lab File: BF117056.D
Acq: 16 Sep 2019 12:11

Instrument :
BNA_F
ClientSampleId :

Tot Ion	Ratio	Lower	Upper
165	100		
63	43.8	35.0	52.4
89	76.1	66.1	99.1
182	2.5	2.0	3.0






Abundance

Ion 165.00 (164.70 to 165.70): E

Ion 63.00 (62.70 to 63.70): BF1


Ion 89.00 (88.70 to 89.70): BF1

Ion 182.00 (181.70 to 182.70): E



#58
Fluorene
Concen: 51.751 ng
RT: 10.40 min Scan# 1413
Delta R.T. -0.01 min
Lab File: BF117056.D
Acq: 16 Sep 2019 12:11

Tgt Ion	Ratio	Lower	Upper
166	100		
165	97.9	78.9	118.3
167	13.3	10.6	15.8





Abundance

Ion 166.00 (165.70 to 166.70): E

Ion 165.00 (164.70 to 165.70): E

Ion 167.00 (166.70 to 167.70): E









#59
2,3,4,6-Tetrachlorophenol
Concen: 53.902 ng
RT: 10.18 min Scan# 1376
Delta R.T. 0.00 min
Lab File: BF117056.D
Acq: 16 Sep 2019 12:11




Instrument :
BNA_F
ClientSampleId :


Tot Ion	Ratio	Lower	Upper
232	100		
131	55.6	45.3	67.9
130	2.6	1.8	2.8
166	34.7	29.4	44.2



#60
Diethylphthalate
Concen: 50.193 ng
RT: 10.27 min Scan# 1391
Delta R.T. 0.00 min
Lab File: BF117056.D
Acq: 16 Sep 2019 12:11

Tgt Ion	Ratio	Lower	Upper
149	100		
177	20.9	15.2	22.8
150	12.9	9.9	14.9









#61
4-Chlorophenyl-phenylether
Concen: 51.094 ng
RT: 10.39 min Scan# 1412
Delta R.T. 0.00 min
Lab File: BF117056.D
Acq: 16 Sep 2019 12:11

Instrument :
BNA_F
ClientSampleId :

Tot Ion:204 Resp: 175734
Ion Ratio Lower Upper
204 100
206 32.0 26.0 39.0
141 68.5 58.2 87.2




Abundance Ion 204.00 (203.70 to 204.70): E
Ion 206.00 (205.70 to 206.70): E
Ion 141.00 (140.70 to 141.70): E






#62
4-Nitroaniline
Concen: 51.017 ng
RT: 10.42 min Scan# 1417
Delta R.T. 0.00 min
Lab File: BF117056.D
Acq: 16 Sep 2019 12:11

Tgt Ion:138 Resp: 117448
Ion Ratio Lower Upper
138 100
92 51.8 32.9 72.9
108 80.1 58.3 98.3



Abundance Ion 138.00 (137.70 to 138.70): E
Ion 92.00 (91.70 to 92.70): BF1
Ion 108.00 (107.70 to 108.70): E





#63

Azobenzene

Concen: 52.164 ng

RT: 10.55 min Scan# 1439

Delta R.T. 0.00 min

Lab File: BF117056.D

Acq: 16 Sep 2019 12:11

Instrument :

BNA_F

ClientSampleId :

Tot Ion: 77 Resp: 451538


Ion Ratio Lower Upper

77 100

182 22.9 0.4 40.4

105 22.1 0.9 40.9

51 23.1 2.8 42.8






Abundance Ion 77.00 (76.70 to 77.70): BF1

Ion 182.00 (181.70 to 182.70): BF1

Ion 105.00 (104.70 to 105.70): BF1

Ion 51.00 (50.70 to 51.70): BF1



#64

Phenanthrene-d10

Concen: 20.000 ng

RT: 11.33 min Scan# 1572

Delta R.T. 0.00 min

Lab File: BF117056.D

Acq: 16 Sep 2019 12:11


Tgt Ion: 188 Resp: 204516

Ion Ratio Lower Upper

188 100

94 9.7 8.2 12.2



80 10.0 9.0 13.6



Abundance Ion 188.00 (187.70 to 188.70): BF1

Ion 94.00 (93.70 to 94.70): BF1

Ion 80.00 (79.70 to 80.70): BF1





#65

4,6-Dinitro-2-methylphenol

Concen: 42.779 ng

RT: 10.45 min Scan# 1422

Delta R.T. -0.00 min

Lab File: BF117056.D

Acq: 16 Sep 2019 12:11

Instrument :

BNA_F

ClientSampleId :


Tot Ion:198 Resp: 39047

Ion Ratio Lower Upper

198 100

51 51.7 18.7 58.7




105 57.4 30.2 70.2



Abundance Ion 198.00 (197.70 to 198.70): E

Ion 51.00 (50.70 to 51.70): BF1

Ion 105.00 (104.70 to 105.70): E



#66

n-Nitrosodiphenylamine

Concen: 51.233 ng

RT: 10.51 min Scan# 1432

Delta R.T. 0.00 min

Lab File: BF117056.D

Acq: 16 Sep 2019 12:11


Tgt Ion:169 Resp: 361869

Ion Ratio Lower Upper

169 100

168 65.8 52.2 78.2



167 36.2 28.8 43.2



Abundance Ion 169.00 (168.70 to 169.70): E

Ion 168.00 (167.70 to 168.70): E

Ion 167.00 (166.70 to 167.70): E









#67
4-Bromophenyl-phenylether
Concen: 49.991 ng
RT: 10.88 min Scan# 1495
Delta R.T. 0.00 min
Lab File: BF117056.D
Acq: 16 Sep 2019 12:11




Instrument :
BNA_F
ClientSampled :

Tot Ion:248 Resp: 104411
Ion Ratio Lower Upper
248 100
250 92.9 76.2 114.4
141 82.3 75.0 112.4



#68
Hexachlorobenzene
Concen: 48.518 ng
RT: 10.95 min Scan# 1507
Delta R.T. -0.01 min
Lab File: BF117056.D
Acq: 16 Sep 2019 12:11

Tgt Ion:284 Resp: 110836
Ion Ratio Lower Upper
284 100
142 43.9 31.4 47.2
249 32.7 26.5 39.7









#69
 Atrazine
 Concen: 65.625 ng
 RT: 11.03 min Scan# 1521
 Delta R.T. 0.00 min
 Lab File: BF117056.D
 Acq: 16 Sep 2019 12:11




Instrument :
 BNA_F
 ClientSampleId :


Tot Ion	Ratio	Lower	Upper
200	100		
173	27.7	6.9	46.9
215	44.8	23.8	63.8



#70
 Pentachlorophenol
 Concen: 105.436 ng
 RT: 11.15 min Scan# 1540
 Delta R.T. 0.00 min
 Lab File: BF117056.D
 Acq: 16 Sep 2019 12:11

Tgt Ion	Ratio	Lower	Upper
266	100		
268	64.6	50.5	75.7
264	61.5	48.8	73.2





#71

Phenanthrene

Concen: 50.919 ng

RT: 11.36 min Scan# 1576

Delta R.T. -0.01 min

Lab File: BF117056.D

Acq: 16 Sep 2019 12:11

Instrument :

BNA_F

ClientSampleId :


Tot Ion:178 Resp: 591492

Ion Ratio Lower Upper

178 100

176 19.1 15.4 23.2




179 15.5 12.4 18.6



Abundance Ion 178.00 (177.70 to 178.70): E

Ion 176.00 (175.70 to 176.70): E

Ion 179.00 (178.70 to 179.70): E



#72

Anthracene

Concen: 51.931 ng

RT: 11.41 min Scan# 1585

Delta R.T. -0.01 min

Lab File: BF117056.D

Acq: 16 Sep 2019 12:11


Tgt Ion:178 Resp: 615871

Ion Ratio Lower Upper

178 100

176 19.1 14.6 21.8



179 15.4 11.6 17.4



Abundance Ion 178.00 (177.70 to 178.70): E

Ion 176.00 (175.70 to 176.70): E

Ion 179.00 (178.70 to 179.70): E









#73
 Carbazole
 Concen: 52.226 ng
 RT: 11.56 min Scan# 1611
 Delta R.T. -0.01 min
 Lab File: BF117056.D
 Acq: 16 Sep 2019 12:11




Instrument :
 BNA_F
 ClientSampleId :

Tot Ion	Ratio	Lower	Upper
167	100		
166	21.6	17.2	25.8
139	13.9	10.6	16.0



#74
 Di-n-butylphthalate
 Concen: 51.843 ng
 RT: 11.89 min Scan# 1667
 Delta R.T. 0.00 min
 Lab File: BF117056.D
 Acq: 16 Sep 2019 12:11

Tgt Ion	Ratio	Lower	Upper
149	100		
150	8.4	7.7	11.5
104	4.9	4.2	6.4









#75
 Fluoranthene
 Concen: 52.743 ng
 RT: 12.54 min Scan# 1778
 Delta R.T. 0.00 min
 Lab File: BF117056.D
 Acq: 16 Sep 2019 12:11

Instrument :
 BNA_F
 ClientSampleId :

Tot Ion	Ratio	Lower	Upper
202	100		
101	10.4	0.0	32.3
203	16.8	0.0	37.1




Abundance Ion 202.00 (201.70 to 202.70): E
 Ion 101.00 (100.70 to 101.70): E
 Ion 203.00 (202.70 to 203.70): E






#76
 Chrysene-d12
 Concen: 20.000 ng
 RT: 13.97 min Scan# 2020
 Delta R.T. -0.01 min
 Lab File: BF117056.D
 Acq: 16 Sep 2019 12:11

Tgt Ion	Ratio	Lower	Upper
240	100		
120	11.1	8.3	12.5
236	26.2	20.3	30.5



Abundance Ion 240.00 (239.70 to 240.70): E
 Ion 120.00 (119.70 to 120.70): E
 Ion 236.00 (235.70 to 236.70): E





#77

Benzidine

Concen: 44.623 ng

RT: 12.66 min Scan# 1798

Delta R.T. 0.00 min

Lab File: BF117056.D

Acq: 16 Sep 2019 12:11

Instrument :

BNA_F

ClientSampleId :


Tot Ion:184 Resp: 231719

Ion Ratio Lower Upper




184 100

185 14.1 11.6 17.4

183 11.4 9.0 13.4



Abundance Ion 184.00 (183.70 to 184.70): E
Ion 185.00 (184.70 to 185.70): E
Ion 183.00 (182.70 to 183.70): E



#78

Pyrene

Concen: 46.487 ng

RT: 12.77 min Scan# 1817

Delta R.T. 0.00 min

Lab File: BF117056.D

Acq: 16 Sep 2019 12:11


Tgt Ion:202 Resp: 628800

Ion Ratio Lower Upper



202 100


200 21.3 17.2 25.8

203 17.5 14.0 21.0



Abundance Ion 202.00 (201.70 to 202.70): E
Ion 200.00 (199.70 to 200.70): E
Ion 203.00 (202.70 to 203.70): E









#79
 Terphenyl-d14
 Concen: 92.245 ng
 RT: 12.92 min Scan# 1841
 Delta R.T. 0.00 min
 Lab File: BF117056.D
 Acq: 16 Sep 2019 12:11




Instrument :
 BNA_F
 ClientSampled :


Tot Ion:244 Resp: 795540
 Ion Ratio Lower Upper
 244 100
 212 7.1 5.8 8.8
 122 12.1 11.3 16.9



#80
 Butylbenzylphthalate
 Concen: 47.749 ng
 RT: 13.38 min Scan# 1920
 Delta R.T. -0.01 min
 Lab File: BF117056.D
 Acq: 16 Sep 2019 12:11

Tgt Ion:149 Resp: 305184
 Ion Ratio Lower Upper
 149 100
 91 75.7 58.8 88.2
 206 16.4 14.1 21.1





#81

Benzo(a)anthracene

Concen: 48.852 ng

RT: 13.96 min Scan# 2018

Delta R.T. -0.01 min

Lab File: BF117056.D

Acq: 16 Sep 2019 12:11

Instrument :

BNA_F

ClientSampleId :


Tot Ion:228 Resp: 526151

Ion Ratio Lower Upper

228 100

226 27.4 22.3 33.5




229 20.1 15.9 23.9



Abundance Ion 228.00 (227.70 to 228.70): E

Ion 226.00 (225.70 to 226.70): E

Ion 229.00 (228.70 to 229.70): E



#82

3,3'-Dichlorobenzidine

Concen: 30.139 ng

RT: 13.92 min Scan# 2011

Delta R.T. -0.01 min

Lab File: BF117056.D

Acq: 16 Sep 2019 12:11


Tgt Ion:252 Resp: 104604

Ion Ratio Lower Upper

252 100

254 62.6 52.6 79.0



126 13.6 10.4 15.6




Abundance Ion 252.00 (251.70 to 252.70): E

Ion 254.00 (253.70 to 254.70): E

Ion 126.00 (125.70 to 126.70): E









#83
Chrysene
Concen: 50.336 ng
RT: 14.00 min Scan# 2025
Delta R.T. 0.00 min
Lab File: BF117056.D
Acq: 16 Sep 2019 12:11

Instrument :
BNA_F
ClientSampleId :

Tot Ion	Ratio	Lower	Upper
228	100		
226	29.8	24.8	37.2
229	19.2	15.8	23.8




Abundance Ion 228.00 (227.70 to 228.70): E
Ion 226.00 (225.70 to 226.70): E
Ion 229.00 (228.70 to 229.70): E





#84
Bis(2-ethylhexyl)phthalate
Concen: 50.530 ng
RT: 13.94 min Scan# 2015
Delta R.T. -0.01 min
Lab File: BF117056.D
Acq: 16 Sep 2019 12:11

Tgt Ion	Ratio	Lower	Upper
149	100		
167	26.7	22.2	33.2
279	3.2	2.8	4.2



Abundance Ion 149.00 (148.70 to 149.70): E
Ion 167.00 (166.70 to 167.70): E
Ion 279.00 (278.70 to 279.70): E





#85

Di-n-octyl phthalate

Concen: 55.016 ng

RT: 14.56 min Scan# 2120

Delta R.T. 0.00 min

Lab File: BF117056.D

Acq: 16 Sep 2019 12:11

Instrument :

BNA_F

ClientSampleId :


Tot Ion:149 Resp: 713591

Ion Ratio Lower Upper

149 100

167 1.5 1.2 1.8




43 7.3 6.2 9.4



Abundance Ion 149.00 (148.70 to 149.70): E

Ion 167.00 (166.70 to 167.70): E

Ion 43.00 (42.70 to 43.70): BF1



#86

Indeno(1,2,3-cd)pyrene

Concen: 54.260 ng

RT: 16.86 min Scan# 2512

Delta R.T. 0.00 min

Lab File: BF117056.D

Acq: 16 Sep 2019 12:11


Tgt Ion:276 Resp: 415837

Ion Ratio Lower Upper

276 100

138 24.7 20.7 31.1



277 25.1 19.8 29.8



Abundance Ion 276.00 (275.70 to 276.70): E

Ion 138.00 (137.70 to 138.70): E

Ion 277.00 (276.70 to 277.70): E









#87
Pervlene-d12
Concen: 20.000 ng
RT: 15.42 min Scan# 2266
Delta R.T. -0.01 min
Lab File: BF117056.D
Acq: 16 Sep 2019 12:11

Instrument :
BNA_F
ClientSampleId :

Tot Ion:264 Resp: 152863




Ion	Ratio	Lower	Upper
264	100		
260	24.1	17.6	26.4
265	19.9	16.6	24.8



#88
Benzo(b)fluoranthene
Concen: 50.825 ng
RT: 15.00 min Scan# 2196
Delta R.T. 0.00 min
Lab File: BF117056.D
Acq: 16 Sep 2019 12:11

Tgt Ion:252 Resp: 470612

Ion	Ratio	Lower	Upper
252	100		
253	21.5	17.6	26.4
125	9.1	8.6	12.8









#89
Benzo(k)fluoranthene
Concen: 49.594 ng
RT: 15.03 min Scan# 2201
Delta R.T. 0.00 min
Lab File: BF117056.D
Acq: 16 Sep 2019 12:11




Instrument :
BNA_F
ClientSampleId :

Tot Ion:252 Resp: 434999
Ion Ratio Lower Upper
252 100
253 21.3 17.4 26.2
125 9.6 8.8 13.2



#90
Benzo(a)pyrene
Concen: 52.198 ng
RT: 15.36 min Scan# 2256
Delta R.T. -0.01 min
Lab File: BF117056.D
Acq: 16 Sep 2019 12:11

Tgt Ion:252 Resp: 424129
Ion Ratio Lower Upper
252 100
253 21.8 16.6 25.0
125 12.2 8.9 13.3









#91
Dibenzo(a,h)anthracene
Concen: 53.706 ng
RT: 16.87 min Scan# 2513
Delta R.T. 0.00 min
Lab File: BF117056.D
Acq: 16 Sep 2019 12:11

Instrument :
BNA_F
ClientSampled :

Tot Ion	Ratio	Lower	Upper
278	100		
139	16.3	13.5	20.3
279	23.4	18.8	28.2




Abundance Ion 278.00 (277.70 to 278.70): E
Ion 139.00 (138.70 to 139.70): E
Ion 279.00 (278.70 to 279.70): E



#92
Benzo(g,h,i)perylene
Concen: 53.586 ng
RT: 17.29 min Scan# 2584
Delta R.T. 0.00 min
Lab File: BF117056.D
Acq: 16 Sep 2019 12:11

Tgt Ion	Ratio	Lower	Upper
276	100		
277	22.5	18.6	27.8
138	20.1	18.0	27.0



Abundance Ion 276.00 (275.70 to 276.70): E
Ion 277.00 (276.70 to 277.70): E
Ion 138.00 (137.70 to 138.70): E

