

Data Path : Z:\SVOASRV\HPCHEM1\BNA F\DATA\BF091619\
 Data File : BF117069.D
 Acq On : 16 Sep 2019 17:58
 Operator : HP/JU
 Sample : K4850-03DL 20X
 Misc :
 ALS Vial : 22 Sample Multiplier: 1

Instrument :
 BNA_F
 ClientSampleId :
 AOC-7-(7)DL

Manual Integrations
 APPROVED

mohammad
 9/17/2019 1:25:51 PM

Quant Time: Sep 17 06:53:52 2019
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA F\METHODS\8270-BF091319.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Fri Sep 13 16:23:17 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	6.82	152	53929	20.00	ng	0.00
21) Naphthalene-d8	8.10	136	232457	20.00	ng	0.00
39) Acenaphthene-d10	9.85	164	119899	20.00	ng	0.00
64) Phenanthrene-d10	11.33	188	169491	20.00	ng	0.00
76) Chrysene-d12	13.97	240	119741	20.00	ng	0.00
87) Perylene-d12	15.42	264	120689	20.00	ng	0.00

System Monitoring Compounds

5) 2-Fluorophenol	5.45	112	10592	3.07	ng	0.01
7) Phenol-d6	6.46	99	16608	3.72	ng	0.00
23) Nitrobenzene-d5	7.39	82	10177	2.30	ng	0.00
42) 2,4,6-Tribromophenol	10.64	330	2949	2.94	ng	0.00
45) 2-Fluorobiphenyl	9.17	172	17792	2.32	ng	0.00
79) Terphenyl-d14	12.91	244	14374	2.24	ng	0.00

Target Compounds

						Qvalue
31) Naphthalene	8.13	128	1149041	102.779	ng	99
37) 2-Methylnaphthalene	8.82	142	485931	64.547	ng	99
38) 1-Methylnaphthalene	8.92	142	292545	41.490	ng	99
46) 1,1'-Biphenyl	9.27	154	63179	6.718	ng	97
49) Acenaphthylene	9.71	152	60761	5.464	ng	# 95
52) Acenaphthene	9.89	154	20674	3.137	ng	96
55) Dibenzofuran	10.06	168	24703	2.540	ng	# 94
58) Fluorene	10.40	166	159482	20.358	ng	100
71) Phenanthrene	11.36	178	578029	60.042	ng	100
72) Anthracene	11.41	178	76818	7.816	ng	99
75) Fluoranthene	12.55	202	188832	19.090	ng	97
78) Pyrene	12.77	202	308438	30.699	ng	99
81) Benzo(a)anthracene	13.96	228	109012	13.626	ng	96
83) Chrysene	13.99	228	99820	12.793	ng	98
86) Indeno(1,2,3-cd)pyrene	16.84	276	20652	3.628	ng	97
88) Benzo(b)fluoranthene	15.00	252	69514m	9.509	ng	
89) Benzo(k)fluoranthene	15.02	252	21470m	3.100	ng	
90) Benzo(a)pyrene	15.36	252	57413	8.950	ng	98
92) Benzo(a,h,i)perylene	17.28	276	21196	4.162	ng	# 93

(#) = qualifier out of range (m) = manual integration (+) = signals summed


Data Path : Z:\SVOASRV\HPCHEM1\BNA F\DATA\BF091619\
Data File : BF117069.D
Acq On : 16 Sep 2019 17:58
Operator : HP/JU
Sample : K4850-03DL 20X
Misc :
ALS Vial : 22 Sample Multiplier: 1


Instrument :
BNA_F
ClientSampleId :
AOC-7-(7)DL

Manual Integrations
APPROVED

mohammad
9/17/2019 1:25:51 PM

Quant Time: Sep 17 06:53:52 2019
Quant Method : Z:\SVOASRV\HPCHEM1\BNA F\METHODS\8270-BF091319.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Fri Sep 13 16:23:17 2019
Response via : Initial Calibration





#1


1,4-Dichlorobenzene-d4
Concen: 20.000 ng
RT: 6.82 min Scan# 805
Delta R.T. 0.00 min
Lab File: BF117069.D
Acq: 16 Sep 2019 17:58

Instrument :
BNA_F
ClientSampleId :
AOC-7(7)DL

Tot Ion	Ratio	Lower	Upper
152	100		
150	158.4	127.5	191.3
115	66.9	55.0	82.4

Manual Integrations
APPROVED

mohammad
9/17/2019 1:25:51 PM






Abundance

Ion 152.00 (151.70 to 152.70): E

Ion 150.00 (149.70 to 150.70): E


Ion 115.00 (114.70 to 115.70): E



#5

2-Fluorophenol
Concen: 3.066 ng
RT: 5.45 min Scan# 572
Delta R.T. 0.01 min
Lab File: BF117069.D
Acq: 16 Sep 2019 17:58

Tgt Ion	Ratio	Lower	Upper
112	100		
64	52.5	45.7	68.5
63	32.9	24.9	37.3






Abundance

Ion 112.00 (111.70 to 112.70): E

Ion 64.00 (63.70 to 64.70): BF1

Ion 63.00 (62.70 to 63.70): BF1





#7

Phenol-d6

Concen: 3.720 ng

RT: 6.46 min Scan# 744

Delta R.T. 0.01 min

Lab File: BF117069.D

Acq: 16 Sep 2019 17:58

Instrument :

BNA_F


ClientSampleId :

AOC-7(7)DL




Tot Ion:	99	Resp:	16608
Ion Ratio	Lower	Upper	
99	100		
42	15.8	12.2	18.2
71	37.4	29.8	44.8

Manual Integrations
APPROVED

mohammad
9/17/2019 1:25:51 PM



Abundance Ion 99.00 (98.70 to 99.70): BF1
Ion 42.00 (41.70 to 42.70): BF1
Ion 71.00 (70.70 to 71.70): BF1



#21

Naphthalene-d8

Concen: 20.000 ng

RT: 8.10 min Scan# 1023




Delta R.T. 0.00 min


Lab File: BF117069.D

Acq: 16 Sep 2019 17:58

Tgt Ion:	136	Resp:	232457
Ion Ratio	Lower	Upper	
136	100		
137	10.7	9.0	13.4
54	5.4	4.5	6.7
68	4.5	4.0	6.0

Abundance Ion 136.00 (135.70 to 136.70): E
Ion 137.00 (136.70 to 137.70): E
Ion 54.00 (53.70 to 54.70): BF1
Ion 68.00 (67.70 to 68.70): BF1






#23
Nitrobenzene-d5
Concen: 2.300 ng
RT: 7.39 min Scan# 901
Delta R.T. 0.00 min
Lab File: BF117069.D
Acq: 16 Sep 2019 17:58

Instrument :
BNA_F
ClientSampleId :
AOC-7(7)DL




Tot Ion	Ratio	Lower	Upper
82	100		
128	38.6	32.3	48.5
54	36.6	33.3	49.9

Manual Integrations
APPROVED

mohammad
9/17/2019 1:25:51 PM




Abundance Ion 82.00 (81.70 to 82.70): BF1
Ion 128.00 (127.70 to 128.70): E
Ion 54.00 (53.70 to 54.70): BF1






#31
Naphthalene
Concen: 102.779 ng
RT: 8.13 min Scan# 1027
Delta R.T. 0.00 min
Lab File: BF117069.D
Acq: 16 Sep 2019 17:58

Tgt Ion	Ratio	Lower	Upper
128	100		
129	11.6	9.0	13.4
127	13.9	10.8	16.2



Abundance Ion 128.00 (127.70 to 128.70): E
Ion 129.00 (128.70 to 129.70): E
Ion 127.00 (126.70 to 127.70): E






#37
2-Methylnaphthalene
Concen: 64.547 ng
RT: 8.82 min Scan# 1144
Delta R.T. 0.00 min
Lab File: BF117069.D
Acq: 16 Sep 2019 17:58

Instrument :
BNA_F
ClientSampleId :
AOC-7(7)DL

Tot Ion	Ratio	Resp	Lower	Upper
142	100	485931		
141	87.9	70.7	106.1	
115	38.7	29.4	44.0	

Manual Integrations
APPROVED

mohammad
9/17/2019 1:25:51 PM






Abundance

Ion 142.00 (141.70 to 142.70): E

Ion 141.00 (140.70 to 141.70): E

Ion 115.00 (114.70 to 115.70): E



#38
1-Methylnaphthalene
Concen: 41.490 ng
RT: 8.92 min Scan# 1161
Delta R.T. 0.00 min
Lab File: BF117069.D
Acq: 16 Sep 2019 17:58




Tgt Ion	Ratio	Resp	Lower	Upper
142	100	292545		
141	90.6	71.9	107.9	
116	4.4	3.2	4.8	


Abundance

Ion 142.00 (141.70 to 142.70): E

Ion 141.00 (140.70 to 141.70): E

Ion 116.00 (115.70 to 116.70): E






#39
Acenaphthene-d10
Concen: 20.000 ng
RT: 9.85 min Scan# 1320
Delta R.T. -0.01 min
Lab File: BF117069.D
Acq: 16 Sep 2019 17:58

Instrument :
BNA_F
ClientSampleId :
AOC-7(7)DL




Tot Ion	Ratio	Resp	Lower	Upper
164	100			
162	102.3	83.4	125.2	
160	44.5	36.4	54.6	

Manual Integrations
APPROVED

mohammad
9/17/2019 1:25:51 PM




Abundance Ion 164.00 (163.70 to 164.70): E
Ion 162.00 (161.70 to 162.70): E
Ion 160.00 (159.70 to 160.70): E






#42
2,4,6-Tribromophenol
Concen: 2.943 ng
RT: 10.64 min Scan# 1454
Delta R.T. -0.01 min
Lab File: BF117069.D
Acq: 16 Sep 2019 17:58

Tgt Ion	Ratio	Resp	Lower	Upper
330	100			
332	92.4	78.0	117.0	
141	84.8	35.8	53.6	



Abundance Ion 329.80 (329.50 to 330.50): E
Ion 331.80 (331.50 to 332.50): E
Ion 141.00 (140.70 to 141.70): E









#45
2-Fluorobiphenyl
Concen: 2.320 ng
RT: 9.17 min Scan# 1204
Delta R.T. -0.01 min
Lab File: BF117069.D
Acq: 16 Sep 2019 17:58

Instrument :
BNA_F
ClientSampled :
AOC-7(7)DL

Tot Ion	Ratio	Lower	Upper
172	100		
171	29.8	28.2	42.2
170	21.1	18.9	28.3




Manual Integrations
APPROVED


mohammad
9/17/2019 1:25:51 PM



#46
1,1'-Biphenyl
Concen: 6.718 ng
RT: 9.27 min Scan# 1222
Delta R.T. 0.00 min
Lab File: BF117069.D
Acq: 16 Sep 2019 17:58

Tgt Ion	Ratio	Lower	Upper
154	100		
153	38.7	20.0	60.0
76	10.5	0.0	32.7






#49
Acenaphthylene
Concen: 5.464 ng
RT: 9.71 min Scan# 1296
Delta R.T. -0.01 min
Lab File: BF117069.D
Acq: 16 Sep 2019 17:58

Instrument :
BNA_F
ClientSampleId :
AOC-7(7)DL

Tot Ion	Ratio	Resp	Lower	Upper
152	100	60761		
151	19.7		16.2	24.2
153	16.9		10.2	15.4

Manual Integrations
APPROVED

mohammad
9/17/2019 1:25:51 PM






Abundance

Ion 152.00 (151.70 to 152.70): E

Ion 151.00 (150.70 to 151.70): E

Ion 153.00 (152.70 to 153.70): E



#52
Acenaphthene
Concen: 3.137 ng
RT: 9.89 min Scan# 1326
Delta R.T. 0.00 min
Lab File: BF117069.D
Acq: 16 Sep 2019 17:58




Tgt Ion	Ratio	Resp	Lower	Upper
154	100	20674		
153	111.2		89.0	133.6
152	60.2		41.9	62.9


Abundance

Ion 154.00 (153.70 to 154.70): E

Ion 153.00 (152.70 to 153.70): E

Ion 152.00 (151.70 to 152.70): E






#55
Dibenzenofuran
Concen: 2.540 ng
RT: 10.06 min Scan# 1355
Delta R.T. -0.01 min
Lab File: BF117069.D
Acq: 16 Sep 2019 17:58

Instrument :
BNA_F
ClientSampled :
AOC-7(7)DL

Tot Ion	Ratio	Lower	Upper
168	100		
139	42.3	32.8	49.2
169	21.0	10.5	15.7

Manual Integrations
APPROVED

mohammad
9/17/2019 1:25:51 PM






Abundance

Ion 168.00 (167.70 to 168.70): E

Ion 139.00 (138.70 to 139.70): E

Ion 169.00 (168.70 to 169.70): E



#58
Fluorene
Concen: 20.358 ng
RT: 10.40 min Scan# 1413
Delta R.T. -0.01 min
Lab File: BF117069.D
Acq: 16 Sep 2019 17:58




Tgt Ion	Ratio	Lower	Upper
166	100		
165	98.9	78.9	118.3
167	13.2	10.6	15.8


Abundance

Ion 166.00 (165.70 to 166.70): E

Ion 165.00 (164.70 to 165.70): E

Ion 167.00 (166.70 to 167.70): E









#64
 Phenanthrene-d10
 Concen: 20.000 ng
 RT: 11.33 min Scan# 1572
 Delta R.T. 0.00 min
 Lab File: BF117069.D
 Acq: 16 Sep 2019 17:58

Instrument :
 BNA_F
 ClientSampled :
 AOC-7(7)DL

Tot Ion:188	Resp: 169491
Ion Ratio Lower Upper	
188 100	
94 8.8	8.2 12.2
80 9.8	9.0 13.6




Manual Integrations
 APPROVED


mohammad
 9/17/2019 1:25:51 PM



#71
 Phenanthrene
 Concen: 60.042 ng
 RT: 11.36 min Scan# 1577
 Delta R.T. 0.00 min
 Lab File: BF117069.D
 Acq: 16 Sep 2019 17:58

Tgt Ion:178	Resp: 578029
Ion Ratio Lower Upper	
178 100	
176 19.2	15.4 23.2
179 15.5	12.4 18.6





#72

Anthracene

Concen: 7.816 ng

RT: 11.41 min Scan# 1585

Delta R.T. -0.01 min

Lab File: BF117069.D

Acq: 16 Sep 2019 17:58

Instrument :

BNA_F





ClientSampleId :

AOC-7(7)DL

Tot Ion:	178	Resp:	76818
Ion Ratio	Lower	Upper	
178	100		
176	18.0	14.6	21.8
179	15.1	11.6	17.4

Manual Integrations
APPROVED

mohammad
9/17/2019 1:25:51 PM



#75

Fluoranthene

Concen: 19.090 ng




RT: 12.55 min Scan# 1778


Delta R.T. 0.00 min

Lab File: BF117069.D

Acq: 16 Sep 2019 17:58

Tgt Ion:	202	Resp:	188832
Ion Ratio	Lower	Upper	
202	100		
101	10.5	0.0	32.3
203	17.7	0.0	37.1





#76

Chrysene-d12

Concen: 20.000 ng

RT: 13.97 min Scan# 2020

Delta R.T. -0.01 min

Lab File: BF117069.D

Acq: 16 Sep 2019 17:58

Instrument :

BNA_F


ClientSampleId :

AOC-7(7)DL




Tot Ion:	240	Resp:	119741
Ion Ratio	Lower	Upper	
240	100		
120	9.9	8.3	12.5
236	26.2	20.3	30.5

Manual Integrations
APPROVED

mohammad
9/17/2019 1:25:51 PM



Abundance Ion 240.00 (239.70 to 240.70): E
Ion 120.00 (119.70 to 120.70): E
Ion 236.00 (235.70 to 236.70): E



#78

Pyrene

Concen: 30.699 ng


RT: 12.77 min Scan# 1817

Delta R.T. 0.00 min



Lab File: BF117069.D


Acq: 16 Sep 2019 17:58

Tgt Ion:	202	Resp:	308438
Ion Ratio	Lower	Upper	
202	100		
200	21.8	17.2	25.8
203	16.6	14.0	21.0



Abundance Ion 202.00 (201.70 to 202.70): E
Ion 200.00 (199.70 to 200.70): E
Ion 203.00 (202.70 to 203.70): E









#79
 Terphenyl-d14
 Concen: 2.244 ng
 RT: 12.91 min Scan# 1840
 Delta R.T. -0.01 min
 Lab File: BF117069.D
 Acq: 16 Sep 2019 17:58

Instrument :
 BNA_F
 ClientSampleId :
 AOC-7(7)DL

Tot Ion	Ratio	Lower	Upper
244	100		
212	7.8	5.8	8.8
122	10.0	11.3	16.9




Manual Integrations
 APPROVED


mohammad
 9/17/2019 1:25:51 PM



#81
 Benzo(a)anthracene
 Concen: 13.626 ng
 RT: 13.96 min Scan# 2018
 Delta R.T. -0.01 min
 Lab File: BF117069.D
 Acq: 16 Sep 2019 17:58

Tgt Ion	Ratio	Lower	Upper
228	100		
226	31.0	22.3	33.5
229	20.0	15.9	23.9









#83
Chrysene
Concen: 12.793 ng
RT: 13.99 min Scan# 2024
Delta R.T. -0.01 min
Lab File: BF117069.D
Acq: 16 Sep 2019 17:58

Instrument :
BNA_F
ClientSampled :
AOC-7(7)DL

Tot Ion:	228	Resp:	99820
Ion Ratio		Lower	Upper
228	100		
226	30.7	24.8	37.2
229	21.4	15.8	23.8




Manual Integrations
APPROVED


mohammad
9/17/2019 1:25:51 PM



#86
Indeno(1,2,3-cd)pyrene
Concen: 3.628 ng
RT: 16.84 min Scan# 2508
Delta R.T. -0.02 min
Lab File: BF117069.D
Acq: 16 Sep 2019 17:58

Tgt Ion:	276	Resp:	20652
Ion Ratio		Lower	Upper
276	100		
138	27.6	20.7	31.1
277	26.2	19.8	29.8





#87

Pervlene-d12

Concen: 20.000 ng

RT: 15.42 min Scan# 2266

Delta R.T. -0.01 min

Lab File: BF117069.D

Acq: 16 Sep 2019 17:58

Instrument :

BNA_F





ClientSampleId :

AOC-7(7)DL

Tot Ion:	264	Resp:	120689
Ion Ratio	Lower	Upper	
264	100		
260	23.9	17.6	26.4
265	22.0	16.6	24.8

Manual Integrations
APPROVED

mohammad
9/17/2019 1:25:51 PM



#88

Benzo(b)fluoranthene

Concen: 9.509 ng m




RT: 15.00 min Scan# 2195


Delta R.T. -0.01 min

Lab File: BF117069.D

Acq: 16 Sep 2019 17:58

Tgt Ion:	252	Resp:	69514
Ion Ratio	Lower	Upper	
252	100		
253	22.1	17.6	26.4
125	12.8	8.6	12.8






#89
Benzo(k)fluoranthene
Concen: 3.100 ng m
RT: 15.02 min Scan# 2199
Delta R.T. -0.01 min
Lab File: BF117069.D
Acq: 16 Sep 2019 17:58

Instrument :
BNA_F
ClientSampled :
AOC-7(7)DL

Tot Ion	Ratio	Lower	Upper
252	100		
253	23.1	17.4	26.2
125	10.6	8.8	13.2

Manual Integrations
APPROVED

mohammad
9/17/2019 1:25:51 PM






Abundance

Ion 252.00 (251.70 to 252.70): E

Ion 253.00 (252.70 to 253.70): E

Ion 125.00 (124.70 to 125.70): E



#90
Benzo(a)pyrene
Concen: 8.950 ng
RT: 15.36 min Scan# 2256
Delta R.T. -0.01 min
Lab File: BF117069.D
Acq: 16 Sep 2019 17:58




Tgt Ion	Ratio	Lower	Upper
252	100		
253	21.1	16.6	25.0
125	9.8	8.9	13.3


Abundance

Ion 252.00 (251.70 to 252.70): E

Ion 253.00 (252.70 to 253.70): E

Ion 125.00 (124.70 to 125.70): E





#92
 Benzo(a,h,i)perylene
 Concen: 4.162 ng
 RT: 17.28 min Scan# 2583
 Delta R.T. -0.01 min
 Lab File: BF117069.D
 Acq: 16 Sep 2019 17:58

Instrument :
 BNA_F
 ClientSampleId :
 AOC-7-(7)DL

Tot Ion:	276	Resp:	21196
Ion Ratio	Lower	Upper	
276	100		
277	28.1	18.6	27.8#
138	20.7	18.0	27.0

Manual Integrations
 APPROVED

mohammad
 9/17/2019 1:25:51 PM

