

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF091622\
 Data File : BF130306.D
 Acq On : 16 Sep 2022 23:34
 Operator : CG\JU
 Sample : N4648-01
 Misc :
 ALS Vial : 16 Sample Multiplier: 1

Instrument :
 BNA_F
 ClientSampleId :
 N48854-1040

Quant Time: Sep 17 03:32:11 2022
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF091422.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Sat Sep 17 03:22:32 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)


Internal Standards						
1) 1,4-Dichlorobenzene-d4	6.669	152	84499	20.000	ng	0.00
21) Naphthalene-d8	7.981	136	274072	20.000	ng	0.03
39) Acenaphthene-d10	9.710	164	130007	20.000	ng	0.00
64) Phenanthrene-d10	11.180	188	179118	20.000	ng	0.00
76) Chrysene-d12	13.810	240	175610	20.000	ng	0.00
86) Perylene-d12	15.198	264	203293	20.000	ng	0.00
System Monitoring Compounds						
5) 2-Fluorophenol	5.263	112	254462	52.232	ng	0.00
7) Phenol-d6	6.322	99	195738	32.111	ng	0.02
23) Nitrobenzene-d5	7.263	82	478294	89.157	ng	0.02
42) 2,4,6-Tribromophenol	10.492	330	122868	98.633	ng	0.00
45) 2-Fluorobiphenyl	9.045	172	740661	82.960	ng	0.01
79) Terphenyl-d14	12.768	244	478929	45.176	ng	0.00
Target Compounds						Qvalue
6) Aniline	6.339	93	99327	13.225	ng	# 1
9) Benzaldehyde	6.263	77	40554	9.932	ng	# 11
10) Phenol	6.345	94	3029516	424.090	ng	94
11) bis(2-Chloroethyl)ether	6.398	93	50579	9.816	ng	# 23
15) Benzyl Alcohol	6.981	79	2809901	581.394	ng	# 1
17) 2-Methylphenol	6.981	107	5889048	1305.401	ng	95
20) 3+4-Methylphenols	7.163	107	7257134	1238.330	ng	# 51
22) Acetophenone	7.157	105	69553	10.380	ng	# 1
24) Nitrobenzene	7.392	77	540113	99.154	ng	# 67
25) Isophorone	7.692	82	41591	4.810	ng	# 1
27) 2,4-Dimethylphenol	7.698	122	8151329	2370.787	ng	# 83
28) bis(2-Chloroethoxy)met...	7.698	93	428994	88.112	ng	# 1
31) Naphthalene	8.004	128	1268452	89.835	ng	100
32) Benzoic acid	7.798	122	4104617	1543.741	ng	# 11
35) Caprolactam	8.428	113	7768	7.192	ng	# 1
36) 4-Chloro-3-methylphenol	8.522	107	873453	209.148	ng	# 21
37) 2-Methylnaphthalene	8.686	142	23124	2.569	ng	91
38) 1-Methylnaphthalene	8.786	142	38469	4.448	ng	93
52) Acenaphthene	9.739	154	20719	2.678	ng	95
54) 2,4-Dinitrophenol	9.863	184	2346	8.377	ng	# 1
55) Dibenzofuran	9.910	168	28942	2.689	ng	97
56) 4-Nitrophenol	9.851	139	26543	17.414	ng	# 20
58) Fluorene	10.251	166	23946	2.721	ng	# 95
67) 4-Bromophenyl-phenylether	10.492	248	7399	3.745	ng	# 1
73) Carbazole	11.416	167	110426	12.608	ng	99
77) Benzidine	12.539	184	59495	11.671	ng	# 48


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF091622\
Data File : BF130306.D
Acq On : 16 Sep 2022 23:34
Operator : CG\JU
Sample : N4648-01
Misc :
ALS Vial : 16 Sample Multiplier: 1

Instrument :
BNA_F
ClientSampleId :
N48854-1040

Quant Time: Sep 17 03:32:11 2022
Quant Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF091422.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Sat Sep 17 03:22:32 2022
Response via : Initial Calibration









#1
 1,4-Dichlorobenzene-d4
 Concen: 20.000 ng
 RT: 6.669 min Scan# 71
 Delta R.T. -0.000 min
 Lab File: BF130306.D
 Acq: 16 Sep 2022 23:34




Instrument :
 BNA_F
 ClientSampleId :
 N48854-1040


Tgt Ion:152 Resp: 84499
 Ion Ratio Lower Upper
 152 100
 150 157.4 122.8 184.2
 115 68.8 53.8 80.8



#5
 2-Fluorophenol
 Concen: 52.232 ng
 RT: 5.263 min Scan# 540
 Delta R.T. 0.000 min
 Lab File: BF130306.D
 Acq: 16 Sep 2022 23:34

Tgt Ion:112 Resp: 254462
 Ion Ratio Lower Upper
 112 100
 64 59.4 51.4 77.0
 63 32.7 28.8 43.2





#6

Aniline

Concen: 13.225 ng

RT: 6.339 min Scan# 71

Delta R.T. 0.006 min

Lab File: BF130306.D

Acq: 16 Sep 2022 23:34

Instrument :

BNA_F

ClientSampleId :

N48854-1040





Tgt Ion: 93 Resp: 99327

Ion Ratio Lower Upper

93 100

66 1319.5 34.2 51.4#

65 911.1 16.6 24.8#



#7

Phenol-d6

Concen: 32.111 ng

RT: 6.322 min Scan# 720

Delta R.T. 0.018 min

Lab File: BF130306.D

Acq: 16 Sep 2022 23:34




Tgt Ion: 99 Resp: 195738


Ion Ratio Lower Upper

99 100

42 27.7 17.0 25.6#

71 39.9 32.3 48.5









#9
Benzaldehyde
Concen: 9.932 ng
RT: 6.263 min Scan# 71
Delta R.T. 0.047 min
Lab File: BF130306.D
Acq: 16 Sep 2022 23:34




Instrument :
BNA_F
ClientSampleId :
N48854-1040


Tgt Ion: 77 Resp: 40554
Ion Ratio Lower Upper
77 100
105 9.3 75.1 115.1#
106 5.6 69.7 109.7#



#10
Phenol
Concen: 424.090 ng
RT: 6.345 min Scan# 724
Delta R.T. 0.024 min
Lab File: BF130306.D
Acq: 16 Sep 2022 23:34

Tgt Ion: 94 Resp: 3029516
Ion Ratio Lower Upper
94 100
65 30.5 14.1 54.1
66 44.3 28.1 68.1









#11
bis(2-Chloroethyl)ether
Concen: 9.816 ng
RT: 6.398 min Scan# 71
Delta R.T. -0.018 min
Lab File: BF130306.D
Acq: 16 Sep 2022 23:34




Instrument :
BNA_F
ClientSampleId :
N48854-1040


Tgt Ion: 93 Resp: 50579
Ion Ratio Lower Upper
93 100
63 6.1 62.5 102.5#
95 0.8 12.5 52.5#



#15
Benzyl Alcohol
Concen: 581.394 ng
RT: 6.981 min Scan# 832
Delta R.T. 0.159 min
Lab File: BF130306.D
Acq: 16 Sep 2022 23:34

Tgt Ion: 79 Resp: 2809901
Ion Ratio Lower Upper
79 100
108 247.4 54.0 81.0#
77 97.0 51.3 76.9#









#17
2-Methylphenol
Concen: 1305.401 ng
RT: 6.981 min Scan# 81
Delta R.T. 0.053 min
Lab File: BF130306.D
Acq: 16 Sep 2022 23:34

Instrument :
BNA_F
ClientSampleId :
N48854-1040

Tgt Ion:107 Resp: 5889048




Ion	Ratio	Lower	Upper
107	100		
108	114.4	89.4	134.0
77	44.9	40.6	60.8
79	46.3	40.2	60.2




#20
3+4-Methylphenols
Concen: 1238.330 ng
RT: 7.163 min Scan# 863
Delta R.T. 0.077 min
Lab File: BF130306.D
Acq: 16 Sep 2022 23:34

Tgt Ion:107 Resp: 7257134

Ion	Ratio	Lower	Upper
107	100		
108	102.3	74.8	114.8
77	36.1	125.8	165.8#
79	38.2	17.1	57.1









#21
Naphthalene-d8
Concen: 20.000 ng
RT: 7.981 min Scan# 1002
Delta R.T. 0.029 min
Lab File: BF130306.D
Acq: 16 Sep 2022 23:34




Instrument :
BNA_F
ClientSampleId :
N48854-1040


Tgt Ion:136 Resp: 274072
Ion Ratio Lower Upper
136 100
137 11.7 9.1 13.7
54 9.4 7.1 10.7
68 6.2 4.5 6.7



#22
Acetophenone
Concen: 10.380 ng
RT: 7.157 min Scan# 862
Delta R.T. 0.071 min
Lab File: BF130306.D
Acq: 16 Sep 2022 23:34

Tgt Ion:105 Resp: 69553
Ion Ratio Lower Upper
105 100
71 6.1 3.9 5.9#
51 1319.8 32.2 48.2#
120 1.9 17.3 25.9#









#23
Nitrobenzene-d5
Concen: 89.157 ng
RT: 7.263 min Scan# 81
Delta R.T. 0.024 min
Lab File: BF130306.D
Acq: 16 Sep 2022 23:34




Instrument :
BNA_F
ClientSampleId :
N48854-1040


Tgt Ion: 82 Resp: 478294
Ion Ratio Lower Upper
82 100
128 41.3 32.6 49.0
54 52.9 43.8 65.6



#24
Nitrobenzene
Concen: 99.154 ng
RT: 7.392 min Scan# 902
Delta R.T. 0.135 min
Lab File: BF130306.D
Acq: 16 Sep 2022 23:34

Tgt Ion: 77 Resp: 540113
Ion Ratio Lower Upper
77 100
123 22.4 34.3 51.5#
65 29.5 12.0 18.0#









#25
Isophorone
Concen: 4.810 ng
RT: 7.692 min Scan# 91
Delta R.T. 0.194 min
Lab File: BF130306.D
Acq: 16 Sep 2022 23:34




Instrument :
BNA_F
ClientSampleId :
N48854-1040


Tgt Ion: 82 Resp: 41591
Ion Ratio Lower Upper
82 100
95 178.7 6.4 9.6#
138 5.8 13.8 20.6#



#27
2,4-Dimethylphenol
Concen: 2370.787 ng
RT: 7.698 min Scan# 954
Delta R.T. 0.082 min
Lab File: BF130306.D
Acq: 16 Sep 2022 23:34

Tgt Ion: 122 Resp: 8151329
Ion Ratio Lower Upper
122 100
107 106.7 100.2 150.2
121 45.7 47.8 71.8#









#28
bis(2-Chloroethoxy)methane
Concen: 88.112 ng
RT: 7.698 min Scan# 91
Delta R.T. -0.018 min
Lab File: BF130306.D
Acq: 16 Sep 2022 23:34




Instrument :
BNA_F
ClientSampleId :
N48854-1040


Tgt Ion: 93 Resp: 428994
Ion Ratio Lower Upper
93 100
95 17.4 26.3 39.5#
123 168.1 9.4 14.2#



#31
Naphthalene
Concen: 89.835 ng
RT: 8.004 min Scan# 1006
Delta R.T. 0.029 min
Lab File: BF130306.D
Acq: 16 Sep 2022 23:34

Tgt Ion: 128 Resp: 1268452
Ion Ratio Lower Upper
128 100
129 11.6 9.1 13.7
127 13.2 10.6 16.0









#32
Benzoic acid
Concen: 1543.741 ng
RT: 7.798 min Scan# 91
Delta R.T. 0.059 min
Lab File: BF130306.D
Acq: 16 Sep 2022 23:34




Instrument :
BNA_F
ClientSampleId :
N48854-1040


Tgt Ion:122 Resp: 4104617
Ion Ratio Lower Upper
122 100
105 3.3 112.9 152.9#
77 40.1 82.8 122.8#



#35
Caprolactam
Concen: 7.192 ng
RT: 8.428 min Scan# 1078
Delta R.T. 0.024 min
Lab File: BF130306.D
Acq: 16 Sep 2022 23:34

Tgt Ion:113 Resp: 7768
Ion Ratio Lower Upper
113 100
55 705.0 155.4 195.4#
56 75.1 120.4 160.4#









#36
 4-Chloro-3-methylphenol
 Concen: 209.148 ng
 RT: 8.522 min Scan# 1109
 Delta R.T. 0.012 min
 Lab File: BF130306.D
 Acq: 16 Sep 2022 23:34




Instrument :
 BNA_F
 ClientSampleId :
 N48854-1040


Tgt Ion:107 Resp: 873453
 Ion Ratio Lower Upper
 107 100
 144 0.0 20.0 30.0#
 142 0.0 61.3 91.9#



#37
 2-Methylnaphthalene
 Concen: 2.569 ng
 RT: 8.686 min Scan# 1122
 Delta R.T. 0.024 min
 Lab File: BF130306.D
 Acq: 16 Sep 2022 23:34

Tgt Ion:142 Resp: 23124
 Ion Ratio Lower Upper
 142 100
 141 96.2 70.0 105.0









#38
1-Methylnaphthalene
Concen: 4.448 ng
RT: 8.786 min Scan# 1103
Delta R.T. 0.024 min
Lab File: BF130306.D
Acq: 16 Sep 2022 23:34




Instrument :
BNA_F
ClientSampleId :
N48854-1040


Tgt Ion:142 Resp: 38469
Ion Ratio Lower Upper
142 100
141 99.2 73.8 110.8



#39
Acenaphthene-d10
Concen: 20.000 ng
RT: 9.710 min Scan# 1296
Delta R.T. 0.006 min
Lab File: BF130306.D
Acq: 16 Sep 2022 23:34

Tgt Ion:164 Resp: 130007
Ion Ratio Lower Upper
164 100
162 106.3 81.4 122.0
160 46.7 36.2 54.4





#42

2,4,6-Tribromophenol

Concen: 98.633 ng

RT: 10.492 min Scan# 1429

Delta R.T. 0.000 min

Lab File: BF130306.D

Acq: 16 Sep 2022 23:34

Instrument :

BNA_F

ClientSampleId :

N48854-1040





Tgt Ion:330 Resp: 122868

Ion Ratio Lower Upper

330 100

332 107.0 83.4 125.2

141 44.3 34.0 51.0



#45

2-Fluorobiphenyl

Concen: 82.960 ng

RT: 9.045 min Scan# 1183

Delta R.T. 0.012 min

Lab File: BF130306.D

Acq: 16 Sep 2022 23:34




Tgt Ion:172 Resp: 740661


Ion Ratio Lower Upper

172 100

171 34.8 28.3 42.5

170 23.5 18.7 28.1









#52
Acenaphthene
Concen: 2.678 ng
RT: 9.739 min Scan# 11
Delta R.T. 0.000 min
Lab File: BF130306.D
Acq: 16 Sep 2022 23:34




Instrument :
BNA_F
ClientSampleId :
N48854-1040


Tgt Ion:154 Resp: 20719
Ion Ratio Lower Upper
154 100
153 108.8 90.6 136.0
152 50.8 43.8 65.6



#54
2,4-Dinitrophenol
Concen: 8.377 ng
RT: 9.863 min Scan# 1322
Delta R.T. 0.100 min
Lab File: BF130306.D
Acq: 16 Sep 2022 23:34

Tgt Ion:184 Resp: 2346
Ion Ratio Lower Upper
184 100
63 1592.0 70.0 105.0#
154 26.0 63.0 94.6#





#55

Dibenzofuran

Concen: 2.689 ng

RT: 9.910 min Scan# 11

Delta R.T. 0.000 min

Lab File: BF130306.D

Acq: 16 Sep 2022 23:34

Instrument :

BNA_F

ClientSampleId :

N48854-1040





Tgt Ion:168 Resp: 28942

Ion Ratio Lower Upper

168 100

139 45.2 35.0 52.4

169 15.6 10.9 16.3



#56

4-Nitrophenol

Concen: 17.414 ng

RT: 9.851 min Scan# 1320

Delta R.T. 0.029 min

Lab File: BF130306.D

Acq: 16 Sep 2022 23:34




Tgt Ion:139 Resp: 26543


Ion Ratio Lower Upper

139 100

109 12.7 72.2 112.2#

65 32.9 97.7 137.7#





#58

Fluorene

Concen: 2.721 ng

RT: 10.251 min Scan# 11

Delta R.T. 0.000 min

Lab File: BF130306.D

Acq: 16 Sep 2022 23:34

Instrument :

BNA_F

ClientSampleId :

N48854-1040





Tgt Ion:166 Resp: 23946

Ion Ratio Lower Upper

166 100

165 93.8 78.8 118.2

167 16.2 10.6 15.8#



#64

Phenanthrene-d10

Concen: 20.000 ng

RT: 11.180 min Scan# 1546

Delta R.T. 0.000 min

Lab File: BF130306.D

Acq: 16 Sep 2022 23:34




Tgt Ion:188 Resp: 179118


Ion Ratio Lower Upper

188 100

94 8.8 7.4 11.2

80 8.6 7.9 11.9









#67
4-Bromophenyl-phenylether
Concen: 3.745 ng
RT: 10.492 min Scan# 1429
Delta R.T. -0.241 min
Lab File: BF130306.D
Acq: 16 Sep 2022 23:34




Instrument :
BNA_F
ClientSampleId :
N48854-1040


Tgt Ion:248 Resp: 7399
Ion Ratio Lower Upper
248 100
250 196.1 80.0 120.0#
141 569.8 68.5 102.7#



#73
Carbazole
Concen: 12.608 ng
RT: 11.416 min Scan# 1586
Delta R.T. 0.000 min
Lab File: BF130306.D
Acq: 16 Sep 2022 23:34

Tgt Ion:167 Resp: 110426
Ion Ratio Lower Upper
167 100
166 21.9 17.4 26.2
139 15.9 11.6 17.4





#76

Chrysene-d12

Concen: 20.000 ng

RT: 13.810 min Scan# 1961

Delta R.T. 0.000 min

Lab File: BF130306.D

Acq: 16 Sep 2022 23:34

Instrument :

BNA_F

ClientSampleId :

N48854-1040





Tgt Ion:240 Resp: 175610

Ion Ratio Lower Upper

240 100

120 9.9 8.7 13.1

236 27.0 20.9 31.3



#77

Benzidine

Concen: 11.671 ng

RT: 12.539 min Scan# 1777

Delta R.T. 0.024 min

Lab File: BF130306.D

Acq: 16 Sep 2022 23:34




Tgt Ion:184 Resp: 59495


Ion Ratio Lower Upper

184 100

185 21.6 11.8 17.6#

183 49.2 9.1 13.7#









#79
 Terphenyl-d14
 Concen: 45.176 ng
 RT: 12.768 min Scan# 11
 Delta R.T. 0.000 min
 Lab File: BF130306.D
 Acq: 16 Sep 2022 23:34

Instrument :
 BNA_F
 ClientSampleId :
 N48854-1040

Tgt Ion:244 Resp: 478929
 Ion Ratio Lower Upper
 244 100
 212 7.9 5.9 8.9
 122 12.8 8.5 12.7#



#86
 Perylene-d12
 Concen: 20.000 ng
 RT: 15.198 min Scan# 2229
 Delta R.T. 0.000 min
 Lab File: BF130306.D
 Acq: 16 Sep 2022 23:34

Tgt Ion:264 Resp: 203293
 Ion Ratio Lower Upper
 264 100
 260 23.9 19.3 28.9
 265 20.9 17.0 25.4

