

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF091622\
 Data File : BF130292.D
 Acq On : 16 Sep 2022 16:09
 Operator : CG\JU
 Sample : SSTDCCC040
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_F
 ClientSampleId :
 SSTDCCC040

Manual Integrations
 APPROVED

Reviewed By : Christian Giraldo 09/19/2022
 Supervised By : Jagrut Upadhyay 09/19/2022

Quant Time: Sep 17 03:24:40 2022
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF091422.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Sat Sep 17 03:22:32 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	6.669	152	109575	20.000	ng	0.00
21) Naphthalene-d8	7.951	136	406750	20.000	ng	0.00
39) Acenaphthene-d10	9.704	164	207745	20.000	ng	0.00
64) Phenanthrene-d10	11.180	188	349026	20.000	ng	0.00
76) Chrysene-d12	13.810	240	205901	20.000	ng	0.00
86) Perylene-d12	15.198	264	170906	20.000	ng	0.00
System Monitoring Compounds						
5) 2-Fluorophenol	5.263	112	530855	84.029	ng	0.00
7) Phenol-d6	6.304	99	653628	82.689	ng	0.00
23) Nitrobenzene-d5	7.239	82	633734	79.598	ng	0.00
42) 2,4,6-Tribromophenol	10.492	330	184823	92.848	ng	0.00
45) 2-Fluorobiphenyl	9.033	172	1141686	80.027	ng	0.00
79) Terphenyl-d14	12.768	244	1076498	86.605	ng	0.00
Target Compounds						Qvalue
2) 1,4-Dioxane	2.304	88	108334	37.339	ng	98
3) Pyridine	2.987	79	281604	37.207	ng	99
4) n-Nitrosodimethylamine	2.951	42	150070	36.456	ng	99
6) Aniline	6.334	93	383742	39.400	ng	98
8) 2-Chlorophenol	6.451	128	283825	39.439	ng	98
9) Benzaldehyde	6.216	77	186686	35.258	ng	97
10) Phenol	6.322	94	363372	39.226	ng	95
11) bis(2-Chloroethyl)ether	6.416	93	259188	38.791	ng	93
12) 1,3-Dichlorobenzene	6.610	146	315819	39.884	ng	99
13) 1,4-Dichlorobenzene	6.686	146	320571	39.699	ng	99
14) 1,2-Dichlorobenzene	6.839	146	299460	39.699	ng	99
15) Benzyl Alcohol	6.822	79	235480	37.573	ng	95
16) 2,2'-oxybis(1-Chloropr...	6.951	45	353112	36.234	ng	98
17) 2-Methylphenol	6.928	107	229074	39.157	ng	96
18) Hexachloroethane	7.181	117	123727	38.871	ng	100
19) n-Nitroso-di-n-propyla...	7.098	70	202728	38.007	ng	93
20) 3+4-Methylphenols	7.086	107	295794	38.922	ng	92
22) Acetophenone	7.086	105	398429	40.065	ng	# 98
24) Nitrobenzene	7.257	77	316667	39.171	ng	99
25) Isophorone	7.498	82	498436	38.842	ng	99
26) 2-Nitrophenol	7.575	139	145215	43.819	ng	# 88
27) 2,4-Dimethylphenol	7.616	122	215023	42.139	ng	95
28) bis(2-Chloroethoxy)met...	7.716	93	289420	40.054	ng	99
29) 2,4-Dichlorophenol	7.816	162	234230	41.889	ng	95
30) 1,2,4-Trichlorobenzene	7.898	180	249386	41.251	ng	95
31) Naphthalene	7.975	128	849199	40.524	ng	100
32) Benzoic acid	7.739	122	161614	40.956	ng	95
33) 4-Chloroaniline	8.028	127	325089	40.789	ng	99
34) Hexachlorobutadiene	8.092	225	160822	41.624	ng	100
35) Caprolactam	8.404	113	67292m	41.978	ng	
36) 4-Chloro-3-methylphenol	8.510	107	253313	40.870	ng	97
37) 2-Methylnaphthalene	8.663	142	538339	40.292	ng	99
38) 1-Methylnaphthalene	8.763	142	522008	40.670	ng	100
40) 1,2,4,5-Tetrachloroben...	8.828	216	237019	41.806	ng	99
41) Hexachlorocyclopentadiene	8.816	237	130361	42.947	ng	97
43) 2,4,6-Trichlorophenol	8.939	196	165474	41.820	ng	97

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF091622\
 Data File : BF130292.D
 Acq On : 16 Sep 2022 16:09
 Operator : CG\JU
 Sample : SSTDCCC040
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_F
 ClientSampleId :
 SSTDCCC040

Manual Integrations
 APPROVED

Reviewed By : Christian Giraldo 09/19/2022
 Supervised By : Jagrut Upadhyay 09/19/2022

Quant Time: Sep 17 03:24:40 2022
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF091422.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Sat Sep 17 03:22:32 2022
 Response via : Initial Calibration


Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
44) 2,4,5-Trichlorophenol	8.980	196	179737	42.095	ng	97
46) 1,1'-Biphenyl	9.127	154	640995	41.314	ng	100
47) 2-Chloronaphthalene	9.151	162	508739	40.534	ng	98
48) 2-Nitroaniline	9.251	65	162268	38.763	ng	98
49) Acenaphthylene	9.563	152	763625	40.579	ng	100
50) Dimethylphthalate	9.439	163	590931	41.179	ng	99
51) 2,6-Dinitrotoluene	9.498	165	126830	42.260	ng	# 82
52) Acenaphthene	9.739	154	498646m	40.330	ng	
53) 3-Nitroaniline	9.663	138	137932	41.246	ng	99
54) 2,4-Dinitrophenol	9.763	184	60192	40.336	ng	99
55) Dibenzofuran	9.910	168	687802	39.994	ng	99
56) 4-Nitrophenol	9.822	139	101279	41.581	ng	90
57) 2,4-Dinitrotoluene	9.898	165	166771	43.481	ng	85
58) Fluorene	10.251	166	567345	40.339	ng	98
59) 2,3,4,6-Tetrachlorophenol	10.027	232	140863	42.149	ng	96
60) Diethylphthalate	10.133	149	583523	40.549	ng	99
61) 4-Chlorophenyl-phenyle...	10.245	204	254409	41.234	ng	99
62) 4-Nitroaniline	10.274	138	139087	41.402	ng	96
63) Azobenzene	10.404	77	566903	38.258	ng	99
65) 4,6-Dinitro-2-methylph...	10.298	198	83830	43.146	ng	97
66) n-Nitrosodiphenylamine	10.369	169	473311	40.581	ng	98
67) 4-Bromophenyl-phenylether	10.733	248	164551	42.737	ng	96
68) Hexachlorobenzene	10.792	284	184491	42.270	ng	99
69) Atrazine	10.892	200	138503	40.635	ng	97
70) Pentachlorophenol	10.986	266	102098	43.586	ng	97
71) Phenanthrene	11.210	178	787945	40.210	ng	99
72) Anthracene	11.257	178	768974	40.664	ng	99
73) Carbazole	11.416	167	678976	39.785	ng	100
74) Di-n-butylphthalate	11.751	149	860385	41.807	ng	100
75) Fluoranthene	12.392	202	775912	40.977	ng	96
77) Benzidine	12.515	184	200031	33.466	ng	99
78) Pyrene	12.616	202	767321	42.600	ng	100
80) Butylbenzylphthalate	13.245	149	296399	44.419	ng	95
81) Benzo(a)anthracene	13.798	228	561457	40.990	ng	99
82) 3,3'-Dichlorobenzidine	13.768	252	172494	46.262	ng	99
83) Chrysene	13.839	228	536461	41.248	ng	99
84) Bis(2-ethylhexyl)phtha...	13.804	149	332184	43.461	ng	99
85) Di-n-octyl phthalate	14.415	149	510328	43.653	ng	97
87) Indeno(1,2,3-cd)pyrene	16.527	276	527494	43.524	ng	99
88) Benzo(b)fluoranthene	14.809	252	452756	40.588	ng	98
89) Benzo(k)fluoranthene	14.839	252	447020	40.738	ng	98
90) Benzo(a)pyrene	15.139	252	367326	41.565	ng	99
91) Dibenzo(a,h)anthracene	16.545	278	438040	44.058	ng	99
92) Benzo(g,h,i)perylene	16.927	276	431829	43.116	ng	99


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Manual Integrations

Reviewed By :Christian Giraldo	09/19/2022
Supervised By :Jagrut Upadhyay	09/19/2022


Quant Time: Sep 17 03:24:40 2022
Quant Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF091422.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Sat Sep 17 03:22:32 2022
Response via : Initial Calibration





#1
1,4-Dichlorobenzene-d4
Concen: 20.000 ng
RT: 6.669 min Scan# 71
Delta R.T. -0.000 min
Lab File: BF130292.D
Acq: 16 Sep 2022 16:09




Instrument :
BNA_F
ClientSampleId :
SSTDCCC040




Tgt Ion:152 Resp: 109575
Ion Ratio Lower Upper
152 100
150 158.9 122.8 184.2
115 66.9 53.8 80.8

Manual Integrations
APPROVED



Reviewed By :Christian Giraldo 09/19/2022
Supervised By :Jagrut Upadhyay 09/19/2022




#2
1,4-Dioxane
Concen: 37.339 ng
RT: 2.304 min Scan# 37
Delta R.T. 0.000 min
Lab File: BF130292.D
Acq: 16 Sep 2022 16:09




Tgt Ion: 88 Resp: 108334
Ion Ratio Lower Upper
88 100
58 76.0 61.3 91.9
43 31.6 23.0 34.6





#3
Pyridine
Concen: 37.207 ng
RT: 2.987 min Scan# 11
Delta R.T. 0.000 min
Lab File: BF130292.D
Acq: 16 Sep 2022 16:09




Instrument :
BNA_F
ClientSampleId :
SSTDCCC040




Tgt Ion: 79 Resp: 281604
Ion Ratio Lower Upper
79 100
52 77.0 62.4 93.6
51 40.3 32.2 48.4

Manual Integrations
APPROVED



Reviewed By :Christian Giraldo 09/19/2022
Supervised By :Jagrut Upadhyay 09/19/2022




#4
n-Nitrosodimethylamine
Concen: 36.456 ng
RT: 2.951 min Scan# 147
Delta R.T. 0.000 min
Lab File: BF130292.D
Acq: 16 Sep 2022 16:09




Tgt Ion: 42 Resp: 150070
Ion Ratio Lower Upper
42 100
74 111.1 88.4 132.6
44 6.1 5.0 7.4





#5
2-Fluorophenol
Concen: 84.029 ng
RT: 5.263 min Scan# 54
Delta R.T. 0.000 min
Lab File: BF130292.D
Acq: 16 Sep 2022 16:09




Instrument :
BNA_F
ClientSampleId :
SSTDCCC040



Tgt Ion:112 Resp: 53085
Ion Ratio Lower Upper
112 100
64 61.9 51.4 77.0
63 33.9 28.8 43.2




Manual Integrations
APPROVED


Reviewed By :Christian Giraldo 09/19/2022
Supervised By :Jagrut Upadhyay 09/19/2022



#6
Aniline
Concen: 39.400 ng
RT: 6.334 min Scan# 722
Delta R.T. 0.000 min
Lab File: BF130292.D
Acq: 16 Sep 2022 16:09


Tgt Ion: 93 Resp: 383742
Ion Ratio Lower Upper
93 100
66 41.2 34.2 51.4
65 21.3 16.6 24.8





#7
Phenol-d6
Concen: 82.689 ng
RT: 6.304 min Scan# 711
Delta R.T. 0.000 min
Lab File: BF130292.D
Acq: 16 Sep 2022 16:09




Instrument :
BNA_F
ClientSampleId :
SSTDCCC040




Tgt Ion: 99 Resp: 653628
Ion Ratio Lower Upper
99 100
42 20.8 17.0 25.6
71 39.1 32.3 48.5

Manual Integrations
APPROVED



Reviewed By :Christian Giraldo 09/19/2022
Supervised By :Jagrut Upadhyay 09/19/2022




#8
2-Chlorophenol
Concen: 39.439 ng
RT: 6.451 min Scan# 742
Delta R.T. 0.000 min
Lab File: BF130292.D
Acq: 16 Sep 2022 16:09




Tgt Ion:128 Resp: 283825
Ion Ratio Lower Upper
128 100
130 32.3 12.4 52.4
64 49.4 31.6 71.6





#9
Benzaldehyde
Concen: 35.258 ng
RT: 6.216 min Scan# 70
Delta R.T. 0.000 min
Lab File: BF130292.D
Acq: 16 Sep 2022 16:09




Instrument :
BNA_F
ClientSampleId :
SSTDCCC040




Tgt Ion: 77 Resp: 186686
Ion Ratio Lower Upper
77 100
105 96.8 75.1 115.1
106 94.5 69.7 109.7

Manual Integrations
APPROVED



Reviewed By :Christian Giraldo 09/19/2022
Supervised By :Jagrut Upadhyay 09/19/2022




#10
Phenol
Concen: 39.226 ng
RT: 6.322 min Scan# 720
Delta R.T. 0.000 min
Lab File: BF130292.D
Acq: 16 Sep 2022 16:09



Tgt Ion: 94 Resp: 363372
Ion Ratio Lower Upper
94 100
65 34.5 14.1 54.1
66 53.4 28.1 68.1









#11
bis(2-Chloroethyl)ether
Concen: 38.791 ng
RT: 6.416 min Scan# 736
Delta R.T. 0.000 min
Lab File: BF130292.D
Acq: 16 Sep 2022 16:09

Instrument :
BNA_F
ClientSampleId :
SSTDCCC040

Tgt Ion: 93 Resp: 259188
Ion Ratio Lower Upper
93 100
63 73.6 62.5 102.5
95 32.9 12.5 52.5




Manual Integrations
APPROVED


Reviewed By :Christian Giraldo 09/19/2022
Supervised By :Jagrut Upadhyay 09/19/2022



#12
1,3-Dichlorobenzene
Concen: 39.884 ng
RT: 6.610 min Scan# 769
Delta R.T. 0.000 min
Lab File: BF130292.D
Acq: 16 Sep 2022 16:09

Tgt Ion:146 Resp: 315819
Ion Ratio Lower Upper
146 100
148 64.2 51.4 77.2
75 33.9 28.8 43.2





#13

1,4-Dichlorobenzene

Concen: 39.699 ng

RT: 6.686 min Scan# 71

Delta R.T. 0.000 min

Lab File: BF130292.D

Acq: 16 Sep 2022 16:09

Instrument :

BNA_F

ClientSampleId :

SSTDCCC040

Tgt Ion:146 Resp: 32057

Ion Ratio Lower Upper

146 100

148 63.6 50.3 75.5





111 46.4 36.3 54.5

Manual Integrations

APPROVED

Reviewed By :Christian Giraldo 09/19/2022

Supervised By :Jagrut Upadhyay 09/19/2022



#14

1,2-Dichlorobenzene

Concen: 39.699 ng

RT: 6.839 min Scan# 808

Delta R.T. 0.000 min

Lab File: BF130292.D

Acq: 16 Sep 2022 16:09




Tgt Ion:146 Resp: 299460


Ion Ratio Lower Upper

146 100

148 65.0 51.6 77.4


111 49.6 39.2 58.8





#15
Benzyl Alcohol
Concen: 37.573 ng
RT: 6.822 min Scan# 801
Delta R.T. 0.000 min
Lab File: BF130292.D
Acq: 16 Sep 2022 16:09




Instrument :
BNA_F
ClientSampleId :
SSTDCCC040




Tgt Ion: 79 Resp: 235486
Ion Ratio Lower Upper
79 100
108 73.0 54.0 81.0
77 65.9 51.3 76.9

Manual Integrations
APPROVED



Reviewed By :Christian Giraldo 09/19/2022
Supervised By :Jagrut Upadhyay 09/19/2022




#16
2,2'-oxybis(1-Chloropropane)
Concen: 36.234 ng
RT: 6.951 min Scan# 827
Delta R.T. 0.000 min
Lab File: BF130292.D
Acq: 16 Sep 2022 16:09




Tgt Ion: 45 Resp: 353112
Ion Ratio Lower Upper
45 100
77 15.6 0.0 34.9
79 12.3 0.0 31.0





#17
2-Methylphenol
Concen: 39.157 ng
RT: 6.928 min Scan# 81
Delta R.T. 0.000 min
Lab File: BF130292.D
Acq: 16 Sep 2022 16:09

Instrument :
BNA_F
ClientSampleId :
SSTDCCC040






Tgt Ion:107 Resp: 229074


Ion	Ratio	Lower	Upper
107	100		
108	115.7	89.4	134.0
77	53.3	40.6	60.8
79	53.7	40.2	60.2

Manual Integrations
APPROVED

Reviewed By :Christian Giraldo 09/19/2022
Supervised By :Jagrut Upadhyay 09/19/2022






#18
Hexachloroethane
Concen: 38.871 ng
RT: 7.181 min Scan# 866
Delta R.T. 0.000 min
Lab File: BF130292.D
Acq: 16 Sep 2022 16:09



Tgt Ion:117 Resp: 123727

Ion	Ratio	Lower	Upper
117	100		
119	97.2	77.8	116.6
201	86.4	68.7	103.1





#19

n-Nitroso-di-n-propylamine

Concen: 38.007 ng

RT: 7.098 min Scan# 81

Delta R.T. 0.000 min

Lab File: BF130292.D

Acq: 16 Sep 2022 16:09

Instrument :

BNA_F

ClientSampleId :

SSTDCCC040

Tgt Ion: 70 Resp: 202728

Ion Ratio Lower Upper

70 100

42 56.8 50.8 76.2

101 10.7 8.2 12.4





130 23.4 16.6 25.0

Manual Integrations

APPROVED

Reviewed By :Christian Giraldo 09/19/2022

Supervised By :Jagrut Upadhyay 09/19/2022



#20

3+4-Methylphenols

Concen: 38.922 ng

RT: 7.086 min Scan# 850

Delta R.T. 0.000 min

Lab File: BF130292.D

Acq: 16 Sep 2022 16:09

Tgt Ion:107 Resp: 295794




Ion Ratio Lower Upper


107 100

108 90.5 74.8 114.8

77 130.9 125.8 165.8


79 34.1 17.1 57.1





#21
Naphthalene-d8
Concen: 20.000 ng
RT: 7.951 min Scan# 94
Delta R.T. 0.000 min
Lab File: BF130292.D
Acq: 16 Sep 2022 16:09




Instrument :
BNA_F
ClientSampleId :
SSTDCCC040



Tgt Ion:136 Resp: 406750
Ion Ratio Lower Upper
136 100
137 10.9 9.1 13.7
54 8.0 7.1 10.7
68 5.1 4.5 6.7




Manual Integrations
APPROVED


Reviewed By :Christian Giraldo 09/19/2022
Supervised By :Jagrut Upadhyay 09/19/2022



#22
Acetophenone
Concen: 40.065 ng
RT: 7.086 min Scan# 850
Delta R.T. 0.000 min
Lab File: BF130292.D
Acq: 16 Sep 2022 16:09

Tgt Ion:105 Resp: 398429
Ion Ratio Lower Upper
105 100
71 3.8 3.9 5.9#
51 38.6 32.2 48.2
120 21.6 17.3 25.9





#23

Nitrobenzene-d5

Concen: 79.598 ng

RT: 7.239 min Scan# 81

Delta R.T. 0.000 min

Lab File: BF130292.D

Acq: 16 Sep 2022 16:09

Instrument :

BNA_F





ClientSampleId :

SSTDCCC040

Tgt Ion: 82 Resp: 633734
 Ion Ratio Lower Upper
 82 100
 128 41.7 32.6 49.0
 54 53.0 43.8 65.6

Manual Integrations
 APPROVED

Reviewed By :Christian Giraldo 09/19/2022
 Supervised By :Jagrut Upadhyay 09/19/2022



#24

Nitrobenzene

Concen: 39.171 ng




RT: 7.257 min Scan# 879


Delta R.T. 0.000 min

Lab File: BF130292.D

Acq: 16 Sep 2022 16:09

Tgt Ion: 77 Resp: 316667
 Ion Ratio Lower Upper
 77 100
 123 43.3 34.3 51.5
 65 14.7 12.0 18.0









#25
Isophorone
Concen: 38.842 ng
RT: 7.498 min Scan# 911
Delta R.T. 0.000 min
Lab File: BF130292.D
Acq: 16 Sep 2022 16:09

Instrument :
BNA_F
ClientSampleId :
SSTDCCC040

Tgt Ion: 82 Resp: 498436
Ion Ratio Lower Upper
82 100
95 8.2 6.4 9.6
138 18.0 13.8 20.6




Manual Integrations
APPROVED


Reviewed By :Christian Giraldo 09/19/2022
Supervised By :Jagrut Upadhyay 09/19/2022



#26
2-Nitrophenol
Concen: 43.819 ng
RT: 7.575 min Scan# 933
Delta R.T. 0.000 min
Lab File: BF130292.D
Acq: 16 Sep 2022 16:09

Tgt Ion:139 Resp: 145215
Ion Ratio Lower Upper
139 100
109 34.7 30.2 45.4
65 48.6 48.9 73.3#





#27

2,4-Dimethylphenol

Concen: 42.139 ng

RT: 7.616 min Scan# 94

Delta R.T. 0.000 min

Lab File: BF130292.D

Acq: 16 Sep 2022 16:09

Instrument :

BNA_F

ClientSampleId :

SSTDCCC040





Tgt Ion:122 Resp: 215023
 Ion Ratio Lower Upper
 122 100
 107 118.3 100.2 150.2
 121 57.7 47.8 71.8

Manual Integrations

APPROVED

Reviewed By :Christian Giraldo 09/19/2022

Supervised By :Jagrut Upadhyay 09/19/2022



#28

bis(2-Chloroethoxy)methane

Concen: 40.054 ng




RT: 7.716 min Scan# 957


Delta R.T. 0.000 min

Lab File: BF130292.D

Acq: 16 Sep 2022 16:09

Tgt Ion: 93 Resp: 289420
 Ion Ratio Lower Upper
 93 100
 95 32.7 26.3 39.5
 123 12.3 9.4 14.2





#29

2,4-Dichlorophenol

Concen: 41.889 ng

RT: 7.816 min Scan# 91

Delta R.T. 0.000 min

Lab File: BF130292.D

Acq: 16 Sep 2022 16:09

Instrument :

BNA_F

ClientSampleId :

SSTDCCC040





Tgt Ion:162 Resp: 234236
 Ion Ratio Lower Upper
 162 100
 164 64.9 42.2 82.2
 98 36.6 21.7 61.7

Manual Integrations

APPROVED

Reviewed By :Christian Giraldo 09/19/2022

Supervised By :Jagrut Upadhyay 09/19/2022



#30

1,2,4-Trichlorobenzene

Concen: 41.251 ng




RT: 7.898 min Scan# 988


Delta R.T. 0.000 min

Lab File: BF130292.D

Acq: 16 Sep 2022 16:09


Tgt Ion:180 Resp: 249386
 Ion Ratio Lower Upper
 180 100
 182 96.9 73.3 109.9
 145 32.7 28.3 42.5





#31
Naphthalene
Concen: 40.524 ng
RT: 7.975 min Scan# 10
Delta R.T. 0.000 min
Lab File: BF130292.D
Acq: 16 Sep 2022 16:09




Instrument :
BNA_F
ClientSampleId :
SSTDCCC040




Tgt Ion:128 Resp: 849199
Ion Ratio Lower Upper
128 100
129 11.3 9.1 13.7
127 13.3 10.6 16.0

Manual Integrations
APPROVED



Reviewed By :Christian Giraldo 09/19/2022
Supervised By :Jagrut Upadhyay 09/19/2022




#32
Benzoic acid
Concen: 40.956 ng
RT: 7.739 min Scan# 961
Delta R.T. 0.000 min
Lab File: BF130292.D
Acq: 16 Sep 2022 16:09



Tgt Ion:122 Resp: 161614
Ion Ratio Lower Upper
122 100
105 127.3 112.9 152.9
77 97.5 82.8 122.8





#33
4-Chloroaniline
Concen: 40.789 ng
RT: 8.028 min Scan# 1010
Delta R.T. 0.000 min
Lab File: BF130292.D
Acq: 16 Sep 2022 16:09

Instrument :

BNA_F

ClientSampleId :





SSTDCCC040

Tgt Ion:127 Resp: 325089
Ion Ratio Lower Upper
127 100
129 32.4 25.6 38.4
65 32.5 27.0 40.6
92 20.0 15.9 23.9

Manual Integrations




APPROVED


Reviewed By :Christian Giraldo 09/19/2022
Supervised By :Jagrut Upadhyay 09/19/2022



#34
Hexachlorobutadiene
Concen: 41.624 ng
RT: 8.092 min Scan# 1021
Delta R.T. 0.000 min
Lab File: BF130292.D
Acq: 16 Sep 2022 16:09


Tgt Ion:225 Resp: 160822
Ion Ratio Lower Upper
225 100
223 63.4 50.3 75.5
227 63.7 51.1 76.7





#35
Caprolactam
Concen: 41.978 ng m
RT: 8.404 min Scan# 1074
Delta R.T. 0.000 min
Lab File: BF130292.D
Acq: 16 Sep 2022 16:09




Instrument :
BNA_F
ClientSampleId :
SSTDCCC040



Tgt Ion:113 Resp: 67292
Ion Ratio Lower Upper
113 100
55 167.4 155.4 195.4
56 137.8 120.4 160.4




Manual Integrations
APPROVED


Reviewed By :Christian Giraldo 09/19/2022
Supervised By :Jagrut Upadhyay 09/19/2022



#36
4-Chloro-3-methylphenol
Concen: 40.870 ng
RT: 8.510 min Scan# 1092
Delta R.T. 0.000 min
Lab File: BF130292.D
Acq: 16 Sep 2022 16:09

Tgt Ion:107 Resp: 253313
Ion Ratio Lower Upper
107 100
144 24.4 20.0 30.0
142 73.8 61.3 91.9





#37

2-Methylnaphthalene

Concen: 40.292 ng

RT: 8.663 min Scan# 1118

Delta R.T. 0.000 min

Lab File: BF130292.D

Acq: 16 Sep 2022 16:09

Instrument :

BNA_F

ClientSampleId :

SSTDCCC040

Tgt Ion:142 Resp: 538339

Ion Ratio Lower Upper

142 100





141 88.8 70.0 105.0

Manual Integrations

APPROVED

Reviewed By :Christian Giraldo 09/19/2022

Supervised By :Jagrut Upadhyay 09/19/2022



#38

1-Methylnaphthalene

Concen: 40.670 ng

RT: 8.763 min Scan# 1135

Delta R.T. 0.000 min

Lab File: BF130292.D




Acq: 16 Sep 2022 16:09


Tgt Ion:142 Resp: 522008

Ion Ratio Lower Upper

142 100

141 92.4 73.8 110.8






#39
Acenaphthene-d10
Concen: 20.000 ng
RT: 9.704 min Scan# 1145
Delta R.T. 0.000 min
Lab File: BF130292.D
Acq: 16 Sep 2022 16:09

Instrument :

BNA_F

ClientSampleId :

SSTDCCC040






Tgt Ion:164 Resp: 207745
Ion Ratio Lower Upper
164 100
162 102.3 81.4 122.0
160 44.8 36.2 54.4


Manual Integrations

APPROVED



Reviewed By :Christian Giraldo 09/19/2022
Supervised By :Jagrut Upadhyay 09/19/2022




#40
1,2,4,5-Tetrachlorobenzene
Concen: 41.806 ng
RT: 8.828 min Scan# 1146
Delta R.T. 0.000 min
Lab File: BF130292.D
Acq: 16 Sep 2022 16:09




Tgt Ion:216 Resp: 237019
Ion Ratio Lower Upper
216 100
214 79.0 62.8 94.2
179 24.0 19.5 29.3
108 22.3 19.1 28.7





#41
Hexachlorocyclopentadiene
Concen: 42.947 ng
RT: 8.816 min Scan# 1112
Delta R.T. 0.000 min
Lab File: BF130292.D
Acq: 16 Sep 2022 16:09




Instrument :
BNA_F
ClientSampleId :
SSTDCCC040




Tgt Ion: 237 Resp: 130361
Ion Ratio Lower Upper
237 100
235 60.1 42.5 82.5
272 12.2 0.0 32.8

Manual Integrations
APPROVED



Reviewed By : Christian Giraldo 09/19/2022
Supervised By : Jagrut Upadhyay 09/19/2022




#42
2,4,6-Tribromophenol
Concen: 92.848 ng
RT: 10.492 min Scan# 1429
Delta R.T. 0.000 min
Lab File: BF130292.D
Acq: 16 Sep 2022 16:09




Tgt Ion: 330 Resp: 184823
Ion Ratio Lower Upper
330 100
332 103.8 83.4 125.2
141 37.9 34.0 51.0





#43
2,4,6-Trichlorophenol
Concen: 41.820 ng
RT: 8.939 min Scan# 1165
Delta R.T. 0.000 min
Lab File: BF130292.D
Acq: 16 Sep 2022 16:09




Instrument :
BNA_F
ClientSampleId :
SSTDCCC040




Tgt Ion:196 Resp: 165474
Ion Ratio Lower Upper
196 100
198 93.5 72.4 108.6
200 30.4 23.8 35.6

Manual Integrations
APPROVED



Reviewed By :Christian Giraldo 09/19/2022
Supervised By :Jagrut Upadhyay 09/19/2022




#44
2,4,5-Trichlorophenol
Concen: 42.095 ng
RT: 8.980 min Scan# 1172
Delta R.T. 0.000 min
Lab File: BF130292.D
Acq: 16 Sep 2022 16:09




Tgt Ion:196 Resp: 179737
Ion Ratio Lower Upper
196 100
198 97.0 74.8 112.2
97 54.8 43.7 65.5
132 30.5 22.8 34.2





#45
2-Fluorobiphenyl
Concen: 80.027 ng
RT: 9.033 min Scan# 1148
Delta R.T. 0.000 min
Lab File: BF130292.D
Acq: 16 Sep 2022 16:09




Instrument :
BNA_F
ClientSampleId :
SSTDCCC040




Tgt Ion:172 Resp: 1141686
Ion Ratio Lower Upper
172 100
171 35.5 28.3 42.5
170 23.5 18.7 28.1

Manual Integrations
APPROVED



Reviewed By :Christian Giraldo 09/19/2022
Supervised By :Jagrut Upadhyay 09/19/2022




#46
1,1'-Biphenyl
Concen: 41.314 ng
RT: 9.127 min Scan# 1197
Delta R.T. 0.000 min
Lab File: BF130292.D
Acq: 16 Sep 2022 16:09



Tgt Ion:154 Resp: 640995
Ion Ratio Lower Upper
154 100
153 40.0 20.0 60.0
76 12.7 0.0 33.3





#47

2-Chloronaphthalene

Concen: 40.534 ng

RT: 9.151 min Scan# 1169

Delta R.T. 0.000 min

Lab File: BF130292.D

Acq: 16 Sep 2022 16:09

Instrument :

BNA_F

ClientSampleId :

SSTDCCC040





Tgt Ion:162 Resp: 508739
 Ion Ratio Lower Upper
 162 100
 127 43.6 35.9 53.9
 164 32.2 26.2 39.2

Manual Integrations

APPROVED

Reviewed By :Christian Giraldo 09/19/2022

Supervised By :Jagrut Upadhyay 09/19/2022



#48

2-Nitroaniline

Concen: 38.763 ng




RT: 9.251 min Scan# 1218


Delta R.T. 0.000 min

Lab File: BF130292.D

Acq: 16 Sep 2022 16:09


Tgt Ion: 65 Resp: 162268
 Ion Ratio Lower Upper
 65 100
 92 67.8 52.2 78.4
 138 97.1 77.0 115.4





#49
Acenaphthylene
Concen: 40.579 ng
RT: 9.563 min Scan# 1239
Delta R.T. 0.000 min
Lab File: BF130292.D
Acq: 16 Sep 2022 16:09




Instrument :
BNA_F
ClientSampleId :
SSTDCCC040




Tgt Ion:152 Resp: 763625
Ion Ratio Lower Upper
152 100
151 20.2 16.2 24.2
153 13.2 10.6 15.8

Manual Integrations
APPROVED



Reviewed By :Christian Giraldo 09/19/2022
Supervised By :Jagrut Upadhyay 09/19/2022




#50
Dimethylphthalate
Concen: 41.179 ng
RT: 9.439 min Scan# 1250
Delta R.T. 0.000 min
Lab File: BF130292.D
Acq: 16 Sep 2022 16:09



Tgt Ion:163 Resp: 590931
Ion Ratio Lower Upper
163 100
194 3.9 2.8 4.2
164 9.9 8.2 12.2





#51

2,6-Dinitrotoluene

Concen: 42.260 ng

RT: 9.498 min Scan# 1301

Delta R.T. 0.000 min

Lab File: BF130292.D

Acq: 16 Sep 2022 16:09

Instrument :

BNA_F

ClientSampleId :

SSTDCCC040

Tgt Ion:165 Resp: 126830





Ion	Ratio	Lower	Upper
165	100		
63	55.5	58.8	88.2
89	55.9	54.5	81.7

Manual Integrations

APPROVED

Reviewed By :Christian Giraldo 09/19/2022

Supervised By :Jagrut Upadhyay 09/19/2022



#52

Acenaphthene

Concen: 40.330 ng m

RT: 9.739 min Scan# 1301




Delta R.T. 0.000 min


Lab File: BF130292.D

Acq: 16 Sep 2022 16:09

Tgt Ion:154 Resp: 498646

Ion	Ratio	Lower	Upper
154	100		
153	112.3	90.6	136.0
152	53.9	43.8	65.6









#53
3-Nitroaniline
Concen: 41.246 ng
RT: 9.663 min Scan# 11
Delta R.T. 0.000 min
Lab File: BF130292.D
Acq: 16 Sep 2022 16:09

Instrument :
BNA_F
ClientSampleId :
SSTDCCC040

Tgt Ion:138 Resp: 137932
Ion Ratio Lower Upper
138 100
108 11.7 9.2 13.8
92 125.3 99.2 148.8




Manual Integrations
APPROVED


Reviewed By :Christian Giraldo 09/19/2022
Supervised By :Jagrut Upadhyay 09/19/2022



#54
2,4-Dinitrophenol
Concen: 40.336 ng
RT: 9.763 min Scan# 1305
Delta R.T. 0.000 min
Lab File: BF130292.D
Acq: 16 Sep 2022 16:09

Tgt Ion:184 Resp: 60192
Ion Ratio Lower Upper
184 100
63 86.8 70.0 105.0
154 77.5 63.0 94.6






#55
Dibenzofuran
Concen: 39.994 ng
RT: 9.910 min Scan# 1310
Delta R.T. 0.000 min
Lab File: BF130292.D
Acq: 16 Sep 2022 16:09

Instrument :

BNA_F

ClientSampleId :

SSTDCCC040






Tgt Ion:168 Resp: 687802
Ion Ratio Lower Upper
168 100
139 43.0 35.0 52.4
169 13.2 10.9 16.3


Manual Integrations

APPROVED



Reviewed By :Christian Giraldo 09/19/2022
Supervised By :Jagrut Upadhyay 09/19/2022




#56
4-Nitrophenol
Concen: 41.581 ng
RT: 9.822 min Scan# 1315
Delta R.T. 0.000 min
Lab File: BF130292.D
Acq: 16 Sep 2022 16:09



Tgt Ion:139 Resp: 101279
Ion Ratio Lower Upper
139 100
109 88.7 72.2 112.2
65 100.5 97.7 137.7





#57

2,4-Dinitrotoluene

Concen: 43.481 ng

RT: 9.898 min Scan# 1328

Delta R.T. 0.000 min

Lab File: BF130292.D


Acq: 16 Sep 2022 16:09

Instrument :

BNA_F

ClientSampleId :

SSTDCCC040



Tgt Ion:165 Resp: 16677

Ion Ratio Lower Upper

165 100

63 47.8 44.7 67.1

89 70.3 69.0 103.4




182 2.6 1.9 2.9

Manual Integrations

APPROVED

Reviewed By :Christian Giraldo 09/19/2022

Supervised By :Jagrut Upadhyay 09/19/2022



#58

Fluorene

Concen: 40.339 ng

RT: 10.251 min Scan# 1388

Delta R.T. 0.000 min

Lab File: BF130292.D

Acq: 16 Sep 2022 16:09




Tgt Ion:166 Resp: 567345


Ion Ratio Lower Upper

166 100

165 96.2 78.8 118.2


167 13.4 10.6 15.8





#59
2,3,4,6-Tetrachlorophenol
Concen: 42.149 ng
RT: 10.027 min Scan# 1317
Delta R.T. 0.000 min
Lab File: BF130292.D
Acq: 16 Sep 2022 16:09




Instrument :
BNA_F
ClientSampleId :
SSTDCCC040




Tgt Ion:232 Resp: 140863
Ion Ratio Lower Upper
232 100
131 54.9 46.4 69.6
130 2.9 2.3 3.5
166 35.2 29.5 44.3

Manual Integrations
APPROVED



Reviewed By :Christian Giraldo 09/19/2022
Supervised By :Jagrut Upadhyay 09/19/2022




#60
Diethylphthalate
Concen: 40.549 ng
RT: 10.133 min Scan# 1368
Delta R.T. 0.000 min
Lab File: BF130292.D
Acq: 16 Sep 2022 16:09



Tgt Ion:149 Resp: 583523
Ion Ratio Lower Upper
149 100
177 20.6 16.1 24.1
150 12.6 9.9 14.9






#61
4-Chlorophenyl-phenylether
Concen: 41.234 ng
RT: 10.245 min Scan# 1387
Delta R.T. 0.000 min
Lab File: BF130292.D
Acq: 16 Sep 2022 16:09

Instrument :

BNA_F

ClientSampleId :

SSTDCCC040






Tgt Ion:204 Resp: 254409
Ion Ratio Lower Upper
204 100
206 32.4 26.6 39.8
141 71.1 57.1 85.7


Manual Integrations

APPROVED



Reviewed By :Christian Giraldo 09/19/2022
Supervised By :Jagrut Upadhyay 09/19/2022




#62
4-Nitroaniline
Concen: 41.402 ng
RT: 10.274 min Scan# 1392
Delta R.T. 0.000 min
Lab File: BF130292.D
Acq: 16 Sep 2022 16:09



Tgt Ion:138 Resp: 139087
Ion Ratio Lower Upper
138 100
92 54.4 36.7 76.7
108 93.7 78.2 118.2









#63
Azobenzene
Concen: 38.258 ng
RT: 10.404 min Scan# 14
Delta R.T. 0.000 min
Lab File: BF130292.D
Acq: 16 Sep 2022 16:09

Instrument :
BNA_F
ClientSampleId :
SSTDCCC040

Tgt Ion: 77 Resp: 566903
Ion Ratio Lower Upper
77 100
182 22.1 2.1 42.1
105 19.3 0.0 39.4
51 32.2 13.1 53.1




Manual Integrations
APPROVED


Reviewed By :Christian Giraldo 09/19/2022
Supervised By :Jagrut Upadhyay 09/19/2022



#64
Phenanthrene-d10
Concen: 20.000 ng
RT: 11.180 min Scan# 1546
Delta R.T. 0.000 min
Lab File: BF130292.D
Acq: 16 Sep 2022 16:09


Tgt Ion:188 Resp: 349026
Ion Ratio Lower Upper
188 100
94 9.1 7.4 11.2
80 9.5 7.9 11.9





#65
4,6-Dinitro-2-methylphenol
Concen: 43.146 ng
RT: 10.298 min Scan# 1364
Delta R.T. 0.000 min
Lab File: BF130292.D
Acq: 16 Sep 2022 16:09




Instrument :
BNA_F
ClientSampleId :
SSTDCCC040



Tgt Ion:198 Resp: 83830
Ion Ratio Lower Upper
198 100
51 59.7 36.9 76.9
105 54.5 33.0 73.0




Manual Integrations
APPROVED


Reviewed By :Christian Giraldo 09/19/2022
Supervised By :Jagrut Upadhyay 09/19/2022



#66
n-Nitrosodiphenylamine
Concen: 40.581 ng
RT: 10.369 min Scan# 1408
Delta R.T. 0.000 min
Lab File: BF130292.D
Acq: 16 Sep 2022 16:09

Tgt Ion:169 Resp: 473311
Ion Ratio Lower Upper
169 100
168 64.5 53.2 79.8
167 35.6 29.2 43.8





#67
4-Bromophenyl-phenylether
Concen: 42.737 ng
RT: 10.733 min Scan# 1437
Delta R.T. 0.000 min
Lab File: BF130292.D
Acq: 16 Sep 2022 16:09

Instrument :

BNA_F

ClientSampleId :





SSTDCCC040

Tgt Ion:248 Resp: 164551
Ion Ratio Lower Upper
248 100
250 96.2 80.0 120.0
141 81.3 68.5 102.7

Manual Integrations




APPROVED


Reviewed By :Christian Giraldo 09/19/2022
Supervised By :Jagrut Upadhyay 09/19/2022



#68
Hexachlorobenzene
Concen: 42.270 ng
RT: 10.792 min Scan# 1480
Delta R.T. 0.000 min
Lab File: BF130292.D
Acq: 16 Sep 2022 16:09


Tgt Ion:284 Resp: 184491
Ion Ratio Lower Upper
284 100
142 34.6 27.2 40.8
249 30.2 23.8 35.8





#69
 Atrazine
 Concen: 40.635 ng
 RT: 10.892 min Scan# 1465
 Delta R.T. 0.000 min
 Lab File: BF130292.D
 Acq: 16 Sep 2022 16:09




Instrument :
 BNA_F
 ClientSampleId :
 SSTDCCC040




Tgt Ion:200 Resp: 138503
 Ion Ratio Lower Upper
 200 100
 173 27.3 6.2 46.2
 215 44.5 22.5 62.5

Manual Integrations
 APPROVED



Reviewed By :Christian Giraldo 09/19/2022
 Supervised By :Jagrut Upadhyay 09/19/2022




#70
 Pentachlorophenol
 Concen: 43.586 ng
 RT: 10.986 min Scan# 1513
 Delta R.T. 0.000 min
 Lab File: BF130292.D
 Acq: 16 Sep 2022 16:09



Tgt Ion:266 Resp: 102098
 Ion Ratio Lower Upper
 266 100
 268 66.6 50.5 75.7
 264 62.6 50.5 75.7









#71
Phenanthrene
Concen: 40.210 ng
RT: 11.210 min Scan# 11
Delta R.T. 0.000 min
Lab File: BF130292.D
Acq: 16 Sep 2022 16:09

Instrument :
BNA_F
ClientSampleId :
SSTDCCC040

Tgt Ion:178 Resp: 787945
Ion Ratio Lower Upper
178 100
176 19.3 15.4 23.0
179 14.9 12.4 18.6




Manual Integrations
APPROVED


Reviewed By :Christian Giraldo 09/19/2022
Supervised By :Jagrut Upadhyay 09/19/2022



#72
Anthracene
Concen: 40.664 ng
RT: 11.257 min Scan# 1559
Delta R.T. 0.000 min
Lab File: BF130292.D
Acq: 16 Sep 2022 16:09


Tgt Ion:178 Resp: 768974
Ion Ratio Lower Upper
178 100
176 18.5 15.0 22.6
179 15.6 12.7 19.1





#73
 Carbazole
 Concen: 39.785 ng
 RT: 11.416 min Scan# 1553
 Delta R.T. 0.000 min
 Lab File: BF130292.D
 Acq: 16 Sep 2022 16:09




Instrument :
 BNA_F
 ClientSampleId :
 SSTDCCC040




Tgt Ion:167 Resp: 678976
 Ion Ratio Lower Upper
 167 100
 166 21.5 17.4 26.2
 139 14.4 11.6 17.4

Manual Integrations
 APPROVED



Reviewed By :Christian Giraldo 09/19/2022
 Supervised By :Jagrut Upadhyay 09/19/2022




#74
 Di-n-butylphthalate
 Concen: 41.807 ng
 RT: 11.751 min Scan# 1643
 Delta R.T. 0.000 min
 Lab File: BF130292.D
 Acq: 16 Sep 2022 16:09




Tgt Ion:149 Resp: 860385
 Ion Ratio Lower Upper
 149 100
 150 9.3 7.4 11.2
 104 5.8 4.8 7.2





#75
Fluoranthene
Concen: 40.977 ng
RT: 12.392 min Scan# 1719
Delta R.T. 0.000 min
Lab File: BF130292.D
Acq: 16 Sep 2022 16:09




Instrument :
BNA_F
ClientSampleId :
SSTDCCC040




Tgt Ion:202 Resp: 775912
Ion Ratio Lower Upper
202 100
101 8.9 0.0 31.8
203 17.3 0.0 37.8

Manual Integrations
APPROVED



Reviewed By :Christian Giraldo 09/19/2022
Supervised By :Jagrut Upadhyay 09/19/2022




#76
Chrysene-d12
Concen: 20.000 ng
RT: 13.810 min Scan# 1993
Delta R.T. 0.000 min
Lab File: BF130292.D
Acq: 16 Sep 2022 16:09




Tgt Ion:240 Resp: 205901
Ion Ratio Lower Upper
240 100
120 10.4 8.7 13.1
236 25.1 20.9 31.3





#77
Benzidine
Concen: 33.466 ng
RT: 12.515 min Scan# 1773
Delta R.T. 0.000 min
Lab File: BF130292.D
Acq: 16 Sep 2022 16:09




Instrument :
BNA_F
ClientSampleId :
SSTDCCC040




Tgt Ion:184 Resp: 200031
Ion Ratio Lower Upper
184 100
185 14.3 11.8 17.6
183 11.1 9.1 13.7

Manual Integrations
APPROVED



Reviewed By :Christian Giraldo 09/19/2022
Supervised By :Jagrut Upadhyay 09/19/2022




#78
Pyrene
Concen: 42.600 ng
RT: 12.616 min Scan# 1790
Delta R.T. 0.000 min
Lab File: BF130292.D
Acq: 16 Sep 2022 16:09




Tgt Ion:202 Resp: 767321
Ion Ratio Lower Upper
202 100
200 21.2 17.2 25.8
203 17.8 14.3 21.5





#79
 Terphenyl-d14
 Concen: 86.605 ng
 RT: 12.768 min Scan# 1816
 Delta R.T. 0.000 min
 Lab File: BF130292.D
 Acq: 16 Sep 2022 16:09




Instrument :
 BNA_F
 ClientSampleId :
 SSTDCCC040




Tgt Ion:244 Resp: 1076498
 Ion Ratio Lower Upper
 244 100
 212 7.3 5.9 8.9
 122 10.6 8.5 12.7

Manual Integrations
 APPROVED



Reviewed By :Christian Giraldo 09/19/2022
 Supervised By :Jagrut Upadhyay 09/19/2022




#80
 Butylbenzylphthalate
 Concen: 44.419 ng
 RT: 13.245 min Scan# 1897
 Delta R.T. 0.000 min
 Lab File: BF130292.D
 Acq: 16 Sep 2022 16:09




Tgt Ion:149 Resp: 296399
 Ion Ratio Lower Upper
 149 100
 91 70.1 59.8 89.8
 206 18.6 13.9 20.9





#81
Benzo(a)anthracene
Concen: 40.990 ng
RT: 13.798 min Scan# 1959
Delta R.T. 0.000 min
Lab File: BF130292.D
Acq: 16 Sep 2022 16:09




Instrument :
BNA_F
ClientSampleId :
SSTDCCC040




Tgt Ion:228 Resp: 561452
Ion Ratio Lower Upper
228 100
226 27.6 21.9 32.9
229 19.7 16.1 24.1

Manual Integrations
APPROVED



Reviewed By :Christian Giraldo 09/19/2022
Supervised By :Jagrut Upadhyay 09/19/2022




#82
3,3'-Dichlorobenzidine
Concen: 46.262 ng
RT: 13.768 min Scan# 1986
Delta R.T. 0.000 min
Lab File: BF130292.D
Acq: 16 Sep 2022 16:09



Tgt Ion:252 Resp: 172494
Ion Ratio Lower Upper
252 100
254 63.5 51.1 76.7
126 12.6 10.4 15.6





#83

Chrysene

Concen: 41.248 ng

RT: 13.839 min Scan# 1998

Delta R.T. 0.000 min

Lab File: BF130292.D

Acq: 16 Sep 2022 16:09

Instrument :

BNA_F

ClientSampleId :





SSTDCCC040

Tgt Ion:228	Resp: 53646
Ion Ratio	Lower Upper
228 100	
226 29.5	23.8 35.8
229 19.3	15.3 22.9

Manual Integrations

APPROVED

Reviewed By :Christian Giraldo 09/19/2022
Supervised By :Jagrut Upadhyay 09/19/2022



#84

Bis(2-ethylhexyl)phthalate

Concen: 43.461 ng




RT: 13.804 min Scan# 1992


Delta R.T. 0.000 min

Lab File: BF130292.D

Acq: 16 Sep 2022 16:09


Tgt Ion:149	Resp: 332184
Ion Ratio	Lower Upper
149 100	
167 25.9	20.5 30.7
279 4.2	2.9 4.3





#85
Di-n-octyl phthalate
Concen: 43.653 ng
RT: 14.415 min Scan# 2063
Delta R.T. 0.000 min
Lab File: BF130292.D
Acq: 16 Sep 2022 16:09




Instrument :
BNA_F
ClientSampleId :
SSTDCCC040




Tgt Ion:149 Resp: 510328
Ion Ratio Lower Upper
149 100
167 1.5 1.2 1.8
43 9.5 8.7 13.1

Manual Integrations
APPROVED



Reviewed By :Christian Giraldo 09/19/2022
Supervised By :Jagrut Upadhyay 09/19/2022




#86
Perylene-d12
Concen: 20.000 ng
RT: 15.198 min Scan# 2229
Delta R.T. 0.000 min
Lab File: BF130292.D
Acq: 16 Sep 2022 16:09



Tgt Ion:264 Resp: 170906
Ion Ratio Lower Upper
264 100
260 24.5 19.3 28.9
265 21.9 17.0 25.4






#87
Indeno(1,2,3-cd)pyrene
Concen: 43.524 ng
RT: 16.527 min Scan# 2423
Delta R.T. 0.000 min
Lab File: BF130292.D
Acq: 16 Sep 2022 16:09

Instrument :

BNA_F

ClientSampleId :

SSTDCCC040






Tgt Ion:276 Resp: 527494
Ion Ratio Lower Upper
276 100
138 21.8 18.5 27.7
277 25.4 20.2 30.4


Manual Integrations

APPROVED



Reviewed By :Christian Giraldo 09/19/2022
Supervised By :Jagrut Upadhyay 09/19/2022




#88
Benzo(b)fluoranthene
Concen: 40.588 ng
RT: 14.809 min Scan# 2163
Delta R.T. 0.000 min
Lab File: BF130292.D
Acq: 16 Sep 2022 16:09




Tgt Ion:252 Resp: 452756
Ion Ratio Lower Upper
252 100
253 21.8 17.7 26.5
125 8.4 8.1 12.1





#89
Benzo(k)fluoranthene
Concen: 40.738 ng
RT: 14.839 min Scan# 2135
Delta R.T. 0.000 min
Lab File: BF130292.D
Acq: 16 Sep 2022 16:09




Instrument :
BNA_F
ClientSampleId :
SSTDCCC040



Tgt Ion:252 Resp: 447020
Ion Ratio Lower Upper
252 100
253 22.1 17.4 26.2
125 8.0 7.9 11.9




Manual Integrations
APPROVED


Reviewed By :Christian Giraldo 09/19/2022
Supervised By :Jagrut Upadhyay 09/19/2022



#90
Benzo(a)pyrene
Concen: 41.565 ng
RT: 15.139 min Scan# 2219
Delta R.T. 0.000 min
Lab File: BF130292.D
Acq: 16 Sep 2022 16:09


Tgt Ion:252 Resp: 367326
Ion Ratio Lower Upper
252 100
253 21.6 17.8 26.6
125 10.8 9.0 13.6





#91
Dibenzo(a,h)anthracene
Concen: 44.058 ng
RT: 16.545 min Scan# 2427
Delta R.T. 0.000 min
Lab File: BF130292.D
Acq: 16 Sep 2022 16:09




Instrument :
BNA_F
ClientSampleId :
SSTDCCC040




Tgt Ion:278 Resp: 438040
Ion Ratio Lower Upper
278 100
139 13.7 11.8 17.8
279 23.9 19.4 29.0

Manual Integrations
APPROVED

Reviewed By :Christian Giraldo 09/19/2022
Supervised By :Jagrut Upadhyay 09/19/2022



#92
Benzo(g,h,i)perylene
Concen: 43.116 ng
RT: 16.927 min Scan# 2523
Delta R.T. 0.000 min
Lab File: BF130292.D
Acq: 16 Sep 2022 16:09



Tgt Ion:276 Resp: 431829
Ion Ratio Lower Upper
276 100
277 23.1 18.6 27.8
138 18.7 14.6 21.8

