

Data Path : Z:\SVOASRV\HPCHEM1\BNA F\DATA\BF091718\
 Data File : BF109056.D
 Acq On : 17 Sep 2018 18:51
 Operator : JU/SJ
 Sample : J4943-01 2X
 Misc :
 ALS Vial : 20 Sample Multiplier: 1

Instrument :
 BNA_F
 ClientSampleId :
 RBR200051

Manual Integrations
 APPROVED

Sohil
 9/18/2018 6:00:13 PM

Quant Time: Sep 18 02:35:26 2018
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA F\METHODS\8270-BF091418.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Fri Sep 14 13:26:52 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	6.72	152	125729	20.00	ng	0.00
21) Naphthalene-d8	8.00	136	434595	20.00	ng	0.00
38) Acenaphthene-d10	9.75	164	187584	20.00	ng	0.00
63) Phenanthrene-d10	11.23	188	361169	20.00	ng	0.00
75) Chrysene-d12	13.85	240	359953	20.00	ng	0.00
86) Perylene-d12	15.26	264	268748	20.00	ng	0.00
System Monitoring Compounds						
5) 2-Fluorophenol	5.32	112	284742	37.61	ng	0.00
7) Phenol-d6	6.35	99	345991	35.45	ng	-0.01
23) Nitrobenzene-d5	7.28	82	200476	25.87	ng	-0.01
41) 2,4,6-Tribromophenol	10.54	330	74142	36.42	ng	0.00
44) 2-Fluorobiphenyl	9.08	172	363022	30.32	ng	0.00
78) Terphenyl-d14	12.81	244	367506	24.20	ng	0.00
Target Compounds						
49) Dimethylphthalate	9.47	163	49882	3.721	ng	Qvalue # 98
74) Fluoranthene	12.44	202	90470	4.840	ng	96
77) Pyrene	12.66	202	78810	2.948	ng	97
82) Chrysene	13.88	228	53401	2.449	ng	96
83) Bis(2-ethylhexyl)phthalate	13.85	149	93630	6.475	ng	99
87) Benzo(b)fluoranthene	14.86	252	67119m	4.514	ng	
89) Benzo(a)pyrene	15.19	252	32915	2.493	ng	# 95
91) Benzo(g,h,i)perylene	17.01	276	23242	2.096	ng	# 91

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\SVOASRV\HPCHEM1\BNA F\DATA\BF091718\
Data File : BF109056.D
Acq On : 17 Sep 2018 18:51
Operator : JU/SJ
Sample : J4943-01 2X
Misc :
ALS Vial : 20 Sample Multiplier: 1

Instrument :
BNA_F
ClientSampleId :
RBR200051

Manual Integrations
APPROVED

Sohil
9/18/2018 6:00:13 PM

Quant Time: Sep 18 02:35:26 2018
Quant Method : Z:\SVOASRV\HPCHEM1\BNA F\METHODS\8270-BF091418.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Fri Sep 14 13:26:52 2018
Response via : Initial Calibration

