

Data Path : Z:\SVOASRV\HPCHEM1\BNA F\DATA\BF091719\
 Data File : BF117111.D
 Acq On : 17 Sep 2019 17:17
 Operator : HP/JU
 Sample : SSTDCCC040
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_F
 ClientSampleId :
 SSTDCCC040

Quant Time: Sep 18 03:51:10 2019
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA F\METHODS\8270-BF091319.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Fri Sep 13 16:23:17 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	6.82	152	75529	20.00	ng	0.00
21) Naphthalene-d8	8.10	136	308406	20.00	ng	0.00
39) Acenaphthene-d10	9.86	164	159482	20.00	ng	0.00
64) Phenanthrene-d10	11.33	188	260157	20.00	ng	0.00
76) Chrysene-d12	13.97	240	178575	20.00	ng	0.00
87) Perylene-d12	15.42	264	144115	20.00	ng	0.00

System Monitoring Compounds

5) 2-Fluorophenol	5.45	112	361111	74.64	ng	0.00
7) Phenol-d6	6.46	99	468899	75.00	ng	0.00
23) Nitrobenzene-d5	7.39	82	447079	76.17	ng	0.00
42) 2,4,6-Tribromophenol	10.65	330	102137	76.63	ng	0.00
45) 2-Fluorobiphenyl	9.17	172	784543	76.92	ng	0.00
79) Terphenyl-d14	12.92	244	756703	79.21	ng	0.00

Target Compounds

						Qvalue
2) 1,4-Dioxane	2.56	88	72566	35.847	ng	97
3) Pyridine	3.31	79	205193	37.033	ng	95
4) n-Nitrosodimethylamine	3.26	42	66084	34.754	ng	98
6) Aniline	6.49	93	300035	37.273	ng	92
8) 2-Chlorophenol	6.62	128	195467	37.453	ng	95
9) Benzaldehyde	6.37	77	129863	42.056	ng	99
10) Phenol	6.48	94	255606	37.085	ng	89
11) bis(2-Chloroethyl)ether	6.56	93	187380	36.990	ng	99
12) 1,3-Dichlorobenzene	6.77	146	219998	37.566	ng	97
13) 1,4-Dichlorobenzene	6.84	146	225476	37.727	ng	100
14) 1,2-Dichlorobenzene	7.00	146	213940	38.448	ng	99
15) Benzyl Alcohol	6.97	79	183103	38.191	ng	100
16) 2,2'-oxybis(1-Chloropropan	7.10	45	183682	36.701	ng	92
17) 2-Methylphenol	7.09	107	162744	37.207	ng	96
18) Hexachloroethane	7.33	117	86570	37.653	ng	99
19) n-Nitroso-di-n-propylamine	7.24	70	144153	37.865	ng	96
20) 3+4-Methylphenols	7.24	107	206346	37.874	ng	# 87
22) Acetophenone	7.23	105	301384	38.463	ng	# 98
24) Nitrobenzene	7.41	77	220465	38.034	ng	93
25) Isophorone	7.65	82	390427	37.567	ng	99
26) 2-Nitrophenol	7.72	139	100201	40.245	ng	98
27) 2,4-Dimethylphenol	7.76	122	150475	38.111	ng	99
28) bis(2-Chloroethoxy)methane	7.85	93	246389	38.480	ng	99
29) 2,4-Dichlorophenol	7.97	162	161232	38.972	ng	96
30) 1,2,4-Trichlorobenzene	8.05	180	171927	38.466	ng	97
31) Naphthalene	8.13	128	567403	38.254	ng	100
32) Benzoic acid	7.89	122	91856	33.172	ng	96
33) 4-Chloroaniline	8.17	127	251492	38.228	ng	98
34) Hexachlorobutadiene	8.25	225	99254	39.141	ng	97
35) Caprolactam	8.55	113	52081	37.192	ng	87
36) 4-Chloro-3-methylphenol	8.66	107	184064	38.156	ng	96
37) 2-Methylnaphthalene	8.82	142	383339	38.380	ng	100
38) 1-Methylnaphthalene	8.92	142	358265	38.298	ng	98
40) 1,2,4,5-Tetrachlorobenzene	8.99	216	158129	38.397	ng	99


Data Path : Z:\SVOASRV\HPCHEM1\BNA F\DATA\BF091719\
 Data File : BF117111.D
 Acq On : 17 Sep 2019 17:17
 Operator : HP/JU
 Sample : SSTDCCC040
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_F
 ClientSampleId :
 SSTDCCC040

Quant Time: Sep 18 03:51:10 2019
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA F\METHODS\8270-BF091319.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Fri Sep 13 16:23:17 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
41) Hexachlorocyclopentadiene	8.97	237	69854	42.457	ng	99
43) 2,4,6-Trichlorophenol	9.10	196	104667	37.451	ng	95
44) 2,4,5-Trichlorophenol	9.14	196	109690	36.735	ng	99
46) 1,1'-Biphenyl	9.27	154	475350	37.998	ng	100
47) 2-Chloronaphthalene	9.30	162	372422	37.722	ng	99
48) 2-Nitroaniline	9.40	65	124381	39.277	ng	94
49) Acenaphthylene	9.72	152	556307	37.613	ng	100
50) Dimethylphthalate	9.57	163	427228	37.834	ng	100
51) 2,6-Dinitrotoluene	9.64	165	90132	39.191	ng	93
52) Acenaphthene	9.89	154	329823	37.619	ng	98
53) 3-Nitroaniline	9.81	138	111654	38.479	ng	97
54) 2,4-Dinitrophenol	9.92	184	33933	39.422	ng	95
55) Dibenzofuran	10.06	168	490631	37.929	ng	98
56) 4-Nitrophenol	9.98	139	74504	39.617	ng	94
57) 2,4-Dinitrotoluene	10.05	165	124407	41.673	ng	95
58) Fluorene	10.40	166	398902	38.282	ng	100
59) 2,3,4,6-Tetrachlorophenol	10.18	232	84173	36.837	ng	96
60) Diethylphthalate	10.27	149	421705	37.959	ng	98
61) 4-Chlorophenyl-phenylether	10.39	204	176426	39.133	ng	99
62) 4-Nitroaniline	10.42	138	115253	38.193	ng	96
63) Azobenzene	10.55	77	423685	37.341	ng	97
65) 4,6-Dinitro-2-methylphenol	10.46	198	44433	38.996	ng	90
66) n-Nitrosodiphenylamine	10.51	169	343718	38.256	ng	100
67) 4-Bromophenyl-phenylether	10.88	248	102141	38.445	ng	96
68) Hexachlorobenzene	10.96	284	110954	38.182	ng	94
69) Atrazine	11.03	200	87350	37.312	ng	99
70) Pentachlorophenol	11.15	266	44064	32.349	ng	96
71) Phenanthrene	11.36	178	554104	37.498	ng	99
72) Anthracene	11.41	178	570744	37.833	ng	99
73) Carbazole	11.57	167	543590	37.961	ng	100
74) Di-n-butylphthalate	11.89	149	655741	38.489	ng	99
75) Fluoranthene	12.54	202	572836	37.728	ng	100
77) Benzidine	12.66	184	231602	40.262	ng	99
78) Pyrene	12.77	202	571740	38.157	ng	100
80) Butylbenzylphthalate	13.38	149	278275	39.304	ng	97
81) Benzo(a)anthracene	13.96	228	449424	37.669	ng	99
82) 3,3'-Dichlorobenzidine	13.92	252	151461	39.395	ng	98
83) Chrysene	14.00	228	445588	38.293	ng	98
84) Bis(2-ethylhexyl)phthalate	13.94	149	365988	40.092	ng	99
85) Di-n-octyl phthalate	14.56	149	583157	40.587	ng	99
86) Indeno(1,2,3-cd)pyrene	16.87	276	279458	32.918	ng	98
88) Benzo(b)fluoranthene	15.00	252	366924	42.032	ng	99
89) Benzo(k)fluoranthene	15.03	252	320774	38.791	ng	99
90) Benzo(a)pyrene	15.36	252	301743	39.390	ng	98
91) Dibenzo(a,h)anthracene	16.87	278	227959	37.355	ng	96
92) Benzo(g,h,i)perylene	17.30	276	218446	35.922	ng	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed




#1




1,4-Dichlorobenzene-d4
Concen: 20.000 ng
RT: 6.82 min Scan# 805
Delta R.T. -0.00 min
Lab File: BF117111.D
Acq: 17 Sep 2019 17:17

Instrument :
BNA_F
ClientSampleId :
SSTDCCC040

Tot Ion:152 Resp: 75529
Ion Ratio Lower Upper
152 100
150 159.8 127.5 191.3
115 68.2 55.0 82.4




Abundance Ion 152.00 (151.70 to 152.70): E
Ion 150.00 (149.70 to 150.70): E
Ion 115.00 (114.70 to 115.70): E





#2


1,4-Dioxane
Concen: 35.847 ng
RT: 2.56 min Scan# 81
Delta R.T. -0.00 min
Lab File: BF117111.D
Acq: 17 Sep 2019 17:17

Tgt Ion: 88 Resp: 72566
Ion Ratio Lower Upper
88 100
58 58.3 49.3 73.9
43 23.7 18.8 28.2



Abundance Ion 88.00 (87.70 to 88.70): BF1
Ion 58.00 (57.70 to 58.70): BF1
Ion 43.00 (42.70 to 43.70): BF1





#3

Pvridine

Concen: 37.033 ng

RT: 3.31 min Scan# 208

Delta R.T. 0.01 min

Lab File: BF117111.D

Acq: 17 Sep 2019 17:17

Instrument :

BNA_F

ClientSampleId :

SSTDCCC040


Tot Ion: 79 Resp: 205193

Ion Ratio Lower Upper




79 100

52 49.9 43.5 65.3

51 25.8 21.8 32.6



Abundance Ion 79.00 (78.70 to 79.70): BF1
Ion 52.00 (51.70 to 52.70): BF1
Ion 51.00 (50.70 to 51.70): BF1



#4

n-Nitrosodimethylamine

Concen: 34.754 ng

RT: 3.26 min Scan# 199

Delta R.T. 0.01 min

Lab File: BF117111.D

Acq: 17 Sep 2019 17:17


Tgt Ion: 42 Resp: 66084

Ion Ratio Lower Upper



42 100


74 155.7 127.1 190.7

44 4.8 4.3 6.5



Abundance Ion 42.00 (41.70 to 42.70): BF1
Ion 74.00 (73.70 to 74.70): BF1
Ion 44.00 (43.70 to 44.70): BF1





#5

2-Fluorophenol

Concen: 74.645 ng

RT: 5.45 min Scan# 571

Delta R.T. 0.01 min

Lab File: BF117111.D

Acq: 17 Sep 2019 17:17

Instrument :

BNA_F

ClientSampleId :

SSTDCCC040


Tot Ion: 112 Resp: 361111

Ion Ratio Lower Upper




112 100

64 55.6 45.7 68.5

63 30.9 24.9 37.3



Abundance Ion 112.00 (111.70 to 112.70): BF1
Ion 64.00 (63.70 to 64.70): BF1
Ion 63.00 (62.70 to 63.70): BF1



#6

Aniline

Concen: 37.273 ng

RT: 6.49 min Scan# 748

Delta R.T. -0.00 min

Lab File: BF117111.D

Acq: 17 Sep 2019 17:17


Tgt Ion: 93 Resp: 300035

Ion Ratio Lower Upper



93 100


66 45.8 32.6 49.0

65 23.9 16.7 25.1



Abundance Ion 93.00 (92.70 to 93.70): BF1
Ion 66.00 (65.70 to 66.70): BF1
Ion 65.00 (64.70 to 65.70): BF1





#7

Phenol-d6

Concen: 74.999 ng

RT: 6.46 min Scan# 744

Delta R.T. 0.01 min

Lab File: BF117111.D

Acq: 17 Sep 2019 17:17

Instrument :

BNA_F

ClientSampleId :

SSTDCCC040


Tot Ion: 99 Resp: 468899

Ion Ratio Lower Upper

99 100

42 16.0 12.2 18.2




71 37.2 29.8 44.8



Abundance Ion 99.00 (98.70 to 99.70): BF1

Ion 42.00 (41.70 to 42.70): BF1

Ion 71.00 (70.70 to 71.70): BF1



#8

2-Chlorophenol

Concen: 37.453 ng

RT: 6.62 min Scan# 770

Delta R.T. 0.01 min

Lab File: BF117111.D

Acq: 17 Sep 2019 17:17


Tgt Ion: 128 Resp: 195467

Ion Ratio Lower Upper

128 100

130 32.8 11.8 51.8



64 40.5 25.0 65.0




Abundance Ion 128.00 (127.70 to 128.70): E

Ion 130.00 (129.70 to 130.70): E

Ion 64.00 (63.70 to 64.70): BF1









#9
Benzaldehyde
Concen: 42.056 ng
RT: 6.37 min Scan# 729
Delta R.T. 0.00 min
Lab File: BF117111.D
Acq: 17 Sep 2019 17:17




Instrument :
BNA_F
ClientSampleId :
SSTDCCC040


Tot Ion: 77 Resp: 129863
Ion Ratio Lower Upper
77 100
105 101.1 81.8 121.8
106 97.0 77.3 117.3



#10
Phenol
Concen: 37.085 ng
RT: 6.48 min Scan# 747
Delta R.T. 0.01 min
Lab File: BF117111.D
Acq: 17 Sep 2019 17:17

Tgt Ion: 94 Resp: 255606
Ion Ratio Lower Upper
94 100
65 36.9 12.5 52.5
66 57.4 28.9 68.9





#11

bis(2-Chloroethyl)ether

Concen: 36.990 ng

RT: 6.56 min Scan# 760

Delta R.T. -0.00 min

Lab File: BF117111.D

Acq: 17 Sep 2019 17:17

Instrument :

BNA_F

ClientSampleId :

SSTDCCC040


Tot Ion: 93 Resp: 187380

Ion Ratio Lower Upper

93 100

63 68.2 49.3 89.3




95 31.7 11.5 51.5



Abundance Ion 93.00 (92.70 to 93.70): BF1

Ion 63.00 (62.70 to 63.70): BF1

Ion 95.00 (94.70 to 95.70): BF1



#12

1,3-Dichlorobenzene

Concen: 37.566 ng

RT: 6.77 min Scan# 796

Delta R.T. 0.01 min

Lab File: BF117111.D

Acq: 17 Sep 2019 17:17


Tgt Ion: 146 Resp: 219998

Ion Ratio Lower Upper

146 100

148 63.8 50.8 76.2



75 31.5 28.7 43.1




Abundance Ion 146.00 (145.70 to 146.70): E

Ion 148.00 (147.70 to 148.70): E

Ion 75.00 (74.70 to 75.70): BF1









#13
 1,4-Dichlorobenzene
 Concen: 37.727 ng
 RT: 6.84 min Scan# 808
 Delta R.T. -0.00 min
 Lab File: BF117111.D
 Acq: 17 Sep 2019 17:17




Instrument :
 BNA_F
 ClientSampleId :
 SSTDCCC040




Tot Ion:146 Resp: 225476
 Ion Ratio Lower Upper
 146 100
 148 63.5 50.6 76.0
 111 44.3 35.0 52.6



#14
 1,2-Dichlorobenzene
 Concen: 38.448 ng
 RT: 7.00 min Scan# 835
 Delta R.T. -0.00 min
 Lab File: BF117111.D
 Acq: 17 Sep 2019 17:17

Tgt Ion:146 Resp: 213940
 Ion Ratio Lower Upper
 146 100
 148 64.4 52.6 78.8
 111 44.7 36.2 54.2





#15

Benzyl Alcohol

Concen: 38.191 ng

RT: 6.97 min Scan# 830

Delta R.T. 0.01 min

Lab File: BF117111.D

Acq: 17 Sep 2019 17:17

Instrument :

BNA_F

ClientSampleId :

SSTDCCC040





Tot Ion: 79 Resp: 183103

Ion Ratio Lower Upper

79 100

108 74.0 59.4 89.0

77 62.8 50.2 75.2



#16

2,2'-oxybis(1-Chloropropane)

Concen: 36.701 ng

RT: 7.10 min Scan# 852

Delta R.T. -0.00 min

Lab File: BF117111.D

Acq: 17 Sep 2019 17:17


Tgt Ion: 45 Resp: 183682


Ion Ratio Lower Upper

45 100

77 26.8 2.7 42.7

79 21.8 0.0 38.9









#17
2-Methylphenol
Concen: 37.207 ng
RT: 7.09 min Scan# 850
Delta R.T. 0.01 min
Lab File: BF117111.D
Acq: 17 Sep 2019 17:17




Instrument :
BNA_F
ClientSampleId :
SSTDCCC040


Tot Ion:	107	Resp:	162744
Ion	Ratio	Lower	Upper
107	100		
108	111.8	87.7	131.5
77	51.2	38.6	58.0
79	50.8	36.8	55.2



#18
Hexachloroethane
Concen: 37.653 ng
RT: 7.33 min Scan# 892
Delta R.T. -0.00 min
Lab File: BF117111.D
Acq: 17 Sep 2019 17:17

Tgt Ion:	117	Resp:	86570
Ion	Ratio	Lower	Upper
117	100		
119	95.7	75.8	113.8
201	77.3	60.9	91.3









#19
n-Nitroso-di-n-propylamine
Concen: 37.865 ng
RT: 7.24 min Scan# 876
Delta R.T. -0.00 min
Lab File: BF117111.D
Acq: 17 Sep 2019 17:17




Instrument :
BNA_F
ClientSampleId :
SSTDCCC040


Tot Ion:	70	Resp:	144153
Ion	Ratio	Lower	Upper
70	100		
42	46.4	34.0	51.0
101	9.9	7.8	11.6
130	21.6	17.8	26.8



#20
3+4-Methylphenols
Concen: 37.874 ng
RT: 7.24 min Scan# 876
Delta R.T. 0.01 min
Lab File: BF117111.D
Acq: 17 Sep 2019 17:17

Tgt Ion:	107	Resp:	206346
Ion	Ratio	Lower	Upper
107	100		
108	92.5	72.6	112.6
77	121.6	132.1	172.1#
79	32.2	12.1	52.1





#21

Naphthalene-d8

Concen: 20.000 ng

RT: 8.10 min Scan# 1023

Delta R.T. -0.00 min

Lab File: BF117111.D

Acq: 17 Sep 2019 17:17

Instrument :

BNA_F

ClientSampleId :

SSTDCCC040

Tot Ion:136 Resp: 308406



Ion Ratio Lower Upper

136 100

137 11.4 9.0 13.4

54 5.9 4.5 6.7

68 4.9 4.0 6.0





Abundance Ion 136.00 (135.70 to 136.70): E

Ion 137.00 (136.70 to 137.70): E

Ion 54.00 (53.70 to 54.70): BF1

Ion 68.00 (67.70 to 68.70): BF1



#22

Acetophenone

Concen: 38.463 ng

RT: 7.23 min Scan# 875

Delta R.T. -0.00 min

Lab File: BF117111.D

Acq: 17 Sep 2019 17:17

Tgt Ion:105 Resp: 301384



Ion Ratio Lower Upper

105 100

71 5.4 5.7 8.5#

51 26.8 22.1 33.1

120 20.4 17.1 25.7





Abundance Ion 105.00 (104.70 to 105.70): E

Ion 71.00 (70.70 to 71.70): BF1

Ion 51.00 (50.70 to 51.70): BF1

Ion 120.00 (119.70 to 120.70): E









#23
 Nitrobenzene-d5
 Concen: 76.168 ng
 RT: 7.39 min Scan# 902
 Delta R.T. 0.01 min
 Lab File: BF117111.D
 Acq: 17 Sep 2019 17:17

Instrument :
 BNA_F
 ClientSampleId :
 SSTDCCC040

Tot Ion	82	Resp	447079
Ion Ratio	Lower	Upper	
82	100		
128	44.1	32.3	48.5
54	40.4	33.3	49.9




Abundance Ion 82.00 (81.70 to 82.70): BF1
 Ion 128.00 (127.70 to 128.70): BF1
 Ion 54.00 (53.70 to 54.70): BF1






#24
 Nitrobenzene
 Concen: 38.034 ng
 RT: 7.41 min Scan# 905
 Delta R.T. 0.01 min
 Lab File: BF117111.D
 Acq: 17 Sep 2019 17:17

Tgt Ion	77	Resp	220465
Ion Ratio	Lower	Upper	
77	100		
123	47.5	33.5	50.3
65	13.1	10.5	15.7



Abundance Ion 77.00 (76.70 to 77.70): BF1
 Ion 123.00 (122.70 to 123.70): BF1
 Ion 65.00 (64.70 to 65.70): BF1





#25

Isophorone

Concen: 37.567 ng

RT: 7.65 min Scan# 945

Delta R.T. -0.00 min

Lab File: BF117111.D

Acq: 17 Sep 2019 17:17

Instrument :

BNA_F

ClientSampleId :

SSTDCCC040





Tot Ion: 82 Resp: 390427

Ion Ratio Lower Upper

82 100

95 7.1 5.8 8.6

138 16.6 13.5 20.3



#26

2-Nitrophenol

Concen: 40.245 ng

RT: 7.72 min Scan# 958

Delta R.T. -0.00 min

Lab File: BF117111.D

Acq: 17 Sep 2019 17:17




Tgt Ion: 139 Resp: 100201


Ion Ratio Lower Upper

139 100

109 30.8 24.6 36.8

65 53.7 40.9 61.3





#27

2,4-Dimethylphenol

Concen: 38.111 ng

RT: 7.76 min Scan# 965

Delta R.T. 0.01 min

Lab File: BF117111.D

Acq: 17 Sep 2019 17:17

Instrument :

BNA_F

ClientSampleId :

SSTDCCC040


Tot Ion:122 Resp: 150475

Ion Ratio Lower Upper

122 100

107 121.7 98.9 148.3




121 56.8 45.3 67.9



Abundance Ion 122.00 (121.70 to 122.70): E

Ion 107.00 (106.70 to 107.70): E

Ion 121.00 (120.70 to 121.70): E



#28

bis(2-Chloroethoxy)methane

Concen: 38.480 ng

RT: 7.85 min Scan# 980

Delta R.T. -0.00 min

Lab File: BF117111.D

Acq: 17 Sep 2019 17:17


Tgt Ion: 93 Resp: 246389

Ion Ratio Lower Upper

93 100

95 33.1 26.2 39.4



123 11.4 9.0 13.6




Abundance Ion 93.00 (92.70 to 93.70): BF1

Ion 95.00 (94.70 to 95.70): BF1

Ion 123.00 (122.70 to 123.70): E






#29
 2,4-Dichlorophenol
 Concen: 38.972 ng
 RT: 7.97 min Scan# 1000
 Delta R.T. 0.01 min
 Lab File: BF117111.D
 Acq: 17 Sep 2019 17:17

Instrument :
 BNA_F
 ClientSampleId :
 SSTDCCC040

Tot Ion	Ratio	Lower	Upper
162	100		
164	64.6	43.8	83.8
98	37.5	22.1	62.1






Abundance

Ion 162.00 (161.70 to 162.70): E


Ion 164.00 (163.70 to 164.70): E

Ion 98.00 (97.70 to 98.70): BF1



#30
 1,2,4-Trichlorobenzene
 Concen: 38.466 ng
 RT: 8.05 min Scan# 1013
 Delta R.T. -0.00 min
 Lab File: BF117111.D
 Acq: 17 Sep 2019 17:17

Tgt Ion	Ratio	Lower	Upper
180	100		
182	90.7	75.4	113.0
145	33.7	26.7	40.1






Abundance

Ion 180.00 (179.70 to 180.70): E

Ion 182.00 (181.70 to 182.70): E

Ion 145.00 (144.70 to 145.70): E









#31
Naphthalene
Concen: 38.254 ng
RT: 8.13 min Scan# 1027
Delta R.T. -0.00 min
Lab File: BF117111.D
Acq: 17 Sep 2019 17:17

Instrument :
BNA_F
ClientSampleId :
SSTDCCC040

Tot Ion:128 Resp: 567403
Ion Ratio Lower Upper
128 100
129 11.1 9.0 13.4
127 13.2 10.8 16.2




Abundance Ion 128.00 (127.70 to 128.70): E
Ion 129.00 (128.70 to 129.70): E
Ion 127.00 (126.70 to 127.70): E






#32
Benzoic acid
Concen: 33.172 ng
RT: 7.89 min Scan# 987
Delta R.T. 0.01 min
Lab File: BF117111.D
Acq: 17 Sep 2019 17:17

Tgt Ion:122 Resp: 91856
Ion Ratio Lower Upper
122 100
105 136.7 114.8 154.8
77 105.6 78.7 118.7



Abundance Ion 122.00 (121.70 to 122.70): E
Ion 105.00 (104.70 to 105.70): E
Ion 77.00 (76.70 to 77.70): BF1









#33
4-Chloroaniline
Concen: 38.228 ng
RT: 8.17 min Scan# 1035
Delta R.T. -0.00 min
Lab File: BF117111.D
Acq: 17 Sep 2019 17:17




Instrument :
BNA_F
ClientSampleId :
SSTDCCC040


Tot Ion:127	Resp:	251492
Ion Ratio	Lower	Upper
127	100	
129	32.0	26.2 39.2
65	30.9	23.5 35.3
92	20.5	15.9 23.9



#34
Hexachlorobutadiene
Concen: 39.141 ng
RT: 8.25 min Scan# 1047
Delta R.T. -0.00 min
Lab File: BF117111.D
Acq: 17 Sep 2019 17:17

Tgt Ion:225	Resp:	99254
Ion Ratio	Lower	Upper
225	100	
223	63.5	49.1 73.7
227	63.2	52.2 78.2









#35
 Caprolactam
 Concen: 37.192 ng
 RT: 8.55 min Scan# 1099
 Delta R.T. 0.01 min
 Lab File: BF117111.D
 Acq: 17 Sep 2019 17:17

Instrument :
 BNA_F
 ClientSampleId :
 SSTDCCC040

Tot Ion:113 Resp: 52081
 Ion Ratio Lower Upper
 113 100
 55 121.4 120.6 160.6
 56 98.6 88.8 128.8




Abundance Ion 113.00 (112.70 to 113.70): E
 Ion 55.00 (54.70 to 55.70): BF1
 Ion 56.00 (55.70 to 56.70): BF1






#36
 4-Chloro-3-methylphenol
 Concen: 38.156 ng
 RT: 8.66 min Scan# 1117
 Delta R.T. -0.00 min
 Lab File: BF117111.D
 Acq: 17 Sep 2019 17:17

Tgt Ion:107 Resp: 184064
 Ion Ratio Lower Upper
 107 100
 144 23.4 19.2 28.8
 142 70.8 59.6 89.4



Abundance Ion 107.00 (106.70 to 107.70): E
 Ion 144.00 (143.70 to 144.70): E
 Ion 142.00 (141.70 to 142.70): E






#37
2-Methylnaphthalene
Concen: 38.380 ng
RT: 8.82 min Scan# 1144
Delta R.T. -0.00 min
Lab File: BF117111.D
Acq: 17 Sep 2019 17:17

Instrument :
BNA_F
ClientSampleId :
SSTDCCC040

Tot Ion	Ratio	Lower	Upper
142	100		
141	88.6	70.7	106.1
115	36.1	29.4	44.0






Abundance

Ion 142.00 (141.70 to 142.70): E


Ion 141.00 (140.70 to 141.70): E

Ion 115.00 (114.70 to 115.70): E



#38
1-Methylnaphthalene
Concen: 38.298 ng
RT: 8.92 min Scan# 1161
Delta R.T. -0.00 min
Lab File: BF117111.D
Acq: 17 Sep 2019 17:17

Tgt Ion	Ratio	Lower	Upper
142	100		
141	92.2	71.9	107.9
116	3.9	3.2	4.8






Abundance

Ion 142.00 (141.70 to 142.70): E

Ion 141.00 (140.70 to 141.70): E

Ion 116.00 (115.70 to 116.70): E









#39
Acenaphthene-d10
Concen: 20.000 ng
RT: 9.86 min Scan# 1321
Delta R.T. -0.00 min
Lab File: BF117111.D
Acq: 17 Sep 2019 17:17

Instrument :
BNA_F
ClientSampleId :
SSTDCCC040

Tot Ion:164 Resp: 159482
Ion Ratio Lower Upper
164 100
162 104.0 83.4 125.2
160 44.7 36.4 54.6




Abundance Ion 164.00 (163.70 to 164.70): E
Ion 162.00 (161.70 to 162.70): E
Ion 160.00 (159.70 to 160.70): E






#40
1,2,4,5-Tetrachlorobenzene
Concen: 38.397 ng
RT: 8.99 min Scan# 1173
Delta R.T. -0.00 min
Lab File: BF117111.D
Acq: 17 Sep 2019 17:17

Tgt Ion:216 Resp: 158129
Ion Ratio Lower Upper
216 100
214 78.2 63.6 95.4
179 23.9 18.3 27.5
108 22.4 17.8 26.6



Abundance Ion 216.00 (215.70 to 216.70): E
Ion 214.00 (213.70 to 214.70): E
Ion 179.00 (178.70 to 179.70): E
Ion 108.00 (107.70 to 108.70): E





#41

Hexachlorocyclopentadiene

Concen: 42.457 ng

RT: 8.97 min Scan# 1170

Delta R.T. 0.00 min

Lab File: BF117111.D

Acq: 17 Sep 2019 17:17

Instrument :

BNA_F

ClientSampleId :

SSTDCCC040


Tot Ion:237 Resp: 69854

Ion Ratio Lower Upper




237 100

235 62.3 43.2 83.2

272 12.9 0.0 31.9



Abundance Ion 236.80 (236.50 to 237.50): E
Ion 234.90 (234.60 to 235.60): E
Ion 271.75 (271.45 to 272.45): E



#42

2,4,6-Tribromophenol

Concen: 76.627 ng

RT: 10.65 min Scan# 1455

Delta R.T. -0.00 min

Lab File: BF117111.D

Acq: 17 Sep 2019 17:17


Tgt Ion:330 Resp: 102137

Ion Ratio Lower Upper



330 100


332 98.3 78.0 117.0

141 44.4 35.8 53.6



Abundance Ion 329.80 (329.50 to 330.50): E
Ion 331.80 (331.50 to 332.50): E
Ion 141.00 (140.70 to 141.70): E






#43
 2,4,6-Trichlorophenol
 Concen: 37.451 ng
 RT: 9.10 min Scan# 1192
 Delta R.T. 0.01 min
 Lab File: BF117111.D
 Acq: 17 Sep 2019 17:17

Instrument :
 BNA_F
 ClientSampleId :
 SSTDCCC040

Tot Ion	Ratio	Lower	Upper
196	100		
198	99.9	75.7	113.5
200	32.0	23.3	34.9






Abundance

Ion 196.00 (195.70 to 196.70): E

Ion 198.00 (197.70 to 198.70): E

Ion 200.00 (199.70 to 200.70): E



#44
 2,4,5-Trichlorophenol
 Concen: 36.735 ng
 RT: 9.14 min Scan# 1199
 Delta R.T. 0.01 min
 Lab File: BF117111.D
 Acq: 17 Sep 2019 17:17

Tgt Ion	Ratio	Lower	Upper
196	100		
198	93.2	74.6	112.0
97	60.3	49.5	74.3
132	33.7	27.4	41.2




Abundance


Ion 196.00 (195.70 to 196.70): E

Ion 198.00 (197.70 to 198.70): E

Ion 97.00 (96.70 to 97.70): BF1

Ion 132.00 (131.70 to 132.70): E






#45
2-Fluorobiphenyl
Concen: 76.917 ng
RT: 9.17 min Scan# 1205
Delta R.T. -0.00 min
Lab File: BF117111.D
Acq: 17 Sep 2019 17:17

Instrument :
BNA_F
ClientSampleId :
SSTDCCC040

Tot Ion	Ratio	Lower	Upper
172	100		
171	35.7	28.2	42.2
170	23.5	18.9	28.3






Abundance

Ion 172.00 (171.70 to 172.70): E


Ion 171.00 (170.70 to 171.70): E

Ion 170.00 (169.70 to 170.70): E



#46
1,1'-Biphenyl
Concen: 37.998 ng
RT: 9.27 min Scan# 1222
Delta R.T. -0.00 min
Lab File: BF117111.D
Acq: 17 Sep 2019 17:17

Tgt Ion	Ratio	Lower	Upper
154	100		
153	39.9	20.0	60.0
76	13.0	0.0	32.7






Abundance

Ion 154.00 (153.70 to 154.70): E

Ion 153.00 (152.70 to 153.70): E

Ion 76.00 (75.70 to 76.70): BF1









#47
2-Chloronaphthalene
Concen: 37.722 ng
RT: 9.30 min Scan# 1227
Delta R.T. -0.00 min
Lab File: BF117111.D
Acq: 17 Sep 2019 17:17




Instrument :
BNA_F
ClientSampleId :
SSTDCCC040


Tot Ion	Ratio	Lower	Upper
162	100		
127	42.0	34.1	51.1
164	32.6	25.9	38.9



#48
2-Nitroaniline
Concen: 39.277 ng
RT: 9.40 min Scan# 1243
Delta R.T. -0.00 min
Lab File: BF117111.D
Acq: 17 Sep 2019 17:17

Tgt Ion	Ratio	Lower	Upper
65	100		
92	70.5	57.5	86.3
138	104.5	90.6	136.0





#49

Acenaphthylene

Concen: 37.613 ng

RT: 9.72 min Scan# 1297

Delta R.T. -0.00 min

Lab File: BF117111.D

Acq: 17 Sep 2019 17:17

Instrument :

BNA_F

ClientSampleId :

SSTDCCC040


Tot Ion:152 Resp: 556307

Ion Ratio Lower Upper

152 100

151 20.1 16.2 24.2




153 13.1 10.2 15.4



Abundance Ion 152.00 (151.70 to 152.70): E

Ion 151.00 (150.70 to 151.70): E

Ion 153.00 (152.70 to 153.70): E



#50

Dimethylphthalate

Concen: 37.834 ng

RT: 9.57 min Scan# 1273

Delta R.T. -0.00 min

Lab File: BF117111.D

Acq: 17 Sep 2019 17:17


Tgt Ion:163 Resp: 427228

Ion Ratio Lower Upper

163 100

194 3.6 2.9 4.3



164 10.1 8.2 12.4




Abundance Ion 163.00 (162.70 to 163.70): E

Ion 194.00 (193.70 to 194.70): E

Ion 164.00 (163.70 to 164.70): E





#51

2,6-Dinitrotoluene

Concen: 39.191 ng

RT: 9.64 min Scan# 1284

Delta R.T. -0.00 min

Lab File: BF117111.D

Acq: 17 Sep 2019 17:17

Instrument :

BNA_F

ClientSampleId :

SSTDCCC040


Tot Ion:165 Resp: 90132

Ion Ratio Lower Upper

165 100

63 58.0 41.8 62.8




89 62.2 46.5 69.7



Abundance Ion 165.00 (164.70 to 165.70): E

Ion 63.00 (62.70 to 63.70): BF1

Ion 89.00 (88.70 to 89.70): BF1



#52

Acenaphthene

Concen: 37.619 ng

RT: 9.89 min Scan# 1326

Delta R.T. -0.00 min

Lab File: BF117111.D

Acq: 17 Sep 2019 17:17


Tgt Ion:154 Resp: 329823

Ion Ratio Lower Upper

154 100

153 109.6 89.0 133.6



152 53.7 41.9 62.9




Abundance Ion 154.00 (153.70 to 154.70): E

Ion 153.00 (152.70 to 153.70): E

Ion 152.00 (151.70 to 152.70): E









#53
3-Nitroaniline
Concen: 38.479 ng
RT: 9.81 min Scan# 1313
Delta R.T. -0.00 min
Lab File: BF117111.D
Acq: 17 Sep 2019 17:17

Instrument :
BNA_F
ClientSampleId :
SSTDCCC040

Tot Ion:138 Resp: 111654
Ion Ratio Lower Upper
138 100
108 11.5 9.5 14.3
92 121.3 94.1 141.1




Abundance Ion 138.00 (137.70 to 138.70): E
Ion 108.00 (107.70 to 108.70): E
Ion 92.00 (91.70 to 92.70): BF1






#54
2,4-Dinitrophenol
Concen: 39.422 ng
RT: 9.92 min Scan# 1332
Delta R.T. 0.01 min
Lab File: BF117111.D
Acq: 17 Sep 2019 17:17

Tgt Ion:184 Resp: 33933
Ion Ratio Lower Upper
184 100
63 84.7 63.8 95.6
154 79.1 60.4 90.6



Abundance Ion 184.00 (183.70 to 184.70): E
Ion 63.00 (62.70 to 63.70): BF1
Ion 154.00 (153.70 to 154.70): E





#55

Dibenzenofuran

Concen: 37.929 ng

RT: 10.06 min Scan# 1355

Delta R.T. -0.01 min

Lab File: BF117111.D

Acq: 17 Sep 2019 17:17

Instrument :

BNA_F

ClientSampleId :

SSTDCCC040


Tot Ion:168 Resp: 490631

Ion Ratio Lower Upper

168 100

139 42.5 32.8 49.2




169 13.3 10.5 15.7



Abundance Ion 168.00 (167.70 to 168.70): E

Ion 139.00 (138.70 to 139.70): E

Ion 169.00 (168.70 to 169.70): E



#56

4-Nitrophenol

Concen: 39.617 ng

RT: 9.98 min Scan# 1342

Delta R.T. 0.01 min

Lab File: BF117111.D

Acq: 17 Sep 2019 17:17


Tgt Ion:139 Resp: 74504

Ion Ratio Lower Upper

139 100

109 75.0 56.4 96.4



65 92.4 82.5 122.5




Abundance Ion 139.00 (138.70 to 139.70): E

Ion 109.00 (108.70 to 109.70): E

Ion 65.00 (64.70 to 65.70): BF1









#57
 2,4-Dinitrotoluene
 Concen: 41.673 ng
 RT: 10.05 min Scan# 1353
 Delta R.T. -0.00 min
 Lab File: BF117111.D
 Acq: 17 Sep 2019 17:17




Instrument :
 BNA_F
 ClientSampleId :
 SSTDCCC040


Tot Ion:165	Resp:	124407
Ion Ratio	Lower	Upper
165	100	
63	47.3	35.0 52.4
89	78.5	66.1 99.1
182	2.2	2.0 3.0



#58
 Fluorene
 Concen: 38.282 ng
 RT: 10.40 min Scan# 1414
 Delta R.T. -0.00 min
 Lab File: BF117111.D
 Acq: 17 Sep 2019 17:17

Tgt Ion:166	Resp:	398902
Ion Ratio	Lower	Upper
166	100	
165	99.1	78.9 118.3
167	13.2	10.6 15.8





#59

2,3,4,6-Tetrachlorophenol

Concen: 36.837 ng

RT: 10.18 min Scan# 1376

Delta R.T. -0.00 min

Lab File: BF117111.D

Acq: 17 Sep 2019 17:17

Instrument :

BNA_F

ClientSampleId :

SSTDCCC040

Tot Ion:232 Resp: 84173



Ion Ratio Lower Upper

232 100

131 61.1 45.3 67.9

130 2.4 1.8 2.8

166 37.1 29.4 44.2





Abundance Ion 232.00 (231.70 to 232.70): E

Ion 131.00 (130.70 to 131.70): E

Ion 130.00 (129.70 to 130.70): E

Ion 166.00 (165.70 to 166.70): E



#60

Diethylphthalate

Concen: 37.959 ng

RT: 10.27 min Scan# 1391

Delta R.T. -0.00 min

Lab File: BF117111.D

Acq: 17 Sep 2019 17:17



Tgt Ion:149 Resp: 421705

Ion Ratio Lower Upper

149 100

177 20.2 15.2 22.8


150 12.6 9.9 14.9




Abundance Ion 149.00 (148.70 to 149.70): E

Ion 177.00 (176.70 to 177.70): E

Ion 150.00 (149.70 to 150.70): E









#61
4-Chlorophenyl-phenylether
Concen: 39.133 ng
RT: 10.39 min Scan# 1412
Delta R.T. -0.00 min
Lab File: BF117111.D
Acq: 17 Sep 2019 17:17




Instrument :
BNA_F
ClientSampleId :
SSTDCCC040


Tot Ion:204 Resp: 176426
Ion Ratio Lower Upper
204 100
206 32.6 26.0 39.0
141 71.1 58.2 87.2



#62
4-Nitroaniline
Concen: 38.193 ng
RT: 10.42 min Scan# 1417
Delta R.T. -0.00 min
Lab File: BF117111.D
Acq: 17 Sep 2019 17:17

Tgt Ion:138 Resp: 115253
Ion Ratio Lower Upper
138 100
92 54.9 32.9 72.9
108 82.4 58.3 98.3





#63

Azobenzene

Concen: 37.341 ng

RT: 10.55 min Scan# 1439

Delta R.T. -0.00 min

Lab File: BF117111.D

Acq: 17 Sep 2019 17:17

Instrument :

BNA_F

ClientSampleId :

SSTDCCC040

Tot Ion: 77 Resp: 423685



Ion Ratio Lower Upper

77 100

182 22.7 0.4 40.4

105 22.4 0.9 40.9

51 23.1 2.8 42.8





Abundance Ion 77.00 (76.70 to 77.70): BF1

Ion 182.00 (181.70 to 182.70): BF1

Ion 105.00 (104.70 to 105.70): BF1

Ion 51.00 (50.70 to 51.70): BF1



#64

Phenanthrene-d10

Concen: 20.000 ng

RT: 11.33 min Scan# 1572

Delta R.T. -0.00 min

Lab File: BF117111.D

Acq: 17 Sep 2019 17:17



Tgt Ion: 188 Resp: 260157

Ion Ratio Lower Upper

188 100

94 9.4 8.2 12.2


80 10.6 9.0 13.6




Abundance Ion 188.00 (187.70 to 188.70): BF1

Ion 94.00 (93.70 to 94.70): BF1

Ion 80.00 (79.70 to 80.70): BF1









#65
4,6-Dinitro-2-methylphenol
Concen: 38.996 ng
RT: 10.46 min Scan# 1423
Delta R.T. 0.00 min
Lab File: BF117111.D
Acq: 17 Sep 2019 17:17




Instrument :
BNA_F
ClientSampleId :
SSTDCCC040


Tot Ion:198 Resp: 44433
Ion Ratio Lower Upper
198 100
51 48.0 18.7 58.7
105 54.5 30.2 70.2



#66
n-Nitrosodiphenylamine
Concen: 38.256 ng
RT: 10.51 min Scan# 1432
Delta R.T. -0.00 min
Lab File: BF117111.D
Acq: 17 Sep 2019 17:17

Tgt Ion:169 Resp: 343718
Ion Ratio Lower Upper
169 100
168 65.5 52.2 78.2
167 36.0 28.8 43.2









#67
4-Bromophenyl-phenylether
Concen: 38.445 ng
RT: 10.88 min Scan# 1495
Delta R.T. -0.00 min
Lab File: BF117111.D
Acq: 17 Sep 2019 17:17

Instrument :
BNA_F
ClientSampleId :
SSTDCCC040

Tot Ion:248 Resp: 102141
Ion Ratio Lower Upper
248 100
250 98.2 76.2 114.4
141 88.9 75.0 112.4




Abundance Ion 248.00 (247.70 to 248.70): E
Ion 250.00 (249.70 to 250.70): E
Ion 141.00 (140.70 to 141.70): E






#68
Hexachlorobenzene
Concen: 38.182 ng
RT: 10.96 min Scan# 1508
Delta R.T. -0.00 min
Lab File: BF117111.D
Acq: 17 Sep 2019 17:17

Tgt Ion:284 Resp: 110954
Ion Ratio Lower Upper
284 100
142 35.1 31.4 47.2
249 30.3 26.5 39.7



Abundance Ion 283.80 (283.50 to 284.50): E
Ion 142.00 (141.70 to 142.70): E
Ion 249.00 (248.70 to 249.70): E





#69

Atrazine

Concen: 37.312 ng

RT: 11.03 min Scan# 1521

Delta R.T. 0.00 min

Lab File: BF117111.D

Acq: 17 Sep 2019 17:17

Instrument :

BNA_F

ClientSampleId :

SSTDCCC040


Tot Ion:200 Resp: 87350

Ion Ratio Lower Upper

200 100

173 28.6 6.9 46.9




215 44.0 23.8 63.8



Abundance Ion 200.00 (199.70 to 200.70): E

Ion 173.00 (172.70 to 173.70): E

Ion 215.00 (214.70 to 215.70): E



#70

Pentachlorophenol

Concen: 32.349 ng

RT: 11.15 min Scan# 1541

Delta R.T. 0.01 min

Lab File: BF117111.D

Acq: 17 Sep 2019 17:17


Tgt Ion:266 Resp: 44064

Ion Ratio Lower Upper

266 100

268 62.0 50.5 75.7



264 66.5 48.8 73.2




Abundance Ion 265.70 (265.40 to 266.40): E

Ion 268.00 (267.70 to 268.70): E

Ion 264.00 (263.70 to 264.70): E









#71
Phenanthrene
Concen: 37.498 ng
RT: 11.36 min Scan# 1577
Delta R.T. -0.00 min
Lab File: BF117111.D
Acq: 17 Sep 2019 17:17

Instrument :
BNA_F
ClientSampleId :
SSTDCCC040

Tot Ion:178 Resp: 554104
Ion Ratio Lower Upper
178 100
176 18.8 15.4 23.2
179 15.2 12.4 18.6




Abundance Ion 178.00 (177.70 to 178.70): E
Ion 176.00 (175.70 to 176.70): E
Ion 179.00 (178.70 to 179.70): E






#72
Anthracene
Concen: 37.833 ng
RT: 11.41 min Scan# 1585
Delta R.T. -0.01 min
Lab File: BF117111.D
Acq: 17 Sep 2019 17:17

Tgt Ion:178 Resp: 570744
Ion Ratio Lower Upper
178 100
176 18.6 14.6 21.8
179 15.2 11.6 17.4



Abundance Ion 178.00 (177.70 to 178.70): E
Ion 176.00 (175.70 to 176.70): E
Ion 179.00 (178.70 to 179.70): E





#73

Carbazole

Concen: 37.961 ng

RT: 11.57 min Scan# 1612

Delta R.T. -0.00 min

Lab File: BF117111.D

Acq: 17 Sep 2019 17:17

Instrument :

BNA_F

ClientSampleId :

SSTDCCC040


Tot Ion:167 Resp: 543590

Ion Ratio Lower Upper

167 100

166 21.4 17.2 25.8




139 13.3 10.6 16.0



Abundance Ion 167.00 (166.70 to 167.70): E

Ion 166.00 (165.70 to 166.70): E

Ion 139.00 (138.70 to 139.70): E



#74

Di-n-butylphthalate

Concen: 38.489 ng

RT: 11.89 min Scan# 1667

Delta R.T. -0.00 min

Lab File: BF117111.D

Acq: 17 Sep 2019 17:17


Tgt Ion:149 Resp: 655741

Ion Ratio Lower Upper

149 100

150 8.9 7.7 11.5



104 5.2 4.2 6.4




Abundance Ion 149.00 (148.70 to 149.70): E

Ion 150.00 (149.70 to 150.70): E

Ion 104.00 (103.70 to 104.70): E









#75
 Fluoranthene
 Concen: 37.728 ng
 RT: 12.54 min Scan# 1778
 Delta R.T. -0.00 min
 Lab File: BF117111.D
 Acq: 17 Sep 2019 17:17




Instrument :
 BNA_F
 ClientSampleId :
 SSTDCCC040


Tot Ion	Ratio	Lower	Upper
202	100		
101	12.4	0.0	32.3
203	17.2	0.0	37.1



#76
 Chrysene-d12
 Concen: 20.000 ng
 RT: 13.97 min Scan# 2021
 Delta R.T. -0.00 min
 Lab File: BF117111.D
 Acq: 17 Sep 2019 17:17

Tgt Ion	Ratio	Lower	Upper
240	100		
120	9.2	8.3	12.5
236	26.1	20.3	30.5





#77

Benzidine

Concen: 40.262 ng

RT: 12.66 min Scan# 1798

Delta R.T. -0.00 min

Lab File: BF117111.D

Acq: 17 Sep 2019 17:17

Instrument :

BNA_F

ClientSampleId :

SSTDCCC040


Tot Ion:184 Resp: 231602

Ion Ratio Lower Upper

184 100

185 14.9 11.6 17.4




183 11.4 9.0 13.4



Abundance Ion 184.00 (183.70 to 184.70): E

Ion 185.00 (184.70 to 185.70): E

Ion 183.00 (182.70 to 183.70): E



#78

Pyrene

Concen: 38.157 ng

RT: 12.77 min Scan# 1817

Delta R.T. -0.00 min

Lab File: BF117111.D

Acq: 17 Sep 2019 17:17


Tgt Ion:202 Resp: 571740

Ion Ratio Lower Upper

202 100

200 21.7 17.2 25.8



203 17.6 14.0 21.0




Abundance Ion 202.00 (201.70 to 202.70): E

Ion 200.00 (199.70 to 200.70): E

Ion 203.00 (202.70 to 203.70): E









#79
 Terphenyl-d14
 Concen: 79.207 ng
 RT: 12.92 min Scan# 1841
 Delta R.T. -0.00 min
 Lab File: BF117111.D
 Acq: 17 Sep 2019 17:17




Instrument :
 BNA_F
 ClientSampleId :
 SSTCCCC040


Tot Ion:244 Resp: 756703
 Ion Ratio Lower Upper
 244 100
 212 7.0 5.8 8.8
 122 12.5 11.3 16.9



#80
 Butylbenzylphthalate
 Concen: 39.304 ng
 RT: 13.38 min Scan# 1920
 Delta R.T. -0.01 min
 Lab File: BF117111.D
 Acq: 17 Sep 2019 17:17

Tgt Ion:149 Resp: 278275
 Ion Ratio Lower Upper
 149 100
 91 76.7 58.8 88.2
 206 16.7 14.1 21.1





#81

Benzo(a)anthracene

Concen: 37.669 ng

RT: 13.96 min Scan# 2019

Delta R.T. -0.00 min

Lab File: BF117111.D

Acq: 17 Sep 2019 17:17

Instrument :

BNA_F

ClientSampleId :

SSTDCCC040


Tot Ion:228 Resp: 449424

Ion Ratio Lower Upper

228 100

226 27.5 22.3 33.5




229 19.7 15.9 23.9



Abundance Ion 228.00 (227.70 to 228.70): E

Ion 226.00 (225.70 to 226.70): E

Ion 229.00 (228.70 to 229.70): E



#82

3,3'-Dichlorobenzidine

Concen: 39.395 ng

RT: 13.92 min Scan# 2012

Delta R.T. -0.00 min

Lab File: BF117111.D

Acq: 17 Sep 2019 17:17


Tgt Ion:252 Resp: 151461

Ion Ratio Lower Upper

252 100

254 64.4 52.6 79.0



126 12.5 10.4 15.6




Abundance Ion 252.00 (251.70 to 252.70): E

Ion 254.00 (253.70 to 254.70): E

Ion 126.00 (125.70 to 126.70): E





#83

Chrysene

Concen: 38.293 ng

RT: 14.00 min Scan# 2025

Delta R.T. -0.00 min

Lab File: BF117111.D

Acq: 17 Sep 2019 17:17

Instrument :

BNA_F

ClientSampleId :

SSTDCCC040


Tot Ion:228 Resp: 445588

Ion Ratio Lower Upper

228 100

226 29.8 24.8 37.2




229 19.2 15.8 23.8



Abundance Ion 228.00 (227.70 to 228.70): E

Ion 226.00 (225.70 to 226.70): E

Ion 229.00 (228.70 to 229.70): E



#84

Bis(2-ethylhexyl)phthalate

Concen: 40.092 ng

RT: 13.94 min Scan# 2015

Delta R.T. -0.01 min

Lab File: BF117111.D

Acq: 17 Sep 2019 17:17


Tgt Ion:149 Resp: 365988

Ion Ratio Lower Upper

149 100

167 27.0 22.2 33.2



279 2.9 2.8 4.2




Abundance Ion 149.00 (148.70 to 149.70): E

Ion 167.00 (166.70 to 167.70): E

Ion 279.00 (278.70 to 279.70): E









#85
Di-n-octyl phthalate
Concen: 40.587 ng
RT: 14.56 min Scan# 2120
Delta R.T. -0.00 min
Lab File: BF117111.D
Acq: 17 Sep 2019 17:17

Instrument :
BNA_F
ClientSampleId :
SSTDCCC040

Tot Ion	Ratio	Lower	Upper
149	100		
167	1.5	1.2	1.8
43	7.4	6.2	9.4




Abundance Ion 149.00 (148.70 to 149.70): E
Ion 167.00 (166.70 to 167.70): E
Ion 43.00 (42.70 to 43.70): BF1






#86
Indeno(1,2,3-cd)pyrene
Concen: 32.918 ng
RT: 16.87 min Scan# 2513
Delta R.T. 0.01 min
Lab File: BF117111.D
Acq: 17 Sep 2019 17:17

Tgt Ion	Ratio	Lower	Upper
276	100		
138	24.3	20.7	31.1
277	24.7	19.8	29.8



Abundance Ion 276.00 (275.70 to 276.70): E
Ion 138.00 (137.70 to 138.70): E
Ion 277.00 (276.70 to 277.70): E









#87
Pervlene-d12
Concen: 20.000 ng
RT: 15.42 min Scan# 2267
Delta R.T. -0.00 min
Lab File: BF117111.D
Acq: 17 Sep 2019 17:17

Instrument :
BNA_F
ClientSampleId :
SSTDCCC040

Tot Ion:264 Resp: 144115




Ion	Ratio	Lower	Upper
264	100		
260	23.1	17.6	26.4
265	21.9	16.6	24.8




#88
Benzo(b)fluoranthene
Concen: 42.032 ng
RT: 15.00 min Scan# 2196
Delta R.T. -0.00 min
Lab File: BF117111.D
Acq: 17 Sep 2019 17:17

Tgt Ion:252 Resp: 366924

Ion	Ratio	Lower	Upper
252	100		
253	21.7	17.6	26.4
125	11.1	8.6	12.8









#89
Benzo(k)fluoranthene
Concen: 38.791 ng
RT: 15.03 min Scan# 2201
Delta R.T. -0.00 min
Lab File: BF117111.D
Acq: 17 Sep 2019 17:17




Instrument :
BNA_F
ClientSampleId :
SSTDCCC040


Tot Ion:252 Resp: 320774
Ion Ratio Lower Upper
252 100
253 21.7 17.4 26.2
125 9.8 8.8 13.2



#90
Benzo(a)pyrene
Concen: 39.390 ng
RT: 15.36 min Scan# 2257
Delta R.T. -0.00 min
Lab File: BF117111.D
Acq: 17 Sep 2019 17:17

Tgt Ion:252 Resp: 301743
Ion Ratio Lower Upper
252 100
253 21.5 16.6 25.0
125 12.1 8.9 13.3





#91

Dibenzo(a,h)anthracene

Concen: 37.355 ng

RT: 16.87 min Scan# 2514

Delta R.T. 0.01 min

Lab File: BF117111.D

Acq: 17 Sep 2019 17:17

Instrument :

BNA_F

ClientSampleId :

SSTDCCC040


Tot Ion:278 Resp: 227959

Ion Ratio Lower Upper




278 100

139 14.8 13.5 20.3

279 21.7 18.8 28.2



Abundance Ion 278.00 (277.70 to 278.70): E
Ion 139.00 (138.70 to 139.70): E
Ion 279.00 (278.70 to 279.70): E



#92

Benzo(g,h,i)perylene

Concen: 35.922 ng

RT: 17.30 min Scan# 2586

Delta R.T. 0.01 min

Lab File: BF117111.D

Acq: 17 Sep 2019 17:17


Tgt Ion:276 Resp: 218446

Ion Ratio Lower Upper

276 100

277 25.4 18.6 27.8

138 22.5 18.0 27.0



Abundance Ion 276.00 (275.70 to 276.70): E
Ion 277.00 (276.70 to 277.70): E
Ion 138.00 (137.70 to 138.70): E

