

Data Path : Z:\SVOASRV\HPCHEM1\BNA F\DATA\BF092019\
 Data File : BF117193.D
 Acq On : 20 Sep 2019 18:43
 Operator : HP/JU
 Sample : PB123268BL
 Misc :
 ALS Vial : 12 Sample Multiplier: 1

Instrument :
 BNA_F
 ClientSampleId :
 PB123268BL

Quant Time: Sep 21 01:02:11 2019
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA F\METHODS\8270-BF091319.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Fri Sep 13 16:23:17 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	6.82	152	39637	20.00	ng	0.00
21) Naphthalene-d8	8.10	136	161848	20.00	ng	0.00
39) Acenaphthene-d10	9.85	164	90682	20.00	ng	0.00
64) Phenanthrene-d10	11.34	188	159987	20.00	ng	0.00
76) Chrysene-d12	13.98	240	112349	20.00	ng	0.00
87) Perylene-d12	15.43	264	89430	20.00	ng	0.01
System Monitoring Compounds						
5) 2-Fluorophenol	5.46	112	290420	114.39	ng	0.02
7) Phenol-d6	6.46	99	402878	122.79	ng	0.00
23) Nitrobenzene-d5	7.38	82	258389	83.88	ng	0.00
42) 2,4,6-Tribromophenol	10.65	330	104518	137.91	ng	0.00
45) 2-Fluorobiphenyl	9.17	172	524607	90.45	ng	0.00
79) Terphenyl-d14	12.93	244	585013	97.33	ng	0.01

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\SVOASRV\HPCHEM1\BNA F\DATA\BF092019\
Data File : BF117193.D
Acq On : 20 Sep 2019 18:43
Operator : HP/JU
Sample : PB123268BL
Misc :
ALS Vial : 12 Sample Multiplier: 1

Instrument :
BNA_F
ClientSampleId :
PB123268BL

Quant Time: Sep 21 01:02:11 2019
Quant Method : Z:\SVOASRV\HPCHEM1\BNA F\METHODS\8270-BF091319.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Fri Sep 13 16:23:17 2019
Response via : Initial Calibration

