

Data Path : Z:\SVOASRV\HPCHEM1\BNA F\DATA\BF092218\
 Data File : BF109247.D
 Acq On : 23 Sep 2018 3:45
 Operator : JU/SJ
 Sample : J5017-02
 Misc :
 ALS Vial : 29 Sample Multiplier: 1

Instrument :
 BNA_F
 ClientSampleId :
 RA-091918-N-FL-056

Quant Time: Sep 24 08:22:35 2018
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA F\METHODS\8270-BF092218.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Sat Sep 22 13:32:18 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	6.70	152	115531	20.00	ng	0.00
21) Naphthalene-d8	7.99	136	421701	20.00	ng	0.00
38) Acenaphthene-d10	9.74	164	175817	20.00	ng	0.00
63) Phenanthrene-d10	11.22	188	238834	20.00	ng	0.00
75) Chrysene-d12	13.86	240	176682	20.00	ng	0.01
86) Perylene-d12	15.26	264	153981	20.00	ng	0.02
System Monitoring Compounds						
5) 2-Fluorophenol	5.32	112	654977	93.38	ng	0.00
7) Phenol-d6	6.35	99	834043	93.18	ng	-0.01
23) Nitrobenzene-d5	7.27	82	517107	72.39	ng	-0.01
41) 2,4,6-Tribromophenol	10.52	330	143581	82.84	ng	-0.01
44) 2-Fluorobiphenyl	9.06	172	893407	76.64	ng	-0.01
78) Terphenyl-d14	12.82	244	490966	68.20	ng	0.01
Target Compounds						Qvalue
49) Dimethylphthalate	9.46	163	208382	16.295	ng	# 99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\SVOASRV\HPCHEM1\BNA F\DATA\BF092218\
Data File : BF109247.D
Acq On : 23 Sep 2018 3:45
Operator : JU/SJ
Sample : J5017-02
Misc :
ALS Vial : 29 Sample Multiplier: 1

Instrument :
BNA_F
ClientSampleId :
RA-091918-N-FL-056

Quant Time: Sep 24 08:22:35 2018
Quant Method : Z:\SVOASRV\HPCHEM1\BNA F\METHODS\8270-BF092218.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Sat Sep 22 13:32:18 2018
Response via : Initial Calibration

