

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF092221\
 Data File : BF125549.D
 Acq On : 22 Sep 2021 9:34
 Operator : JU/CG
 Sample : SSTDCCC040
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_F
 LabSampleId :
 SSTDCCC040

Quant Time: Sep 22 10:48:15 2021
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_F\METHODS\8270-BF092121.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Tue Sep 21 19:21:36 2021
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	109	0.00
2	1,4-Dioxane	0.510	0.466	8.6	107	0.00
3	Pyridine	1.366	1.253	8.3	106	0.00
4	n-Nitrosodimethylamine	0.539	0.473	12.2	103	0.00
5 S	2-Fluorophenol	1.265	1.150	9.1	108	0.00
6	Aniline	1.877	1.721	8.3	107	0.00
7 S	Phenol-d6	1.631	1.479	9.3	108	0.00
8	2-Chlorophenol	1.444	1.334	7.6	108	0.00
9	Benzaldehyde	0.912	0.794	12.9	108	0.00
10 C	Phenol	1.660	1.506	9.3	108	0.00
11	bis(2-Chloroethyl)ether	1.327	1.209	8.9	108	0.00
12	1,3-Dichlorobenzene	1.552	1.429	7.9	107	0.00
13 C	1,4-Dichlorobenzene	1.569	1.440	8.2	108	0.00
14	1,2-Dichlorobenzene	1.465	1.346	8.1	108	0.00
15	Benzyl Alcohol	1.167	1.066	8.7	107	0.00
16	2,2'-oxybis(1-Chloropropane	1.826	1.618	11.4	104	0.00
17	2-Methylphenol	1.153	1.058	8.2	108	0.00
18	Hexachloroethane	0.559	0.513	8.2	107	0.00
19 P	n-Nitroso-di-n-propylamine	0.976	0.877	10.1	109	0.00
20	3+4-Methylphenols	1.433	1.296	9.6	107	0.00
21 I	Naphthalene-d8	1.000	1.000	0.0	109	0.00
22	Acetophenone	0.463	0.427	7.8	109	0.00
23 S	Nitrobenzene-d5	0.314	0.315	-0.3	113	0.00
24	Nitrobenzene	0.332	0.330	0.6	112	0.00
25	Isophorone	0.675	0.622	7.9	109	0.00
26 C	2-Nitrophenol	0.150	0.169	-12.7	122	0.00
27	2,4-Dimethylphenol	0.300	0.284	5.3	112	0.00
28	bis(2-Chloroethoxy)methane	0.388	0.359	7.5	108	0.00
29 C	2,4-Dichlorophenol	0.303	0.291	4.0	111	0.00
30	1,2,4-Trichlorobenzene	0.319	0.299	6.3	109	0.00
31	Naphthalene	1.027	0.949	7.6	109	0.00
32	Benzoic acid	0.182	0.210	-15.4	131	0.00
33	4-Chloroaniline	0.421	0.393	6.7	109	0.00
34 C	Hexachlorobutadiene	0.180	0.170	5.6	110	0.00
35	Caprolactam	0.083	0.080	3.6	112	0.00
36 C	4-Chloro-3-methylphenol	0.298	0.283	5.0	111	0.00
37	2-Methylnaphthalene	0.696	0.649	6.8	111	0.00
38	1-Methylnaphthalene	0.653	0.615	5.8	111	0.00
39 I	Acenaphthene-d10	1.000	1.000	0.0	112	0.00
40	1,2,4,5-Tetrachlorobenzene	0.539	0.501	7.1	111	0.00
41 P	Hexachlorocyclopentadiene	0.301	0.277	8.0	107	0.00
42 S	2,4,6-Tribromophenol	0.196	0.192	2.0	116	0.00
43 C	2,4,6-Trichlorophenol	0.406	0.387	4.7	113	0.00
44	2,4,5-Trichlorophenol	0.426	0.413	3.1	111	0.00
45 S	2-Fluorobiphenyl	1.323	1.119	15.4	110	0.00
46	1,1'-Biphenyl	1.507	1.393	7.6	111	0.00
47	2-Chloronaphthalene	1.252	1.169	6.6	111	0.00

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF092221\
 Data File : BF125549.D
 Acq On : 22 Sep 2021 9:34
 Operator : JU/CG
 Sample : SSTDCCC040
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_F
 LabSampleId :
 SSTDCCC040

Quant Time: Sep 22 10:48:15 2021
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_F\METHODS\8270-BF092121.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Tue Sep 21 19:21:36 2021
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
48	2-Nitroaniline	0.267	0.293	-9.7	121	0.00
49	Acenaphthylene	1.839	1.726	6.1	112	0.00
50	Dimethylphthalate	1.376	1.288	6.4	113	0.00
51	2,6-Dinitrotoluene	0.274	0.294	-7.3	119	0.00
52 C	Acenaphthene	1.161	1.080	7.0	111	0.00
53	3-Nitroaniline	0.291	0.307	-5.5	118	0.00
54 P	2,4-Dinitrophenol	0.099	0.116	-17.2	136	0.00
55	Dibenzofuran	1.723	1.588	7.8	112	0.00
56 P	4-Nitrophenol	0.236	0.250	-5.9	117	0.00
57	2,4-Dinitrotoluene	0.328	0.371	-13.1	121	0.00
58	Fluorene	1.325	1.161	12.4	112	0.00
59	2,3,4,6-Tetrachlorophenol	0.320	0.314	1.9	115	0.00
60	Diethylphthalate	1.357	1.291	4.9	113	0.00
61	4-Chlorophenyl-phenylether	0.589	0.544	7.6	115	0.00
62	4-Nitroaniline	0.263	0.278	-5.7	120	0.00
63	Azobenzene	1.318	1.201	8.9	109	0.00
64 I	Phenanthrene-d10	1.000	1.000	0.0	112	0.00
65	4,6-Dinitro-2-methylphenol	0.097	0.111	-14.4	127	0.00
66 c	n-Nitrosodiphenylamine	0.726	0.677	6.7	111	0.00
67	4-Bromophenyl-phenylether	0.229	0.214	6.6	114	0.00
68	Hexachlorobenzene	0.258	0.240	7.0	111	0.00
69	Atrazine	0.195	0.193	1.0	114	0.00
70 C	Pentachlorophenol	0.147	0.148	-0.7	114	0.00
71	Phenanthrene	1.123	1.041	7.3	113	0.00
72	Anthracene	1.120	1.042	7.0	112	0.00
73	Carbazole	1.059	0.971	8.3	111	0.00
74	Di-n-butylphthalate	1.257	1.159	7.8	111	0.00
75 C	Fluoranthene	1.102	0.993	9.9	109	0.00
76 I	Chrysene-d12	1.000	1.000	0.0	112	0.00
77	Benzydine	0.567	0.558	1.6	107	0.00
78	Pyrene	1.874	1.806	3.6	111	0.00
79 S	Terphenyl-d14	1.239	1.137	8.2	109	0.00
80	Butylbenzylphthalate	0.757	0.739	2.4	112	0.00
81	Benzo(a)anthracene	1.361	1.271	6.6	112	0.00
82	3,3'-Dichlorobenzidine	0.417	0.400	4.1	113	0.00
83	Chrysene	1.375	1.308	4.9	113	0.00
84	Bis(2-ethylhexyl)phthalate	0.926	0.917	1.0	114	0.00
85 c	Di-n-octyl phthalate	1.539	1.546	-0.5	114	0.00
86 I	Perylene-d12	1.000	1.000	0.0	116	0.00
87	Indeno(1,2,3-cd)pyrene	1.436	1.424	0.8	114	0.00
88	Benzo(b)fluoranthene	1.463	1.365	6.7	114	0.00
89	Benzo(k)fluoranthene	1.338	1.211	9.5	113	0.00
90 C	Benzo(a)pyrene	1.273	1.186	6.8	114	0.00
91	Dibenzo(a,h)anthracene	1.217	1.208	0.7	114	0.00
92	Benzo(g,h,i)perylene	1.201	1.190	0.9	114	0.00

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF092221\
Data File : BF125549.D
Acq On : 22 Sep 2021 9:34
Operator : JU/CG
Sample : SSTDCCC040
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
BNA_F
LabSampleId :
SSTDCCC040

Quant Time: Sep 22 10:48:15 2021
Quant Method : Z:\SVOASRV\HPCHEM1\BNA_F\METHODS\8270-BF092121.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Tue Sep 21 19:21:36 2021
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

Compound	AvgRF	CCRF	%Dev Area	% Dev(min)
----------	-------	------	-----------	------------

(#) = Out of Range

SPCC's out = 0 CCC's out = 0