




Quant Time: Sep 23 06:04:12 2023
Quant Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF091123.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Tue Sep 12 00:19:07 2023
Response via : Initial Calibration






TIC: BF135400.D\data.ms

response **198241**

Ion	Exp%	Act%
265.70	100.00	100.00
268.00	65.10	63.37
264.00	63.50	64.58
0.00	0.00	0.00

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF092223\
Data File : BF135400.D
Acq On : 22 Sep 2023 14:44
Operator : CG\JU
Sample : DFTPP
Misc :
ALS Vial : 1 Sample Multiplier: 1

Quant Time: Sep 23 06:04:12 2023
Quant Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF091123.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Tue Sep 12 00:19:07 2023
Response via : Initial Calibration



TIC: BF135400.D\data.ms

(77) Benzidine**12.639min (+ 0.000) 621392.11 ng****response 1107541**

Ion	Exp%	Act%
184.00	100.00	100.00
185.00	14.60	14.33
183.00	10.30	10.63
0.00	0.00	0.00