

Data Path : Z:\SVOASRV\HPCHEM1\BNA F\DATA\BF092319\
 Data File : BF117211.D
 Acq On : 23 Sep 2019 18:06
 Operator : HP/JU
 Sample : K4665-01 2X
 Misc :
 ALS Vial : 5 Sample Multiplier: 1

Instrument :
 BNA_F
 ClientSampleId :
 EO-01-091919

Quant Time: Sep 24 12:29:36 2019
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA F\METHODS\8270-BF091319.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Fri Sep 13 16:23:17 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	6.81	152	25230	20.00	ng	-0.01
21) Naphthalene-d8	8.09	136	105689	20.00	ng	-0.01
39) Acenaphthene-d10	9.85	164	55310	20.00	ng	-0.01
64) Phenanthrene-d10	11.33	188	82490	20.00	ng	0.00
76) Chrysene-d12	13.97	240	51424	20.00	ng	0.00
87) Perylene-d12	15.43	264	53602	20.00	ng	0.00
System Monitoring Compounds						
5) 2-Fluorophenol	5.43	112	67523	41.78	ng	0.00
7) Phenol-d6	6.45	99	94069	45.04	ng	0.00
23) Nitrobenzene-d5	7.37	82	52240	25.97	ng	-0.02
42) 2,4,6-Tribromophenol	10.64	330	18297	39.58	ng	0.00
45) 2-Fluorobiphenyl	9.17	172	108616	30.70	ng	0.00
79) Terphenyl-d14	12.91	244	86949	31.61	ng	0.00
Target Compounds						
50) Dimethylphthalate	9.56	163	15274	3.900	ng	Qvalue # 93
75) Fluoranthene	12.55	202	13498	2.804	ng	97
78) Pyrene	12.77	202	13180	3.055	ng	97

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\SVOASRV\HPCHEM1\BNA F\DATA\BF092319\
Data File : BF117211.D
Acq On : 23 Sep 2019 18:06
Operator : HP/JU
Sample : K4665-01 2X
Misc :
ALS Vial : 5 Sample Multiplier: 1

Instrument :
BNA_F
ClientSampleId :
EO-01-091919

Quant Time: Sep 24 12:29:36 2019
Quant Method : Z:\SVOASRV\HPCHEM1\BNA F\METHODS\8270-BF091319.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Fri Sep 13 16:23:17 2019
Response via : Initial Calibration

