


Data Path : Z:\SVOASRV\HPCHEM1\BNA_F\DATA\BF092319\
 Data File : BF117207.D
 Acq On : 23 Sep 2019 16:12
 Operator : HP/JU
 Sample : DFTPP
 Misc :
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 BNA_F
 ClientSampleId :
 DFTPP

Integration File: rteint.p

Method : Z:\SVOASRV\HPCHEM1\BNA_F\METHODS\8270-BF091319.M
 Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 Last Update : Fri Sep 13 16:23:17 2019



AutoFind: Scans 1614, 1615, 1616; Background Corrected with Scan 1610

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
51	198	10	80	38.0	7737	PASS
68	69	0.00	2	0.0	0	PASS
69	198	0.00	100	48.3	9833	PASS
70	69	0.00	2	0.0	0	PASS
127	198	10	80	57.1	11631	PASS
197	198	0.00	2	0.0	0	PASS
198	198	100	100	100.0	20356	PASS
199	198	5	9	7.1	1450	PASS
275	198	10	60	25.0	5089	PASS
365	198	1	100	2.8	560	PASS
441	198	0.01	100	10.7	2177	PASS
442	198	50	100	73.1	14881	PASS
443	442	15	24	18.3	2724	PASS