

Data Path : Z:\SVOASRV\HPCHEM1\BNA F\DATA\BF092418\
 Data File : BF109256.D
 Acq On : 24 Sep 2018 10:20
 Operator : JU/SJ
 Sample : J4993-06
 Misc :
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 BNA_F
 ClientSampleId :
 NS-OILY-WATER

Quant Time: Sep 24 13:30:11 2018
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA F\METHODS\8270-BF092218.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Sat Sep 22 13:32:18 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	6.70	152	121852	20.00	ng	0.00
21) Naphthalene-d8	7.99	136	474014	20.00	ng	0.00
38) Acenaphthene-d10	9.74	164	229740	20.00	ng	0.00
63) Phenanthrene-d10	11.22	188	426000	20.00	ng	0.00
75) Chrysene-d12	13.84	240	293756	20.00	ng	0.00
86) Perylene-d12	15.24	264	310560	20.00	ng	0.00
System Monitoring Compounds						
5) 2-Fluorophenol	5.31	112	192371	26.00	ng	0.00
7) Phenol-d6	6.34	99	155032	16.42	ng	-0.02
23) Nitrobenzene-d5	7.26	82	400412	49.87	ng	-0.02
41) 2,4,6-Tribromophenol	10.52	330	147921	65.31	ng	-0.01
44) 2-Fluorobiphenyl	9.06	172	749221	49.18	ng	-0.01
78) Terphenyl-d14	12.80	244	683036	57.06	ng	0.00
Target Compounds						Qvalue
49) Dimethylphthalate	9.46	163	91894	5.499	ng	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\SVOASRV\HPCHEM1\BNA F\DATA\BF092418\
Data File : BF109256.D
Acq On : 24 Sep 2018 10:20
Operator : JU/SJ
Sample : J4993-06
Misc :
ALS Vial : 3 Sample Multiplier: 1

Instrument :
BNA_F
ClientSampleId :
NS-OILY-WATER

Quant Time: Sep 24 13:30:11 2018
Quant Method : Z:\SVOASRV\HPCHEM1\BNA F\METHODS\8270-BF092218.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Sat Sep 22 13:32:18 2018
Response via : Initial Calibration

