

Data Path : Z:\SVOASRV\HPCHEM1\BNA F\DATA\BF092519\
 Data File : BF117248.D
 Acq On : 25 Sep 2019 18:52
 Operator : JU
 Sample : K5014-02
 Misc :
 ALS Vial : 14 Sample Multiplier: 1

Instrument :
 BNA_F
 ClientSampleId :

Quant Time: Sep 26 02:30:18 2019
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA F\METHODS\8270-BF091319.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Fri Sep 13 16:23:17 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	6.81	152	36365	20.00	ng	-0.01
21) Naphthalene-d8	8.10	136	145600	20.00	ng	0.00
39) Acenaphthene-d10	9.85	164	67954	20.00	ng	0.00
64) Phenanthrene-d10	11.33	188	94310	20.00	ng	0.00
76) Chrysene-d12	13.98	240	59930	20.00	ng	0.00
87) Perylene-d12	15.43	264	67234	20.00	ng	0.01

System Monitoring Compounds						
5) 2-Fluorophenol	5.45	112	218619	93.86	ng	0.00
7) Phenol-d6	6.46	99	298601	99.20	ng	0.00
23) Nitrobenzene-d5	7.37	82	189111	68.24	ng	-0.01
42) 2,4,6-Tribromophenol	10.64	330	51492	90.66	ng	0.00
45) 2-Fluorobiphenyl	9.17	172	322467	74.20	ng	0.00
79) Terphenyl-d14	12.92	244	240581	75.04	ng	0.00


Target Compounds				Qvalue		
9) Benzaldehyde	6.46	77	468	Below Cal	#	1
12) 1,3-Dichlorobenzene	6.83	146	10184	3.612	ng	98
13) 1,4-Dichlorobenzene	6.83	146	10184	3.539	ng	96
14) 1,2-Dichlorobenzene	6.83	146	10184	3.801	ng	94
19) n-Nitroso-di-n-propylamine	7.37	70	25475	13.898	ng	# 79
25) Isophorone	7.37	82	189109	38.543	ng	# 66
31) Naphthalene	8.12	128	293959	41.979	ng	98
32) Benzoic acid	8.12	122	754	7.056	ng	# 1
33) 4-Chloroaniline	8.12	127	37974	12.227	ng	# 34
35) Caprolactam	8.52	113	1972	2.983	ng	# 73
37) 2-Methylnaphthalene	8.81	142	52011	11.030	ng	98
38) 1-Methylnaphthalene	8.90	142	29963	6.785	ng	97
50) Dimethylphthalate	9.56	163	37610	7.817	ng	96
51) 2,6-Dinitrotoluene	9.60	165	14716	15.017	ng	# 27
60) Diethylphthalate	10.26	149	184997	39.081	ng	99
65) 4,6-Dinitro-2-methylphenol	10.53	198	232	7.365	ng	# 1
67) 4-Bromophenyl-phenylether	10.64	248	2889	3.000	ng	# 1
74) Di-n-butylphthalate	11.89	149	113850	18.434	ng	99
84) Bis(2-ethylhexyl)phthalate	13.94	149	19256	6.285	ng	# 99


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\SVOASRV\HPCHEM1\BNA F\DATA\BF092519\
Data File : BF117248.D
Acq On : 25 Sep 2019 18:52
Operator : JU
Sample : K5014-02
Misc :
ALS Vial : 14 Sample Multiplier: 1

Instrument :
BNA_F
ClientSampleId :

Quant Time: Sep 26 02:30:18 2019
Quant Method : Z:\SVOASRV\HPCHEM1\BNA F\METHODS\8270-BF091319.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Fri Sep 13 16:23:17 2019
Response via : Initial Calibration






#1




1,4-Dichlorobenzene-d4
Concen: 20.000 ng
RT: 6.81 min Scan# 803
Delta R.T. -0.01 min
Lab File: BF117248.D
Acq: 25 Sep 2019 18:52

Instrument :
BNA_F
ClientSampleId :

Tot Ion:152 Resp: 36365
Ion Ratio Lower Upper
152 100
150 163.1 127.5 191.3
115 67.0 55.0 82.4




Abundance Ion 152.00 (151.70 to 152.70): E
Ion 150.00 (149.70 to 150.70): E
Ion 115.00 (114.70 to 115.70): E





#5


2-Fluorophenol
Concen: 93.859 ng
RT: 5.45 min Scan# 571
Delta R.T. 0.01 min
Lab File: BF117248.D
Acq: 25 Sep 2019 18:52

Tgt Ion:112 Resp: 218619
Ion Ratio Lower Upper
112 100
64 56.7 45.7 68.5
63 31.2 24.9 37.3



Abundance Ion 112.00 (111.70 to 112.70): E
Ion 64.00 (63.70 to 64.70): BF1
Ion 63.00 (62.70 to 63.70): BF1





#7

Phenol-d6

Concen: 99.196 ng

RT: 6.46 min Scan# 743

Delta R.T. 0.00 min

Lab File: BF117248.D

Acq: 25 Sep 2019 18:52

Instrument :

BNA_F

ClientSampleId :


Tot Ion: 99 Resp: 298601

Ion Ratio Lower Upper




99 100

42 16.1 12.2 18.2

71 38.4 29.8 44.8



Abundance Ion 99.00 (98.70 to 99.70): BF1
Ion 42.00 (41.70 to 42.70): BF1
Ion 71.00 (70.70 to 71.70): BF1



#9

Benzaldehyde

Concen: Below Cal

RT: 6.46 min Scan# 744

Delta R.T. 0.09 min

Lab File: BF117248.D

Acq: 25 Sep 2019 18:52


Tgt Ion: 77 Resp: 468

Ion Ratio Lower Upper



77 100


105 0.0 81.8 121.8#

106 0.0 77.3 117.3#



Abundance Ion 77.00 (76.70 to 77.70): BF1
Ion 105.00 (104.70 to 105.70): E
Ion 106.00 (105.70 to 106.70): E





#12

1,3-Dichlorobenzene

Concen: 3.612 ng

RT: 6.83 min Scan# 807

Delta R.T. 0.07 min

Lab File: BF117248.D

Acq: 25 Sep 2019 18:52

Instrument :

BNA_F

ClientSampleId :


Tot Ion:146 Resp: 10184

Ion Ratio Lower Upper

146 100

148 62.3 50.8 76.2




75 34.6 28.7 43.1



Abundance Ion 146.00 (145.70 to 146.70): E

Ion 148.00 (147.70 to 148.70): E

Ion 75.00 (74.70 to 75.70): BF1



#13

1,4-Dichlorobenzene

Concen: 3.539 ng

RT: 6.83 min Scan# 807

Delta R.T. -0.01 min

Lab File: BF117248.D

Acq: 25 Sep 2019 18:52


Tgt Ion:146 Resp: 10184

Ion Ratio Lower Upper

146 100

148 62.3 50.6 76.0



111 39.1 35.0 52.6




Abundance Ion 146.00 (145.70 to 146.70): E

Ion 148.00 (147.70 to 148.70): E

Ion 111.00 (110.70 to 111.70): E









#14
1,2-Dichlorobenzene
Concen: 3.801 ng
RT: 6.83 min Scan# 807
Delta R.T. -0.16 min
Lab File: BF117248.D
Acq: 25 Sep 2019 18:52

Instrument :
BNA_F
ClientSampleId :

Tot Ion:146 Resp: 10184
Ion Ratio Lower Upper
146 100
148 62.3 52.6 78.8
111 39.1 36.2 54.2




Abundance Ion 146.00 (145.70 to 146.70): E
Ion 148.00 (147.70 to 148.70): E
Ion 111.00 (110.70 to 111.70): E






#19
n-Nitroso-di-n-propylamine
Concen: 13.898 ng
RT: 7.37 min Scan# 899
Delta R.T. 0.14 min
Lab File: BF117248.D
Acq: 25 Sep 2019 18:52

Tgt Ion: 70 Resp: 25475
Ion Ratio Lower Upper
70 100
42 36.7 34.0 51.0
101 0.0 7.8 11.6#
130 2.0 17.8 26.8#



Abundance Ion 70.00 (69.70 to 70.70): BF1
Ion 42.00 (41.70 to 42.70): BF1
Ion 101.00 (100.70 to 101.70): E
Ion 130.00 (129.70 to 130.70): E





#21

Naphthalene-d8

Concen: 20.000 ng

RT: 8.10 min Scan# 1022

Delta R.T. -0.01 min

Lab File: BF117248.D

Acq: 25 Sep 2019 18:52

Instrument :

BNA_F

ClientSampleId :

Tot Ion:136 Resp: 145600



Ion Ratio Lower Upper

136 100

137 10.4 9.0 13.4

54 5.3 4.5 6.7

68 4.3 4.0 6.0





Abundance Ion 136.00 (135.70 to 136.70): E

Ion 137.00 (136.70 to 137.70): E

Ion 54.00 (53.70 to 54.70): BF1

Ion 68.00 (67.70 to 68.70): BF1



#23

Nitrobenzene-d5

Concen: 68.244 ng

RT: 7.37 min Scan# 899

Delta R.T. -0.01 min

Lab File: BF117248.D

Acq: 25 Sep 2019 18:52



Tgt Ion: 82 Resp: 189111

Ion Ratio Lower Upper

82 100

128 40.7 32.3 48.5


54 43.1 33.3 49.9




Abundance Ion 82.00 (81.70 to 82.70): BF1

Ion 128.00 (127.70 to 128.70): E

Ion 54.00 (53.70 to 54.70): BF1





#25

Isophorone

Concen: 38.543 ng

RT: 7.37 min Scan# 899

Delta R.T. -0.27 min

Lab File: BF117248.D

Acq: 25 Sep 2019 18:52

Instrument :

BNA_F

ClientSampled :





Tot Ion: 82 Resp: 189109

Ion Ratio Lower Upper

82 100

95 0.0 5.8 8.6#

138 0.0 13.5 20.3#



#31

Naphthalene

Concen: 41.979 ng

RT: 8.12 min Scan# 1025

Delta R.T. -0.01 min

Lab File: BF117248.D

Acq: 25 Sep 2019 18:52




Tgt Ion: 128 Resp: 293959


Ion Ratio Lower Upper

128 100

129 10.8 9.0 13.4

127 12.7 10.8 16.2









#32
Benzoic acid
Concen: 7.056 ng
RT: 8.12 min Scan# 1025
Delta R.T. 0.23 min
Lab File: BF117248.D
Acq: 25 Sep 2019 18:52

Instrument :
BNA_F
ClientSampleId :

Tot Ion:122 Resp: 754
Ion Ratio Lower Upper
122 100
105 34.6 114.8 154.8#
77 1466.8 78.7 118.7#




Abundance Ion 122.00 (121.70 to 122.70): E
Ion 105.00 (104.70 to 105.70): E
Ion 77.00 (76.70 to 77.70): BF1






#33
4-Chloroaniline
Concen: 12.227 ng
RT: 8.12 min Scan# 1025
Delta R.T. -0.06 min
Lab File: BF117248.D
Acq: 25 Sep 2019 18:52

Tgt Ion:127 Resp: 37974
Ion Ratio Lower Upper
127 100
129 84.9 26.2 39.2#
65 0.0 23.5 35.3#
92 1.3 15.9 23.9#



Abundance Ion 127.00 (126.70 to 127.70): E
Ion 129.00 (128.70 to 129.70): E
Ion 65.00 (64.70 to 65.70): BF1
Ion 92.00 (91.70 to 92.70): BF1









#35
 Caprolactam
 Concen: 2.983 ng
 RT: 8.52 min Scan# 1093
 Delta R.T. -0.02 min
 Lab File: BF117248.D
 Acq: 25 Sep 2019 18:52

Instrument :
 BNA_F
 ClientSampled :

Tot Ion:113 Resp: 1972
 Ion Ratio Lower Upper
 113 100
 55 169.6 120.6 160.6#
 56 141.2 88.8 128.8#




Abundance Ion 113.00 (112.70 to 113.70): E
 Ion 55.00 (54.70 to 55.70): BF1
 Ion 56.00 (55.70 to 56.70): BF1






#37
 2-Methylnaphthalene
 Concen: 11.030 ng
 RT: 8.81 min Scan# 1143
 Delta R.T. -0.01 min
 Lab File: BF117248.D
 Acq: 25 Sep 2019 18:52

Tgt Ion:142 Resp: 52011
 Ion Ratio Lower Upper
 142 100
 141 85.5 70.7 106.1
 115 36.3 29.4 44.0



Abundance Ion 142.00 (141.70 to 142.70): E
 Ion 141.00 (140.70 to 141.70): E
 Ion 115.00 (114.70 to 115.70): E









#38
1-Methylnaphthalene
Concen: 6.785 ng
RT: 8.90 min Scan# 1159
Delta R.T. -0.01 min
Lab File: BF117248.D
Acq: 25 Sep 2019 18:52




Instrument :
BNA_F
ClientSampleId :


Tot Ion:142 Resp: 29963
Ion Ratio Lower Upper
142 100
141 92.5 71.9 107.9
116 4.2 3.2 4.8



#39
Acenaphthene-d10
Concen: 20.000 ng
RT: 9.85 min Scan# 1320
Delta R.T. -0.01 min
Lab File: BF117248.D
Acq: 25 Sep 2019 18:52

Tgt Ion:164 Resp: 67954
Ion Ratio Lower Upper
164 100
162 102.3 83.4 125.2
160 45.1 36.4 54.6









#42
 2,4,6-Tribromophenol
 Concen: 90.665 ng
 RT: 10.64 min Scan# 1454
 Delta R.T. -0.01 min
 Lab File: BF117248.D
 Acq: 25 Sep 2019 18:52

Instrument :
 BNA_F
 ClientSampleId :

Tot Ion:330 Resp: 51492
 Ion Ratio Lower Upper
 330 100
 332 99.1 78.0 117.0
 141 47.6 35.8 53.6




Abundance Ion 329.80 (329.50 to 330.50): E
 Ion 331.80 (331.50 to 332.50): E
 Ion 141.00 (140.70 to 141.70): E






#45
 2-Fluorobiphenyl
 Concen: 74.197 ng
 RT: 9.17 min Scan# 1205
 Delta R.T. 0.00 min
 Lab File: BF117248.D
 Acq: 25 Sep 2019 18:52

Tgt Ion:172 Resp: 322467
 Ion Ratio Lower Upper
 172 100
 171 34.0 28.2 42.2
 170 22.7 18.9 28.3



Abundance Ion 172.00 (171.70 to 172.70): E
 Ion 171.00 (170.70 to 171.70): E
 Ion 170.00 (169.70 to 170.70): E






#50
Dimethylphthalate
Concen: 7.817 ng
RT: 9.56 min Scan# 1271
Delta R.T. -0.01 min
Lab File: BF117248.D
Acq: 25 Sep 2019 18:52

Instrument :
BNA_F
ClientSampleId :

Tot Ion	Ratio	Lower	Upper
163	100		
194	4.1	2.9	4.3
164	11.9	8.2	12.4






Abundance

Ion 163.00 (162.70 to 163.70): E


Ion 194.00 (193.70 to 194.70): E

Ion 164.00 (163.70 to 164.70): E



#51
2,6-Dinitrotoluene
Concen: 15.017 ng
RT: 9.60 min Scan# 1278
Delta R.T. -0.04 min
Lab File: BF117248.D
Acq: 25 Sep 2019 18:52

Tgt Ion	Ratio	Lower	Upper
165	100		
63	2.1	41.8	62.8#
89	2.0	46.5	69.7#






Abundance

Ion 165.00 (164.70 to 165.70): E

Ion 63.00 (62.70 to 63.70): BF1

Ion 89.00 (88.70 to 89.70): BF1





#60

Diethylphthalate

Concen: 39.081 ng

RT: 10.26 min Scan# 1390

Delta R.T. -0.01 min

Lab File: BF117248.D

Acq: 25 Sep 2019 18:52


Instrument :

BNA_F

ClientSampleId :

Tot Ion:149 Resp: 184997

Ion	Ratio	Lower	Upper
149	100		
177	18.7	15.2	22.8
150	12.0	9.9	14.9






Abundance

Ion 149.00 (148.70 to 149.70): E

Ion 177.00 (176.70 to 177.70): E

Ion 150.00 (149.70 to 150.70): E



#64

Phenanthrene-d10

Concen: 20.000 ng

RT: 11.33 min Scan# 1572


Delta R.T. 0.00 min

Lab File: BF117248.D

Acq: 25 Sep 2019 18:52

Tgt Ion:188 Resp: 94310

Ion	Ratio	Lower	Upper
188	100		
94	9.8	8.2	12.2
80	10.5	9.0	13.6






Abundance

Ion 188.00 (187.70 to 188.70): E

Ion 94.00 (93.70 to 94.70): BF1

Ion 80.00 (79.70 to 80.70): BF1









#65
 4,6-Dinitro-2-methylphenol
 Concen: 7.365 ng
 RT: 10.53 min Scan# 1435
 Delta R.T. 0.07 min
 Lab File: BF117248.D
 Acq: 25 Sep 2019 18:52




Instrument :
 BNA_F
 ClientSampled :


Tot Ion:198 Resp: 232
 Ion Ratio Lower Upper
 198 100
 51 0.0 18.7 58.7#
 105 491.7 30.2 70.2#



#67
 4-Bromophenyl-phenylether
 Concen: 3.000 ng
 RT: 10.64 min Scan# 1454
 Delta R.T. -0.24 min
 Lab File: BF117248.D
 Acq: 25 Sep 2019 18:52

Tgt Ion:248 Resp: 2889
 Ion Ratio Lower Upper
 248 100
 250 207.4 76.2 114.4#
 141 763.4 75.0 112.4#





#74

Di-n-butylphthalate

Concen: 18.434 ng

RT: 11.89 min Scan# 1667

Delta R.T. 0.00 min

Lab File: BF117248.D

Acq: 25 Sep 2019 18:52

Instrument :

BNA_F

ClientSampleId :


Tot Ion:149 Resp: 113850

Ion Ratio Lower Upper

149 100

150 9.1 7.7 11.5




104 5.2 4.2 6.4



Abundance Ion 149.00 (148.70 to 149.70): E

Ion 150.00 (149.70 to 150.70): E

Ion 104.00 (103.70 to 104.70): E



#76

Chrysene-d12

Concen: 20.000 ng

RT: 13.98 min Scan# 2022

Delta R.T. 0.01 min

Lab File: BF117248.D

Acq: 25 Sep 2019 18:52


Tgt Ion:240 Resp: 59930

Ion Ratio Lower Upper

240 100

120 10.1 8.3 12.5



236 24.2 20.3 30.5




Abundance Ion 240.00 (239.70 to 240.70): E

Ion 120.00 (119.70 to 120.70): E

Ion 236.00 (235.70 to 236.70): E









#79
 Terphenyl-d14
 Concen: 75.037 ng
 RT: 12.92 min Scan# 1841
 Delta R.T. 0.00 min
 Lab File: BF117248.D
 Acq: 25 Sep 2019 18:52




Instrument :
 BNA_F
 ClientSampleId :


Tot Ion:244 Resp: 240581
 Ion Ratio Lower Upper
 244 100
 212 6.9 5.8 8.8
 122 13.0 11.3 16.9



#84
 Bis(2-ethylhexyl)phthalate
 Concen: 6.285 ng
 RT: 13.94 min Scan# 2016
 Delta R.T. 0.00 min
 Lab File: BF117248.D
 Acq: 25 Sep 2019 18:52

Tgt Ion:149 Resp: 19256
 Ion Ratio Lower Upper
 149 100
 167 27.8 22.2 33.2
 279 4.9 2.8 4.2#






#87
Pervlene-d12
Concen: 20.000 ng
RT: 15.43 min Scan# 2269
Delta R.T. 0.01 min
Lab File: BF117248.D
Acq: 25 Sep 2019 18:52

Instrument :
BNA_F
ClientSampleId :

Tot Ion	Ratio	Lower	Upper
264	100		
260	22.5	17.6	26.4
265	21.4	16.6	24.8



Abundance

Ion 264.00 (263.70 to 264.70): E

Ion 260.00 (259.70 to 260.70): E

Ion 265.00 (264.70 to 265.70): E

