

Data Path : Z:\SVOASRV\HPCHEM1\BNA F\DATA\BF092519\
 Data File : BF117248.D
 Acq On : 25 Sep 2019 18:52
 Operator : JU
 Sample : K5014-02
 Misc :
 ALS Vial : 14 Sample Multiplier: 1

Instrument :
 BNA_F
 ClientSampleId :
 WALL-1

Quant Time: Sep 26 03:51:26 2019
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA F\METHODS\8270-BF091319.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Fri Sep 13 16:23:17 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	6.81	152	36365	20.00	ng	-0.01
21) Naphthalene-d8	8.10	136	145600	20.00	ng	0.00
39) Acenaphthene-d10	9.85	164	67954	20.00	ng	0.00
64) Phenanthrene-d10	11.33	188	94310	20.00	ng	0.00
76) Chrysene-d12	13.98	240	59930	20.00	ng	0.00
87) Perylene-d12	15.43	264	67234	20.00	ng	0.01
System Monitoring Compounds						
5) 2-Fluorophenol	5.45	112	218619	93.86	ng	0.00
7) Phenol-d6	6.46	99	298601	99.20	ng	0.00
23) Nitrobenzene-d5	7.37	82	189111	68.24	ng	-0.01
42) 2,4,6-Tribromophenol	10.64	330	51492	90.66	ng	0.00
45) 2-Fluorobiphenyl	9.17	172	322467	74.20	ng	0.00
79) Terphenyl-d14	12.92	244	240581	75.04	ng	0.00
Target Compounds						
13) 1,4-Dichlorobenzene	6.83	146	10184	3.539	ng	96
31) Naphthalene	8.12	128	293959	41.979	ng	98
37) 2-Methylnaphthalene	8.81	142	52011	11.030	ng	98
38) 1-Methylnaphthalene	8.90	142	29963	6.785	ng	97
50) Dimethylphthalate	9.56	163	37610	7.817	ng	96
60) Diethylphthalate	10.26	149	184997	39.081	ng	99
74) Di-n-butylphthalate	11.89	149	113850	18.434	ng	99
84) Bis(2-ethylhexyl)phthalate	13.94	149	19256	6.285	ng	# 99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\SVOASRV\HPCHEM1\BNA F\DATA\BF092519\
Data File : BF117248.D
Acq On : 25 Sep 2019 18:52
Operator : JU
Sample : K5014-02
Misc :
ALS Vial : 14 Sample Multiplier: 1

Instrument :
BNA_F
ClientSampleId :
WALL-1

Quant Time: Sep 26 03:51:26 2019
Quant Method : Z:\SVOASRV\HPCHEM1\BNA F\METHODS\8270-BF091319.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Fri Sep 13 16:23:17 2019
Response via : Initial Calibration

