

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF092525\
 Data File : BF143808.D
 Acq On : 25 Sep 2025 17:21
 Operator : RC/JU
 Sample : Q3188-01 10X
 Misc :
 ALS Vial : 8 Sample Multiplier: 1

Instrument :
 BNA_F
 ClientSampleId :
 4135

Quant Time: Sep 25 17:53:46 2025
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF092325.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Wed Sep 24 05:00:00 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	6.887	152	81296	20.000	ng	0.00
21) Naphthalene-d8	8.169	136	374000	20.000	ng	0.00
39) Acenaphthene-d10	9.939	164	178882	20.000	ng	0.01
64) Phenanthrene-d10	11.433	188	341921	20.000	ng	0.02
76) Chrysene-d12	14.069	240	413505	20.000	ng	0.00
86) Perylene-d12	15.551	264	559999	20.000	ng	0.01
System Monitoring Compounds						
5) 2-Fluorophenol	5.493	112	33043	6.604	ng	0.00
7) Phenol-d6	6.545	99	37907	6.557	ng	0.04
23) Nitrobenzene-d5	7.440	82	18282	2.992	ng	-0.01
42) 2,4,6-Tribromophenol	10.733	330	18560	7.007	ng	0.02
45) 2-Fluorobiphenyl	9.239	172	46533	4.059	ng	0.00
79) Terphenyl-d14	13.004	244	91607	4.506	ng	0.00
Target Compounds						
37) 2-Methylnaphthalene	8.881	142	39143	3.429	ng	Qvalue 96
74) Di-n-butylphthalate	11.980	149	84876	4.794	ng	# 96
84) Bis(2-ethylhexyl)phtha...	14.045	149	267308	25.928	ng	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF092525\
Data File : BF143808.D
Acq On : 25 Sep 2025 17:21
Operator : RC/JU
Sample : Q3188-01 10X
Misc :
ALS Vial : 8 Sample Multiplier: 1

Instrument :
BNA_F
ClientSampleId :
4135

Quant Time: Sep 25 17:53:46 2025
Quant Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF092325.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Wed Sep 24 05:00:00 2025
Response via : Initial Calibration

