

Data Path : Z:\SVOASRV\HPCHEM1\BNA F\DATA\BF092718\
 Data File : BF109426.D
 Acq On : 28 Sep 2018 10:58
 Operator : JU/SJ
 Sample : J5133-06
 Misc :
 ALS Vial : 47 Sample Multiplier: 1

Instrument :
 BNA_F
 ClientSampleId :
 092618-SIDI-F-U-S

Quant Time: Sep 28 12:08:22 2018
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA F\METHODS\8270-BF092218.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Sat Sep 22 13:32:18 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	6.70	152	106010	20.00	ng	0.00
21) Naphthalene-d8	7.99	136	399859	20.00	ng	0.00
38) Acenaphthene-d10	9.73	164	184115	20.00	ng	0.00
63) Phenanthrene-d10	11.21	188	303740	20.00	ng	0.00
75) Chrysene-d12	13.84	240	244997	20.00	ng	0.00
86) Perylene-d12	15.24	264	199785	20.00	ng	0.00
System Monitoring Compounds						
5) 2-Fluorophenol	5.31	112	264791	41.14	ng	0.00
7) Phenol-d6	6.34	99	220854	26.89	ng	-0.02
23) Nitrobenzene-d5	7.27	82	686641	101.37	ng	-0.01
41) 2,4,6-Tribromophenol	10.52	330	155935	85.92	ng	-0.01
44) 2-Fluorobiphenyl	9.07	172	1225517	100.39	ng	0.00
78) Terphenyl-d14	12.80	244	977480	97.92	ng	0.00
Target Compounds						
10) Phenol	6.35	94	22252	2.392	ng	97
49) Dimethylphthalate	9.46	163	35811	2.674	ng	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\SVOASRV\HPCHEM1\BNA F\DATA\BF092718\
Data File : BF109426.D
Acq On : 28 Sep 2018 10:58
Operator : JU/SJ
Sample : J5133-06
Misc :
ALS Vial : 47 Sample Multiplier: 1

Instrument :
BNA_F
ClientSampleId :
092618-SIDI-F-U-S

Quant Time: Sep 28 12:08:22 2018
Quant Method : Z:\SVOASRV\HPCHEM1\BNA F\METHODS\8270-BF092218.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Sat Sep 22 13:32:18 2018
Response via : Initial Calibration

